

Login Checklist

RQ- 2023 - 05 - 22 - 24

Login Station ID: # 2

Shipping Method: FedEx | UPS | HD | GreyhoundDate/Time of Receipt: **07/20/2023 / 09:05**[illegible]

COOLER CHECK

***All samples received on ice unless otherwise noted. HNO3 and Formalin preserved samples do NOT have a temperature requirement.**

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. **NCR#** _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. **NCR#** _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. **NCR#** _____
- If coolers or samples are received late; identify the affected samples in an NCR. **NCR#** _____

Chain of Custody/ Field Sheet(s) Included?

Yes x No

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes No **x**

CONTAINER CHECK

Evidence Tape on Bottles? Yes _____ No _____ NA x Evidence Tape intact? Yes _____ No _____

If Criminal, photographs taken? Yes No Containers intact? Yes x No

Caps on tight? Yes x No

Sufficient Sample Volume: Yes x No

If **NO** is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? (Blue dot on container) Yes x No Init: **ABS** /

Chlorine detected? (One container checked per Field ID) Yes _____ No x Init: **ABS** / _____

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

** Acid Preserved Samples: pH < 2.0? Yes x No NA Init: ABS /

W-1-4-DIOX (Green dot on container) preserved to pH<4.0? Yes _____ No _____ NA **X Init: **ABS** /

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: **ABS** / _____ Date: **07/20/2023** Event contains NCRs (Y/N)? **y**

EventID: NW - DIST - 2023 - 07 - 20 - 01

Login Checklist

Login Preservation Equipment Lot # & Exp: Date Additional Comments:

Infrared Thermometer ID:	IRTG # 2 EXP:10/17/2023
pH Test Strips 0 – 6	LOT#223819AV EXP:8/30/23
pH Paper 0 – 3	N/A
pH Paper 0 – 13	LOT#207621 EXP:3/15/2024
Chlorine Test Strips	LOT#223819AV EXP:8/30/23 1196 1/16/23/ 02

1196/2024/02 abs/04/03/23

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: ABS

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 7/19/2023
Customer: NW-DIST
RQ: RQ-2023-05-22-24
Collected By: Frank Butera, Myranda Butler
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
HURRICANE LAKE (CENTER)	 G4NW0033

RELINQUISHED BY(SIGNATURE):

Myranda Butler

DATE/TIME: 7/19/2023
1600 CEN

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By: A.SHAIK

Rec'd/Inspected
Date: 07/20/2023 / 09:05



RQ-2023-05-22-24	Collected By: Frank Butera, Myranda Butler
Customer: FDEP	Field Report Prepared by: Frank Butera, Myranda Butler
Project ID: SMP	Sampling Agency: FDEP

WIN ID/Field ID:  WIN ID/Field ID: G4NW0033	Comments:
---	-----------

Location: HURRICANE LAKE (CENTER)

Matrix: Fresh

(Check One)

☒ Grab☐ Composite

Date and Time {CEN}	DO (mg/L)	DO (%SAT)	Temp. °C	pH	Sample Depth (m)	Secchi Depth	Total Depth (m)	Sp COND (µmhos /cm)	Salinity (PPT _h)
7/19/2023 1235	8.48	116.7	32.3	6.92	0.3	0.5	5.6	17.5	0.01

Parameter Suite	Parameter Methods						Preservation			pH Check		# Bottles sent to Lab	Bottle Group			
Preserved Nutrients (P-500ML)	X	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP					X	Ice	X	H2SO4*	X	<2	1	A		
	Lot #		SA3013020		Exp:		01/25/2024									
Metals (P-500ML)	X	W-ICPMS		X	W-ICP		X	W-HARD	X	Ice	X	HNO3*	X	<2	1	A
	Lot #		NA2230040		Exp:		09/01/2023									
Filtered Nutrients (P125ML)	X	W-PO4-F		X	Field Filtered 0.45 um Filter			X	Ice				1	A		
	Filter Lot #		17460495		Syringe Lot #		136314									
BOD (BP-1L)		OV-BODUNIN							Ice							
Chlorophyll (BP-1L)	X	CHLSUITE-W						X	Ice				1	A		
Physical/Aggregate (Fresh) (P-1L)	X	Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W- TDS						X	Ice				1	A		
Physical/Aggregate (Marine) (P-1L)		Alkalinity / W-F / W-TSS / TURBIDITY							Ice							
Macroinvertebrates (PJ-2L)		MI-FW-QLDC		Formalin Lot #				Formalin								
Periphyton (PT-50ML)		ALGAE-ID							Ice							
Microbiology (Contract Lab Bottle)		E Coli			F Coli			Enterococci		Ice						
	Contract Lab															
Molecular (QPCR-P-500ML)		PCR-HF183			PCR-BacR			PCR-DG3		Ice						
		PCR-GULL2			PCR-GFD											
Tracers (BG-500ML)		W-E8321-DI			W-E8321-MS				Ice							
Pesticides (BG-500ML) (BG-1L)	X	W-E8321-DI	X	W-E8321-MS		X	W-PSNP-TQ		X	Ice	MCAA Exp.	MCAA Lot#	5	A		
		W-PCL-TQ-R		W-CARB-AA (5 ml of MCAA Buffer Solution)		W-CT-CL-R										
	X	W-TR-SQ-R			W-PCB-TQ-R											
Phytoplankton (BP-1L)		DTY-QN-C			PKD-QN-BV				Ice							
DOC (P-500ML)		W-DOC		Field Filtered 0.45 um Filter				Ice		H2SO4*	<2					
	Acid Preservative Expiration Date															
	Acid Preservative Lot Number															

Date of Request: 18-APR-2023
 Created By: SLADE_C On: 18-APR-2023
 Modified By: JASROTIA_P On: 04-MAY-2023
 Customer: NW-DIST
 Project: SMP
 Division: Division of Environmental Assessment and Restoration
 District: Northwest District
 Event Description: Hurricane Lake LVI

Send Coolers To:

Phone: 850-595-0567
 160 W Government Street
 Suite 208
 Pensacola, FL 32502-5794
 Attn: Stopher Slade
 Cooler Ship EMail: christopher.slade@dep.state.fl.us

Priority: 3
 Event Status: R
 Criminal Investigation: NO
 Chemistry Request Reviewed By: AYRES_J On: 04-MAY-2023
 Biology Request Reviewed By: JASROTIA_P On: 04-MAY-2023
 Sampling Kit Required: YES Ship on: 12-MAY-2023
 Sampling Kit Shipped: YES On: 08-MAY-2023
 Sampling Kit Packed By: PENDLETO_C
 Preservatives Needed: NO
 Date to Receive Samples: 22-MAY-2023
 Logged In By: SHAIK_A On: 20-JUL-2023

Send Final Report To:

160 W Government Street
 Suite 208
 Pensacola, FL 32502-5794
 Attn: Myranda K Butler
 Report Notification EMail:
 christopher.slade@dep.state.fl.us;myranda.k.butler@dep.state.fl.us

Comments: Group A: (WBID 83A) Nutrients w/ PO4, Metals, Traces, Limited Pesticides

Packing notes;

FedEx labels

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO ₃ /L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(DEP SOP: NU-094-1) True Color measured at 450 nm (not valid for NPDES monitoring)
Color (true)		
W-F	Template: DEFAULT	(SM 4500-F- C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2015) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2015) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L	Number of Bottles: 1	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10150 B) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: BG-500ML	Number of Bottles: 1	Preserved With: ICE
W-E8321-DI	Template: DEFAULT	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
W-E8321-MS	Template: SMP	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO ₃
W-HARD	Template: DEFAULT	(SM 2340 B-2011) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO ₃
Hardness		
W-ICP	Template: SMP-FRESH	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Calcium		
Iron		
Magnesium		

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: HNO3
 W-ICP Template: SMP-FRESH (EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects

Potassium
 Sodium
 Strontium
 Tin
 Titanium
 Vanadium
 Zinc

W-ICPMS Template: DEFAULT (EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

Aluminum
 Antimony
 Arsenic
 Barium
 Beryllium
 Boron
 Cadmium
 Chromium
 Cobalt
 Copper
 Lead
 Manganese
 Molybdenum
 Nickel
 Selenium
 Silver
 Thallium

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: ICE And H2SO4
 W-NH3 Template: DEFAULT (EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
 W-NO2NO3 Template: DEFAULT (EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
 W-S-A-TP Template: DEFAULT (EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
 W-TKN Template: DEFAULT (EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
 W-TOC Template: DEFAULT (SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Type: P-125ML Number of Bottles: 1 Preserved With: FILTER-ICE
 W-PO4-F Template: DEFAULT (EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Bottle Type: BG-500ML Number of Bottles: 4 Preserved With: ICE
 W-PSNP-TQ Template: DEFAULT (EPA 8270E) Organonitrogen and phosphorus pesticides in water matrices by GC/MS/MS
 W-TR-SQ-R Template: DEFAULT (EPA 8270E) Analysis of Sunscreen Compounds and Flame Retardants in Water by GC/MS-SIM

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.