

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: CHERRY

RQ: SEE COMMENTS

Project: SMP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 9/8/23

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	K TURMAN	20 SEP 23	0722	22.5	761.2	8.7	8.69	100.4	-	093306	P / F	L / F
ICV	"	↓	0725	22.5	761.2	8.7	8.67	100.1	-	-	P / F	L / F
CCV	Hannah Henning	↓	1355	22.5	7600	8.7	8.63	99.8			P / F	L / F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	K TURMAN	20 SEP 23	0728	430679	6/25	1000	998	P / F	L / F
ICV	"	↓	0730	210987	9/23	100	99.1	P / F	L / F
CCV	Hannah Henning	↓	1400	221064	10/24	500	495.0	P / F	L / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	K TURMAN	20 SEP 23	0733	224883	9/24	7.01	22.4	7.01	-223	P / F	L / F
Calibr.	"	↓	0736	219096	2/24	4.00	22.5	3.97	157.8	P / F	L / F
Calibr.	"	↓	0738	213755	6/23	10.02	22.3	10.02	-195.6	P / F	L / F
ICV	"	↓	0740	213755	6/23	10.00	22.2	10.03	-	P / F	L / F
CCV	Hannah Henning	↓	1359	224883	9/24	7	22.0	7.08		P / F	L / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50;

mV pH 4 Range +180 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50;

If mV are recorded: slope from 7 to 10 173.3, slope from 4 to 7 180.1 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: #2; Date of last verification: 3/14/23

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: x3 RQ's

G2SW0087 = RQ 2023.06.2606

G3SW0100 = RQ 2023.08.28.41

G2SW0053 } $\rightarrow$ RQ 2023.09.18.34

G3SW0043 }

Form effective August 1, 2018