

Chemical Analysis Report

SW-DIST-2023-05-04-01

Florida DEP Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **LVI 2023**
Request ID: **RQ-2023-04-24-13**
Customer: **SW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
13051 N. Telecom Parkway
Temple Terrace, FL 33637-0926
Attn: Ben Paswater


For additional information please contact
Colin Wright, Ph.D.
Chris Bowen, B.S.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

This Report replaces the previous Report Serial Number: 0146724

Report Comment: Missing I qualifier was added to the octinoxate result.

Revision certified by: Colin Wright, Program Administrator

Date Certified: 28-JUN-2023 13:52



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals, Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: PLATT LAKE CENTER

Collection Date/Time: 05/03/2023 11:30

Field ID: G1SW0083

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2405693	DEP SOP: NU-094-1	Color (true)	41		PCU	P429422	
	EPA 180.1 Rev. 2.0	Turbidity	0.95	I	NTU	P429436	
	EPA 300.0 Rev. 2.1	Chloride	31	A	mg Cl/L	P429662	
		Sulfate	7.3	A	mg SO4/L	P429664	
	SM 2320 B-2011	Total Alkalinity	33		mg CaCO3/L	P429832	
	SM 2540 C-2015	TDS	123		mg/L	P429834	
	SM 2540 D-2015	TSS	2	I	mg/L	P429764	
	SM 4500-F- C-2011	Fluoride	0.079	I	mg F/L	P429697	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.008		mg N/L	P429609	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.94		mg N/L	P429573	
2405694	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P429812	
	EPA 365.1 Rev. 2.0	Total-P	0.020		mg P/L	P429536	
	SM 5310 B-2011	Organic Carbon	14		mg C/L	P429705	
	EPA 365.1 Rev. 2.0	O-Phosphate-P	0.004	U	mg P/L	P429425	
	dissolved						
2405697	EPA 8321B	2,4-D	0.0020	U	ug/L	P429412	
		Acesulfame K	0.10	U	ug/L	P429323	
		2,4,5-T	0.0040	U	ug/L	P429412	
		AMPA	0.10	U	ug/L	P429323	
		Acetaminophen	0.0080	U	ug/L	P429412	
		Acetamiprid	0.0020	U	ug/L	P429412	
		Endothall	0.25	U	ug/L	P429323	
		Glufosinate	0.10	U	ug/L	P429323	
		Glyphosate	0.10	U	ug/L	P429323	
		Afidopyropen	0.0020	U	ug/L	P429412	
		Sucralose	0.51		ug/L	P429323	
		Bentazon	0.26		ug/L	P429412	
		Benzovindiflupyr	0.0020	U	ug/L	P429412	
		Carbamazepine	8.0E-04	U	ug/L	P429412	
		Clothianidin	0.0020	U	ug/L	P429412	
		Dinotefuran	0.0020	U	ug/L	P429412	
		Diuron	0.0053	I	ug/L	P429412	
		Fenuron	0.032	U	ug/L	P429412	
		Fluridone	0.0022	I	ug/L	P429412	
		Imazapyr	0.021		ug/L	P429412	
		Hydrocodone	0.0020	U	ug/L	P429412	
		Ibuprofen	0.080	U	ug/L	P429412	
		Imidacloprid	0.0030	I	ug/L	P429412	
		Linuron	0.0040	U	ug/L	P429412	
		Mandestrobin	0.0020	U	ug/L	P429412	
		MCPP	0.0020	U	ug/L	P429412	
		Naproxen	0.0080	U	ug/L	P429412	
		Primidone	0.0040	U	ug/L	P429412	
		Pyraclostrobin	0.0020	U	ug/L	P429412	

MS

Field ID: G1SW0083

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2405697	EPA 8321B	Silvex	0.0040	U	ug/L	P429412	
		Thiamethoxam	0.0020	U	ug/L	P429412	
		Tolfenpyrad	0.0020	U	ug/L	P429412	
		Triclopyr	0.0040	U	ug/L	P429412	
2405704	EPA 200.7 Rev. 4.4	Calcium	16.7		mg/L	P429542	
		Iron	30	U	ug/L	P429542	
		Magnesium	1.69		mg/L	P429542	
		Potassium	2.1		mg/L	P429542	
		Sodium	18.1		mg/L	P429542	
		Strontium	40.7		ug/L	P429542	
		Tin	3.0	U	ug/L	P429542	
		Titanium	0.75	U	ug/L	P429542	
		Vanadium	2.0	U	ug/L	P429542	
		Zinc	5.0	U	ug/L	P429542	
	EPA 200.8 Rev. 5.4	Aluminum	27		ug/L	P429542	
		Antimony	0.14	I	ug/L	P429542	
		Arsenic	2.69		ug/L	P429542	
		Barium	12.5		ug/L	P429542	
		Beryllium	0.010	U	ug/L	P429542	
		Boron	36.1		ug/L	P429542	
		Cadmium	0.020	U	ug/L	P429542	
		Chromium	0.40	U	ug/L	P429542	
		Cobalt	0.020	U	ug/L	P429542	
		Copper	0.63	I	ug/L	P429542	
2405711	SM 2340 B-2011	Lead	0.20	U	ug/L	P429542	
		Manganese	5.61		ug/L	P429542	
		Molybdenum	0.15	U	ug/L	P429542	
		Nickel	0.50	U	ug/L	P429542	
		Selenium	0.20	U	ug/L	P429542	
		Silver	0.010	U	ug/L	P429542	
		Thallium	0.10	U	ug/L	P429542	
		Hardness	48.6		mg/L	P429542	
	EPA 8270E	Oxybenzone**	54		ng/L	P429507	
		Acetochlor	0.22	U	ng/L	P429507	
		Octinoxate**	38	IV	ng/L	P429507	MS
		Alachlor	0.43	U	ng/L	P429507	
		Avobenzone**	110	U	ng/L	P429507	
		Ametryn	1.2	U	ng/L	P429507	
		Atrazine	130		ng/L	P429507	
		PBDE-47**	4.3	U	ng/L	P429507	
		Atrazine Desethyl	16		ng/L	P429507	
		PBDE-99**	5.4	U	ng/L	P429507	
		Azinphos Methyl	3.2	U	ng/L	P429507	
		Tri-o-cresyl Phosphate**	6.5	U	ng/L	P429507	
		Azoxystrobin	0.43	U	ng/L	P429507	

Field ID: G1SW0083

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2405711	EPA 8270E	Bromacil	0.65	U	ng/L	P429507	
		2-Ethylhexyl	13	U	ng/L	P429507	
		2,3,4,5-tetrabromobenzoate (EHTBB)**					
		Butylate	0.22	U	ng/L	P429507	
		Dechlorane Plus**	32	U	ng/L	P429507	
		Caffeine**	380		ng/L	P429507	
		Dechlorane 602**	5.4	U	ng/L	P429507	
		Carbophenothion	0.54	U	ng/L	P429507	
		Dechlorane 603**	4.3	U	ng/L	P429507	
		Carfentrazone Ethyl	0.16	U	ng/L	P429507	
		Hexabromobenzene**	6.5	U	ng/L	P429507	
		Chlorfenapyr	0.32	U	ng/L	P429507	
		PBB-153**	8.6	U	ng/L	P429507	
		Chlorpyrifos Ethyl	0.32	U	ng/L	P429507	
		Pentabromotoluene**	3.2	U	ng/L	P429507	
		Chlorpyrifos Methyl	0.11	U	ng/L	P429507	
		Tris(2-chloroethyl) phosphate (TCEP)**	76	U	ng/L	P429507	
		Coumaphos	1.1	U	ng/L	P429507	
		Tris(1-chloro-2-propyl) phosphate (TCPP)**	46	I	ng/L	P429507	
		Cyanazine	5.4	U	ng/L	P429507	
		Tris(1,3-dichloro-2-propyl) phosphate (TDCPP)**	20		ng/L	P429507	
		Cyprodinil	0.22	U	ng/L	P429507	
		Tetradecachloro-m-terphenyl**	8.1	UJ	ng/L	P429507	LCS, MS
		Demeton	3.8	U	ng/L	P429507	
		Diazinon	0.13	U	ng/L	P429507	
		Difenoconazole	0.65	U	ng/L	P429507	
		Dimethenamid	0.11	U	ng/L	P429507	
		Dimethomorph	0.32	U	ng/L	P429507	
		Disulfoton	0.65	U	ng/L	P429507	
		Dithiopyr	1.1	U	ng/L	P429507	
		EPTC	0.54	U	ng/L	P429507	
		Ethion	0.32	U	ng/L	P429507	
		Ethoprop	0.13	U	ng/L	P429507	
		Etridiazole	0.11	U	ng/L	P429507	
		Fenamiphos	0.49	U	ng/L	P429507	
		Fenbuconazole	0.54	U	ng/L	P429507	
		Fipronil	0.22	U	ng/L	P429507	
		Fipronil Desulfinyl**	0.64		ng/L	P429507	
		Fipronil Sulfide	0.19	I	ng/L	P429507	
		Fipronil Sulfone	0.47	I	ng/L	P429507	
		Fluazifop-P-butyl	0.11	U	ng/L	P429507	
		Fludioxonil	0.11	U	ng/L	P429507	
		Flumioxazin	3.2	U	ng/L	P429507	
		Flutolanil	0.11	U	ng/L	P429507	
		Fluxapyroxad	0.27	U	ng/L	P429507	

Field ID: G1SW0083

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2405711	EPA 8270E	Fonofos	0.16	U	ng/L	P429507	LCS, MS
		Hexazinone	0.54	U	ng/L	P429507	
		Imazalil	0.70	U	ng/L	P429507	
		Iprodione	0.54	UJ	ng/L	P429507	
		Loratadine**	0.32	U	ng/L	P429507	
		Malathion	0.38	U	ng/L	P429507	
		Metalaxyl	0.54	U	ng/L	P429507	
		Metolachlor	0.81	U	ng/L	P429507	
		Metribuzin	0.43	U	ng/L	P429507	
		Mevinphos	1.1	U	ng/L	P429507	
		Molinate	0.27	U	ng/L	P429507	
		Myclobutanil	1.3	I	ng/L	P429507	
		Naled	2.7	U	ng/L	P429507	
		Napropamide	0.76	U	ng/L	P429507	
		Norflurazon	1.1	U	ng/L	P429507	
		Oxadiazon	0.52	I	ng/L	P429507	
		Oxyfluorfen	0.22	U	ng/L	P429507	
		Parathion Ethyl	0.16	U	ng/L	P429507	
		Parathion Methyl	0.22	U	ng/L	P429507	
		Pendimethalin	0.81	U	ng/L	P429507	MS
		Phenytol**	5.7	U	ng/L	P429507	
		Phorate	0.27	U	ng/L	P429507	
		Prodiamine	0.22	U	ng/L	P429507	
		Prometon	3.2	U	ng/L	P429507	
		Prometryn	0.38	U	ng/L	P429507	
		Pronamide	0.11	U	ng/L	P429507	
		Propanil	0.11	U	ng/L	P429507	
		Propargite	0.86	U	ng/L	P429507	
		Propiconazole	0.65	U	ng/L	P429507	
		Pyraflufen Ethyl	0.32	U	ng/L	P429507	LCS
		Pyridaben	1.5	U	ng/L	P429507	
		Pyriproxyfen	0.27	UJ	ng/L	P429507	
		Simazine	8.8		ng/L	P429507	
		Stirophos	0.43	U	ng/L	P429507	
		Sulfentrazone	1.5	U	ng/L	P429507	
		Tebuconazole	1.5	U	ng/L	P429507	
		Terbufos	0.22	U	ng/L	P429507	
		Terbuthylazine	0.22	U	ng/L	P429507	
		Thiobencarb	0.54	U	ng/L	P429507	
		Triadimefon	0.16	U	ng/L	P429507	

Ref. Method and Comment:

SM 2540 D-2015: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 8270E: V - Due to the presence of octinoxate in both the Method Blank and the sample. Refer to the Lab Analysis Report for an explanation of QC Codes.

EPA 8270E: MS accuracy for bromacil and norflurazon not assessed due to high content of these parameters in the sample spiked. Refer to the Lab Analysis Report for an explanation of QC Codes.

Quality Assurance Report Method Blank Results

Reference Method: DEP SOP: NU-094-1
Batch ID: P429422

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P429436

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P429542

Component	Result	Code	Units
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Strontium	2.0	U	ug/L
Tin	3.0	U	ug/L
Titanium	0.75	U	ug/L
Vanadium	2.0	U	ug/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P429542

Component	Result	Code	Units
Aluminum	5.0	U	ug/L
Antimony	0.050	U	ug/L
Arsenic	0.050	U	ug/L
Barium	0.20	U	ug/L
Beryllium	0.010	U	ug/L
Boron	2.0	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Cobalt	0.020	U	ug/L
Copper	0.40	U	ug/L
Lead	0.20	U	ug/L
Manganese	0.10	U	ug/L
Molybdenum	0.15	U	ug/L
Nickel	0.50	U	ug/L
Selenium	0.20	U	ug/L
Silver	0.010	U	ug/L
Thallium	0.10	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P429662

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P429664

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P429609

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P429573

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P429812

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P429536

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P429425

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8270E
Batch ID: P429507

Component	Result	Code	Units
2-Ethylhexyl 2,3,4,5-tetrabromobenzoate (EHTBB)	12	U	ng/L
2-Ethylhexyl 2,3,4,5-tetrabromobenzoate (EHTBB)	12	U	ng/L
Acetochlor	0.20	U	ng/L
Acetochlor	0.20	U	ng/L
Alachlor	0.40	U	ng/L
Alachlor	0.40	U	ng/L
Ametryn	1.1	U	ng/L
Ametryn	1.1	U	ng/L
Atrazine	0.20	U	ng/L
Atrazine	0.20	U	ng/L
Atrazine Desethyl	1.3	U	ng/L
Atrazine Desethyl	1.3	U	ng/L
Avobenzone	100	U	ng/L
Avobenzone	100	U	ng/L
Azinphos Methyl	3.0	U	ng/L
Azinphos Methyl	3.0	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E

Batch ID: P429507

Component	Result	Code	Units
Azoxystrobin	0.40	U	ng/L
Azoxystrobin	0.40	U	ng/L
Bromacil	0.60	U	ng/L
Bromacil	0.60	U	ng/L
Butylate	0.20	U	ng/L
Butylate	0.20	U	ng/L
Caffeine	7.5	U	ng/L
Caffeine	7.5	U	ng/L
Carbophenothion	0.50	U	ng/L
Carbophenothion	0.50	U	ng/L
Carfentrazone Ethyl	0.15	U	ng/L
Carfentrazone Ethyl	0.15	U	ng/L
Chlorfenapyr	0.30	U	ng/L
Chlorfenapyr	0.30	U	ng/L
Chlorpyrifos Ethyl	0.30	U	ng/L
Chlorpyrifos Ethyl	0.30	U	ng/L
Chlorpyrifos Methyl	0.10	U	ng/L
Chlorpyrifos Methyl	0.10	U	ng/L
Coumaphos	1.0	U	ng/L
Coumaphos	1.0	U	ng/L
Cyanazine	5.0	U	ng/L
Cyanazine	5.0	U	ng/L
Cyprodinil	0.20	U	ng/L
Cyprodinil	0.20	U	ng/L
Dechlorane 602	5.0	U	ng/L
Dechlorane 602	5.0	U	ng/L
Dechlorane 603	4.0	U	ng/L
Dechlorane 603	4.0	U	ng/L
Dechlorane Plus	30	U	ng/L
Dechlorane Plus	30	U	ng/L
Demeton	3.5	U	ng/L
Demeton	3.5	U	ng/L
Diazinon	0.12	U	ng/L
Diazinon	0.12	U	ng/L
Difenoconazole	0.60	U	ng/L
Difenoconazole	0.60	U	ng/L
Dimethenamid	0.10	U	ng/L
Dimethenamid	0.10	U	ng/L
Dimethomorph	0.30	U	ng/L
Dimethomorph	0.30	U	ng/L
Disulfoton	0.60	U	ng/L
Disulfoton	0.60	U	ng/L
Dithiopyr	1.0	U	ng/L
Dithiopyr	1.0	U	ng/L
EPTC	0.50	U	ng/L
EPTC	0.50	U	ng/L
Ethion	0.30	U	ng/L
Ethion	0.30	U	ng/L
Ethoprop	0.12	U	ng/L
Ethoprop	0.12	U	ng/L
Etridiazole	0.10	U	ng/L
Etridiazole	0.10	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E

Batch ID: P429507

Component	Result	Code	Units
Fenamiphos	0.45	U	ng/L
Fenamiphos	0.45	U	ng/L
Fenbuconazole	0.50	U	ng/L
Fenbuconazole	0.50	U	ng/L
Fipronil	0.20	U	ng/L
Fipronil	0.20	U	ng/L
Fipronil Desulfinyl	0.10	U	ng/L
Fipronil Desulfinyl	0.10	U	ng/L
Fipronil Sulfide	0.10	U	ng/L
Fipronil Sulfide	0.10	U	ng/L
Fipronil Sulfone	0.10	U	ng/L
Fipronil Sulfone	0.10	U	ng/L
Fluazifop-P-butyl	0.10	U	ng/L
Fluazifop-P-butyl	0.10	U	ng/L
Fludioxonil	0.10	U	ng/L
Fludioxonil	0.10	U	ng/L
Flumioxazin	3.0	U	ng/L
Flumioxazin	3.0	U	ng/L
Flutolanil	0.10	U	ng/L
Flutolanil	0.10	U	ng/L
Fluxapyroxad	0.25	U	ng/L
Fluxapyroxad	0.25	U	ng/L
Fonofos	0.15	U	ng/L
Fonofos	0.15	U	ng/L
Hexabromobenzene	6.0	U	ng/L
Hexabromobenzene	6.0	U	ng/L
Hexazinone	0.50	U	ng/L
Hexazinone	0.50	U	ng/L
Imazalil	0.65	U	ng/L
Imazalil	0.65	U	ng/L
Iprodione	0.50	U	ng/L
Iprodione	0.50	U	ng/L
Loratadine	0.30	U	ng/L
Loratadine	0.30	U	ng/L
Malathion	0.35	U	ng/L
Malathion	0.35	U	ng/L
Metalaxyl	0.50	U	ng/L
Metalaxyl	0.50	U	ng/L
Metolachlor	0.75	U	ng/L
Metolachlor	0.75	U	ng/L
Metribuzin	0.40	U	ng/L
Metribuzin	0.40	U	ng/L
Mevinphos	1.0	U	ng/L
Mevinphos	1.0	U	ng/L
Molinate	0.25	U	ng/L
Molinate	0.25	U	ng/L
Myclobutanil	0.35	U	ng/L
Myclobutanil	0.35	U	ng/L
Naled	2.5	U	ng/L
Naled	2.5	U	ng/L
Napropamide	0.70	U	ng/L
Napropamide	0.70	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E

Batch ID: P429507

Component	Result	Code	Units
Norflurazon	1.0	U	ng/L
Norflurazon	1.0	U	ng/L
Octinoxate	92		ng/L
Octinoxate	42	I	ng/L
Oxadiazon	0.35	U	ng/L
Oxadiazon	0.35	U	ng/L
Oxybenzone	10	U	ng/L
Oxybenzone	10	U	ng/L
Oxyfluorfen	0.20	U	ng/L
Oxyfluorfen	0.20	U	ng/L
Parathion Ethyl	0.15	U	ng/L
Parathion Ethyl	0.15	U	ng/L
Parathion Methyl	0.20	U	ng/L
Parathion Methyl	0.20	U	ng/L
PBB-153	8.0	U	ng/L
PBB-153	8.0	U	ng/L
PBDE-47	4.0	U	ng/L
PBDE-47	4.0	U	ng/L
PBDE-99	5.0	U	ng/L
PBDE-99	5.0	U	ng/L
Pendimethalin	0.75	U	ng/L
Pendimethalin	0.75	U	ng/L
Pentabromotoluene	3.0	U	ng/L
Pentabromotoluene	3.0	U	ng/L
Phenytoin	5.3	U	ng/L
Phenytoin	5.3	U	ng/L
Phorate	0.25	U	ng/L
Phorate	0.25	U	ng/L
Prodiamine	0.20	U	ng/L
Prodiamine	0.20	U	ng/L
Prometon	3.0	U	ng/L
Prometon	3.0	U	ng/L
Prometryn	0.35	U	ng/L
Prometryn	0.35	U	ng/L
Pronamide	0.10	U	ng/L
Pronamide	0.10	U	ng/L
Propanil	0.10	U	ng/L
Propanil	0.10	U	ng/L
Propargite	0.80	U	ng/L
Propargite	0.80	U	ng/L
Propiconazole	0.60	U	ng/L
Propiconazole	0.60	U	ng/L
Pyraflufen Ethyl	0.30	U	ng/L
Pyraflufen Ethyl	0.30	U	ng/L
Pyridaben	1.4	U	ng/L
Pyridaben	1.4	U	ng/L
Pyriproxyfen	0.25	U	ng/L
Pyriproxyfen	0.25	U	ng/L
Simazine	0.25	U	ng/L
Simazine	0.25	U	ng/L
Stirophos	0.40	U	ng/L
Stirophos	0.40	U	ng/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 8270E

Batch ID: P429507

Component	Result	Code	Units
Sulfentrazone	1.4	U	ng/L
Sulfentrazone	1.4	U	ng/L
Tebuconazole	1.4	U	ng/L
Tebuconazole	1.4	U	ng/L
Terbufos	0.20	U	ng/L
Terbufos	0.20	U	ng/L
Terbutylazine	0.20	U	ng/L
Terbutylazine	0.20	U	ng/L
Tetradecachloro-m-terphenyl	7.5	U	ng/L
Tetradecachloro-m-terphenyl	7.5	U	ng/L
Thiobencarb	0.50	U	ng/L
Thiobencarb	0.50	U	ng/L
Tri-o-cresyl Phosphate	6.0	U	ng/L
Tri-o-cresyl Phosphate	6.0	U	ng/L
Triadimefon	0.15	U	ng/L
Triadimefon	0.15	U	ng/L
Tris(1-chloro-2-propyl) phosphate (TCPP)	15	U	ng/L
Tris(1-chloro-2-propyl) phosphate (TCPP)	15	U	ng/L
Tris(1,3-dichloro-2-propyl) phosphate (TDCPP)	2.5	U	ng/L
Tris(1,3-dichloro-2-propyl) phosphate (TDCPP)	2.5	U	ng/L
Tris(2-chloroethyl) phosphate (TCEP)	70	U	ng/L
Tris(2-chloroethyl) phosphate (TCEP)	70	U	ng/L

Reference Method: EPA 8321B

Batch ID: P429323

Component	Result	Code	Units
Acesulfame K	0.10	U	ug/L
AMPA	0.10	U	ug/L
Endothall	0.25	U	ug/L
Glufosinate	0.10	U	ug/L
Glyphosate	0.10	U	ug/L
Sucralose	0.010	U	ug/L

Reference Method: EPA 8321B

Batch ID: P429412

Component	Result	Code	Units
2,4-D	0.0020	U	ug/L
2,4,5-T	0.0040	U	ug/L
Acetaminophen	0.0080	U	ug/L
Acetamiprid	0.0020	U	ug/L
Afidopyopen	0.0020	U	ug/L
Bentazon	8.0E-04	U	ug/L
Benzovindiflupyr	0.0020	U	ug/L
Carbamazepine	8.0E-04	U	ug/L
Clothianidin	0.0020	U	ug/L
Dinotefuran	0.0020	U	ug/L
Diuron	0.0020	U	ug/L
Fenuron	0.032	U	ug/L
Fluridone	8.0E-04	U	ug/L
Hydrocodone	0.0020	U	ug/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8321B
Batch ID: P429412

Component	Result	Code	Units
Ibuprofen	0.080	U	ug/L
Imazapyr	0.0040	U	ug/L
Imidacloprid	0.0020	U	ug/L
Linuron	0.0040	U	ug/L
Mandestrobin	0.0020	U	ug/L
MCPP	0.0020	U	ug/L
Naproxen	0.0080	U	ug/L
Primidone	0.0040	U	ug/L
Pyraclostrobin	0.0020	U	ug/L
Silvex	0.0040	U	ug/L
Thiamethoxam	0.0020	U	ug/L
Tolfenpyrad	0.0020	U	ug/L
Triclopyr	0.0040	U	ug/L

Reference Method: SM 2320 B-2011
Batch ID: P429832

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-2015
Batch ID: P429834

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2015
Batch ID: P429764

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500-F- C-2011
Batch ID: P429697

Component	Result	Code	Units
Fluoride	0.025	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P429705

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: DEP SOP: NU-094-1
Batch ID: P429422

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	98.4		P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P429436

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Turbidity	99.4		P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P429542

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Calcium	100		P	80 - 120
Iron	100		P	80 - 120
Magnesium	101		P	80 - 120
Potassium	100		P	80 - 120
Sodium	99.5		P	80 - 120
Strontium	102		P	80 - 120
Tin	102		P	80 - 120
Titanium	100		P	80 - 120
Vanadium	99.7		P	80 - 120
Zinc	102		P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P429542

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aluminum	102		P	85 - 115
Antimony	108		P	85 - 115
Arsenic	102		P	85 - 115
Barium	105		P	85 - 115
Beryllium	98.9		P	85 - 115
Boron	103		P	85 - 115
Cadmium	102		P	85 - 115
Chromium	98.6		P	85 - 115
Cobalt	103		P	85 - 115
Copper	98.5		P	85 - 115
Lead	105		P	85 - 115
Manganese	99.7		P	85 - 115
Molybdenum	101		P	85 - 115
Nickel	98.0		P	85 - 115
Selenium	102		P	85 - 115
Silver	106		P	85 - 115
Thallium	101		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P429662

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.7		P	90 - 110

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P429664

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	97.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P429609

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.0		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P429573

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	92.4		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P429812

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P429536

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P429425

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.2		P	90 - 110

Reference Method: EPA 8270E
Batch ID: P429507

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2-Ethylhexyl 2,3,4,5-tetrabromobenzoate (EHTBB)	123		P	42 - 162
Acetochlor	71.8		P	69 - 123
Alachlor	100		P	65 - 118
Ametryn	135		P	58 - 141
Atrazine	94.6		P	73 - 118
Atrazine Desethyl	109		P	32 - 125
Avobenzone	170		P	40 - 203
Azinphos Methyl	193		P	36 - 197
Azoxystrobin	98.7		P	58 - 226
Bromacil	94.8		P	48 - 120
Butylate	57.6		P	27 - 85.0
Caffeine	104		P	40 - 137
Carbophenothion	110		P	54 - 132
Carfentrazone Ethyl	103		P	60 - 142
Chlorfenapyr	71.6		P	53 - 117
Chlorpyrifos Ethyl	73.2		P	59 - 116

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8270E

Batch ID: P429507

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chlorpyrifos Methyl	82.0		P	66 - 119
Coumaphos	99.4		P	11 - 234
Cyanazine	210		P	61 - 248
Cyprodinil	101		P	50 - 147
Dechlorane 602	94.0		P	50 - 150
Dechlorane 603	138		P	50 - 150
Dechlorane Plus	95.2		P	47 - 182
Demeton	81.6		P	30 - 138
Diazinon	111		P	67 - 126
Difenoconazole	86.3		P	38 - 205
Dimethenamid	103		P	57 - 111
Dimethomorph	152		P	16 - 212
Disulfoton	98.2		P	46 - 126
Dithiopyr	106		P	62 - 115
EPTC	46.5		P	28 - 80.0
Ethion	81.4		P	24 - 122
Ethoprop	82.2		P	62 - 113
Etridiazole	60.6		P	28 - 92.0
Fenamiphos	68.3		P	57 - 128
Fenbuconazole	67.7		P	52 - 155
Fipronil	76.6		P	63 - 130
Fipronil Desulfinyl	87.3		P	61 - 130
Fipronil Sulfide	91.0		P	56 - 117
Fipronil Sulfone	82.4		P	52 - 118
Fluazifop-P-butyl	104		P	61 - 128
Fludioxonil	82.5		P	59 - 129
Flumioxazin	137		P	30 - 250
Flutolanil	102		P	53 - 127
Fluxapyroxad	81.5		P	42 - 132
Fonofos	109		P	61 - 110
Hexabromobenzene	128		P	50 - 150
Hexazinone	87.3		P	46 - 139
Imazalil	86.0		P	40 - 139
Iprodione	31.2		F	40 - 146
Loratadine	168		P	10 - 224
Malathion	80.7		P	61 - 135
Metaxyl	91.3		P	58 - 121
Metolachlor	77.2		P	64 - 118
Metribuzin	109		P	45 - 245
Mevinphos	63.1		P	49 - 118
Molinate	58.1		P	42 - 92.0
Myclobutanil	99.8		P	55 - 127
Naled	35.0		P	10 - 114
Napropamide	82.8		P	39 - 121
Norflurazon	96.9		P	55 - 129
Octinoxate	166		P	26 - 166
Oxadiazon	88.9		P	54 - 115
Oxybenzone	118		P	49 - 135
Oxyfluorfen	79.7		P	64 - 139
Parathion Ethyl	90.3		P	70 - 111
Parathion Methyl	89.5		P	76 - 136
PBB-153	145		P	50 - 150

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8270E

Batch ID: P429507

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
PBDE-47	116		P	41 - 148
PBDE-99	114		P	38 - 147
Pendimethalin	85.9		P	52 - 118
Pentabromotoluene	126		P	50 - 150
Phenytoin	136		P	10 - 162
Phorate	99.9		P	40 - 122
Prodiamine	55.9		P	26 - 141
Prometon	105		P	54 - 122
Prometryn	124		P	61 - 128
Pronamide	110		P	72 - 121
Propanil	106		P	45 - 144
Propargite	66.0		P	10 - 199
Propiconazole	75.9		P	56 - 127
Pyraflufen Ethyl	98.0		P	56 - 140
Pyridaben	77.6		P	26 - 211
Pyriproxyfen	221		F	33 - 194
Simazine	82.6		P	67 - 121
Stirophos	58.9		P	20 - 142
Sulfentrazone	62.2		P	21 - 144
Tebuconazole	68.2		P	10 - 122
Terbufos	119		P	60 - 129
Terbutylazine	85.3		P	69 - 113
Tetradecachloro-m-terphenyl	233		F	70 - 200
Thiobencarb	105		P	51 - 120
Tri-o-cresyl Phosphate	118		P	49 - 150
Triadimefon	94.9		P	56 - 114
Tris(1-chloro-2-propyl) phosphate (TCPP)	98.5		P	50 - 150
Tris(1,3-dichloro-2-propyl) phosphate (TDCPP)	113		P	50 - 150
Tris(2-chloroethyl) phosphate (TCEP)	119		P	70 - 180

Reference Method: EPA 8321B

Batch ID: P429323

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Acesulfame K	144		P	40 - 150
AMPA	97.2		P	40 - 150
Endothall	77.8		P	40 - 150
Glufosinate	79.1		P	40 - 150
Glyphosate	96.7		P	40 - 150
Sucralose	118		P	40 - 150

Reference Method: EPA 8321B

Batch ID: P429412

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2,4-D	94.0		P	40 - 150
2,4,5-T	91.9		P	40 - 150
Acetaminophen	74.1		P	30 - 150
Acetamiprid	82.4		P	40 - 150
Afidopyropen	68.1		P	30 - 150
Bentazon	139		P	30 - 150
Benzovindiflupyr	79.3		P	40 - 150

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8321B

Batch ID: P429412

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Carbamazepine	72.2		P	40 - 150
Clothianidin	84.7		P	40 - 150
Dinotefuran	94.6		P	40 - 150
Diuron	70.0		P	40 - 150
Fenuron	81.9		P	40 - 150
Fluridone	73.3		P	40 - 150
Hydrocodone	78.2		P	40 - 150
Ibuprofen	73.6		P	40 - 150
Imazapyr	88.6		P	40 - 150
Imidacloprid	83.3		P	40 - 150
Linuron	68.3		P	40 - 150
Mandestrobin	82.6		P	40 - 150
MCPP	98.2		P	40 - 150
Naproxen	75.7		P	40 - 150
Primidone	84.1		P	40 - 150
Pyraclostrobin	77.6		P	40 - 150
Silvex	87.8		P	40 - 150
Thiamethoxam	79.7		P	40 - 150
Tolfenpyrad	51.9		P	30 - 150
Triclopyr	91.7		P	40 - 150

Reference Method: SM 2320 B-2011

Batch ID: P429832

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	98.6		P	94 - 106

Reference Method: SM 2540 C-2015

Batch ID: P429834

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
TDS	96.5		P	90 - 110

Reference Method: SM 2540 D-2015

Batch ID: P429764

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
TSS	95.3		P	85 - 106

Reference Method: SM 4500-F- C-2011

Batch ID: P429697

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.5		P	96 - 104

Reference Method: SM 5310 B-2011

Batch ID: P429705

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	100		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P429542

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405704	Calcium	99.2		P	80 - 120
2405704	Iron	105	103	P/P	80 - 120
2405704	Magnesium	104	102	P/P	80 - 120
2405704	Potassium	104	103	P/P	80 - 120
2405704	Sodium	98.9		P	80 - 120
2405704	Strontium	104	103	P/P	80 - 120
2405704	Tin	101	99.4	P/P	80 - 120
2405704	Titanium	101	102	P/P	80 - 120
2405704	Vanadium	101	102	P/P	80 - 120
2405704	Zinc	104	103	P/P	80 - 120
2406025	Calcium	99.2		P	80 - 120
2406025	Iron	102		P	80 - 120
2406025	Magnesium	100		P	80 - 120
2406025	Potassium	100		P	80 - 120
2406025	Sodium	98.9		P	80 - 120
2406025	Strontium	101		P	80 - 120
2406025	Tin	97.8		P	80 - 120
2406025	Titanium	101		P	80 - 120
2406025	Vanadium	100		P	80 - 120
2406025	Zinc	98.7		P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P429542

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405704	Aluminum	98.9	100	P/P	80 - 120
2405704	Antimony	107	107	P/P	80 - 120
2405704	Arsenic	102	101	P/P	80 - 120
2405704	Barium	103	104	P/P	80 - 120
2405704	Beryllium	99.1	98.6	P/P	80 - 120
2405704	Boron	100	102	P/P	80 - 120
2405704	Cadmium	103	103	P/P	80 - 120
2405704	Chromium	97.5	97.4	P/P	80 - 120
2405704	Cobalt	101	101	P/P	80 - 120
2405704	Copper	96.8	96.6	P/P	80 - 120
2405704	Lead	104	104	P/P	80 - 120
2405704	Manganese	97.9	97.1	P/P	80 - 120
2405704	Molybdenum	103	101	P/P	80 - 120
2405704	Nickel	97.5	96.8	P/P	80 - 120
2405704	Selenium	102	101	P/P	80 - 120
2405704	Silver	103	102	P/P	80 - 120
2405704	Thallium	103	103	P/P	80 - 120
2406025	Aluminum	98.8		P	80 - 120
2406025	Antimony	107		P	80 - 120
2406025	Arsenic	100		P	80 - 120
2406025	Barium	103		P	80 - 120
2406025	Beryllium	96.8		P	80 - 120
2406025	Boron	103		P	80 - 120
2406025	Cadmium	101		P	80 - 120
2406025	Chromium	96.8		P	80 - 120
2406025	Cobalt	101		P	80 - 120
2406025	Copper	96.2		P	80 - 120

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.8 Rev. 5.4

Batch ID: P429542

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2406025	Lead	105		P	80 - 120
2406025	Manganese	98.0		P	80 - 120
2406025	Molybdenum	100		P	80 - 120
2406025	Nickel	96.5		P	80 - 120
2406025	Selenium	101		P	80 - 120
2406025	Silver	104		P	80 - 120
2406025	Thallium	104		P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P429662

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405693	Chloride	91.5		P	90 - 110
2406008	Chloride	98.0		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P429664

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405693	Sulfate	94.1		P	90 - 110
2406008	Sulfate	95.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P429609

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405647	Ammonia-N	97.8	99.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P429573

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405920	Kjeldahl Nitrogen	96.3	97.7	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P429812

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405057	NO2NO3-N	103	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P429536

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405694	Total-P	101	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P429425

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405438	O-Phosphate-P	93.0	93.7	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270E

Batch ID: P429507

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405052	2-Ethylhexyl 2,3,4,5-tetrabromobenzoate (EHTBB)	110	119	P/P	26 - 165
2405052	Acetochlor	70.2	72.8	P/P	67 - 124
2405052	Alachlor	82.7	81.1	P/P	60 - 120
2405052	Ametryn	136	137	P/P	54 - 216
2405052	Atrazine	86.2	103	P/P	66 - 128
2405052	Atrazine Desethyl	109	99.7	P/P	15 - 122
2405052	Avobenzene	149	163	P/P	37 - 178
2405052	Azinphos Methyl	162	163	P/P	40 - 292
2405052	Azoxystrobin	106	122	P/P	35 - 220
2405052	Butylate	57.3	59.0	P/P	20 - 83.0
2405052	Caffeine	99.5	95.8	P/P	10 - 194
2405052	Carbophenothion	92.7	98.7	P/P	57 - 127
2405052	Carfentrazone Ethyl	92.7	94.9	P/P	51 - 141
2405052	Chlorfenapyr	62.5	67.9	P/P	46 - 111
2405052	Chlorpyrifos Ethyl	61.8	60.7	P/P	52 - 120
2405052	Chlorpyrifos Methyl	81.9	89.1	P/P	63 - 119
2405052	Coumaphos	80.9	84.9	P/P	10 - 228
2405052	Cyanazine	234	235	P/P	59 - 241
2405052	Cyprodinil	97.6	107	P/P	47 - 146
2405052	Dechlorane 602	79.1	82.3	P/P	50 - 150
2405052	Dechlorane 603	118	128	P/P	50 - 150
2405052	Dechlorane Plus	82.0	81.2	P/P	50 - 150
2405052	Demeton	83.4	86.4	P/P	40 - 136
2405052	Diazinon	112	113	P/P	69 - 132
2405052	Difenoconazole	68.4	75.9	P/P	57 - 258
2405052	Dimethenamid	91.6	95.9	P/P	54 - 110
2405052	Dimethomorph	166	171	P/P	28 - 210
2405052	Disulfoton	96.7	93.9	P/P	43 - 133
2405052	Dithiopyr	85.4	90.2	P/P	39 - 116
2405052	EPTC	44.9	46.4	P/P	20 - 78.0
2405052	Ethion	92.1	102	P/P	17 - 119
2405052	Ethoprop	89.9	96.6	P/P	66 - 120
2405052	Etridiazole	56.8	57.2	P/P	28 - 96.0
2405052	Fenamiphos	69.8	69.4	P/P	41 - 126
2405052	Fenbuconazole	54.4	57.0	P/P	42 - 169
2405052	Fipronil	65.7	68.9	P/P	61 - 132
2405052	Fipronil Desulfinyl	78.4	77.4	P/P	56 - 132
2405052	Fipronil Sulfide	71.8	69.9	P/P	48 - 119
2405052	Fipronil Sulfone	72.8	76.4	P/P	42 - 131
2405052	Fluazifop-P-butyl	103	105	P/P	53 - 131
2405052	Fludioxonil	80.7	79.2	P/P	56 - 127
2405052	Flumioxazin	116	127	P/P	19 - 300
2405052	Flutolanil	93.8	102	P/P	41 - 127
2405052	Fluxapyroxad	71.0	74.8	P/P	42 - 172
2405052	Fonofos	110	110	P/P	59 - 113
2405052	Hexabromobenzene	120	121	P/P	50 - 150
2405052	Hexazinone	94.9	92.6	P/P	66 - 122
2405052	Imazalil	59.5	58.2	P/P	21 - 283
2405052	Iprodione	34.0	42.2	F/F	43 - 150
2405052	Loratadine	148	146	P/P	23 - 201
2405052	Malathion	75.2	77.7	P/P	58 - 136
2405052	Metalaxyl	80.0	73.5	P/P	55 - 120

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270E

Batch ID: P429507

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405052	Metolachlor	76.8	76.4	P/P	57 - 121
2405052	Metribuzin	107	108	P/P	58 - 294
2405052	Mevinphos	67.3	71.4	P/P	50 - 126
2405052	Molinate	59.9	61.6	P/P	42 - 93.0
2405052	Myclobutanil	96.0	95.7	P/P	55 - 125
2405052	Naled	41.0	50.6	P/P	18 - 200
2405052	Napropamide	72.7	79.1	P/P	39 - 110
2405052	Octinoxate	161	177	F/F	21 - 159
2405052	Oxadiazon	83.4	82.8	P/P	43 - 112
2405052	Oxybenzone	107	111	P/P	41 - 142
2405052	Oxyfluorfen	89.0	84.4	P/P	64 - 153
2405052	Parathion Ethyl	93.6	90.8	P/P	69 - 114
2405052	Parathion Methyl	90.8	101	P/P	79 - 139
2405052	PBB-153	126	140	P/P	50 - 150
2405052	PBDE-47	103	110	P/P	27 - 144
2405052	PBDE-99	95.1	101	P/P	18 - 150
2405052	Pendimethalin	85.9	95.6	P/P	50 - 139
2405052	Pentabromotoluene	116	124	P/P	50 - 150
2405052	Phenytoin	156	162	F/F	10 - 146
2405052	Phorate	107	121	P/P	54 - 161
2405052	Prodiamine	61.2	61.7	P/P	30 - 161
2405052	Prometon	94.9	96.1	P/P	45 - 134
2405052	Prometryn	112	110	P/P	45 - 137
2405052	Pronamide	101	107	P/P	65 - 133
2405052	Propanil	114	112	P/P	49 - 142
2405052	Propargite	59.8	60.3	P/P	10 - 203
2405052	Propiconazole	70.7	73.3	P/P	38 - 146
2405052	Pyraflufen Ethyl	83.6	77.2	P/P	34 - 153
2405052	Pyridaben	47.7	55.5	P/P	12 - 192
2405052	Pyriproxyfen	128	171	P/P	33 - 180
2405052	Simazine	56.6	66.9	P/P	51 - 157
2405052	Stiropfos	40.1	42.0	P/P	10 - 128
2405052	Sulfentrazone	73.4	59.3	P/P	53 - 186
2405052	Tebuconazole	63.7	63.2	P/P	10 - 133
2405052	Terbufos	115	120	P/P	65 - 139
2405052	Terbutylazine	81.3	85.4	P/P	66 - 117
2405052	Tetradecachloro-m-terphenyl	227	251	F/F	70 - 200
2405052	Thiobencarb	92.0	95.8	P/P	51 - 119
2405052	Tri-o-cresyl Phosphate	102	109	P/P	31 - 144
2405052	Triadimefon	86.2	90.0	P/P	48 - 119
2405052	Tris(1-chloro-2-propyl) phosphate (TCPP)	100	102	P/P	50 - 150
2405052	Tris(1,3-dichloro-2-propyl) phosphate (TDCPP)	103	109	P/P	50 - 150
2405052	Tris(2-chloroethyl) phosphate (TCEP)	106	108	P/P	70 - 180

Reference Method: EPA 8321B

Batch ID: P429323

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405697	Acesulfame K	133	135	P/P	40 - 150
2405697	AMPA	91.2	90.8	P/P	40 - 150
2405697	Endothall	77.2	76.0	P/P	40 - 150
2405697	Glufosinate	67.6	69.6	P/P	40 - 150

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8321B
Batch ID: P429323

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405697	Glyphosate	82.8	91.7	P/P	40 - 150
2405697	Sucralose	109	136	P/P	40 - 150

Reference Method: EPA 8321B
Batch ID: P429412

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405330	2,4-D	113	110	P/P	40 - 150
2405330	2,4,5-T	115	109	P/P	40 - 150
2405330	Acetaminophen	85.7	84.1	P/P	30 - 150
2405330	Acetamiprid	98.6	96.6	P/P	40 - 150
2405330	Afidopyropen	99.1	95.3	P/P	30 - 150
2405330	Bentazon	164	174	F/F	30 - 150
2405330	Benzovindiflupyr	81.9	87.3	P/P	40 - 150
2405330	Carbamazepine	82.6	84.1	P/P	40 - 150
2405330	Clothianidin	99.7	99.2	P/P	40 - 150
2405330	Dinotefuran	112	116	P/P	40 - 150
2405330	Diuron	72.5	73.5	P/P	40 - 150
2405330	Fenuron	77.3	84.5	P/P	40 - 150
2405330	Fluridone	81.8	96.1	P/P	40 - 150
2405330	Hydrocodone	80.7	85.4	P/P	40 - 150
2405330	Ibuprofen	86.5	78.2	P/P	40 - 150
2405330	Imazapyr	143	114	P/P	40 - 150
2405330	Imidacloprid	104	102	P/P	40 - 150
2405330	Linuron	74.5	74.5	P/P	40 - 150
2405330	Mandestrobin	94.1	90.4	P/P	40 - 150
2405330	MCPP	126	110	P/P	40 - 150
2405330	Naproxen	93.3	73.3	P/P	40 - 150
2405330	Primidone	98.8	95.8	P/P	40 - 150
2405330	Pyraclostrobin	85.1	85.7	P/P	40 - 150
2405330	Silvex	112	117	P/P	40 - 150
2405330	Thiamethoxam	98.5	95.4	P/P	40 - 150
2405330	Tolfenpyrad	51.2	49.7	P/P	30 - 150
2405330	Triclopyr	105	103	P/P	40 - 150

Reference Method: SM 4500-F- C-2011
Batch ID: P429697

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405628	Fluoride	102	101	P/P	94 - 104

Reference Method: SM 5310 B-2011
Batch ID: P429705

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2405454	Organic Carbon	97.0	98.8	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: DEP SOP: NU-094-1
Batch ID: P429422

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405633	Color (true)	0.462	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P429436

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405646	Turbidity	10.8	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P429542

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405704	Calcium	1.17	Spike	P	0 - 20
2405704	Iron	1.58	Spike	P	0 - 20
2405704	Magnesium	1.48	Spike	P	0 - 20
2405704	Potassium	0.505	Spike	P	0 - 20
2405704	Sodium	0.341	Spike	P	0 - 20
2405704	Strontium	1.30	Spike	P	0 - 20
2405704	Tin	1.73	Spike	P	0 - 20
2405704	Titanium	0.741	Spike	P	0 - 20
2405704	Vanadium	0.432	Spike	P	0 - 20
2405704	Zinc	0.175	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P429542

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405704	Aluminum	1.15	Spike	P	0 - 20
2405704	Antimony	0.0221	Spike	P	0 - 20
2405704	Arsenic	0.678	Spike	P	0 - 20
2405704	Barium	0.369	Spike	P	0 - 20
2405704	Beryllium	0.549	Spike	P	0 - 20
2405704	Boron	0.948	Spike	P	0 - 20
2405704	Cadmium	0.285	Spike	P	0 - 20
2405704	Chromium	0.202	Spike	P	0 - 20
2405704	Cobalt	0.0975	Spike	P	0 - 20
2405704	Copper	0.165	Spike	P	0 - 20
2405704	Lead	0.260	Spike	P	0 - 20
2405704	Manganese	0.716	Spike	P	0 - 20
2405704	Molybdenum	1.46	Spike	P	0 - 20
2405704	Nickel	0.800	Spike	P	0 - 20
2405704	Selenium	1.12	Spike	P	0 - 20
2405704	Silver	0.598	Spike	P	0 - 20
2405704	Thallium	0.226	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P429662

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405693	Chloride	0.314	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P429664

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405693	Sulfate	1.07	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P429609

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405647	Ammonia-N	1.19	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P429573

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405920	Kjeldahl Nitrogen	1.19	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P429812

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405057	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P429536

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405694	Total-P	0.609	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P429425

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405438	O-Phosphate-P	0.750	Spike	P	0 - 20

Reference Method: EPA 8270E
Batch ID: P429507

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405052	2-Ethylhexyl 2,3,4,5-tetrabromobenzoate (EHTBB)	7.35	Spike	P	0 - 37
2405052	Acetochlor	3.60	Spike	P	0 - 28

Quality Assurance Report Precision

Reference Method: EPA 8270E
Batch ID: P429507

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405052	Alachlor	1.98	Spike	P	0 - 30
2405052	Ametryn	0.661	Spike	P	0 - 32
2405052	Atrazine	8.90	Spike	P	0 - 41
2405052	Atrazine Desethyl	7.87	Spike	P	0 - 36
2405052	Avobenzone	9.11	Spike	P	0 - 36
2405052	Azinphos Methyl	0.0844	Spike	P	0 - 75
2405052	Azoxystrobin	13.2	Spike	P	0 - 39
2405052	Bromacil	0.648	Spike	P	0 - 33
2405052	Butylate	2.81	Spike	P	0 - 44
2405052	Caffeine	3.26	Spike	P	0 - 74
2405052	Carbophenothion	6.29	Spike	P	0 - 25
2405052	Carfentrazone Ethyl	2.31	Spike	P	0 - 27
2405052	Chlorfenapyr	8.39	Spike	P	0 - 24
2405052	Chlorpyrifos Ethyl	1.69	Spike	P	0 - 26
2405052	Chlorpyrifos Methyl	8.35	Spike	P	0 - 26
2405052	Coumaphos	4.82	Spike	P	0 - 28
2405052	Cyanazine	0.632	Spike	P	0 - 48
2405052	Cyprodinil	9.56	Spike	P	0 - 29
2405052	Dechlorane 602	4.00	Spike	P	0 - 30
2405052	Dechlorane 603	8.10	Spike	P	0 - 30
2405052	Dechlorane Plus	0.976	Spike	P	0 - 30
2405052	Demeton	3.54	Spike	P	0 - 33
2405052	Diazinon	0.719	Spike	P	0 - 29
2405052	Difenoconazole	10.4	Spike	P	0 - 34
2405052	Dimethenamid	4.57	Spike	P	0 - 39
2405052	Dimethomorph	2.72	Spike	P	0 - 51
2405052	Disulfoton	2.96	Spike	P	0 - 30
2405052	Dithiopyr	5.39	Spike	P	0 - 26
2405052	EPTC	3.37	Spike	P	0 - 43
2405052	Ethion	10.7	Spike	P	0 - 79
2405052	Ethoprop	7.12	Spike	P	0 - 29
2405052	Etridiazole	0.606	Spike	P	0 - 33
2405052	Fenamiphos	0.499	Spike	P	0 - 40
2405052	Fenbuconazole	4.16	Spike	P	0 - 74
2405052	Fipronil	4.82	Spike	P	0 - 29
2405052	Fipronil Desulfinyl	1.15	Spike	P	0 - 32
2405052	Fipronil Sulfide	2.73	Spike	P	0 - 30
2405052	Fipronil Sulfone	4.79	Spike	P	0 - 33
2405052	Fluazifop-P-butyl	2.08	Spike	P	0 - 38
2405052	Fludioxonil	1.84	Spike	P	0 - 29
2405052	Flumioxazin	9.05	Spike	P	0 - 51
2405052	Flutolanil	8.36	Spike	P	0 - 25
2405052	Fluxapyroxad	5.17	Spike	P	0 - 45
2405052	Fonofos	0.467	Spike	P	0 - 27
2405052	Hexabromobenzene	0.424	Spike	P	0 - 30
2405052	Hexazinone	1.67	Spike	P	0 - 30
2405052	Imazalil	2.21	Spike	P	0 - 100
2405052	Iprodione	21.6	Spike	P	0 - 33
2405052	Loratadine	1.03	Spike	P	0 - 56
2405052	Malathion	3.26	Spike	P	0 - 37
2405052	Metalaxyl	1.94	Spike	P	0 - 40

Quality Assurance Report Precision

Reference Method: EPA 8270E

Batch ID: P429507

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405052	Metolachlor	0.492	Spike	P	0 - 29
2405052	Metribuzin	0.842	Spike	P	0 - 45
2405052	Mevinphos	5.89	Spike	P	0 - 29
2405052	Molinate	2.79	Spike	P	0 - 33
2405052	Myclobutanil	0.312	Spike	P	0 - 26
2405052	Naled	21.0	Spike	P	0 - 37
2405052	Napropamide	8.36	Spike	P	0 - 44
2405052	Norflurazon	0.371	Spike	P	0 - 30
2405052	Octinoxate	9.10	Spike	P	0 - 45
2405052	Oxadiazon	0.712	Spike	P	0 - 28
2405052	Oxybenzone	3.89	Spike	P	0 - 46
2405052	Oxyfluorfen	5.35	Spike	P	0 - 28
2405052	Parathion Ethyl	3.04	Spike	P	0 - 31
2405052	Parathion Methyl	10.4	Spike	P	0 - 29
2405052	PBB-153	10.1	Spike	P	0 - 30
2405052	PBDE-47	6.55	Spike	P	0 - 36
2405052	PBDE-99	6.06	Spike	P	0 - 40
2405052	Pendimethalin	10.7	Spike	P	0 - 33
2405052	Pentabromotoluene	7.15	Spike	P	0 - 30
2405052	Phenytoin	3.48	Spike	P	0 - 100
2405052	Phorate	11.7	Spike	P	0 - 36
2405052	Prodiamine	0.880	Spike	P	0 - 71
2405052	Prometon	1.35	Spike	P	0 - 36
2405052	Prometryn	1.38	Spike	P	0 - 28
2405052	Pronamide	5.59	Spike	P	0 - 24
2405052	Propanil	2.46	Spike	P	0 - 28
2405052	Propargite	0.974	Spike	P	0 - 43
2405052	Propiconazole	3.50	Spike	P	0 - 33
2405052	Pyraflufen Ethyl	7.95	Spike	P	0 - 29
2405052	Pyridaben	8.96	Spike	P	0 - 30
2405052	Pyriproxyfen	29.1	Spike	P	0 - 79
2405052	Simazine	5.80	Spike	P	0 - 29
2405052	Stirophos	4.53	Spike	P	0 - 61
2405052	Sulfentrazone	21.3	Spike	P	0 - 33
2405052	Tebuconazole	0.696	Spike	P	0 - 69
2405052	Terbufos	4.50	Spike	P	0 - 29
2405052	Terbuthylazine	4.89	Spike	P	0 - 32
2405052	Tetradecachloro-m-terphenyl	10.0	Spike	P	0 - 30
2405052	Thiobencarb	4.02	Spike	P	0 - 26
2405052	Tri-o-cresyl Phosphate	7.00	Spike	P	0 - 33
2405052	Triadimefon	4.30	Spike	P	0 - 28
2405052	Tris(1-chloro-2-propyl) phosphate (TCPP)	1.59	Spike	P	0 - 30
2405052	Tris(1,3-dichloro-2-propyl) phosphate (TDCPP)	6.05	Spike	P	0 - 30
2405052	Tris(2-chloroethyl) phosphate (TCEP)	2.22	Spike	P	0 - 30

Reference Method: EPA 8321B

Batch ID: P429323

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405697	Acesulfame K	1.22	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8321B

Batch ID: P429323

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405697	AMPA	0.419	Spike	P	0 - 30
2405697	Endothall	1.55	Spike	P	0 - 30
2405697	Glufosinate	3.03	Spike	P	0 - 30
2405697	Glyphosate	10.1	Spike	P	0 - 30
2405697	Sucralose	19.1	Spike	P	0 - 30

Reference Method: EPA 8321B

Batch ID: P429412

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405330	2,4-D	2.74	Spike	P	0 - 30
2405330	2,4,5-T	5.85	Spike	P	0 - 30
2405330	Acetaminophen	1.85	Spike	P	0 - 30
2405330	Acetamiprid	2.02	Spike	P	0 - 30
2405330	Afidopyropen	3.85	Spike	P	0 - 30
2405330	Bentazon	5.03	Spike	P	0 - 30
2405330	Benzovindiflupyr	6.39	Spike	P	0 - 30
2405330	Carbamazepine	1.79	Spike	P	0 - 30
2405330	Clothianidin	0.475	Spike	P	0 - 30
2405330	Dinotefuran	3.54	Spike	P	0 - 30
2405330	Diuron	1.32	Spike	P	0 - 30
2405330	Fenuron	8.83	Spike	P	0 - 30
2405330	Fluridone	16.1	Spike	P	0 - 30
2405330	Hydrocodone	5.60	Spike	P	0 - 30
2405330	Ibuprofen	9.98	Spike	P	0 - 30
2405330	Imazapyr	20.0	Spike	P	0 - 30
2405330	Imidacloprid	2.17	Spike	P	0 - 30
2405330	Linuron	0.0137	Spike	P	0 - 30
2405330	Mandestrobin	3.98	Spike	P	0 - 30
2405330	MCP	13.5	Spike	P	0 - 30
2405330	Naproxen	23.9	Spike	P	0 - 30
2405330	Primidone	3.10	Spike	P	0 - 30
2405330	Pyraclostrobin	0.684	Spike	P	0 - 30
2405330	Silvex	3.93	Spike	P	0 - 30
2405330	Thiamethoxam	3.27	Spike	P	0 - 30
2405330	Tolfenpyrad	2.86	Spike	P	0 - 30
2405330	Triclopyr	2.18	Spike	P	0 - 30

Reference Method: SM 2320 B-2011

Batch ID: P429832

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2406052	Total Alkalinity	0.346	Sample	P	0 - 3.8

Reference Method: SM 2540 C-2015

Batch ID: P429834

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
--------------------------	-----------	-----------	-------------------	-----------	----------------

Quality Assurance Report Precision

Reference Method: SM 2540 C-2015
Batch ID: P429834

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2404876	TDS	2.69	Sample	P	0 - 10

Reference Method: SM 4500-F- C-2011
Batch ID: P429697

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405628	Fluoride	0.480	Spike	P	0 - 1.6

Reference Method: SM 5310 B-2011
Batch ID: P429705

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2405454	Organic Carbon	1.44	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Surrogates

Lab Sample ID: 2405697
Field Sample ID: G1SW0083

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	2,4-D-d3	113	P	30 - 160
EPA 8321B	Acetaminophen-d4	92.2	P	30 - 160
EPA 8321B	Carbamazepine-d10	70.1	P	30 - 160
EPA 8321B	Diuron-d6	61.6	P	30 - 160
EPA 8321B	Endothall-d6	97.2	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	88.4	P	30 - 160
EPA 8321B	Ibuprofen-d3	78.1	P	30 - 160
EPA 8321B	Primidone-d5	84.1	P	30 - 160
EPA 8321B	Sucralose-d6	110	P	30 - 160

Lab Sample ID: 2405711
Field Sample ID: G1SW0083

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270E	Caffeine-d9	96.8	P	20 - 200
EPA 8270E	Simazine-d10	101	P	62 - 142
EPA 8270E	Terbufos-d10	90.1	P	58 - 132
EPA 8270E	Terphenyl-d14	86.4	P	52 - 137

Quality Assurance Report Calibration Verification

Reference Method: DEP SOP: NU-094-1

Run ID: A118955

Included Lab Sample IDs: 2405693

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	101	99.5	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A118956

Included Lab Sample IDs: 2405695

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	97.9	99.3	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A118961

Included Lab Sample IDs: 2405693

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	99.5	P/P	95 - 105

Reference Method: EPA 8321B

Run ID: A118997

Included Lab Sample IDs: 2405697

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acesulfame K	153	155	P/P	60 - 160
Sucralose	114	103	P/P	60 - 160

Reference Method: EPA 8321B

Run ID: A118999

Included Lab Sample IDs: 2405697

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
2,4-D	92.2	84.6	P/P	50 - 150
2,4,5-T	85.6	94.0	P/P	50 - 150
Acetaminophen	86.4	86.6	P/P	60 - 160
Acetamiprid	106	105	P/P	50 - 150
Afidopyropen	96.4	108	P/P	50 - 150
Bentazon	141	118	P/P	50 - 160
Benzovindiflupyr	106	95.9	P/P	50 - 150
Carbamazepine	97.7	108	P/P	60 - 150
Clothianidin	106	99.3	P/P	60 - 150
Dinotefuran	113	113	P/P	50 - 150
Diuron	102	107	P/P	60 - 160
Fenuron	101	99.0	P/P	60 - 150
Fluridone	118	106	P/P	60 - 160
Hydrocodone	104	109	P/P	60 - 150
Ibuprofen	86.9	72.3	P/P	50 - 160
Imazapyr	91.0	102	P/P	50 - 150
Imidacloprid	96.9	95.2	P/P	60 - 150
Linuron	110	89.1	P/P	60 - 150
Mandestrobin	111	101	P/P	50 - 150
MCPP	91.4	89.1	P/P	50 - 150
Naproxen	120	98.8	P/P	50 - 160
Primidone	101	102	P/P	60 - 150

Quality Assurance Report Calibration Verification

Reference Method: EPA 8321B
Run ID: A118999
Included Lab Sample IDs: 2405697

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Pyraclostrobin	101	103	P/P	60 - 150
Silvex	101	87.6	P/P	50 - 150
Thiamethoxam	92.3	93.5	P/P	50 - 150
Tolfenpyrad	95.3	93.6	P/P	60 - 150
Triclopyr	87.5	85.2	P/P	50 - 150

Reference Method: EPA 8321B
Run ID: A119009
Included Lab Sample IDs: 2405697

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
AMPA	94.2	97.1	P/P	60 - 160
Endothall	81.9	79.9	P/P	60 - 160
Glufosinate	83.1	79.4	P/P	60 - 160
Glyphosate	97.2	97.4	P/P	60 - 160

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A119045
Included Lab Sample IDs: 2405694

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.1	98.5	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A119048
Included Lab Sample IDs: 2405694

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	98.5	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4
Run ID: A119053
Included Lab Sample IDs: 2405704

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aluminum	101	101	P/P	90 - 110
Antimony	106	105	P/P	90 - 110
Arsenic	101	101	P/P	90 - 110
Barium	101	100	P/P	90 - 110
Beryllium	97.1	98.3	P/P	90 - 110
Boron	97.2	98.3	P/P	90 - 110
Cadmium	101	102	P/P	90 - 110
Chromium	97.9	99.6	P/P	90 - 110
Cobalt	102	103	P/P	90 - 110
Copper	96.9	96.1	P/P	90 - 110
Lead	103	103	P/P	90 - 110
Manganese	97.7	101	P/P	90 - 110
Molybdenum	99.9	100	P/P	90 - 110
Nickel	98.3	98.0	P/P	90 - 110
Selenium	102	103	P/P	90 - 110
Silver	103	104	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A119053

Included Lab Sample IDs: 2405704

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Thallium	101	102	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A119070

Included Lab Sample IDs: 2405693

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	99.9	97.6	P/P	90 - 110
Sulfate	98.6	96.6	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A119075

Included Lab Sample IDs: 2405694

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	101	95.8	P/P	90 - 110

Reference Method: SM 4500-F- C-2011

Run ID: A119080

Included Lab Sample IDs: 2405693

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	102	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A119082

Included Lab Sample IDs: 2405694

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	99.2	98.2	P/P	90 - 110

Reference Method: EPA 8270E

Run ID: A119083

Included Lab Sample IDs: 2405711

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
2-Ethylhexyl 2,3,4,5-tetrabromobenzoate (EHTBB)	101		P	80 - 120
Avobenzene	105		P	80 - 120
Dechlorane 602	103		P	80 - 120
Dechlorane 603	105		P	80 - 120
Dechlorane Plus	89.6		P	80 - 120
Hexabromobenzene	102		P	80 - 120
Octinoxate	97.5		P	80 - 120
Oxybenzone	95.5		P	80 - 120
PBB-153	103		P	80 - 120
PBDE-47	96.9		P	80 - 120
PBDE-99	104		P	80 - 120
Pentabromotoluene	102		P	80 - 120
Tetradecachloro-m-terphenyl	105		P	80 - 120
Tri-o-cresyl Phosphate	97.6		P	80 - 120
Tris(1-chloro-2-propyl) phosphate (TCPP)	81.1		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270E
Run ID: A119083
Included Lab Sample IDs: 2405711

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Tris(1,3-dichloro-2-propyl) phosphate (TDCPP)	96.9		P	80 - 120
Tris(2-chloroethyl) phosphate (TCEP)	97.9		P	80 - 120

Reference Method: EPA 200.7 Rev. 4.4
Run ID: A119091
Included Lab Sample IDs: 2405704

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Calcium	98.9	99.2	P/P	90 - 110
Iron	101	102	P/P	90 - 110
Magnesium	100	102	P/P	90 - 110
Potassium	99.9	102	P/P	90 - 110
Sodium	99.1	99.5	P/P	90 - 110
Strontium	101	102	P/P	90 - 110
Tin	98.8	98.1	P/P	90 - 110
Titanium	99.2	99.5	P/P	90 - 110
Vanadium	98.9	98.7	P/P	90 - 110
Zinc	99.3	96.2	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A119129
Included Lab Sample IDs: 2405694

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	103	103	P/P	90 - 110

Reference Method: SM 2320 B-2011
Run ID: A119143
Included Lab Sample IDs: 2405693

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	97.2	99.4	P/P	90 - 110

Reference Method: EPA 8270E
Run ID: A119206
Included Lab Sample IDs: 2405711

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acetochlor	87.5		P	80 - 120
Alachlor	85.5		P	80 - 120
Ametryn	96.1		P	80 - 120
Atrazine Desethyl	87.0		P	80 - 120
Azinphos Methyl	111		P	80 - 120
Azoxystrobin	101		P	80 - 120
Bromacil	92.7		P	80 - 120
Butylate	98.2		P	80 - 120
Caffeine	102		P	80 - 120
Carbophenothion	98.5		P	80 - 120
Carfentrazone Ethyl	95.2		P	80 - 120
Chlorfenapyr	84.5		P	80 - 120
Chlorpyrifos Ethyl	86.1		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270E
Run ID: A119206
Included Lab Sample IDs: 2405711

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chlorpyrifos Methyl	102		P	80 - 120
Coumaphos	98.8		P	80 - 120
Cyanazine	103		P	80 - 120
Cyprodinil	99.7		P	80 - 120
Demeton	102		P	80 - 120
Diazinon	90.4		P	80 - 120
Difenoconazole	104		P	80 - 120
Dimethenamid	97.4		P	80 - 120
Dimethomorph	94.2		P	80 - 120
Disulfoton	94.2		P	80 - 120
Dithiopyr	97.2		P	80 - 120
EPTC	97.1		P	80 - 120
Ethion	91.9		P	80 - 120
Ethoprop	101		P	80 - 120
Etridiazole	94.2		P	80 - 120
Fenamiphos	90.2		P	80 - 120
Fenbuconazole	88.6		P	80 - 120
Fipronil	86.7		P	80 - 120
Fipronil Desulfinyl	102		P	80 - 120
Fipronil Sulfide	98.9		P	80 - 120
Fipronil Sulfone	99.9		P	80 - 120
Fluazifop-P-butyl	99.8		P	80 - 120
Fludioxonil	103		P	80 - 120
Flumioxazin	103		P	80 - 120
Flutolanil	105		P	80 - 120
Fluxapyroxad	98.5		P	80 - 120
Fonofos	94.7		P	80 - 120
Hexazinone	102		P	80 - 120
Imazalil	96.5		P	80 - 120
Iprodione	89.3		P	80 - 120
Loratadine	87.7		P	80 - 120
Malathion	89.0		P	80 - 120
Metaxyl	86.3		P	80 - 120
Metolachlor	102		P	80 - 120
Metribuzin	98.6		P	80 - 120
Mevinphos	103		P	80 - 120
Molinate	88.5		P	80 - 120
Myclobutanil	90.5		P	80 - 120
Naled	95.2		P	80 - 120
Napropamide	100		P	80 - 120
Norflurazon	95.7		P	80 - 120
Oxadiazon	97.0		P	80 - 120
Oxyfluorfen	103		P	80 - 120
Parathion Ethyl	98.0		P	80 - 120
Parathion Methyl	88.4		P	80 - 120
Pendimethalin	112		P	80 - 120
Phenytol	98.4		P	80 - 120
Phorate	98.8		P	80 - 120
Prodiamine	105		P	80 - 120
Prometon	99.9		P	80 - 120
Prometryn	101		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270E
Run ID: A119206
Included Lab Sample IDs: 2405711

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Pronamide	97.3		P	80 - 120
Propanil	99.9		P	80 - 120
Propargite	98.6		P	80 - 120
Propiconazole	87.3		P	80 - 120
Pyraflufen Ethyl	90.3		P	80 - 120
Pyridaben	95.3		P	80 - 120
Pyriproxyfen	98.3		P	80 - 120
Simazine	80.9		P	80 - 120
Stirophos	88.1		P	80 - 120
Sulfentrazone	97.1		P	80 - 120
Tebuconazole	97.6		P	80 - 120
Terbufos	96.2		P	80 - 120
Terbuthylazine	87.7		P	80 - 120
Thiobencarb	96.1		P	80 - 120
Triadimefon	99.9		P	80 - 120

Reference Method: EPA 8270E
Run ID: A119344
Included Lab Sample IDs: 2405711

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Atrazine	85.7		P	80 - 120

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			LCS	Precision SMP	MS
DEP SOP: NU-094-1	Color (true)	98.4					0.462	
EPA 180.1 Rev. 2.0	Turbidity	99.4					10.8	
EPA 200.7 Rev. 4.4	Calcium	100	99.2	99.2				1.17
	Iron	100	105	103	102			1.58
	Magnesium	101	104	102	100			1.48
	Potassium	100	104	103	100			0.505
	Sodium	99.5	98.9	98.9				0.341
	Strontium	102	104	103	101			1.30
	Tin	102	101	99.4	97.8			1.73
	Titanium	100	101	102	101			0.741
	Vanadium	99.7	101	102	100			0.432
	Zinc	102	104	103	98.7			0.175
EPA 200.8 Rev. 5.4	Aluminum	102	98.9	100	98.8			1.15
	Antimony	108	107	107	107			0.0221
	Arsenic	102	102	101	100			0.678
	Barium	105	103	104	103			0.369
	Beryllium	98.9	99.1	98.6	96.8			0.549
	Boron	103	100	102	103			0.948
	Cadmium	102	103	103	101			0.285
	Chromium	98.6	97.5	97.4	96.8			0.202
	Cobalt	103	101	101	101			0.0975
	Copper	98.5	96.8	96.6	96.2			0.165
	Lead	105	104	104	105			0.260
	Manganese	99.7	97.9	97.1	98.0			0.716
	Molybdenum	101	103	101	100			1.46
	Nickel	98.0	97.5	96.8	96.5			0.800
	Selenium	102	102	101	101			1.12
	Silver	106	103	102	104			0.598
	Thallium	101	103	103	104			0.226
EPA 300.0 Rev. 2.1	Chloride	98.7	91.5	98.0			0.314	
	Sulfate	97.3	94.1	95.3			1.07	
EPA 350.1 Rev. 2.0	Ammonia-N	97.0	97.8	99.4				1.19
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	92.4	96.3	97.7				1.19
EPA 353.2 Rev. 2.0	NO2NO3-N	104	103	103				0.0
EPA 365.1 Rev. 2.0	Total-P	101	101	102				0.609
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.2	93.0	93.7				0.750
EPA 8270E	2-Ethylhexyl 2,3,4,5-tetrabromobenzoate (EHTBB)	123	110	119				7.35
	Acetochlor	71.8	70.2	72.8				3.60
	Alachlor	100	82.7	81.1				1.98
	Ametryn	135	136	137				0.661
	Atrazine	94.6	86.2	103				8.90
	Atrazine Desethyl	109	109	99.7				7.87
	Avobenzone	170	149	163				9.11
	Azinphos Methyl	193	162	163				0.0844
	Azoxystrobin	98.7	106	122				13.2
	Bromacil	94.8						0.648
	Butylate	57.6	57.3	59.0				2.81
	Caffeine	104	99.5	95.8				3.26
	Carbophenothion	110	92.7	98.7				6.29
	Carfentrazone Ethyl	103	92.7	94.9				2.31
	Chlorfenapyr	71.6	62.5	67.9				8.39
	Chlorpyrifos Ethyl	73.2	61.8	60.7				1.69

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 8270E	Chlorpyrifos Methyl	82.0	81.9	89.1			8.35
	Coumaphos	99.4	80.9	84.9			4.82
	Cyanazine	210	234	235			0.632
	Cyprodinil	101	97.6	107			9.56
	Dechlorane 602	94.0	79.1	82.3			4.00
	Dechlorane 603	138	118	128			8.10
	Dechlorane Plus	95.2	82.0	81.2			0.976
	Demeton	81.6	83.4	86.4			3.54
	Diazinon	111	112	113			0.719
	Difenoconazole	86.3	68.4	75.9			10.4
	Dimethenamid	103	91.6	95.9			4.57
	Dimethomorph	152	166	171			2.72
	Disulfoton	98.2	96.7	93.9			2.96
	Dithiopyr	106	85.4	90.2			5.39
	EPTC	46.5	44.9	46.4			3.37
	Ethion	81.4	92.1	102			10.7
	Ethoprop	82.2	89.9	96.6			7.12
	Etridiazole	60.6	56.8	57.2			0.606
	Fenamiphos	68.3	69.8	69.4			0.499
	Fenbuconazole	67.7	54.4	57.0			4.16
	Fipronil	76.6	65.7	68.9			4.82
	Fipronil Desulfinyl	87.3	78.4	77.4			1.15
	Fipronil Sulfide	91.0	71.8	69.9			2.73
	Fipronil Sulfone	82.4	72.8	76.4			4.79
	Fluazifop-P-butyl	104	103	105			2.08
	Fludioxonil	82.5	80.7	79.2			1.84
	Flumioxazin	137	116	127			9.05
	Flutolanil	102	93.8	102			8.36
	Fluxapyroxad	81.5	71.0	74.8			5.17
	Fonofos	109	110	110			0.467
	Hexabromobenzene	128	120	121			0.424
	Hexazinone	87.3	94.9	92.6			1.67
	Imazalil	86.0	59.5	58.2			2.21
	Iprodione	31.2	34.0	42.2			21.6
	Loratadine	168	148	146			1.03
	Malathion	80.7	75.2	77.7			3.26
	Metalaxyl	91.3	80.0	73.5			1.94
	Metolachlor	77.2	76.8	76.4			0.492
	Metribuzin	109	107	108			0.842
	Mevinphos	63.1	67.3	71.4			5.89
	Molinate	58.1	59.9	61.6			2.79
	Myclobutanil	99.8	96.0	95.7			0.312
	Naled	35.0	41.0	50.6			21.0
	Napropamide	82.8	72.7	79.1			8.36
	Norflurazon	96.9					0.371
	Octinoxate	166	161	177			9.10
	Oxadiazon	88.9	83.4	82.8			0.712
	Oxybenzone	118	107	111			3.89
	Oxyfluorfen	79.7	89.0	84.4			5.35
	Parathion Ethyl	90.3	93.6	90.8			3.04
	Parathion Methyl	89.5	90.8	101			10.4
	PBB-153	145	126	140			10.1
	PBDE-47	116	103	110			6.55
	PBDE-99	114	95.1	101			6.06

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 8270E	Pendimethalin	85.9	85.9	95.6			10.7
	Pentabromotoluene	126	116	124			7.15
	Phenytoin	136	156	162			3.48
	Phorate	99.9	107	121			11.7
	Prodiamine	55.9	61.2	61.7			0.880
	Prometon	105	94.9	96.1			1.35
	Prometryn	124	112	110			1.38
	Pronamide	110	101	107			5.59
	Propanil	106	114	112			2.46
	Propargite	66.0	59.8	60.3			0.974
	Propiconazole	75.9	70.7	73.3			3.50
	Pyraflufen Ethyl	98.0	83.6	77.2			7.95
	Pyridaben	77.6	47.7	55.5			8.96
	Pyriproxyfen	221	128	171			29.1
	Simazine	82.6	56.6	66.9			5.80
	Stiophos	58.9	40.1	42.0			4.53
	Sulfentrazone	62.2	73.4	59.3			21.3
	Tebuconazole	68.2	63.7	63.2			0.696
	Terbufos	119	115	120			4.50
	Terbutylazine	85.3	81.3	85.4			4.89
	Tetradecachloro-m-terphenyl	233	227	251			10.0
	Thiobencarb	105	92.0	95.8			4.02
	Tri-o-cresyl Phosphate	118	102	109			7.00
	Triadimefon	94.9	86.2	90.0			4.30
	Tris(1-chloro-2-propyl) phosphate (TCPP)	98.5	100	102			1.59
	Tris(1,3-dichloro-2-propyl) phosphate (TDCPP)	113	103	109			6.05
	Tris(2-chloroethyl) phosphate (TCEP)	119	106	108			2.22
EPA 8321B	2,4-D	94.0	113	110			2.74
	2,4,5-T	91.9	115	109			5.85
	Acesulfame K	144	133	135			1.22
	Acetaminophen	74.1	85.7	84.1			1.85
	Acetamiprid	82.4	98.6	96.6			2.02
	Afidopyropen	68.1	99.1	95.3			3.85
	AMPA	97.2	91.2	90.8			0.419
	Bentazon	139	164	174			5.03
	Benzovindiflupyr	79.3	81.9	87.3			6.39
	Carbamazepine	72.2	82.6	84.1			1.79
	Clothianidin	84.7	99.7	99.2			0.475
	Dinotefuran	94.6	112	116			3.54
	Diuron	70.0	72.5	73.5			1.32
	Endothall	77.8	77.2	76.0			1.55
	Fenuron	81.9	77.3	84.5			8.83
	Fluridone	73.3	81.8	96.1			16.1
	Glufosinate	79.1	67.6	69.6			3.03
	Glyphosate	96.7	82.8	91.7			10.1
	Hydrocodone	78.2	80.7	85.4			5.60
	Ibuprofen	73.6	86.5	78.2			9.98
	Imazapyr	88.6	143	114			20.0
	Imidacloprid	83.3	104	102			2.17
	Linuron	68.3	74.5	74.5			0.0137
	Mandestrobin	82.6	94.1	90.4			3.98

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision SMP	MS
EPA 8321B	MCP	98.2	126	110			13.5
	Naproxen	75.7	93.3	73.3			23.9
	Primidone	84.1	98.8	95.8			3.10
	Pyraclostrobin	77.6	85.1	85.7			0.684
	Silvex	87.8	112	117			3.93
	Sucralose	118	109	136			19.1
	Thiamethoxam	79.7	98.5	95.4			3.27
	Tolfenpyrad	51.9	51.2	49.7			2.86
	Triclopyr	91.7	105	103			2.18
SM 2320 B-2011	Total Alkalinity	98.6				0.346	
SM 2540 C-2015	TDS	96.5				2.69	
SM 2540 D-2015	TSS	95.3					
SM 4500-F- C-2011	Fluoride	99.5	102	101			0.480
SM 5310 B-2011	Organic Carbon	100	97.0	98.8			1.44

Reference Method Descriptions

Method	Description	Associated Samples
DEP SOP: NU-094-1	True Color measured at 450 nm (not valid for NPDES monitoring)	2405693
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2405693
EPA 200.7 Rev. 4.4	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2405704
EPA 200.8 Rev. 5.4	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2405704
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2405693
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2405693
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2405694
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2405694
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2405694
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2405694
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2405695
EPA 8270E	Analysis of Sunscreen Compounds and Flame Retardants in Water by GC/MS-SIM	2405711
EPA 8270E	Organonitrogen and phosphorus pesticides in water matrices by GC/MS/MS	2405711
EPA 8321B	Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS	2405697
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2405693
SM 2340 B-2011	Hardness by calculation, in aqueous matrices, reported as mg/L CaCO ₃	2405704
SM 2540 C-2015	Total Dissolved Solids in aqueous matrices	2405693
SM 2540 D-2015	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2405693
SM 4500-F- C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2405693
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2405694

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
DEP SOP: NU-094-1	05/04/2023			05/04/2023 14:13	Chandler E Cox	2405693
EPA 180.1 Rev. 2.0	05/04/2023			05/04/2023 13:36	Khloe J Berg	2405693
EPA 200.7 Rev. 4.4	05/04/2023	05/08/2023 12:00	George D Taylor	05/10/2023 20:49	McKinley C Carter	2405704
EPA 200.8 Rev. 5.4	05/04/2023	05/08/2023 12:00	George D Taylor	05/09/2023 14:13	Emily M Philpot	2405704
EPA 300.0 Rev. 2.1	05/04/2023			05/09/2023 21:36	James L Waggeberby	2405693
EPA 350.1 Rev. 2.0	05/04/2023			05/09/2023 10:35	Khloe J Berg	2405694
EPA 351.2 Rev. 2.0	05/04/2023	05/09/2023 12:02	Dana G. Hughes	05/10/2023 14:22	Samantha L Bates	2405694
EPA 353.2 Rev. 2.0	05/04/2023			05/12/2023 09:50	Beverly Y. Sanders	2405694
EPA 365.1 Rev. 2.0	05/04/2023	05/08/2023 13:00	Alexis Price	05/09/2023 14:24	Simonne.V. Calhoun	2405694
EPA 365.1 Rev. 2.0 dissolved	05/04/2023			05/04/2023 14:15	Elise M. Gillis	2405695
EPA 8270E	05/04/2023	05/08/2023 08:00	Rasheda Ghaffari	05/10/2023 18:32	Arthur Maknenko	2405711
	05/04/2023	05/08/2023 08:00	Rasheda Ghaffari	05/12/2023 17:50	Vandana T. Nagar	2405711
	05/04/2023	05/08/2023 08:00	Rasheda Ghaffari	05/14/2023 20:17	Vandana T. Nagar	2405711
EPA 8321B	05/04/2023	05/05/2023 09:00	Hoor Shaik	05/05/2023 21:03	Juyoung Kim	2405697
	05/04/2023	05/05/2023 09:00	Hoor Shaik	05/08/2023 10:22	Juyoung Kim	2405697
	05/04/2023	05/05/2023 13:00	Manjita Shrestha	05/05/2023 18:51	Manjita Shrestha	2405697
	05/04/2023	05/05/2023 13:00	Manjita Shrestha	05/06/2023 18:40	Manjita Shrestha	2405697
SM 2320 B-2011	05/04/2023			05/11/2023 16:44	Dale L. Simmons	2405693
SM 2340 B-2011	05/04/2023			05/10/2023 20:49	McKinley C Carter	2405704
SM 2540 C-2015	05/04/2023			05/05/2023 16:02	Yijie Li	2405693
SM 2540 D-2015	05/04/2023			05/05/2023 16:02	Yijie Li	2405693
SM 4500-F- C-2011	05/04/2023			05/10/2023 03:29	Dale L. Simmons	2405693
SM 5310 B-2011	05/04/2023			05/10/2023 03:19	Elise M. Gillis	2405694