

ANALYTICAL TEST REPORT

THESE RESULTS MEET NELAP STANDARDS

Submission Number : 23061119

F D E P/S W D
13051 N Telecom Pkwy.
Tampa, FL 33637

Project Name : RQ-2023-06-19-11
Date Received : 06/20/2023
Time Received : 14:32

Erica Peck

Submission Number: 23061119 Sample Date: 06/20/2023
Sample Number: 001 Sample Time: 09:45
Sample Description: G2SW0162 Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
E- COLI BY IDEXX QUANTITRAY	275	#/100 ML	10	10	SM9223B	06/20/2023 15:08	06/21/2023 15:54	SM

Submission Number: 23061119 Sample Date: 06/20/2023
Sample Number: 002 Sample Time: 10:20
Sample Description: G1SW0138 Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
E- COLI BY IDEXX QUANTITRAY	305	#/100 ML	10	10	SM9223B	06/20/2023 15:08	06/21/2023 15:54	SM

Submission Number: 23061119 Sample Date: 06/20/2023
Sample Number: 003 Sample Time: 10:45
Sample Description: G1SW0092 Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
ENTEROCOCCI	85	#/100 ML	10	10	ENTEROLERT	06/20/2023 14:56	06/21/2023 15:09	SM

Submission Number: 23061119 Sample Date: 06/20/2023
Sample Number: 004 Sample Time: 11:20
Sample Description: G1SW0139 Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
ENTEROCOCCI	10 U	#/100 ML	10	10	ENTEROLERT	06/20/2023 14:56	06/21/2023 15:09	SM
E- COLI BY IDEXX QUANTITRAY	84	#/100 ML	10	10	SM9223B	06/20/2023 15:08	06/21/2023 15:54	SM

BENCHMARK

EnviroAnalytical, Inc.

NELAP Certification #E84167

Submission Number: 23061119	Sample Date: 06/20/2023
Sample Number: 005	Sample Time: 11:50
Sample Description: G1SW0129	Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
E- COLI BY IDEXX QUANTITRAY	4352	#/100 ML	10	10	SM9223B	06/20/2023 15:08	06/21/2023 15:54	SM


Submission Number: 23061119	Sample Date: 06/20/2023
Sample Number: 006	Sample Time: 12:30
Sample Description: G2SW0138	Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
ENTEROCOCCI	677	#/100 ML	10	10	ENTEROLERT	06/20/2023 14:56	06/21/2023 15:09	SM

Submission Number: 23061119	Sample Date: 06/20/2023
Sample Number: 007	Sample Time: 13:30
Sample Description: G1SW0140	Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
E- COLI BY IDEXX QUANTITRAY	717	#/100 ML	10	10	SM9223B	06/20/2023 15:08	06/21/2023 15:54	SM

BENCHMARK



EnviroAnalytical, Inc.

NELAP Certification #E84167

Haley Rin

06/26/2023

Dale D. Dixon / Laboratory Director

Date

Kathleen Gauthier / Haley Richardson - QC Officers

DATA QUALIFIERS THAT MAY APPLY:

A = Value reported is an average of two or more determinations.
B = Results based upon colony counts outside the ideal range.
H = Value based on field kit determination. Results may not be accurate.
I = Reported value is between the laboratory MDL and the PQL.
J1 = Estimated value. Surrogate recovery limits exceeded.
J2 = Estimated value. No quality control criteria exists for component.
J3 = Estimated value. Quality control criteria for precision or accuracy not met.
J4 = Estimated value. Sample matrix interference suspected.
J5 = Estimated value. Data questionable due to improper lab or field protocols.
K = Off-scale low. Value is known to be < the value reported.
L = Off-scale high. Value is known to be > the value reported.
N = Presumptive evidence of presence of material.
O = Sampled, but analysis lost or not performed.

Q = Sample held beyond accepted hold time.

T = Value reported is < MDL. Reported for informational purposes only and shall not be used in statistical analysis.

U = Analyte analyzed but not detected at the value indicated.

V = Analyte detected in sample and method blank. Results for this analyte in associated samples may be biased high. Standard, Duplicate and Spike values are within control limits. Reported data are usable.

Y = Analysis performed on an improperly preserved sample. Data may be inaccurate.

Z = Too many colonies were present (TNTC). The numeric value represents the filtration volume.

I = Data deviate from historically established concentration ranges.

? = Data rejected and should not be used. Some or all of QC data were outside criteria, and the presence or absence of the analyte cannot be determined from the data.

* = Not reported due to interference.

Oil & Grease - If client does not send sufficient sample quantity for spike evaluation surface water samples are supplied by the laboratory.

COMMENTS:

1 (10 mL) dilution on each Enterococci and E. coli sample.

NOTES:

MBAS calculated as LAS; molecular weight = 340.

PQL = 4xMDL.

ND = Not detected at or above the adjusted reporting limit.

X = Value exceeds MCL.

G1 = Accuracy standard does not meet method control limits, but does meet lab control limits that are in agreement with USEPA generated data. USEPA letter available upon request.

For questions or comments regarding these results, please contact us at (941) 723-9986.

Results relate only to the samples.

Benchmark EnviroAnalytical, Inc
1711 12th Street East
Palmetto, FL 34221
941-723-9986
941-723-6061 Fax

PO# C00EF5

Client Information: FDEP

Address: 13051 N. Telecom Parkway

Tampa FL 33604

Phone # 815-470-5878

Report to: benjamin.paswater@dep.state.fl.us
DEP-Overflomicro@dep.state.fl.us

Invoice to: FDEP Micro Overflow

Project Name: RQ-2023- 06-19-11

Laboratory Submission #

23061089

Sample Name	Sample Type ¹ / Sample Matrix ²	Collection		Container			Preservative ⁴	Parameters for Analysis	Laboratory Sample #
		Date	Time	Qty	Capacity	Type ³			
G2SW0162	SW	6/20/2023	945	1	250mL	P	Plain (on Ice)	E-Coli	1
G1SW0138	SW	✓	1020	1	250mL	P	Plain (on Ice)	E-Coli	2
G1SW0092	SW	✓	1045	1	250mL	P	Plain (on Ice)	Enteroc	3
G1SW0139	SW	✓	1120	2	250mL	P	Plain (on Ice)	E-Coli and Enteroc	4
G1SW0129	SW	✓	1150	1	250mL	P	Plain (on Ice)	E-Coli	5
G2SW0138	SW	✓	1230	1	250mL	P	Plain (on Ice)	BP E-Coli Enteroc	6
G1SW0140	SW	✓	1330		250mL	P	Plain (on Ice)	E. Coli	7

1 "Sample Type" is used to indicate whether the sample was a grab (G) or whether it was a composite (C).

2 "Sample Matrix" is used to indicate whether the sample is being discharged to drinking water (DW), groundwater (GW), surface water (SW), fresh surface water (FSW), saline surface water (SSW), soil, sediment (SD/MNT), or sludge (SLDG).

3 "Container Type" is used to indicate whether the container is plastic (P) or glass (G).

4 Under "Preservative," list any preservatives that were added to the sample container.

Under "Preservative," list any preservatives that were added to the sample container.

Instructions:

1. Each bottle has a label identifying sample ID, premeasured preservative contained in the bottle, sample type, client ID, and parameters for analysis.

2. The following information should be added to each bottle label after collection with permanent black ink: date and time of collection, sampler's name or initials, and any field number or ID.

3. All bottles not containing preservative may be rinsed with appropriate sample prior to collection.

4. The client is responsible for documentation of the sampling event. Please note special sampling events on the sample custody form.

1	Collected / Relinquished By:	Date	Time	Received By:	Date	Time
		6/20/23	1430	 Nathan Handsell	6-20-23	1432
2	Relinquished By:	Date	Time	Received By:	Date	Time

Laboratory Sample Acceptability:

Temperature: 1.1°C