

Login Checklist

RQ-2023 - 09 - 18 - 66

Login Station ID: # 2

Shipping Method: FedEx | UPS | HD | Greyhound

Date/Time of Receipt: 09/21/2023 / 09:24

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
					Present?		*Intact?		
	YES	HNO ₃	Formalin		Yes	No	Yes	No	
1.8C				9		X			6484 8458 0359
2.1C				9		X			6484 8458 0360

COOLER CHECK

*All samples received on ice unless otherwise noted. HNO3 and Formalin preserved samples do NOT have a temperature requirement.

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. NCR# _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. NCR# _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. NCR# _____
- If coolers or samples are received late; identify the affected samples in an NCR. NCR# _____

Chain of Custody/ Field Sheet(s) Included?

Yes X No _____

Micro-Biology Overflow Chain of Custody/ FieldSheet(s) Included?

Yes _____ No X

CONTAINER CHECK

Evidence Tape on Bottles?

Yes _____ No _____ NA X

Evidence Tape intact? Yes _____ No _____

If Criminal, photographs taken?

Yes _____ No _____

Containers intact? Yes X No _____

Caps on tight?

Yes X No _____

Sufficient Sample Volume:

Yes X No _____

If **NO** is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? (Blue dot on container)

Yes X No _____ Init: ABS

Chlorine detected? (One container checked per Field ID)

Yes _____ No X Init: ABS

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

**Acid Preserved Samples: pH< 2.0?

Yes X No _____ NA _____ Init: ABS

**W-1-4-DIOX (Green dot on container) preserved to pH< 4.0?

Yes _____ No _____ NA X Init: ABS

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: ABS /

Date: 09/21/2023

Event contains NCRs (Y/N)? N

Event ID: SW - DIST- 2023 - 09 - 21 - 02

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Additional Comments:

Infrared Thermometer ID:	IRTG#2 EXP:10/17/23
pH Test Strips 0 – 6	LOT#223819BVEXP8/15/24
pH Paper 0 – 3	N/A
pH Paper 0 – 13	LOT207621EXP:3/15/24
Chlorine Test Strips	2313 2025/07

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: ABS

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

**FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION****STRATEGIC MONITORING PLAN Field Sheet**

July 2022

Meter
Passed End
of Day CCV
Yes / No

WIN Station Name:	OLD TOWN DRAINAGE DITCH AT OLD TOWN CREEK RD				Sample Date:	9/20/2023	
WIN / Field ID:	G3SW0121	ORG ID:	21FLTPA	ENR:	-		
County:	HARDEE	WBID:	1805	STREAM			
Receiving Body of Water:	Charlie Creek				RQ-	2023-09-18-66	
Latitude:	27.61288			Longitude:	-81.62068		

Sampling Team Members	WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
Ben Paswater		x		x	x
Devin Mathias	x		x		

Sampling Equipment			
x	Sample Container	Carboy	
	Dipper	Van Dorn	
	Syringe	Pole	
	Syringe Lot#	Filter Lot#	
	Secchi Equipment ID:	Secchi #3	
Collection Method			
	Wading	Canoe/Kayak	
x	From Shore	Electric Motor	
		Gasoline Motor	
Depth Measured With		Secchi	

Water Level:	Normal
Flow:	Flowing
Matrix:	Fresh
Tide Stage:	
High Tide:	
Low Tide:	
Photos:	No
Meter ID:	Leia
Depth Gauge ID:	depth_gauge

Time Zone	Time	Total Depth (m)	Sample Depth (m)	DO (mg/L)	DO (%SAT)	Temp. (°C)	pH	Secchi Disk (m)	Sp COND (µmhos/cm)	Salinity (PPT)
EST	1020	0.2	0.1	0.88	10.7	25.4	6.3	0.2	153.8	0.07
								VOB (s)		

Biological Samples Collected: NoneSBIO Visit ID: HA ID: RPS ID: Macrophyte ID: LIMS Sample ID
In The Macrophyte Module:

Parameter Suite	Parameter Methods (LIMS Test Codes)	Type	Lot#	Expiration Date	pH Check
Microbiology	E. coli Contract Lab Benchmark	ICE			
Chlorophyll	CHLSUITE-W	ICE			
Physical/Aggregate Nutrients	ALKALINITY, TURBIDITY, W-F, W-TSS W-CL-IC, W-COLOR, W-SO4-IC, W-TDS	ICE			
Filtered Nutrients					
Acid Preserved Nutrients	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	ICE	H2SO4*	143722	01/22/2024
Metals					
Organics (LC)					
Macroinvertebrates					
Organics (GC)					
Molecular					
Periphyton					
Phytoplankton					
BOD					
DOC					
Other(s)					
Field Comments:	Cloudy, sunny				
Comments on COC:	Microbiology sample(s) submitted to contract laboratory.				
Relinquish Date:	09/20/2023	Signature:			

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:

Save Field Sheets on Network: Migrate Data to WIN:

(Initials/Date)

(Initials/Date)

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 9/20/2023
Customer: SW-DIST
RQ: RQ-2023-09-18-66
Collected By: Ben Paswater, Devin Mathias
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 2

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
OLD TOWN DRAINAGE DITCH AT OLD TOWN CREEK RD	 G3SW0121
WHITAKER BAYOU AT RINGLING COLLEGE OF ART	 G3SW0060
MAPLE CREEK AT WAUCHULA RD	 G3SW0087

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 9/20/2023
1700EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By: A.SHAIK

Rec'd/Inspected Date: 09/21/2023



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION

STRATEGIC MONITORING PLAN Field Sheet

July 2022

Meter
Passed End
of Day CCV
Yes / No

WIN Station Name:	MAPLE CREEK AT WAUCHULA RD				Sample Date:	9/20/2023	
WIN / Field ID:	G3SW0087	ORG ID:	21FLTPA	ENR:	-	Latitude:	27.38689
County:	MANATEE	WBID:	1935	STREAM		Longitude:	-82.13884
Receiving Body of Water:	Myakka River Upper Segment				RQ-	2023-09-18-66	

Sampling Team Members	WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
Ben Paswater		x		x	x
Devin Mathias	x		x		

Sampling Equipment			
x	Sample Container	Carboy	
	Dipper	Van Dorn	
	Syringe	Pole	
	Syringe Lot#	Filter Lot#	
	Secchi Equipment ID:	Secchi #3	
Collection Method			
	Wading	Canoe/Kayak	
x	From Shore	Electric Motor	
		Gasoline Motor	
Depth Measured With		Secchi	

Water Level:	Normal
Flow:	Flowing
Matrix:	Fresh
Tide Stage:	
High Tide:	
Low Tide:	
Photos:	No
Meter ID:	Leia
Depth Gauge ID:	depth_gauge

Time Zone	Time	Total Depth (m)	Sample Depth (m)	DO (mg/L)	DO (%SAT)	Temp. (°C)	pH	Secchi Disk (m)	Sp COND (µmhos/cm)	Salinity (PPT)
EST	1120	1.2	0.3	3.1	37.5	25	6.31	0.5	455.3	0.22

Biological Samples Collected: None

SBIO Visit ID: HA ID: RPS ID: Macrophyte ID: LIMS Sample ID
In The Macrophyte Module:

Parameter Suite	Parameter Methods (LIMS Test Codes)			Preservation			
	Type	Lot#	Expiration Date	pH Check			
Microbiology	E. coli						
	Contract Lab	Benchmark		ICE			
Chlorophyll	CHLSUITE-W			ICE			
Physical/Aggregate	ALKALINITY, TURBIDITY, W-F, W-TSS			ICE			
Nutrients	W-CL-IC, W-COLOR, W-SO4-IC, W-TDS			ICE			
Filtered Nutrients					Syringe Lot#; Filter Lot#:		
Acid Preserved Nutrients	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	ICE	H2SO4*	143722	01/22/2024	H2SO4 pH<2	
Metals							
Organics (LC)	W-E8321-DI	W-E8321-MS	W-CARB-AA**	ICE	MCAA***	143355	07/13/2024
Macroinvertebrates							
Organics (GC)	W-CT-CI-R	W-PCL-TQ-R	W-PCB-TQ-R	Ice	*2.0 ml of H2SO4 Used Unless Noted Otherwise ***5 ml MCAA (Monochloroacetic Acid Buffer Solution)		
	W-PSNP-TQ	W-TR-SQ-R					
Molecular							
Periphyton							
Phytoplankton							
BOD							
DOC					SyringeLot#	FilterLot#	
Other(s)							
Field Comments:	Cloudy, sunny						
Comments on COC:	Microbiology sample(s) submitted to contract laboratory.						
Relinquish Date:	09/20/2023	Signature:					

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:

Save Field Sheets on Network: (Initials/Date) Migrate Data to WIN: (Initials/Date)

(Initials/Date) (Initials/Date)

**FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION**

STRATEGIC MONITORING PLAN Field Sheet

July 2022

Meter
Passed End
of Day CCV
Yes / No

WIN Station Name:	WHITAKER BAYOU AT RINGLING COLLEGE OF ART				Sample Date:	9/20/2023	
WIN / Field ID:	G3SW0060	ORG ID:	21FLTPA	ENR:	-		
County:	SARASOTA	WBID:	1936	ESTUARY			
Receiving Body of Water:	Sarasota Bay				RQ-	2023-09-18-66	

Sampling Team Members	WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
Ben Paswater		x		x	x
Devin Mathias	x		x		

Sampling Equipment	
x	Sample Container
	Dipper
	Syringe
	Syringe Lot#
	Secchi Equipment ID:
	Carboy
	Van Dorn
	Pole
	Filter Lot#
	Secchi #3
Collection Method	
	Wading
x	From Shore
	Canoe/Kayak
	Electric Motor
	Gasoline Motor
Depth Measured With	
Secchi	

Water Level:	Normal
Flow:	Flowing
Matrix:	Marine
Tide Stage:	Rising
High Tide:	1717
Low Tide:	1036
Photos:	No
Meter ID:	Leia
Depth Gauge ID:	depth_gauge

Time Zone	Time	Total Depth (m)	Sample Depth (m)	DO (mg/L)	DO (%SAT)	Temp. (°C)	pH	Secchi Disk (m)	Sp COND (µmhos/cm)	Salinity (PPT _h)
EST	1240	0.4	0.2	3.29	54.4	32.5	7.98	0.4	51106	33.26
VOB (s)										

Biological Samples Collected: None

SBIO Visit ID: HA ID: RPS ID: Macrophyte ID: LIMS Sample ID In The Macrophyte Module:

Parameter Suite	Parameter Methods (LIMS Test Codes)			Preservation			
				Type	Lot#	Expiration Date	pH Check
Microbiology	Contract Lab	Enterococci	Benchmark	ICE			
Chlorophyll							
Physical/Aggregate							
Nutrients							
Filtered Nutrients							
Acid Preserved							
Nutrients							
Metals							
Organics (LC)							
Macroinvertebrates							
Organics (GC)	W-CT-CI-R	W-PCL-TQ-R	W-PCB-TQ-R	Ice			
Molecular							
Periphyton							
Phytoplankton							
BOD							
DOC							
Other(s)							
Field Comments:	Cloudy, sunny						
Comments on COC:	Microbiology sample(s) submitted to contract laboratory.						
Relinquish Date:	09/20/2023	Signature:					

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID in LIMS DMT: QC Station ID, Field Parameters in LIMS DMT:

Save Field Sheets on Network: Migrate Data to WIN:

Benchmark EnviroAnalytical, Inc
1711 12th Street East
Palmetto, FL 34221
941-723-9986
941-723-6061 Fax

Client Information: **FDEP**

Address: 13051 N. Telecom Parkway
Tampa FL 33604
Phone # 815-470-5878

Report to: benjamin.paswater@dep.state.fl.us
DEP-Overflomicro@dep.state.fl.us

PO# **CIEFCB**

Project Name: **RQ-2023-09-18-CSE**

Invoice to: **FDEP Micro Overflow**

Laboratory Submission #

23091086

Sample Name	Sample Type / Sample Matrix	Collection		Container			Preservative ⁴	Parameters for Analysis	Laboratory Sample #
		Date	Time	Qty	Capacity	Type ³			
G35M0121	SW	9/20/23	1020	1	250mL	P	Plain (on Ice)	E. coli	1
G35M0087	SW		1120	1	250mL	P	Plain (on Ice)	T	2
G35M0050	SW		1240	1	250mL	P	Plain (on Ice)	Enteroc	3
	SW				250mL	P	Plain (on Ice)		
	SW				250mL	P	Plain (on Ice)		
	SW				250mL	P	Plain (on Ice)		
	SW				250mL	P	Plain (on Ice)		

¹ Sample Type¹ is used to indicate whether the sample was a grab (G) or whether it was a composite (C).
² Sample Matrix² is used to indicate whether the sample is Surface Water (SW), Groundwater (GW), Surface Water (SW), or other (OT).
³ Container Type³ is used to indicate whether the sample is plastic (P) or glass (G).
⁴ Preservative⁴ is used to indicate whether the sample is preserved or not. The temperature during storage should be less than or equal to 4°C (40°F).
⁵ Under "Preservative," list any preservatives that were added to the sample container.
⁶ All bottles are labeled with the sample ID, preservation preservative contained in the bottle, sample type, date and time of collection, and any field number or ID.
⁷ The following information is required for documentation of the sample: sample ID, preservation preservative, sample type, date and time of collection, and any field number or ID.
⁸ The client is responsible for documentation of the sample. Please refer to the sample ID, sample type, date and time of collection, and any field number or ID.

Laboratory Sample Acceptability:
Temperature: **18°C**

1	Collected/Relinquished By:	Date	Time	Received By:	Date	Time
2	Relinquished By:	Date	Time	Received By:	Date	Time

Date of Request: 19-SEP-2023
 Created By: PASWATER_B On: 19-SEP-2023
 Modified By: AYRES_J On: 21-SEP-2023
 Customer: SW-DIST
 Project: SMP
 Division: Division of Environmental Assessment and Restoration
 District: Southwest District
 Event Description: Maple Unnamed & Whikater Bayou

Send Coolers To:

Phone: 813-632-7600
 FL Dept. of Environmental Protection
 13051 N. Telecom Parkway
 Temple Terrace, FL 33637-0926
 Attn: Ben Paswater
 Cooler Ship EMail: benjamin.paswater@dep.state.fl.us

Priority: 3
 Event Status: R
 Criminal Investigation: NO
 Chemistry Request Reviewed By: AYRES_J On: 21-SEP-2023
 Biology Request Reviewed By: AYRES_J On: 21-SEP-2023
 Sampling Kit Required: NO Ship on:
 Sampling Kit Shipped:
 Sampling Kit Packed By:
 Preservatives Needed: NO
 Date to Receive Samples: 19-SEP-2023
 Logged In By: SHAIK_A On: 21-SEP-2023

Send Final Report To:

13051 N. Telecom Parkway
 Temple Terrace, FL 33637-0926
 Attn: Ben Paswater
 Report Notification EMail:
 benjamin.paswater@dep.state.fl.us;sarah.meyer@dep.state.fl.us
 ;Devin.Mathias@Floridadep.gov

Comments: No Kit Required

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO ₃ /L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(DEP SOP: NU-094-1) True Color measured at 450 nm (not valid for NPDES monitoring)
Color (true)		
W-F	Template: DEFAULT	(SM 4500-F- C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO ₄ -IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2015) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2015) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L	Number of Bottles: 1	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10150 B) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: P-250ML-WI-THI	Number of Bottles: 1	Preserved With: ICE
SW-BEA-ECQ	Template: DEFAULT	(SM 9223 B Quanti-Tray) E. coli by Colilert/QT by BEA in Palmetto
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: ICE And H ₂ SO ₄
W-NH ₃	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO ₂ NO ₃	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Group B (Water) With 1 Samples:

Bottle Type: BG-500ML	Number of Bottles: 1	Preserved With: ICE
W-E8321-DI	Template: DEFAULT	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
W-E8321-MS	Template: SMP	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: ICE And H ₂ SO ₄
W-NH ₃	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L

Bottle Group B (Water) With 1 Samples:

Bottle Type: P-500ML		Number of Bottles: 1	Preserved With: ICE And H2SO4
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L	
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L	
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices	
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices	
Bottle Type: BG-500ML		Number of Bottles: 4	Preserved With: ICE
W-PSNP-TQ	Template: DEFAULT	(EPA 8270E) Organonitrogen and phosphorus pesticides in water matrices by GC/MS/MS	
W-TR-SQ-R	Template: DEFAULT	(EPA 8270E) Analysis of Sunscreen Compounds and Flame Retardants in Water by GC/MS-SIM	
Bottle Type: P-1L		Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO3/L	
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices	
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices	
W-COLOR	Template: DEFAULT	(DEP SOP: NU-094-1) True Color measured at 450 nm (not valid for NPDES monitoring)	
Color (true)			
W-F	Template: DEFAULT	(SM 4500-F- C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices	
W-TDS	Template: DEFAULT	(SM 2540 C-2015) Total Dissolved Solids in aqueous matrices	
W-TSS	Template: DEFAULT	(SM 2540 D-2015) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	
Bottle Type: BP-1L		Number of Bottles: 1	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10150 B) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry	
Bottle Type: P-250ML-WI-THI		Number of Bottles: 1	Preserved With: ICE
SW-BEA-ECQ	Template: DEFAULT	(SM 9223 B Quanti-Tray) E. coli by Colilert/QT by BEA in Palmetto	
Bottle Type: BG-500ML		Number of Bottles: 1	Preserved With: CH2ClCOOH and Ice
W-CARB-AA	Template: DEFAULT	(EPA 8321B) Carbamates in water matrices by HPLC/MS/MS	
Bottle Type: BG-500ML FILL TO		Number of Bottles: 4	Preserved With: ICE
W-CT-CI-R	Template: DEFAULT	(EPA 8276) Chlordane and Toxaphene in water matrices by GC/MS with Negative Chemical Ionization	
W-PCB-TQ-R	Template: DEFAULT	(EPA 8270E) PCBs in water matrices by GC/MS/MS	
W-PCL-TQ-R	Template:	(EPA 8270E) Organochlorine pesticides, pyrethroids and tracers in water matrices by GC/MS/MS	

Bottle Group C (Water) With 1 Samples:

Bottle Type: P-250ML-WI-THI		Number of Bottles: 1	Preserved With: ICE
SW-BEA-ENQ	Template: DEFAULT	(Enterolert/QT) Enterococci by Enterolert/QT by BEA in Palmetto	
Bottle Type: BG-500ML FILL TO		Number of Bottles: 2	Preserved With: ICE
W-CT-CI-R	Template: DEFAULT	(EPA 8276) Chlordane and Toxaphene in water matrices by GC/MS with Negative Chemical Ionization	
W-PCB-TQ-R	Template: DEFAULT	(EPA 8270E) PCBs in water matrices by GC/MS/MS	
W-PCL-TQ-R	Template:	(EPA 8270E) Organochlorine pesticides, pyrethroids and tracers in water matrices by GC/MS/MS	

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.