

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: The One

RQ- 202404-29-05
202404-01-43

Project: Punta Pass

Boldly "X" this box if there are qualified data on this page.

- Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 3/28/24

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass Fail	Lab Field
Calibr.	Samantha Galinas	4/30/24	706	20.2	761.1	9.1	9.07	100.1	—	1.047	P/F	L/F
ICV	"	"	709	19.9	761.0	9.1	9.19	101.0	—	—	P/F	L/F
CCV	Samantha Galinas	4/30/24	1602	22.7	759.7	8.6	8.65	100.3	—	—	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass Fail	Lab Field
Calibr.	Samantha Galinas	4/30/24	655	2309J21	9/25	2500	2500	P/F	L/F
ICV	"	"	656	2306J21	6/25	100	104.5	P/F	L/F
CCV	Samantha Galinas	4/30/24	1557	2311L37	11/25	50,000	49572	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass Fail	Lab Field
Calibr.	Samantha Galinas	4/30/24	700	2309H95	4/25	7.01	21.4	7.02	-61.3	P/F	L/F
Calibr.	"	"	702	2309H15	8/25	4.01	21.2	4.00	118.9	P/F	L/F
Calibr.	"	"	704	2308P48	2/25	10.05	21.4	10.04	-229.3	P/F	L/F
ICV	"	"	705	2309H15	8/25	4.01	21.0	3.98	—	P/F	L/F
CCV	Samantha Galinas	4/30/24	1600	2308P48	2/25	10.06	19.5	10.05	—	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;
If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)
If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____
If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: slippery disk Date of last verification: 11/7/24

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass/ Fail	Lab/ Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: