

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

Hodor

RQ-

2024 07-01-22

Project:

SMP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

06/12/2024

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Drew Litter	07/08/24	0804	23.9	761.5	8.4	8.44	100.2	—	1.04	P/F	L/F
ICV	Drew Litter	07/08/24	0806	23.9	761.5	8.4	8.44	100.4	—	1.04	P/F	L/F
CCV	Drew Litter	07/08/24	1640	27.1	760.1	8.0	8.13	102.4	—	1.08	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Drew Litter	07/08/24	0808	240209B	02/25	1000	1000	P/F	L/F
ICV	Drew Litter	07/08/24	0810	240209B	02/25	100	98.6	P/F	L/F
CCV	Drew Litter	07/08/24	1644	230714E	07/24	50,000	50956	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU 7.00	mV	Pass / Fail	Lab / Field
Calibr.	Drew Litter	07/08/24	0812	240117B	07/25	7.00	24.8	7.00	-22.9	P/F	L/F
Calibr.	Drew Litter	07/08/24	0814	240209C	08/25	4.01	24.8	4.01	153.9	P/F	L/F
Calibr.	Drew Litter	07/08/24	0816	240117C	07/25	10.00	25.2	10.00	-193.4	P/F	L/F
ICV	Drew Litter	07/08/24	0818	240117C	07/25	10.00	25.2	10.00	-193.5	P/F	L/F
CCV	Drew Litter	07/08/24	1646	240209C	08/24	4.01	24.7	4.00	149.7	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50;

mV pH 4 Range +180 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 06/12/2024

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Drew Litter	07/08/24	0820	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: