



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

SPRINGS - SAMPLING FIELD SHEET

STATION INFORMATION

SPRING NAME Wacissa head spring #2 STATION ID _____ DATE 10/22/14 TIME(EST) 10:45 (24HR)
SAMPLE DEPTH(meters) 2.0 BOTTOM VISIBLE? ☒ YES NO WEATHER: ☒ CLEAR PARTLY-CLOUDY CLOUDY RAIN

WATER QUALITY MEASUREMENTS

Analytes collected for this site are as listed per the container inventory, including any QA/QC samples

☒ YES

NO

CHEMICAL STABILITY MONITORING

Time	Temperature (°C)	Conductivity (µs/cm)	DO (% sat)	DO (mg/l)	pH (su)
10:34	20.87	280	42.1	3.76	7.54
10:39	20.86	280	37.5	3.35	7.56
10:44	20.86	280	36.1	3.23	7.56

Stability Criteria: Temp $\pm 0.2^{\circ}\text{C}$ Sp Cond $\pm 5\%$ of reading pH ± 0.2 SU DO $\leq 20\%$ or 0.2 mg/l Turbidity $\leq 20\%$ or 5 NTUs

FINAL FIELD READING

Meter ID	Temperature (°C)	Conductivity (µs/cm)	D.O. (% sat)	D.O. (mg/l)	pH (su)
128647	20.85	280	35.9	3.20	7.57

SAMPLE PRESERVATION

Were all samples collected using the appropriate 0.45-micron filter (when applicable) &

Preserved with the appropriate type and amount of acid (sulfuric for nutrients & nitric for metals)?

☒ YES

NO

Nitric Acid Lot # _____ Sulfuric Acid Lot # _____ Filter Lot # _____

Were all collected samples properly preserved and stored on wet ice at 4°C within 15 minutes?

☒ YES

NO

RQ LABEL

RQ-2014-10-20-62
SPRS1425

SITE LABEL



WACISSA HEAD SPRING #2
SPRS1425 WACISSA HEAD SPRING #2 SIPO

QA/QC Taken?

YES

☒ NO

TIME(EST) _____ (24HR)

BLANK TYPE

EQUIPMENT

FIELD

PERISTALTIC PUMP ID # 1

QA/QC LABEL

(if applicable)

(QA/QC ID)

COMMENTS

Staff gage = 28.9

SAMPLER NAMES & SIGNATURES (All team members present)

Edgar Wadde

Edgar Wadde

K. Eller
K. Eller

Kristina Bridger

Kristina Bridger