



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 10901

PRE-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 1/22/14 TIME 7:40

POST-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 1/23/14 TIME 8:30

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.60	17.88	90.6	F
CALIBRATION	9.49	17.87	100.0	P
PRE-VERIFICATION	9.50	17.85	100.0	P
POST-VERIFICATION	9.39	18.51	100.7	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	70	F
CALIBRATION	100			100	P
PRE-VERIFICATION	100			100	P
POST-VERIFICATION	100			97	P

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7	132908	4/15	6.97	17.48	P
CALIBRATION	7			7.00	17.48	P
PRE-VERIFICATION	7			7.00	17.42	P
POST-VERIFICATION	7			7.02	16.70	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES NO

RQ & SITE IDs

RQ-2014-01-27-28

SPRG1401



QAQC BLANK 1

SPRG1401



ICHETUCKNEE HEAD SPRING

SPRG1401 ICHETUCKNEE HEAD SPRING



MILL POND SPRINGS (COLUMBIA)

SPRG1401 MILL POND SPRINGS (COLUMBIA)

RQ-2014-01-27-27

SPRG1401




MISSION SPRINGS

SPRG1401 MISSION SPRINGS



BLUE HOLE SPRING (COLUMBIA)

SPRG1401 BLUE HOLE SPRING (COLUMBIA)



DEVILS EYE SPRING (GILCHRIST)

SPRG1401 DEVILS EYE SPRING (GILCHRIST)

SAMPLER NAMES & SIGNATURES (All team members present)

K. Eller
K. Eller