



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 101901

PRE-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 2/5/14 TIME 7:15

POST-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 2/5/14 TIME 17:45

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.95	20.72	99.8	P
CALIBRATION	8.96	20.74	100.0	P
PRE-VERIFICATION	8.96	20.74	100.0	P
POST-VERIFICATION	9.10	19.86	99.7	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	101	P
CALIBRATION	100	2301231	6/14	100	P
PRE-VERIFICATION	100	2301231	6/14	100	P
POST-VERIFICATION	100	2301231	6/14	102	P

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7.00	135752	9/15	7.04	20.21	P
CALIBRATION	7.00	135752	9/15	7.00	20.25	P
PRE-VERIFICATION	7.00	135752	9/15	7.00	20.26	P
POST-VERIFICATION	7.00	135752	9/15	6.97	19.94	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. ± 5% pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES NO

RQ & SITE IDS

RQ-2014-02-03-20

SPRG1401



WAKULLA TUNNEL B

SPRG1401 WAKULLA TUNNEL B



WAKULLA TUNNEL C

SPRG1401 WAKULLA TUNNEL C



WAKULLA D TUNNEL

SPRG1401 WAKULLA D TUNNEL

SPRG1401
QA/QC BLANK 2



WAKULLA SPRING

SPRG1401 WAKULLA SPRING



WAKULLA TUBING A/D TUNNEL

SPRG1401 WAKULLA TUBING A/D TUNNEL SIPO



WAKULLA TUBING A/K TUNNEL

SPRG1401 WAKULLA TUBING A/K TUNNEL SIPO



WAKULLA K TUNNEL

SPRG1401 WAKULLA K TUNNEL

SAMPLER NAMES & SIGNATURES (All team members present)

Edgar Wade

Gary Maddox