



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 101901

PRE-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 3/30/14 TIME 19:30

POST-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 3/31/14 TIME 11:20

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.79	21.62	99.8	P
CALIBRATION	8.79	21.62	99.9	P
PRE-VERIFICATION	8.79	21.62	99.7	P
POST-VERIFICATION	8.75	22.48	101.0	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	91	P
CALIBRATION	100	2301231	6/14	100	P
PRE-VERIFICATION	100	2301231	6/14	100	P
POST-VERIFICATION	100	2301231	6/14	100	P

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7	135752	9/15	6.95	21.12	P
CALIBRATION	7	135752	9/15	7.00	21.12	P
PRE-VERIFICATION	7	135752	9/15	7.00	21.12	P
POST-VERIFICATION	7	135752	9/15	6.97	22.17	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES ☒ NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES ☒ NO

RQ & SITE IDs

RQ-2014-03-03-50

SPRG1401



PONCE DE LEON SPRINGS (HOL)

SPRG1401 PONCE DE LEON SPRINGS (HOLMES)

SAMPLER NAMES & SIGNATURES (All team members present)

K. Eiler
H. Eiler