



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # YSI #12

PRE-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 6/10/14 TIME 8:15

POST-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 6/10/14 TIME 18:45

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.64	23.87	102.4	P
CALIBRATION	8.44	23.86	100.0	P
PRE-VERIFICATION	8.73	23.85	99.9	P
POST-VERIFICATION	8.57	23.58	99.8	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	101	P
CALIBRATION	100	2301231	6/14	100	P
PRE-VERIFICATION	100	2301231	6/14	100	P
POST-VERIFICATION	100	2301231	6/14		

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7.00	135752	9/15	7.06	23.14	P
CALIBRATION	7.00	135752	9/15	7.00	23.08	P
PRE-VERIFICATION	7.00	135752	9/15	7.00	23.07	P
POST-VERIFICATION	7.00	13575	9/15	7.02	23.11	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES NO

RQ & SITE IDs

RQ-2014-06-09-29

SPRG1404



CYPRESS SPRING

SPRG1404 CYPRESS SPRING



GAINER SPRING #1C

SPRG1404 GAINER SPRING #1C

SAMPLER NAMES & SIGNATURES (All team members present)

Edgar Wade

Edgar Wade

Kirstin Ellor

Kirstin Ellor