



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # _____

PRE-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 6/12/14 TIME 8:15

POST-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 6/12/14 TIME 19:00

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.74	23.64	103.1	P
CALIBRATION	8.47	23.65	100.0	P
PRE-VERIFICATION	8.46	23.67	99.9	P
POST-VERIFICATION	8.42	23.76	100.1	P

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	100	P
CALIBRATION	100	2301231	6/14	100	P
PRE-VERIFICATION	100	2301231	6/14	100	P
POST-VERIFICATION	100	2301231	6/14	101	P

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7.00	135752	9/15	7.04	23.26	P
CALIBRATION	7.00	135752	9/15	7.00	23.26	P
PRE-VERIFICATION	7.00	135752	9/15	7.00	23.26	P
POST-VERIFICATION	7.00	135752	9/15	7.02	23.35	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES NO

RQ & SITE IDs

RQ-2014-06-02-54	SPRS1405	LOG SPRING	THOMAS SPRING (JEFFERSON)	CASSIDY SPRINGS SIP	BIG BLUE SPRING (JEFFERSON)	WR-1 WR-2 WR-3 WR-4
------------------	----------	------------	---------------------------	---------------------	-----------------------------	------------------------------

WACISSA SPRING #2

SPRG1404 WACISSA SPRING #2

SAMPLER NAMES & SIGNATURES (All team members present)

Edgar Wade

Edgar Wade