



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # YSI 12

PRE-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 6/24/14 TIME 8:30

POST-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 6/25/14 TIME 6:30 (18:30)

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.49	24.16	101.2	P
CALIBRATION	8.40	24.15	100.0	P
PRE-VERIFICATION	8.46	23.76	100.0	P
POST-VERIFICATION				

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	101	P
CALIBRATION	100	2301231	6/14	100	P
PRE-VERIFICATION	100	2301231	6/14	100	P
POST-VERIFICATION	100	2301231	6/14	100	P

pH

Reading Type	Standard	Lot #	Exp. Date	6.09	23.54	Temp °C	Pass or Fail
INITIAL	7.00	135752	9/15	8.44	24.15		P
CALIBRATION	7.00	135752	9/15	7.00	23.54		P
PRE-VERIFICATION	7.00	135752	9/15	7.00	23.54		P
POST-VERIFICATION	7.00	135752	9/15	7.09	23.42		P

Acceptance Criteria: DO $\pm$ 0.3 mg/L Sp. Cond. $\pm$ 5% pH $\pm$ 0.2 SU Temperature $\pm$ 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES NO

RQ & SITE IDs

RQ-2014-04-21-44
SPRG1404



DEVILS EYE SPRING (GILCHRIST)
SPRG1404 DEVILS EYE SPRING (GILCHRIST)

SAMPLER NAMES & SIGNATURES (All team members present)

K. Eller Edgar Wade
[Signature] [Signature]