



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # YSI #12

PRE-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 6/26/14 TIME 8:30

POST-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 6/26/14 TIME 15:48

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.41	23.20	98.5	P
CALIBRATION	8.54	23.23	100.0	P
PRE-VERIFICATION	8.34	23.24	100.0	P
POST-VERIFICATION	8.31	24.73	100.0	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2301231	6/14	100	P
CALIBRATION	100	2301231	6/14	100	P
PRE-VERIFICATION	100	2301231	6/14	100	P
POST-VERIFICATION	100	2301231	6/14	100	P

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7	135752	9/15	7.2	23.02	P
CALIBRATION	7	135752	9/15	7.0	23.03	P
PRE-VERIFICATION	7	135752	9/15	7.0	23.01	P
POST-VERIFICATION	7	135752	9/15	6.9	23.22	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. ± 5% pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES ☒ NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES ☒ NO

RQ & SITE IDS

RQ-2014-05-26-45

SPRS1405



QAQC BLANK 2

SPRG1404



BLUE SPRING (MADISON)

SPRS1313 MADISON BLUE SPRING- STATUS NETWORK

SAMPLER NAMES & SIGNATURES (All team members present)

K. Eller
[Signature]

Edgar Wadde
[Signature]