



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 118153

PRE-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 9/3/14 TIME 9/3/14 7:10

POST-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 9/3/14 TIME 9/3/14 18:20

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.57	23.11	100.1	P
CALIBRATION	8.56	23.11	100.0	P
PRE-VERIFICATION	8.56	23.11	100.0	P
POST-VERIFICATION	8.59	23.14	100.2	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	100	2407C09	07/16	99	P
CALIBRATION	100	2407C09	07/16	100	P
PRE-VERIFICATION	100	2407C09	07/16	100	P
POST-VERIFICATION	1000	2405350	10/15	996	P

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7.00	143306	05/16	7.03	23.01	P
CALIBRATION	7.00	143308	05/16	7.00	23.02	P
PRE-VERIFICATION	7.00	143308	05/16	7.00	23.02	P
POST-VERIFICATION	10.00	135751	09/15	10.06	23.05	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES ☐ NO ☒

Was the Dissolved Oxygen Membrane changed? (circle one) YES ☐ NO ☒

RQ & SITE IDS

RQ-2014-09-01-58

Wacissa #2 Spring
Big Blue Spring

RQ-2014-09-01-59

Cassidy Spring
Log Spring
Thomas Spring

RQ-2014-09-01-60

WR-1
WR-2
WR-3
WR-4

SAMPLER NAMES & SIGNATURES (All team members present)

Edgar Wade