



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # 128647

PRE-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 10/22/14 TIME 8:00

POST-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 10/22/14 TIME 17:30

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	7.93	27.43	100.3	P
CALIBRATION	7.91	27.41	100.0	P
PRE-VERIFICATION	7.91	27.41	100.0	P
POST-VERIFICATION	7.99	26.97	100.1	P

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	1000	2405350	10/15	1004	P
CALIBRATION	1000	" "	" "	1000	P
PRE-VERIFICATION	1000	" "	" "	1000	P
POST-VERIFICATION	1000	" "	" "	1002	P

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	10.00	135751	9/15	10.02	27.63	P
CALIBRATION	10.00	135751	9/15	10.00	27.60	P
PRE-VERIFICATION	10.00	135751	9/15	10.00	27.60	P
POST-VERIFICATION	4.00	135444	10/15	3.98	27.53	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES NO

RQ & SITE IDs

<u>RQ-2014-10-20-63</u> WR-1 WR-2 WR-3 WR-4	<u>RQ-2014-10-20-61</u> Big Blue Spring QA/QC Blank 1	<u>RQ-2014-10-20-62</u> wacissa Head Spring #2 Log Spring Thomas Spring Cassidy Spring
---	---	--

SAMPLER NAMES & SIGNATURES (All team members present)

Edgar Wade
Edgar Wade