



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID #

PRE-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 11/12/14 TIME 8:00

POST-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 11/13/14 TIME 8:00

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	8.69	21.630	98.7	P
CALIBRATION	8.80	21.624	99.9	P
PRE-VERIFICATION	8.80	21.624	99.9	P
POST-VERIFICATION	8.72	22.716	101.1	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	1000	2405350	10/15	999.1	P
CALIBRATION	1000	2405350	10/15	999.6	P
PRE-VERIFICATION	1000	2405350	10/15	999.6	P
POST-VERIFICATION	1000	2405350	10/15	999.7	P

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	10	135751	9/15	10.02	22.375	P
CALIBRATION	10	135751	9/15	10.00	22.356	P
PRE-VERIFICATION	10	135751	9/15	10.00	22.354	P
POST-VERIFICATION	10	135751	9/15	10.03	22.843	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES NO

RQ & SITE IDS



MORRISON SPRING
SPRG1410 MORRISON SPRING

RQ-2014-11-10-26

SPRG1410



CYPRESS SPRING

RQ-2014-11-10-25

SPRG1410

SAMPLER NAME

SPRG1410 CYPRESS SPRING

(List all members present)

K. Keller
K. Keller

Edgar Wade
Edgar Wade