



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID #

PRE-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 11/20/14 TIME 7:30

POST-SAMPLE CALIBRATION (circle one) ☒ IN-LAB ☐ ON-SITE ON-SITE LOCATION _____

DATE 11/20/14 TIME 7:30

DISSOLVED OXYGEN

Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	9.29	19.155	106.1	P
CALIBRATION	9.31	19.118	100.1	P
PRE-VERIFICATION	9.31	18.912	100.4	P
POST-VERIFICATION	9.48	18.881	101.8	P

SPECIFIC CONDUCTANCE

Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	1000	2402978	8/15	993.0	P
CALIBRATION	1000	2402978	8/15	1000.0	P
PRE-VERIFICATION	1000	2402978	8/15	1000.2	P
POST-VERIFICATION	1000	2402978	8/15	997.2	P

pH

Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7	140782	2/2014	7.09	18.382	P
CALIBRATION	7	140782	2/2014	7.00	18.385	P
PRE-VERIFICATION	7	140782	2/2014	7.00	18.397	P
POST-VERIFICATION	7	140782	2/2014	7.03	18.20975	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. $\pm 5\%$ pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES ☒ NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES ☒ NO

RQ & SITE IDS

RQ-2014-11-17-25

SPRG1410



BLUE HOLE SPRING (COLUMBIA)

SPRG1410 BLUE HOLE SPRING (COLUMBIA)

RQ-2014-11-17-26

SPRG1410



MISSION SPRINGS

SPRG1410 MISSION SPRINGS



MILL POND SPRINGS (COLUMBIA)

SPRG1410 MILL POND SPRINGS (COLUMBIA)

RQ2014-11-17-SP
Ginnie Springs
Gilchrist Blue Spring

SAMPLER NAMES & SIGNATURES (All team members present)

K. Eiler
K. Eiler