



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
GROUND WATER MANAGEMENT

CALIBRATION LOG

METER ID # _____

PRE-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 12/17/14 TIME 8:15

POST-SAMPLE CALIBRATION (circle one) IN-LAB ON-SITE ON-SITE LOCATION _____

DATE 12/18/14 TIME 8:15

DISSOLVED OXYGEN				
Reading Type	mg/L	Temp °C	% Sat	Pass or Fail
INITIAL	9.02	20.202	95.8	P
CALIBRATION	ND	20.308	100.3	P
PRE-VERIFICATION	9.06	20.329	100.3	P
POST-VERIFICATION	9.07	20.057	99.8	P

SPECIFIC CONDUCTANCE					
Reading Type	Standard	Lot #	Exp. Date	µS/cm	Pass or Fail
INITIAL	1000	CC12019	1/29/15	994.4	P
CALIBRATION	1000	CC12019	1/29/15	1000.0	P
PRE-VERIFICATION	1000	CC12019	1/29/15	999.9	P
POST-VERIFICATION	1000	CC12019	1/29/15	998.4	P

pH						
Reading Type	Standard	Lot #	Exp. Date	SU	Temp °C	Pass or Fail
INITIAL	7	143079	4/1/10	7.03	20.032	P
CALIBRATION	7	143079	4/1/10	7.00	19.864	P
PRE-VERIFICATION	7	143079	4/1/10	7.00	19.788	P
POST-VERIFICATION	7	143079	4/1/10	7.06	20.250	P

Acceptance Criteria: DO ± 0.3 mg/L Sp. Cond. ± 5% pH ± 0.2 SU Temperature ± 0.2 °C

Was the Specific Conductance Probe cleaned? (circle one) YES NO

Was the Dissolved Oxygen Membrane changed? (circle one) YES NO NA

RQ & SITE IDs

RQ-2014-11-10-26 SPRG1410

ICHETUCKNEE HEAD SPRING

SPRG1410 ICHETUCKNEE HEAD SPRING

RQ-2014-12-8-17

drainfield stations

ichetucknee

SAMPLER NAMES & SIGNATURES (All team members present)

A. Eller

Miller