

CALIBRATION

LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

Desi

RQ-

2023-01-02-48

Project:

Algae Res. FL

- Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

11/14/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	BD	1/12/23	841	21.2	764.8	8.9	8.9	100.6	✓	1.01	P/F	L/F
ICV	↓	↓	843	21.3	764.8	8.9	8.91	100.4	✓	1.01	P/F	L/F
CCV	↓	↓	1503	21.1	761.3	8.9	8.88	99.9	✓	1.01	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	BD	1/12/23	826	221028A	11/23	1000	1000	P/F	L/F
ICV	↓	↓	827	221028B	11/23	1413	1374	P/F	L/F
CCV	↓	↓	1505	221028A	11/23	1000	998	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	BD	1/12/23	829	M118-15	5/24	7.02	21.6	7.02	-21.2	P/F	L/F
Calibr.	↓	↓	833	M188-01	7/24	4.00	21.6	4.00	154.4	P/F	L/F
Calibr.	↓	↓	836	M103-01	4/24	10.03	21.5	10.03	-193.6	P/F	L/F
ICV	↓	↓	838	M118-15	5/24	7.01	21.6	7.03	-20.6	P/F	L/F
CCV	↓	↓	1507	M103-01	4/24	10.03	21.7	10.01	-193.2	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 172.4, slope from 4 to 7 175.6 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 11/14/22

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: ; Date of last verification:

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	BD	1/12/23	842	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2023-01-02-48				Collected By:		Brian Davis, Katrin Villinger					
Customer:		WQAP		Field Report Prepared By:		Brian Davis					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Black Creek @ SR-17				Comments: Windy							
Field ID: NEHAB0035				Coordinates:		30.04217, -81.71166					
WIN Secondary Resource Type: River/Stream				Collection Equip:		WaterBottle					
WIN ID: NEHAB0035				High Tide Time:				Low Tide Time:			
Matrix:		Fresh		Grab		Tide:		☐ Rising ☒ Falling ☐ Slack ☐ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
01/12/2023	10:50	68.9	6.95	14.9	6.47	0.3	3.4	0.7	217.5	0.1	
	10:55	68.7	6.97	14.7	6.5	2.9			235.3	0.11	
Bloom Observations											
Algal Bloom Observed? ☐ Yes ☒ No											
Water Clarity		☐ Clear		☐ Slightly Turbid		☒ Turbid		☐ Opaque			
Bloom Color		☐ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☐ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☐ Specks		☐ Scum		☐ Glob			
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☐ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algal ID (PT-50mL)		☐ ALGAL-ID				☐ Ice					
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	B	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	B	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	B	
		Acid Lot:		SA2229010		Exp:		08/30/23			
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	B	
				Filter Lot # 175533							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped				Received By:		Date/Time	
Brian Davis		01/12/2023 16:30		1							



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2023-01-02-48				Collected By:		Brian Davis, Katrin Villinger					
Customer:		WQAP		Field Report Prepared By:		Brian Davis					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Dr's Lake @ Mill Cove				Comments: Windy							
Field ID: NEHAB0057				Coordinates:		30.12496, -81.72740					
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle					
WIN ID: NEHAB0057				High Tide Time:				Low Tide Time:			
Matrix:		Fresh		Grab		Tide:		☐ Rising ☒ Falling ☐ Slack ☐ N/A			
Water Quality											
Date	Time ■ EST □ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
01/12/2023	12:00	92.3	9.1	16.1	7.39	0.3	2.6	0.7	860	0.42	
	00:02	80.2	7.97	15.5	7.23	2.1			876	0.43	
Bloom Observations											
Algal Bloom Observed? ☐ Yes ☒ No											
Water Clarity		☐ Clear		☐ Slightly Turbid		☒ Turbid		☐ Opaque			
Bloom Color		☐ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☐ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☐ Specks		☐ Scum		☐ Glob			
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☐ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algal ID (PT-50mL)		☐ ALGAL-ID				☐ Ice					
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	B	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	B	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP Acid Lot: SA2145160 Exp: 6/7/23				☒ Ice ☒ H2SO4		☒ <2	1	B	
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	B	
				Filter Lot # 175533							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time			
Brian Davis		01/12/2023 16:30		1							



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2023-01-02-48				Collected By:		Brian Davis, Katrin Villinger					
Customer:		WQAP		Field Report Prepared By:		Brian Davis					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Dr's Lake at Camp Echokotee				Comments: Windy							
Field ID: NEHAB0085				Coordinates:		30.15038, -81.72463					
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle					
WIN ID:NEHAB0085				High Tide Time:				Low Tide Time:			
Matrix:		Fresh		Grab		Tide:		☐ Rising ☒ Falling ☐ Slack ☐ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
01/12/2023	11:35	88.5	8.58	16.9	7.37	0.3	1.4	0.6	544	0.26	
Bloom Observations											
Algal Bloom Observed? ☐ Yes ☒ No											
Water Clarity		☒ Clear ☐ Slightly Turbid ☐ Turbid ☐ Opaque									
Bloom Color		☐ Green ☐ Red ☐ Brown ☐ Other:									
Algae Type		☐ Planktonic ☐ Filamentous ☐ Other:									
Algae Accumulation		☐ Mat ☐ Specks ☐ Scum ☐ Glob									
Algae Distribution		☐ Surface-Entire ☐ Surf-Localized ☐ Surf-Windblown ☐ Suspended ☐ Benthic ☐ Epiphytic									
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algal ID (PT-50mL)		☐ ALGAL-ID				☐ Ice					
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	B	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	B	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	B	
		Acid Lot:		SA2229010		Exp:		6/30/23			
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter			☒ Ice		1	B	
				Filter Lot # 175533							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped			Received By:		Date/Time		
Brian Davis		01/12/2023 16:30		1							



Cyanobacteria Bloom Response Field Data Sheet (2022-08-15 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2023-01-02-48				Collected By:		Brian Davis, Katrin Villinger					
Customer:		WQAP		Field Report Prepared By:		Brian Davis					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Swimming pen Cr @ Whitey's Fish Camp				Comments: Windy							
Field ID: NEHAB0074				Coordinates:		30.10103, -81.74641					
WIN Secondary Resource Type: River/Stream				Collection Equip:		WaterBottle					
WIN ID:NEHAB0074				High Tide Time:				Low Tide Time:			
Matrix:		Fresh		Grab		Tide:		☐ Rising ☒ Falling ☐ Slack ☐ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
01/12/2023	12:35	82.2	8.04	16.3	7.24	0.3	1	1	963	0.48	
Bloom Observations											
Algal Bloom Observed? ☐ Yes ☒ No											
Water Clarity		☒ Clear ☐ Slightly Turbid ☐ Turbid ☐ Opaque									
Bloom Color		☐ Green ☐ Red ☐ Brown ☐ Other:									
Algae Type		☐ Planktonic ☐ Filamentous ☐ Other:									
Algae Accumulation		☐ Mat ☐ Specks ☐ Scum ☐ Glob									
Algae Distribution		☐ Surface-Entire ☐ Surf-Localized ☐ Surf-Windblown ☐ Suspended ☐ Benthic ☐ Epiphytic									
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algal ID (PT-50mL)		☐ ALGAL-ID				☐ Ice					
Bloom ID (PT-50mL)		☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	B	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	B	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	B	
		Acid Lot:		SA2145160		Exp:		6/2/23			
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	B	
				Filter Lot # 175533							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time			
Brian Davis		01/12/2023 16:30		1							