

From: [Jeremy Bosso](#)
To: [Bunch, Cheryl](#); [DEP-Overflowmicro](#); [Dana Morton](#)
Subject: microbiology report for RQ-2021-01-11-56 (Bayou Chico BMAP) samples collected 01/11/2021
Date: Tuesday, January 12, 2021 6:41:31 PM
Attachments: [FF 2021 01 11A.txt](#)
[STF 2021 01 11A.pdf](#)
[DEP Reports 2021 Bayou Chico BMAP.xlsx](#)

Good afternoon. I've attached the microbiology report for RQ-2021-01-11-56 (Bayou Chico BMAP) samples collected 01/11/2021.

Samples analyzed by **Enterolert (QT)** were incubated from 01/11/2021 (16:35) to 01/12/2021 (17:00).

Samples analyzed by **SM 9223B (Colilert-18 QT)** were incubated from 01/11/2021 (16:45) to 01/12/2021 (12:15).

Note: The "J2" qualifier on sample 20-47299 (site 33020JD4 SW Arm Bayou Chico N of Mandalay Dr) is due to our lab duplicate for that sample being outside of our internal precision acceptance criteria. I believe this is most likely due to random error or because the duplicate results were both relatively low concentrations (8.4 versus 2.0 MPN/100 mL before dilution factor). We always report the higher number of duplicate analyses.

Thank you,

Jeremy Bosso, MPH
Technical Director, Wetlands Research Laboratory
University of West Florida
11000 University Pkwy
Pensacola, FL 32514
(850) 857-6096

Confidentiality Notice: This e-mail communication and any attachments may contain confidential and privileged information for the use of the designated recipient(s) named above only and is not to be passed on to other persons, unless otherwise stated. If you are not the intended recipient, you are hereby notified that you have received this communication in error and that any review, disclosure, dissemination, distribution or copying of it or its contents is prohibited. If you have received this communication in error, please delete the communication and *notify the sender that the message was received in error.*