

Boldly "X" this box if there are qualified data on this page.

RQ-

Project: CCHMA

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Date of Last Temperature Verification

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. **Steady-state & Galvanic Sensors:** DO Gain & Charge N/A.

Conductivity Acceptance criteria $\pm 5\%$

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

If **NO**, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: 33123; Date of last verification: 2/11/21

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or $\pm 0.05\text{m}$, whichever is greater.

Form effective August 1, 2018