

Certificate of Analysis List

For request number 1937787

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
2660501 STANDARD 800 NTU	4016	N/A	N/A	STABLCAL sup TS sup0 FORMAZIN

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **STABLCAL|sup|TS|sup0 FORMAZIN STANDARD 800 NTU**COMMODITY NUMBER: **2660501**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A4016****2/1/2024****2/5/2024**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	760 to 840 NTU	818.0 NTU

The expiration date is Apr 2025

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist