Chemical Analysis Report

CEN-DIST-2016-02-23-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Hart Branch SCI**
Request ID: **RQ-2016-02-22-52**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Timothy W. Fitzpatrick
Liang-Tsair Lin, Ph.D.
Colin Wright, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator Date Certified: 14-MAR-2016 16:07

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

A - Value reported is the mean of two or more determinations.
B - Results based on colony counts outside the acceptable range.
I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
J - Estimated value and/or the analysis did not meet established quality control criteria.
K - Actual value is known to be less than value given.
L - Actual value is known to be greater than value given.
N - Presumptive evidence of presence of material.
O - Sampled, but analysis lost or not performed.
Q - Sample held beyond normal holding time.
T - Value reported is less than the criterion of detection.
U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
V - Analyte was detected in both sample and method blank.
X - Too few individuals to calculate SCI value.
Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: SWEETWATER CREEK @ 75M E OF MARVILLE CR **Collection Date/Time: 02/22/2016 10:00**

Field ID: G5CE0051 Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1773674	EPA 180.1 Rev. 2.0	Turbidity	2.2	A	NTU	P299104	
	EPA 300.0 Rev. 2.1	Chloride	61		mg Cl/L	P299332	
		Sulfate	12		mg SO4/L	P299335	
	SM 2120 B	Color (true)	69	A	PCU	P299084	
	SM 2320 B-97	Total Alkalinity	179		mg CaCO3/L	P299660	
	SM 2540 C-97	TDS	310		mg/L	P299493	
	SM 2540 D-97	TSS	3	I	mg/L	P299430	
	SM 4500 F-C-97	Fluoride	0.13		mg F/L	P299331	
1773675	EPA 350.1 Rev. 2.0	Ammonia-N	0.041		mg N/L	P299633	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.85		mg N/L	P299508	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.34		mg N/L	P299497	
	EPA 365.1 Rev. 2.0	Total-P	0.13		mg P/L	P299425	
	SM 5310 B-00	Organic Carbon	17		mg C/L	P299377	
1773676	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.092		mg P/L	P299091	

Ref. Method and Comment:
SM 2540 D-97: The precision data is unavailable due to the low analyte concentration in the QC sample.

Quality Assurance Report
Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P299104

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P299332

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P299335

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P299633

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P299508

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report
Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P299497

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P299425

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P299091

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P299084

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P299660

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P299493

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P299430

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P299331

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P299377

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report
Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P299332

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	103		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P299335

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	99.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P299633

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.2		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P299508

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	97.8		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P299497

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P299425

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	97.2		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P299091

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	101		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P299660

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	95 - 105

Reference Method: SM 4500 F-C-97
Batch ID: P299331

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.3		P	94 - 108

Quality Assurance Report
Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
Batch ID: P299377

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	101		P	85 - 115

Quality Assurance Report
Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P299332

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1773691	Chloride	102		P	90 - 110
1773692	Chloride	103		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P299335

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1773691	Sulfate	100		P	90 - 110
1773692	Sulfate	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P299633

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1773675	Ammonia-N	101	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P299508

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1773418	Kjeldahl Nitrogen	95.5	97.5	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P299497

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1773675	NO2NO3-N	98.8		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P299425

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1773463	Total-P	101	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P299091

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1773702	O-Phosphate-P	103	103	P/P	90 - 110

Quality Assurance Report
Matrix Spike Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P299331

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1773459	Fluoride	96.5	98.9	P/P	94 - 107

Reference Method: SM 5310 B-00
Batch ID: P299377

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1773348	Organic Carbon	106	105	P/P	85 - 115

Quality Assurance Report
Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P299104

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1773674	Turbidity	1.83	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P299332

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1773692	Chloride	0.153	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P299335

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1773692	Sulfate	0.876	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P299633

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1773675	Ammonia-N	0.515	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P299508

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1773418	Kjeldahl Nitrogen	1.89	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P299497

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits

Quality Assurance Report
Precision

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P299497

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1773675	NO2NO3-N	0.368	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P299425

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1773463	Total-P	1.12	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P299091

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1773702	O-Phosphate-P	0.291	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P299084

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1773674	Color (true)	2.01	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P299660

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1773459	Total Alkalinity	0.851	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P299493

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1773450	TDS	1.63	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P299331

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1773459	Fluoride	1.45	Spike	P	0 - 3.7

Reference Method: SM 5310 B-00
Batch ID: P299377

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1773348	Organic Carbon	0.482	Spike	P	0 - 20

Quality Assurance Report
Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report
Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A67740
Included Lab Sample IDs: 1773674

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	101	P/P	90 - 110
Sulfate	99.1	97.7	P/P	90 - 110

Reference Method: SM 2120 B
Run ID: A67756
Included Lab Sample IDs: 1773674

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	94.3	95.2	P/P	85 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A67763
Included Lab Sample IDs: 1773676

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.9	99.1	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A67767
Included Lab Sample IDs: 1773674

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: SM 4500 F-C-97
Run ID: A67857
Included Lab Sample IDs: 1773674

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	98.5	99.0	P/P	90 - 110

Reference Method: SM 5310 B-00
Run ID: A67869
Included Lab Sample IDs: 1773675

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	100	99.8	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A67904
Included Lab Sample IDs: 1773675

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	102	P/P	90 - 110

Quality Assurance Report
Calibration Verification

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A67914
Included Lab Sample IDs: 1773675

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A67955
Included Lab Sample IDs: 1773675

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	100	99.2	P/P	90 - 110

Reference Method: SM 2320 B-97
Run ID: A67973
Included Lab Sample IDs: 1773674

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	104	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A67981
Included Lab Sample IDs: 1773675

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	95.4	98.0	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report
Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision		
					LCS	SMP	MS
EPA 180.1 Rev. 2.0	Turbidity					1.83	
EPA 300.0 Rev. 2.1	Chloride	103	103	102		0.153	
	Sulfate	99.7	101	100		0.876	
EPA 350.1 Rev. 2.0	Ammonia-N	99.2	101	102			0.515
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	97.8	95.5	97.5			1.89
EPA 353.2 Rev. 2.0	NO2NO3-N	100	98.8				0.368
EPA 365.1 Rev. 2.0	Total-P	97.2	101	102			1.12
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	101	103	103			0.291
SM 2120 B	Color (true)					2.01	
SM 2320 B-97	Total Alkalinity	102				0.851	
SM 2540 C-97	TDS					1.63	
SM 4500 F-C-97	Fluoride	96.3	96.5	98.9			1.45
SM 5310 B-00	Organic Carbon	101	106	105			0.482

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1773674
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1773674
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1773674
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1773675
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1773675
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1773675
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1773675
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1773676
SM 2120 B / E31780	True Color measured at 450 nm	1773674
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	1773674
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1773674
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1773674
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1773674
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1773675

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	02/23/2016			02/23/2016 15:58	Sima Feizi	1773674
EPA 300.0 Rev. 2.1	02/23/2016			02/25/2016	Ping Hua	1773674
EPA 350.1 Rev. 2.0	02/23/2016			03/02/2016	Diana M. Turner	1773675
EPA 351.2 Rev. 2.0	02/23/2016	03/01/2016	Sofia Elnemr	03/03/2016	Rochelle Y Jackson	1773675
EPA 353.2 Rev. 2.0	02/23/2016			03/01/2016	Peter D. Kaus	1773675
EPA 365.1 Rev. 2.0	02/23/2016	02/29/2016	Christopher Armour	02/29/2016	Lipi Saha	1773675
EPA 365.1 Rev. 2.0 dissolved	02/23/2016			02/23/2016 15:14	Bryan A. Werth	1773676
SM 2120 B	02/23/2016			02/23/2016 13:13	Rochelle Y Jackson	1773674
SM 2320 B-97	02/23/2016			03/03/2016	Dale L. Simmons	1773674
SM 2540 C-97	02/23/2016			02/24/2016	Yijie Li	1773674
SM 2540 D-97	02/23/2016			02/24/2016	Yijie Li	1773674
SM 4500 F-C-97	02/23/2016			02/24/2016	Dale L. Simmons	1773674
SM 5310 B-00	02/23/2016			02/25/2016	Sofia Elnemr	1773675