Chemical Analysis Report

CEN-DIST-2016-07-12-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Lake Eustis+LVI**
Request ID: **RQ-2016-07-04-50**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Timothy W. Fitzpatrick
Liang-Tsair Lin, Ph.D.
Colin Wright, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Environmental Administrator

Date Certified: 05-AUG-2016 15:06



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

A - Value reported is the mean of two or more determinations.
B - Results based on colony counts outside the acceptable range.
I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
J - Estimated value and/or the analysis did not meet established quality control criteria.
K - Actual value is known to be less than value given.
L - Actual value is known to be greater than value given.
N - Presumptive evidence of presence of material.
O - Sampled, but analysis lost or not performed.
Q - Sample held beyond normal holding time.
T - Value reported is less than the criterion of detection.
U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
V - Analyte was detected in both sample and method blank.
X - Too few individuals to calculate SCI value.
Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Lake Eustis Middle **Collection Date/Time: 07/11/2016 11:01**

Field ID: 20020368 Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1815186	EPA 180.1 Rev. 2.0	Turbidity	2.9		NTU	P307794	
	EPA 300.0 Rev. 2.1	Chloride	33		mg Cl/L	P308872	
		Sulfate	15		mg SO4/L	P308873	
	SM 2120 B	Color (true)	7.8		PCU	P307750	
	SM 2320 B-97	Total Alkalinity	123		mg CaCO3/L	P308094	
	SM 2540 C-97	TDS	226		mg/L	P308584	
	SM 2540 D-97	TSS	9	I	mg/L	P308394	
	SM 4500 F-C-97	Fluoride	0.20		mg F/L	P308099	
1815187	EPA 350.1 Rev. 2.0	Ammonia-N	0.006		mg N/L	P308060	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.1		mg N/L	P307994	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P308291	
	EPA 365.1 Rev. 2.0	Total-P	0.015		mg P/L	P307920	
	SM 5310 B-00	Organic Carbon	15		mg C/L	P308409	
1815188	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P307756	

Ref. Method and Comment:
SM 2540 C-97: The reported result is consistent with the laboratory conductivity, 372 umhos/cm.
SM 2540 D-97: The precision data is unavailable due to the low analyte concentration in the QC sample.

Quality Assurance Report
Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P307794

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P308872

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P308873

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P308060

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P307994

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report
Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P308291

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P307920

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P307756

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P307750

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P308094

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P308584

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P308394

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P308099

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P308409

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report
Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P308872

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	99.4		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P308873

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	95.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P308060

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	98.9		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P307994

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	97.0		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P308291

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	106		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P307920

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	95.9		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P307756

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	101		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P308094

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	104		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P308099

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.4		P	92 - 103

Quality Assurance Report
Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
Batch ID: P308409

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	98.0		P	85 - 115

Quality Assurance Report
Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P308872

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1814875	Chloride	94.9		P	90 - 110
1819272	Chloride	97.5	98.3	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P308873

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1814875	Sulfate	92.8		P	90 - 110
1819272	Sulfate	94.0	94.4	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P308060

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1815187	Ammonia-N	100	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P307994

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1815165	Kjeldahl Nitrogen	102	98.7	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P308291

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1815064	NO2NO3-N	98.1	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P307920

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1815166	Total-P	97.4	96.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P307756

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1815188	O-Phosphate-P	98.0	97.5	P/P	90 - 110

Quality Assurance Report
Matrix Spike Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P308099

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1814878	Fluoride	98.1	98.1	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P308409

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1815004	Organic Carbon	97.3	97.4	P/P	85 - 115

Quality Assurance Report
Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P307794

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1815118	Turbidity	2.52	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P308872

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1819272	Chloride	0.654	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P308873

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1819272	Sulfate	0.371	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P308060

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1815187	Ammonia-N	1.22	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P307994

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1815165	Kjeldahl Nitrogen	1.95	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P308291

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits

Quality Assurance Report
Precision

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P308291

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1815064	NO2NO3-N	4.37	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P307920

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1815166	Total-P	1.30	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P307756

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1815188	O-Phosphate-P	0.512	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P307750

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1815164	Color (true)	0.0208	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P308094

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1814878	Total Alkalinity	0.511	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P308584

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1815179	TDS	5.23	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P308099

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1814878	Fluoride	0.0	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P308409

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1815004	Organic Carbon	0.0848	Spike	P	0 - 20

Quality Assurance Report
Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report
Calibration Verification

Reference Method: SM 2120 B
Run ID: A70507
Included Lab Sample IDs: 1815186

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	96.4	98.1	P/P	90 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A70508
Included Lab Sample IDs: 1815188

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.1	101	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A70517
Included Lab Sample IDs: 1815186

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	101	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A70575
Included Lab Sample IDs: 1815187

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	97.5	98.2	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A70586
Included Lab Sample IDs: 1815187

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	102	102	P/P	90 - 110

Reference Method: SM 2320 B-97
Run ID: A70594
Included Lab Sample IDs: 1815186

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	104	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Run ID: A70594
Included Lab Sample IDs: 1815186

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	101	P/P	90 - 110

Quality Assurance Report
Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A70602
Included Lab Sample IDs: 1815187

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	94.3	96.6	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A70649
Included Lab Sample IDs: 1815187

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110

Reference Method: SM 5310 B-00
Run ID: A70692
Included Lab Sample IDs: 1815187

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.9	102	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A70802
Included Lab Sample IDs: 1815186

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	101	P/P	90 - 110
Sulfate	99.0	96.6	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report
Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			Precision		
						LCS	SMP	MS
EPA 180.1 Rev. 2.0	Turbidity						2.52	
EPA 300.0 Rev. 2.1	Chloride	99.4	97.5	98.3	94.9			0.654
	Sulfate	95.7	94.0	94.4	92.8			0.371
EPA 350.1 Rev. 2.0	Ammonia-N	98.9	100	102				1.22
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	97.0	102	98.7				1.95
EPA 353.2 Rev. 2.0	NO2NO3-N	106	98.1	103				4.37
EPA 365.1 Rev. 2.0	Total-P	95.9	97.4	96.0				1.30
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	101	98.0	97.5				0.512
SM 2120 B	Color (true)						0.0208	
SM 2320 B-97	Total Alkalinity	104					0.511	
SM 2540 C-97	TDS						5.23	
SM 4500 F-C-97	Fluoride	99.4	98.1	98.1				0.0
SM 5310 B-00	Organic Carbon	98.0	97.3	97.4				0.0848

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1815186
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1815186
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1815186
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1815187
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1815187
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1815187
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1815187
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1815188
SM 2120 B / E31780	True Color measured at 450 nm	1815186
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	1815186
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1815186
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1815186
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1815186
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1815187

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	07/12/2016			07/12/2016 15:57	Blanca Fach	1815186
EPA 300.0 Rev. 2.1	07/12/2016			07/29/2016	Ping Hua	1815186
EPA 350.1 Rev. 2.0	07/12/2016			07/15/2016	Diana M. Turner	1815187
EPA 351.2 Rev. 2.0	07/12/2016	07/15/2016	Sofia Elnemr	07/18/2016	Christopher Armour	1815187
EPA 353.2 Rev. 2.0	07/12/2016			07/20/2016	Beverly Y. Sanders	1815187
EPA 365.1 Rev. 2.0	07/12/2016	07/14/2016	Vijayalakshmi Reddy	07/15/2016	Lipi Saha	1815187
EPA 365.1 Rev. 2.0 dissolved	07/12/2016			07/12/2016 14:51	Julia Storbeck	1815188
SM 2120 B	07/12/2016			07/12/2016 15:27	Rochelle Y Jackson	1815186
SM 2320 B-97	07/12/2016			07/16/2016	Dale L. Simmons	1815186
SM 2540 C-97	07/12/2016			07/15/2016	Yijie Li	1815186
SM 2540 D-97	07/12/2016			07/15/2016	Yijie Li	1815186
SM 4500 F-C-97	07/12/2016			07/16/2016	Dale L. Simmons	1815186
SM 5310 B-00	07/12/2016			07/20/2016	Sofia Elnemr	1815187