

Chemical Analysis Report

CEN-DIST-2016-12-02-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Lake Van**
Request ID: **RQ-2016-12-05-61**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Environmental Administrator

Date Certified: 27-DEC-2016 13:00

A handwritten signature in black ink, appearing to read "Colin Wright", is displayed on a light gray rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: LAKE VAN @ NW CORNER OF W LOBE

Collection Date/Time: 12/01/2016 10:56

Field ID: G4CE0100

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1854096	EPA 180.1 Rev. 2.0	Turbidity	2.4	A	NTU	P316108	
	EPA 300.0 Rev. 2.1	Chloride	30		mg Cl/L	P316908	
		Sulfate	32		mg SO4/L	P316905	
	SM 2120 B	Color (true)	81		PCU	P316103	
	SM 2320 B-97	Total Alkalinity	15		mg CaCO3/L	P316305	
	SM 2540 C-97	TDS	119		mg/L	P316695	
	SM 2540 D-97	TSS	3	I	mg/L	P316682	
	SM 4500 F-C-97	Fluoride	0.29		mg F/L	P316307	
1854099	EPA 350.1 Rev. 2.0	Ammonia-N	0.011		mg N/L	P316473	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.74		mg N/L	P316461	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.013		mg N/L	P316399	
	EPA 365.1 Rev. 2.0	Total-P	0.072		mg P/L	P316340	
	SM 5310 B-00	Organic Carbon	12		mg C/L	P316151	
1854102	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.039		mg P/L	P316110	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Chloride: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: LAKE VAN @ SE CORNER OF W LOBE

Collection Date/Time: 12/01/2016 11:09

Field ID: G4CE0102

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1854097	EPA 180.1 Rev. 2.0	Turbidity	2.4		NTU	P316108	
	EPA 300.0 Rev. 2.1	Chloride	29		mg Cl/L	P316908	
		Sulfate	31		mg SO4/L	P316905	
	SM 2120 B	Color (true)	82		PCU	P316103	
	SM 2320 B-97	Total Alkalinity	15		mg CaCO3/L	P316306	
	SM 2540 C-97	TDS	132		mg/L	P316695	
	SM 2540 D-97	TSS	4	I	mg/L	P316682	
	SM 4500 F-C-97	Fluoride	0.30		mg F/L	P316308	
1854100	EPA 350.1 Rev. 2.0	Ammonia-N	0.015		mg N/L	P316473	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.70		mg N/L	P316461	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.015		mg N/L	P316399	
	EPA 365.1 Rev. 2.0	Total-P	0.072		mg P/L	P316340	
	SM 5310 B-00	Organic Carbon	12		mg C/L	P316151	
1854103	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.038		mg P/L	P316110	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Chloride: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: LAKE VAN @ SW CORNER OF W LOBE

Collection Date/Time: 12/01/2016 11:23

Field ID: G4CE0099

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
-----------	-------------	-----------	--------	------	-------	----------	----------

Field ID: G4CE0099

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1854098	EPA 180.1 Rev. 2.0	Turbidity	2.8		NTU	P316108	
	EPA 300.0 Rev. 2.1	Chloride	30		mg Cl/L	P316908	
		Sulfate	32		mg SO4/L	P316905	
	SM 2120 B	Color (true)	82		PCU	P316103	
	SM 2320 B-97	Total Alkalinity	15		mg CaCO3/L	P316306	
	SM 2540 C-97	TDS	148	A	mg/L	P316695	
	SM 2540 D-97	TSS	4	I	mg/L	P316682	
	SM 4500 F-C-97	Fluoride	0.30		mg F/L	P316308	
1854101	EPA 350.1 Rev. 2.0	Ammonia-N	0.012		mg N/L	P316473	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.71		mg N/L	P316461	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.015		mg N/L	P316399	
	EPA 365.1 Rev. 2.0	Total-P	0.071		mg P/L	P316340	
	SM 5310 B-00	Organic Carbon	12		mg C/L	P316151	
1854104	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.038		mg P/L	P316110	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Chloride: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P316108

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P316905

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P316908

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P316473

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P316461

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316399

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P316340

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P316110

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P316103

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P316305

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2320 B-97
Batch ID: P316306

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P316695

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P316682

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P316307

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Quality Assurance Report Method Blank Results

Reference Method: SM 4500 F-C-97
Batch ID: P316308

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P316151

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316905

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	97.0		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316908

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P316473

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	103		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316461

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	103		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316399

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P316340

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P316110

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P316110

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.5		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P316305

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	98.0		P	95 - 108

Reference Method: SM 2320 B-97
Batch ID: P316306

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	98.8		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P316307

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.3		P	92 - 103

Reference Method: SM 4500 F-C-97
Batch ID: P316308

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.3		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P316151

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316905

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1853877	Sulfate	95.4		P	90 - 110
1853878	Sulfate	96.2		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316908

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1853878	Chloride	101		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P316473

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1854099	Ammonia-N	99.8	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316461

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1854100	Kjeldahl Nitrogen	99.6	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316399

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1854099	NO2NO3-N	99.5	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P316340

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1854074	Total-P	96.2	93.7	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P316110

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1854113	O-Phosphate-P	95.9	95.9	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P316307

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1854131	Fluoride	102	101	P/P	91 - 105

Reference Method: SM 4500 F-C-97
Batch ID: P316308

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1854155	Fluoride	99.4	97.6	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P316151

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1853758	Organic Carbon	94.4	94.9	P/P	85 - 115

Quality Assurance Report Precision

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P316108

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1854096	Turbidity	1.26	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316905

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1853878	Sulfate	0.474	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316908

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1853878	Chloride	0.238	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P316473

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1854099	Ammonia-N	1.69	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316461

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1854100	Kjeldahl Nitrogen	1.34	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316399

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1854099	NO2NO3-N	3.25	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P316340

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1854074	Total-P	2.29	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P316110

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1854113	O-Phosphate-P	0.0	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B
Batch ID: P316103

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1854072	Color (true)	1.38	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P316305

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1854131	Total Alkalinity	0.287	Sample	P	0 - 4.8

Reference Method: SM 2320 B-97
Batch ID: P316306

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1854155	Total Alkalinity	0.388	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P316695

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1854098	TDS	4.05	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P316307

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1854131	Fluoride	0.503	Spike	P	0 - 3.0

Reference Method: SM 4500 F-C-97
Batch ID: P316308

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1854155	Fluoride	1.55	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P316151

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1853758	Organic Carbon	0.386	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B

Run ID: A73144

Included Lab Sample IDs: 1854096, 1854097, 1854098

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	99.3	101	P/P	85 - 115

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A73146

Included Lab Sample IDs: 1854096, 1854097, 1854098

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A73147

Included Lab Sample IDs: 1854102, 1854103, 1854104

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.4	98.8	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A73154

Included Lab Sample IDs: 1854099, 1854100, 1854101

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	99.1	97.8	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A73217

Included Lab Sample IDs: 1854096, 1854097, 1854098

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	97.3	98.2	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A73217

Included Lab Sample IDs: 1854096, 1854097, 1854098

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A73247

Included Lab Sample IDs: 1854099, 1854100, 1854101

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	99.5	98.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A73254

Included Lab Sample IDs: 1854099, 1854100, 1854101

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	103	103	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A73263

Included Lab Sample IDs: 1854099, 1854100, 1854101

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	100	98.4	P/P	90 - 110
Ammonia-N	98.4	99.6	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A73301

Included Lab Sample IDs: 1854099, 1854100, 1854101

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	101	97.3	P/P	90 - 110
Kjeldahl Nitrogen	97.3	98.2	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A73375

Included Lab Sample IDs: 1854096, 1854097, 1854098

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	103	104	P/P	90 - 110
Sulfate	99.6	101	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				1.26	
EPA 300.0 Rev. 2.1	Chloride	101	101		0.238	
	Sulfate	97.0	96.2 95.4		0.474	
EPA 350.1 Rev. 2.0	Ammonia-N	103	99.8 102			1.69
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	103	99.6 102			1.34
EPA 353.2 Rev. 2.0	NO2NO3-N	102	99.5 103			3.25
EPA 365.1 Rev. 2.0	Total-P	103	96.2 93.7			2.29
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.5	95.9 95.9			0.0
SM 2120 B	Color (true)				1.38	
SM 2320 B-97	Total Alkalinity	98.0			0.287	
	Total Alkalinity	98.8			0.388	
SM 2540 C-97	TDS				4.05	
SM 4500 F-C-97	Fluoride	98.3	102 101			0.503
	Fluoride	98.3	99.4 97.6			1.55
SM 5310 B-00	Organic Carbon	96.0	94.4 94.9			0.386

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1854096, 1854097, 1854098

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1854096, 1854097, 1854098
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1854096, 1854097, 1854098
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1854099, 1854100, 1854101
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1854099, 1854100, 1854101
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1854099, 1854100, 1854101
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1854099, 1854100, 1854101
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1854102, 1854103, 1854104
SM 2120 B / E31780	True Color measured at 450 nm	1854096, 1854097, 1854098
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1854096, 1854097, 1854098
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1854096, 1854097, 1854098
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1854096, 1854097, 1854098
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1854096, 1854097, 1854098
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1854099, 1854100, 1854101

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	12/02/2016			12/02/2016 14:41	Sima Feizi	1854096, 1854097, 1854098
EPA 300.0 Rev. 2.1	12/02/2016			12/15/2016	Ping Hua	1854096, 1854097, 1854098
EPA 350.1 Rev. 2.0	12/02/2016			12/08/2016	Sofia Elnemr	1854099, 1854100, 1854101
EPA 351.2 Rev. 2.0	12/02/2016	12/09/2016	Martin Acosta	12/12/2016	Joseph Suarez	1854099, 1854100, 1854101
EPA 353.2 Rev. 2.0	12/02/2016			12/08/2016	Beverly Y. Sanders	1854099, 1854100, 1854101
EPA 365.1 Rev. 2.0	12/02/2016	12/07/2016	Vijayalakshmi Reddy	12/08/2016	Lipi Saha	1854099, 1854100, 1854101
EPA 365.1 Rev. 2.0 dissolved	12/02/2016			12/02/2016 15:44	Julia Storbeck	1854102
	12/02/2016			12/02/2016 15:45	Julia Storbeck	1854103
	12/02/2016			12/02/2016 15:46	Julia Storbeck	1854104
SM 2120 B	12/02/2016			12/02/2016 14:29	Jared I. Manley	1854096, 1854097
	12/02/2016			12/02/2016 14:30	Jared I. Manley	1854098
SM 2320 B-97	12/02/2016			12/06/2016	Dale L. Simmons	1854096
	12/02/2016			12/07/2016	Dale L. Simmons	1854097, 1854098
SM 2540 C-97	12/02/2016			12/07/2016	Yijie Li	1854096, 1854097, 1854098
SM 2540 D-97	12/02/2016			12/07/2016	Yijie Li	1854096, 1854097, 1854098
SM 4500 F-C-97	12/02/2016			12/06/2016	Dale L. Simmons	1854096
	12/02/2016			12/07/2016	Dale L. Simmons	1854097, 1854098
SM 5310 B-00	12/02/2016			12/02/2016	Julie Michelle Frank	1854099, 1854100, 1854101