

Chemical Analysis Report

NW-DIST-2016-11-22-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Eglin Run WBID 468A 495 432 476 445 361**
Request ID: **RQ-2016-11-14-23**
Customer: **NW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
160 W Government St
Suite 208A
Pensacola, FL 32502-5794
Attn: Michael King

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 15-DEC-2016 10:39

A handwritten signature in black ink, appearing to read "Liang Lin", is shown on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: TURKEY CREEK ABV EGLIN RD 232

Collection Date/Time: 11/21/2016 10:50

Field ID: G3NW0093

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1851891	EPA 180.1 Rev. 2.0	Turbidity	1.0	A	NTU	P315620	
	EPA 300.0 Rev. 2.1	Chloride	2.9		mg Cl/L	P316113	
		Sulfate	0.50	I	mg SO4/L	P316117	
	SM 2120 B	Color (true)	5.9	I	PCU	P315640	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P315809	
	SM 2540 C-97	TDS	15	U	mg/L	P316226	
	SM 2540 D-97	TSS	2	I	mg/L	P316221	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P316309	
1851897	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P315895	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P316246	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.12		mg N/L	P316014	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P316082	
	SM 5310 B-00	Organic Carbon	0.74	I	mg C/L	P315820	
1851903	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P315623	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

SM 2120 B: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: JUNIPER CREEK @ HWY 85

Collection Date/Time: 11/21/2016 11:15

Field ID: G3NW0095

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1851892	EPA 180.1 Rev. 2.0	Turbidity	1.0		NTU	P315620	
	EPA 300.0 Rev. 2.1	Chloride	2.6		mg Cl/L	P316113	
		Sulfate	0.65		mg SO4/L	P316117	
	SM 2120 B	Color (true)	12		PCU	P315640	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P315809	
	SM 2540 C-97	TDS	15	U	mg/L	P316226	
	SM 2540 D-97	TSS	3	I	mg/L	P316221	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P316309	
1851898	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P315895	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.12	I	mg N/L	P316247	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.079		mg N/L	P316015	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P316082	
	SM 5310 B-00	Organic Carbon	1.1		mg C/L	P315820	
1851904	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P315623	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

SM 2120 B: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: LITTLE ROCKY CREEK ABV EGLIN RD 515

Collection Date/Time: 11/21/2016 11:50

Field ID: G3NW0103

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
-----------	-------------	-----------	--------	------	-------	----------	----------

Field ID: G3NW0103

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1851893	EPA 180.1 Rev. 2.0	Turbidity	0.83		NTU	P315620	
	EPA 300.0 Rev. 2.1	Chloride	2.3		mg Cl/L	P316113	
		Sulfate	0.77		mg SO4/L	P316117	
	SM 2120 B	Color (true)	12		PCU	P315640	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P315809	
	SM 2540 C-97	TDS	15	U	mg/L	P316226	
	SM 2540 D-97	TSS	2	I	mg/L	P316221	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P316309	
1851899	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P315895	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P316247	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.097		mg N/L	P316015	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P316082	
	SM 5310 B-00	Organic Carbon	1.1		mg C/L	P315820	
1851905	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P315623	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

SM 2120 B: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: OPEN BRANCH ABOVE ER214-374 C-72

Collection Date/Time: 11/21/2016 12:50

Field ID: G3NW0105

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1851894	EPA 180.1 Rev. 2.0	Turbidity	0.69		NTU	P315620	
	EPA 300.0 Rev. 2.1	Chloride	1.6		mg Cl/L	P316113	
		Sulfate	0.46	I	mg SO4/L	P316117	
	SM 2120 B	Color (true)	10		PCU	P315640	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P315809	
	SM 2540 C-97	TDS	15	U	mg/L	P316226	
	SM 2540 D-97	TSS	2	I	mg/L	P316221	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P316309	
1851900	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P315895	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.15	I	mg N/L	P316290	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.022		mg N/L	P316015	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P316082	
	SM 5310 B-00	Organic Carbon	1.1		mg C/L	P315820	
1851906	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P315623	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

SM 2120 B: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: ROCKY CREEK@ 214

Collection Date/Time: 11/21/2016 13:10

Field ID: G3NW0060

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1851895	EPA 180.1 Rev. 2.0	Turbidity	1.7		NTU	P315620	

Field ID: G3NW0060

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1851895	EPA 300.0 Rev. 2.1	Chloride	1.9		mg Cl/L	P316113	
		Sulfate	0.66		mg SO4/L	P316117	
	SM 2120 B	Color (true)	8.5		PCU	P315640	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P315809	
	SM 2540 C-97	TDS	15	U	mg/L	P316226	
	SM 2540 D-97	TSS	6	I	mg/L	P316221	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P316309	
1851901	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P315895	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P316290	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.088		mg N/L	P316015	
	EPA 365.1 Rev. 2.0	Total-P	0.002	I	mg P/L	P316082	
	SM 5310 B-00	Organic Carbon	0.72	I	mg C/L	P315820	
1851907	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P315623	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

SM 2120 B: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: ROCKY CREEK AT 200/201

Collection Date/Time: 11/21/2016 13:30

Field ID: G3NW0059

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1851896	EPA 180.1 Rev. 2.0	Turbidity	0.82		NTU	P315620	
	EPA 300.0 Rev. 2.1	Chloride	2.1		mg Cl/L	P316113	
		Sulfate	0.59		mg SO4/L	P316117	
	SM 2120 B	Color (true)	8.4		PCU	P315640	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P315809	
	SM 2540 C-97	TDS	15	I	mg/L	P316227	
	SM 2540 D-97	TSS	4	I	mg/L	P316222	
1851902	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P316309	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P315895	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P316290	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.11		mg N/L	P316015	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P316083	
1851908	SM 5310 B-00	Organic Carbon	0.78	I	mg C/L	P315820	
	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P315623	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

SM 2120 B: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P315620

Component	Result	Code	Units
-----------	--------	------	-------

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P315620

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316113

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316117

Component	Result	Code	Units
Sulfate	0.20	U	mg SO ₄ /L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P315895

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316246

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316247

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316290

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316014

Component	Result	Code	Units
NO ₂ NO ₃ -N	0.004	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316015

Component	Result	Code	Units
NO ₂ NO ₃ -N	0.004	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P316082

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P316083

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P315623

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P315640

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P315809

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P316226

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 C-97
Batch ID: P316227

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P316221

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P316222

Component	Result	Code	Units
TSS	2	U	mg/L

Quality Assurance Report Method Blank Results

Reference Method: SM 4500 F-C-97
Batch ID: P316309

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P315820

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316113

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.4		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316117

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	94.0		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P315895

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	105		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316246

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	102		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316247

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	104		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316290

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	104		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316014

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316014

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	99.0		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316015

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	97.0		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P316082

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	106		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P316083

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	105		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P315623

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.7		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P315809

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P316309

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.8		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P315820

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.2		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316113

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
---------------	-----------	---------	---------	-----------	----------------

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316113

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851937	Chloride	96.5		P	90 - 110
1851938	Chloride	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P315895

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1849691	Ammonia-N	97.3	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316246

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851555	Kjeldahl Nitrogen	97.0	94.7	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316247

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1852079	Kjeldahl Nitrogen	104	110	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316290

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851901	Kjeldahl Nitrogen	110	107	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316014

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851556	NO2NO3-N	94.5		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316015

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851898	NO2NO3-N	97.8	96.4	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P316082

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851899	Total-P	101	106	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P316083

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1852023	Total-P	103	107	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P315623

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851903	O-Phosphate-P	99.1	99.8	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P316309

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1852055	Fluoride	96.1	97.8	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P315820

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1851897	Organic Carbon	94.2	95.8	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P315620

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851891	Turbidity	9.52	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316113

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851937	Chloride	0.00177	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P316117

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851937	Sulfate	0.0167	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P315895

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1849691	Ammonia-N	2.10	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316246

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851555	Kjeldahl Nitrogen	2.11	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316247

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1852079	Kjeldahl Nitrogen	5.00	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P316290

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851901	Kjeldahl Nitrogen	2.38	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316014

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851556	NO2NO3-N	0.209	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P316015

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851898	NO2NO3-N	1.10	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P316082

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851899	Total-P	4.55	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P316083

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1852023	Total-P	3.66	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P315623

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851903	O-Phosphate-P	0.704	Spike	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P315809

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1852055	Total Alkalinity	0.568	Sample	P	0 - 4.8

Quality Assurance Report Precision

Reference Method: SM 2540 C-97
Batch ID: P316226

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851937	TDS	6.32	Sample	P	0 - 10

Reference Method: SM 2540 C-97
Batch ID: P316227

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1852039	TDS	4.22	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P316309

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1852055	Fluoride	1.55	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P315820

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1851897	Organic Carbon	1.62	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A72966
Included Lab Sample IDs: 1851891, 1851892, 1851893, 1851894, 1851895, 1851896

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A72967
Included Lab Sample IDs: 1851903, 1851904, 1851905, 1851906, 1851907, 1851908

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	99.9	P/P	90 - 110
O-Phosphate-P	99.5	99.4	P/P	90 - 110
O-Phosphate-P	99.9	99.0	P/P	90 - 110

Reference Method: SM 2120 B
Run ID: A72971
Included Lab Sample IDs: 1851891, 1851892, 1851893, 1851894, 1851895, 1851896

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
-----------	---------	---------	------------	----------------

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B

Run ID: A72971

Included Lab Sample IDs: 1851891, 1851892, 1851893, 1851894, 1851895, 1851896

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	101	102	P/P	85 - 115

Reference Method: SM 2320 B-97

Run ID: A73035

Included Lab Sample IDs: 1851891, 1851892, 1851893, 1851894, 1851895, 1851896

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	103	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A73041

Included Lab Sample IDs: 1851897, 1851898, 1851899, 1851900, 1851901, 1851902

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	95.4	94.7	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A73074

Included Lab Sample IDs: 1851897, 1851898, 1851899, 1851900, 1851901, 1851902

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	99.6	P/P	90 - 110
Ammonia-N	101	101	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A73111

Included Lab Sample IDs: 1851897, 1851898, 1851899, 1851900, 1851901, 1851902

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	92.5	91.5	P/P	90 - 110
NO2NO3-N	95.5	93.0	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A73131

Included Lab Sample IDs: 1851891, 1851892, 1851893, 1851894, 1851895, 1851896

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	103	P/P	90 - 110
Chloride	103	102	P/P	90 - 110
Sulfate	97.1	99.9	P/P	90 - 110
Sulfate	99.4	97.1	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A73187

Included Lab Sample IDs: 1851897, 1851898, 1851899, 1851900, 1851901, 1851902

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	98.0	98.2	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 4500 F-C-97

Run ID: A73217

Included Lab Sample IDs: 1851891, 1851892, 1851893, 1851894, 1851895, 1851896

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	99.5	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A73234

Included Lab Sample IDs: 1851897, 1851898, 1851899

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	100	97.3	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A73249

Included Lab Sample IDs: 1851900, 1851901, 1851902

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	101	98.7	P/P	90 - 110
Kjeldahl Nitrogen	103	101	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				9.52	
EPA 300.0 Rev. 2.1	Chloride	98.4	96.5 101		0.0017	
	Sulfate	94.0			0.0167	
EPA 350.1 Rev. 2.0	Ammonia-N	105	97.3 100			2.10
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	102	97.0 94.7			2.11
	Kjeldahl Nitrogen	104	104 110			5.00
	Kjeldahl Nitrogen	104	110 107			2.38
EPA 353.2 Rev. 2.0	NO2NO3-N	99.0	94.5			0.209
	NO2NO3-N	97.0	97.8 96.4			1.10
EPA 365.1 Rev. 2.0	Total-P	106	101 106			4.55
	Total-P	105	103 107			3.66
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.7	99.1 99.8			0.704
SM 2320 B-97	Total Alkalinity	103			0.568	
SM 2540 C-97	TDS				6.32	
	TDS				4.22	
SM 4500 F-C-97	Fluoride	98.8	96.1 97.8			1.55
SM 5310 B-00	Organic Carbon	99.2	94.2 95.8			1.62

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1851891, 1851892, 1851893, 1851894, 1851895, 1851896

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1851891, 1851892, 1851893, 1851894, 1851895, 1851896
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1851891, 1851892, 1851893, 1851894, 1851895, 1851896
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1851897, 1851898, 1851899, 1851900, 1851901, 1851902
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1851897, 1851898, 1851899, 1851900, 1851901, 1851902
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1851897, 1851898, 1851899, 1851900, 1851901, 1851902
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1851897, 1851898, 1851899, 1851900, 1851901, 1851902
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1851903, 1851904, 1851905, 1851906, 1851907, 1851908
SM 2120 B / E31780	True Color measured at 450 nm	1851891, 1851892, 1851893, 1851894, 1851895, 1851896
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1851891, 1851892, 1851893, 1851894, 1851895, 1851896
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1851891, 1851892, 1851893, 1851894, 1851895, 1851896
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1851891, 1851892, 1851893, 1851894, 1851895, 1851896
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1851891, 1851892, 1851893, 1851894, 1851895, 1851896
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1851897, 1851898, 1851899, 1851900, 1851901, 1851902

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	11/22/2016			11/22/2016 15:04	Sima Feizi	1851891, 1851892, 1851893, 1851894, 1851895, 1851896
EPA 300.0 Rev. 2.1	11/22/2016			12/02/2016	Ping Hua	1851891, 1851892, 1851893, 1851894, 1851895, 1851896
EPA 350.1 Rev. 2.0	11/22/2016			11/29/2016	Sofia Elnemr	1851897, 1851898, 1851899, 1851900, 1851901, 1851902
EPA 351.2 Rev. 2.0	11/22/2016	12/06/2016	Adrian Shelby	12/07/2016	Joseph Suarez	1851897, 1851898, 1851899
	11/22/2016	12/07/2016	Adrian Shelby	12/08/2016	Joseph Suarez	1851900, 1851901, 1851902
EPA 353.2 Rev. 2.0	11/22/2016			12/01/2016	Beverly Y. Sanders	1851897, 1851898, 1851899, 1851900, 1851901, 1851902
EPA 365.1 Rev. 2.0	11/22/2016	12/02/2016	Adrian Shelby	12/05/2016	Lipi Saha	1851897, 1851898, 1851899, 1851900, 1851901, 1851902
EPA 365.1 Rev. 2.0 dissolved	11/22/2016			11/22/2016 14:45	Julia Storbeck	1851903
	11/22/2016			11/22/2016 15:15	Julia Storbeck	1851904
	11/22/2016			11/22/2016 15:16	Julia Storbeck	1851905
	11/22/2016			11/22/2016 15:17	Julia Storbeck	1851906
	11/22/2016			11/22/2016 15:22	Julia Storbeck	1851907
	11/22/2016			11/22/2016 15:23	Julia Storbeck	1851908
	11/22/2016			11/22/2016 15:42	Jared I. Manley	1851891
SM 2120 B	11/22/2016			11/22/2016 15:43	Jared I. Manley	1851892, 1851893
	11/22/2016			11/22/2016 15:44	Jared I. Manley	1851894
	11/22/2016			11/22/2016 15:45	Jared I. Manley	1851895
	11/22/2016			11/22/2016 15:46	Jared I. Manley	1851896
	11/22/2016			11/29/2016	Dale L. Simmons	1851891, 1851892, 1851893, 1851894, 1851895, 1851896
SM 2320 B-97	11/22/2016			11/26/2016	Yijie Li	1851891, 1851892, 1851893, 1851894, 1851895, 1851896
SM 2540 C-97	11/22/2016			11/26/2016	Yijie Li	1851891, 1851892, 1851893, 1851894, 1851895, 1851896
SM 2540 D-97	11/22/2016			12/07/2016	Dale L. Simmons	1851891, 1851892, 1851893, 1851894, 1851895, 1851896
SM 4500 F-C-97	11/22/2016			11/28/2016	Julie Michelle Frank	1851897, 1851898, 1851899, 1851900, 1851901, 1851902
SM 5310 B-00	11/22/2016					