Chemical Analysis Report

NE-DIST-2018-08-09-04

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Nassau East**
Request ID: **RQ-2018-08-06-18**
Customer: **NE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
8800 Baymeadows Way West
Jacksonville, FL 32256-7577
Attn: Jeremy Parrish

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator Date Certified: 23-AUG-2018 10:42

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

A - Value reported is the mean of two or more determinations.
B - Results based on colony counts outside the acceptable range.
I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
J - Estimated value and/or the analysis did not meet established quality control criteria.
K - Actual value is known to be less than value given.
L - Actual value is known to be greater than value given.
N - Presumptive evidence of presence of material.
O - Sampled, but analysis lost or not performed.
Q - Sample held beyond normal holding time.
T - Value reported is less than the criterion of detection.
U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
V - Analyte was detected in both sample and method blank.
X - Too few individuals to calculate SCI value.
Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: UNNAMED TRIB TO LITTLE MILLS CR @ US 1 Collection Date/Time: 08/08/2018 10:10

Field ID: 19020069 Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2015866	EPA 180.1 Rev. 2.0	Turbidity	6.7		NTU	P349867	
	EPA 300.0 Rev. 2.1	Chloride	18	A	mg Cl/L	P350126	
		Sulfate	21	A	mg SO4/L	P350127	
	SM 2120 B	Color (true)	450		PCU	P349873	
	SM 2320 B-97	Total Alkalinity	102		mg CaCO3/L	P350148	
	SM 2540 C-97	TDS	251		mg/L	P350351	
	SM 2540 D-97	TSS	5	I	mg/L	P350301	
	SM 4500 F-C-97	Fluoride	0.12		mg F/L	P350151	
2015867	EPA 350.1 Rev. 2.0	Ammonia-N	0.25		mg N/L	P350238	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	2.0		mg N/L	P350156	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.081		mg N/L	P350255	
	EPA 365.1 Rev. 2.0	Total-P	0.49		mg P/L	P350352	
	SM 5310 B-00	Organic Carbon	42		mg C/L	P350250	
2015868	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.39		mg P/L	P349932	

Ref. Method and Comment:
SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report
Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P349867

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350126

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350127

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P350238

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P350156

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P350255

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P350352

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P349932

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P349873

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report
Method Blank Results

Reference Method: SM 2320 B-97
Batch ID: P350148

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P350351

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P350301

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P350151

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P350250

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report
Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350126

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.1		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350127

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	97.6		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P350238

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	98.0		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P350156

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	104		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P350255

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P350352

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	97.4		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P349932

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.0		P	90 - 110

Reference Method: SM 2120 B
Batch ID: P349873

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	96.9		P	85 - 115

Reference Method: SM 2320 B-97
Batch ID: P350148

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	98 - 107

Quality Assurance Report
Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P350151

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.5		P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P350250

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.8		P	85 - 115

Quality Assurance Report
Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350126

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015866	Chloride	96.2		P	90 - 110
2016079	Chloride	99.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350127

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015866	Sulfate	98.5		P	90 - 110
2016079	Sulfate	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P350238

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015853	Ammonia-N	97.8	99.0	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P350156

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015778	Kjeldahl Nitrogen	93.4	93.9	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P350255

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015852	NO2NO3-N	97.6	97.6	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P350352

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015886	Total-P	99.5	96.2	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P349932

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015701	O-Phosphate-P	95.0	95.8	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P350151

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2016260	Fluoride	99.4	98.0	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P350250

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2015852	Organic Carbon	105		P	85 - 115

Quality Assurance Report
Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P349867

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015773	Turbidity	2.06	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350126

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015866	Chloride	1.12	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350127

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015866	Sulfate	1.79	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P350238

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015853	Ammonia-N	0.683	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P350156

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015778	Kjeldahl Nitrogen	0.266	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P350255

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015852	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P350352

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015886	Total-P	3.00	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P349932

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015701	O-Phosphate-P	0.599	Spike	P	0 - 20

Quality Assurance Report
Precision

Reference Method: SM 2120 B
Batch ID: P349873

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015883	Color (true)	0.606	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P350148

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2016260	Total Alkalinity	0.622	Sample	P	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P350351

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015770	TDS	0.731	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P350151

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2016260	Fluoride	0.953	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P350250

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2015852	Organic Carbon	0.603	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report
Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A85783
Included Lab Sample IDs: 2015866

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	101	P/P	95 - 105

Reference Method: SM 2120 B
Run ID: A85785
Included Lab Sample IDs: 2015866

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	99.8	P/P	90 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A85809
Included Lab Sample IDs: 2015868

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	97.5	97.7	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A85845
Included Lab Sample IDs: 2015866

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	96.8	102	P/P	90 - 110
Sulfate	96.6	102	P/P	90 - 110

Reference Method: SM 2320 B-97
Run ID: A85893
Included Lab Sample IDs: 2015866

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	102	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Run ID: A85893
Included Lab Sample IDs: 2015866

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A85928
Included Lab Sample IDs: 2015867

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.2	99.4	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A85932
Included Lab Sample IDs: 2015867

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.0	99.0	P/P	90 - 110

Quality Assurance Report
Calibration Verification

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A85934
Included Lab Sample IDs: 2015867

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	101	P/P	90 - 110

Reference Method: SM 5310 B-00
Run ID: A85935
Included Lab Sample IDs: 2015867

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	96.6	96.4	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A85998
Included Lab Sample IDs: 2015867

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	99.9	102	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report
Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision		
					LCS	SMP	MS
EPA 180.1 Rev. 2.0	Turbidity					2.06	
EPA 300.0 Rev. 2.1	Chloride	98.1	96.2	99.8		1.12	
	Sulfate	97.6	98.5	100		1.79	
EPA 350.1 Rev. 2.0	Ammonia-N	98.0	97.8	99.0			0.683
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	104	93.4	93.9			0.266
EPA 353.2 Rev. 2.0	NO2NO3-N	102	97.6	97.6			0.0
EPA 365.1 Rev. 2.0	Total-P	97.4	99.5	96.2			3.00
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.0	95.0	95.8			0.599
SM 2120 B	Color (true)	96.9				0.606	
SM 2320 B-97	Total Alkalinity	103				0.622	
SM 2540 C-97	TDS					0.731	
SM 4500 F-C-97	Fluoride	98.5	99.4	98.0			0.953
SM 5310 B-00	Organic Carbon	95.8	105				0.603

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2015866
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2015866
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2015866
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2015867
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2015867
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2015867
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2015867
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2015868
SM 2120 B / E31780	True Color measured at 450 nm	2015866
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	2015866
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2015866
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2015866
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2015866
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2015867

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/09/2018			08/09/2018 16:23	Sima Feizi	2015866
EPA 300.0 Rev. 2.1	08/09/2018			08/14/2018 15:24	Lipi Saha	2015866
EPA 350.1 Rev. 2.0	08/09/2018			08/16/2018 12:25	Ping Hua	2015867
EPA 351.2 Rev. 2.0	08/09/2018	08/15/2018 10:30	Adrian Shelby	08/16/2018 12:46	Joseph Suarez	2015867
EPA 353.2 Rev. 2.0	08/09/2018			08/16/2018 13:38	Lauren Bottorf	2015867
EPA 365.1 Rev. 2.0	08/09/2018	08/17/2018 14:01	Sima Feizi	08/20/2018 09:24	Beverly Y. Sanders	2015867
EPA 365.1 Rev. 2.0 dissolved	08/09/2018			08/09/2018 15:02	Megan Treadwell	2015868
SM 2120 B	08/09/2018			08/09/2018 16:17	Tanya Welborn	2015866
SM 2320 B-97	08/09/2018			08/15/2018 07:10	Dale L. Simmons	2015866
SM 2540 C-97	08/09/2018			08/13/2018 17:05	William L. Brown IV	2015866
SM 2540 D-97	08/09/2018			08/13/2018 17:05	William L. Brown IV	2015866
SM 4500 F-C-97	08/09/2018			08/15/2018 07:10	Dale L. Simmons	2015866
SM 5310 B-00	08/09/2018			08/16/2018 01:14	Megan Treadwell	2015867