



September 21, 2018

Service Request No:J1806723

Mr. Jeremy Parrish
Florida DEP
7825 Baymeadows Way, Suite 200-B
Jacksonville, FL 32256-7577

Laboratory Results for: Pellicer WQ

Dear Mr.Parrish,

Enclosed are the results of the sample(s) submitted to our laboratory September 19, 2018
For your reference, these analyses have been assigned our service request number **J1806723**.

All analyses were performed according to our laboratory's quality assurance program. The test results meet requirements of the NELAP standards except as noted in the case narrative report. All results are intended to be considered in their entirety, and ALS Environmental is not responsible for use of less than the complete report. Results apply only to the items submitted to the laboratory for analysis and individual items (samples) analyzed, as listed in the report. In accordance to the NELAC 2003 Standard, a statement on the estimated uncertainty of measurement of any quantitative analysis will be supplied upon request.

Respectfully submitted,

ALS Group USA, Corp. dba ALS Environmental

Wendie Lodge
Project Manager

ADDRESS 9143 Philips Highway, Suite 200, Jacksonville, FL 32256
PHONE +1 904 739 2277 | FAX +1 904 739 2011
ALS Group USA, Corp.
dba ALS Environmental



Narrative Documents

ALS Environmental—Jacksonville Laboratory
9143 Philips Highway, Suite 200, Jacksonville, FL 32256
Phone (904) 739-2277 Fax (904) 739-2011
www.alsglobal.com

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28
Sample Matrix: Water

Service Request: J1806723
Date Received: 9/19/18

CASE NARRATIVE

All analyses were performed consistent with the quality assurance program of ALS Environmental. This report contains analytical results for samples designated for Tier I data deliverables. When appropriate to the procedure, method blank results have been reported with each analytical test. Analytical procedures performed by the lab are validated in accordance with NELAC standards. Parameters that are included in the NELAC Fields of Testing but are not included in the lab's NELAC accreditation are identified in the discussion of each analytical procedure.

Sample Receipt

Ten water samples were received for analysis at ALS Environmental on 9/19/18. The samples were received in good condition and consistent with the accompanying chain of custody form. Samples are refrigerated at $\leq 6^{\circ}\text{C}$ upon receipt at the lab except for aqueous samples designated for metals analyses, which are stored at room temperature.

Microbiology Analyses:

Method ASTM D6503: Samples were removed from the incubator and read on 9/20/18 at 17:16.

Method SM9222D: Samples J1806723 (1, 2, 7, 8, 9) were removed from the incubator and read on 9/20/18 at 15:12. Samples J1806723 (3, 4, 5, 6, 10) were removed from the incubator and read on 9/20/18 at 17:04.

Method SM9222D: Results for samples J1806723 (1, 10) are based on colony counts outside the optimal range of 20-60 CFU. The data has been qualified. This is an informational flag and does not impact the quality of the data.

Approved by Wende Lodge Date 9/21/2018

SAMPLE DETECTION SUMMARY

CLIENT ID: PC7	Lab ID: J1806723-001
-----------------------	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Coliform, Fecal	10	B	10	10	CFU/100m L	SM 9222 D
Enterococcus	31		10	10	MPN/100m L	ASTM D6503-99

CLIENT ID: PC6	Lab ID: J1806723-002
-----------------------	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Enterococcus	210		10	10	MPN/100m L	ASTM D6503-99
Coliform, Fecal	200		10	10	CFU/100m L	SM 9222 D

CLIENT ID: PC5	Lab ID: J1806723-003
-----------------------	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Enterococcus	95		10	10	MPN/100m L	ASTM D6503-99
Coliform, Fecal	280		10	10	CFU/100m L	SM 9222 D

CLIENT ID: PC4	Lab ID: J1806723-004
-----------------------	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Coliform, Fecal	420		10	10	CFU/100m L	SM 9222 D
Enterococcus	190		10	10	MPN/100m L	ASTM D6503-99

CLIENT ID: CB	Lab ID: J1806723-005
----------------------	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Enterococcus	310		10	10	MPN/100m L	ASTM D6503-99
Coliform, Fecal	240		10	10	CFU/100m L	SM 9222 D

CLIENT ID: PC3	Lab ID: J1806723-006
-----------------------	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Enterococcus	490		10	10	MPN/100m L	ASTM D6503-99
Coliform, Fecal	430		10	10	CFU/100m L	SM 9222 D

CLIENT ID: PC2	Lab ID: J1806723-007
-----------------------	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Coliform, Fecal	310		10	10	CFU/100m L	SM 9222 D
Enterococcus	420		10	10	MPN/100m L	ASTM D6503-99

CLIENT ID: PC1	Lab ID: J1806723-008
-----------------------	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Coliform, Fecal	340		10	10	CFU/100m L	SM 9222 D

SAMPLE DETECTION SUMMARY

CLIENT ID: PC1	Lab ID: J1806723-008
-----------------------	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Enterococcus	330		10	10	MPN/100m L	ASTM D6503-99

CLIENT ID: SB	Lab ID: J1806723-009
----------------------	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Coliform, Fecal	440		10	10	CFU/100m L	SM 9222 D
Enterococcus	280		10	10	MPN/100m L	ASTM D6503-99

CLIENT ID: DB	Lab ID: J1806723-010
----------------------	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Coliform, Fecal	90	B	10	10	CFU/100m L	SM 9222 D
Enterococcus	63		10	10	MPN/100m L	ASTM D6503-99



Sample Receipt Information

ALS Environmental—Jacksonville Laboratory
9143 Philips Highway, Suite 200, Jacksonville, FL 32256
Phone (904) 739-2277 Fax (904) 739-2011
www.alsglobal.com

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28

Service Request:J1806723

SAMPLE CROSS-REFERENCE

<u>SAMPLE #</u>	<u>CLIENT SAMPLE ID</u>	<u>DATE</u>	<u>TIME</u>
J1806723-001	PC7	9/19/2018	0931
J1806723-002	PC6	9/19/2018	1010
J1806723-003	PC5	9/19/2018	1043
J1806723-004	PC4	9/19/2018	1113
J1806723-005	CB	9/19/2018	1135
J1806723-006	PC3	9/19/2018	1206
J1806723-007	PC2	9/19/2018	0937
J1806723-008	PC1	9/19/2018	1033
J1806723-009	SB	9/19/2018	0940
J1806723-010	DB	9/19/2018	1124

Cooler Receipt Form

Client: FDEP Service Request #: J1806723
 Project: Pellicer WQ Shipping paid by ALS? Yes ☐ No ☒ N/A
 Cooler received on 9/19/18 and opened on 9/19/18 by BB
 COURIER: ALS UPS FEDEX DHL ☒ Client Other ☐ Airbill # _____

- 1 Were custody seals on outside of cooler? Yes ☐ No ☒
 If yes, how many and where? #: _____ on lid other _____
- 2 Were seals intact and signature and date correct? Yes ☐ No ☒ N/A
- 3 Were custody papers properly filled out? Yes ☒ No ☐ N/A
- 4 Temperature of cooler(s) upon receipt (Should be 0°C and ≤ 6°C) 3.2°C
- 5 Thermometer ID 1124
- 6 Temperature Blank Present? Yes ☐ No ☒
- 7 Were Ice or Ice Packs present Ice ☒ Ice Packs ☐ No ☐
- 8 Did all bottles arrive in good condition (unbroken, etc....)? Yes ☒ No ☐ N/A
- 9 Type of packing material present Netting ☐ Vial Holder ☐ Bubble Wrap ☐
 Paper ☐ Styrofoam ☐ Other ☒ N/A
- 10 Were all bottle labels complete (sample ID, preservation, etc....)? Yes ☒ No ☐ N/A
- 11 Did all bottle labels and tags agree with custody papers? Yes ☒ No ☐ N/A
- 12 Were the correct bottles used for the tests indicated? Yes ☒ No ☐ N/A
- 13 Were all of the preserved bottles received with the appropriate preservative?
 HNO3 pH<2 H2SO4 pH<2 ZnAc2/NaOH pH>9 NaOH pH>12 HCl pH<2
 Preservative additions noted below Yes ☒ No ☐ N/A
- 14 Were all samples received within analysis holding times? Yes ☒ No ☐ N/A
- 15 Were VOA vials free of air bubbles greater than 6mm? If present, note below Yes ☐ No ☒ N/A
- 16 Where did the bottles originate? ALS ☒ Client ☐

Sample ID	Reagent	Lot #	ml added	Initials Date/Time

Additional comments and/or explanation of all discrepancies noted above:

provided Field Blank, not on CoC

Client approval to run samples if discrepancies noted:

Date:

CHAIN OF CUSTODY/LABORATORY ANALYSIS REQUEST FORM

J1806723

 Florida DEP
Pellicer WQ

5

9143 Philips Highway, Ste 200 • Jacksonville, FL 32256 (904) 739-2277 • 800-695-7222 x06 • FAX (904) 739-2011 PAGE _____ OF _____



Project Name Pellicer WQ		Project Number RQ-2018-09-03-L8		ANALYSIS REQUESTED (Include Method Number and Container Preservative)																			
Project Manager Nikki Dix		Email Address nikki.dix@floridadep.gov																					
Company/Address DEP/GTM NERR				NUMBER OF CONTAINERS <i>Entero</i> <i>Fecal Coli</i>	PRESERVATIVE																Preservative Key 0. NONE 1. HCL 2. HNO ₃ 3. H ₂ SO ₄ 4. NaOH 5. Zn. Acetate 6. MeOH 7. NaHSO ₄ 8. Other _____		
SOS At Guna River Rd.																							
Ponte Votra Beach 32082																							
Phone # 904-823-4500 FAX # _____																							
Sampler's Signature _____				Sampler's Printed Name _____		REMARKS/ ALTERNATE DESCRIPTION																	
CLIENT SAMPLE ID		LAB ID		SAMPLING DATE TIME MATRIX																			
PC7				9/19/18 9:31 Liquid																			
PC6				10:10																			
PC5				10:43																			
PC4				11:13																			
CB				11:35																			
PC3				12:06																			
PC2				9:37																			
PC1				10:33																			
SR				9:40																			
DB				11:24																			
SPECIAL INSTRUCTIONS/COMMENTS 3.2⁰⁰						TURNAROUND REQUIREMENTS _____ RUSH (SURCHARGES APPLY) _____ STANDARD REQUESTED FAX DATE _____ REQUESTED REPORT DATE _____						REPORT REQUIREMENTS _____ I. Results Only _____ II. Results + QC Summaries (LCS, DUP, MS/MSD as required) _____ III. Results + QC and Calibration Summaries _____ IV. Data Validation Report with Raw Data _____ V. Specialized Forms / Custom Report Edata _____ Yes _____ No						INVOICE INFORMATION PO # _____ BILL TO: _____					
SAMPLE RECEIPT: CONDITION/COOLER TEMP: _____ CUSTODY SEALS: Y N						RECEIVED BY						RECEIVED BY											
RELINQUISHED BY		RECEIVED BY		RELINQUISHED BY																			
Signature _____		Signature _____		Signature _____																			
Printed Name _____		Printed Name _____		Printed Name _____																			
Firm _____		Firm _____		Firm _____		RECEIVED BY						RECEIVED BY											
Date/Time _____		Date/Time _____		Date/Time _____																			



Miscellaneous Forms

ALS Environmental—Jacksonville Laboratory
9143 Philips Highway, Suite 200, Jacksonville, FL 32256
Phone (904) 739-2277 Fax (904) 739-2011
www.alsglobal.com



FLORIDA DEP DATA QUALIFIERS

- B Results based upon colony counts outside the acceptable range.
- D Measurement was made in the field.
- H Value based on field kit determination; results may not be accurate.
- I The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J Estimated value (one of the following reasons is discussed in the project case narrative).
1. The result may be inaccurate because the surrogate recovery limits have been exceeded.
 2. No known quality control criteria exists for the component.
 3. The reported value failed to meet the established quality control criteria for either precision or accuracy.
 4. The sample matrix interfered with the ability to make any accurate determination (e.g., primary and confirmation results show greater than 40% RPD).
 5. The data is questionable because of improper laboratory or field protocols (e.g., GC/MS Tune did not meet method criteria).
- K Off scale low. The value is less than the lowest calibration standard but greater than the method reporting limit (MRL).
- L Off scale high. The analyte is above the upper limit of the linear calibration range.
- M The MDL/MRL has been elevated because the analyte could not be accurately quantified due to matrix interference.
- N Presumptive evidence of the analyte. Confirmation was not performed.
- Q Sample held beyond the accepted holding time.
- T Value reported is less than the laboratory method detection limit. The value is reported for informational purposes only.
- U Indicates that the compound was analyzed for but not detected.
- V Indicates that the analyte was detected in both the sample and the associated method blank.
- Y The laboratory analysis was from an improperly preserved sample.
- Z Too many colonies were present (TNTC). The numeric value represents the filtration volume.



Jacksonville Lab ID # for State Certifications¹

Agency	Number	Expiration Date
Department of Defense	66206	6/30/2020
Florida Department of Health	E82502	6/30/2019
Georgia Department of Natural Resources	958	6/30/2019
Kentucky Division of Waste Management	123042	6/30/2019
Louisiana Department of Environmental Quality	02086	6/30/2019
Maine Department of Health and Human Services	2017003	2/3/2019
North Carolina Department of Environment and Natural Resources	527	12/31/2018
South Carolina Department of Health and Environmental Control	96021001	6/30/2018
Texas Commission on Environmental Quality	T104704197-18-10	5/31/2019
Virginia Environmental Accreditation Program	460191	12/14/2018

¹ Analyses were performed according to our laboratory's NELAP-approved quality assurance program and any applicable state or agency requirements. The test results meet requirements of the current NELAP/TNI standards or state or agency requirements, where applicable, except as noted in the laboratory case narrative provided. For a specific list of accredited analytes, refer to <http://www.alsglobal.com/en/Our-Services/Life-Sciences/Environmental/Downloads/North-America-Downloads>



ACRONYMS

ASTM	American Society for Testing and Materials
A2LA	American Association for Laboratory Accreditation
CARB	California Air Resources Board
CAS Number	Chemical Abstract Service registry Number
CFC	Chlorofluorocarbon
CFU	Colony-Forming Unit
DEC	Department of Environmental Conservation
DEQ	Department of Environmental Quality
DHS	Department of Health Services
DOE	Department of Ecology
DOH	Department of Health
EPA	U. S. Environmental Protection Agency
ELAP	Environmental Laboratory Accreditation Program
GC	Gas Chromatography
GC/MS	Gas Chromatography/Mass Spectrometry
LUFT	Leaking Underground Fuel Tank
M	Modified
MCL	Maximum Contaminant Level is the highest permissible concentration of a substance allowed in drinking water as established by the USEPA.
MDL	Method Detection Limit
MPN	Most Probable Number
MRL	Method Reporting Limit
NA	Not Applicable
NC	Not Calculated
NCASI	National Council of the Paper Industry for Air and Stream Improvement
ND	Not Detected
NIOSH	National Institute for Occupational Safety and Health
PQL	Practical Quantitation Limit
RCRA	Resource Conservation and Recovery Act
SIM	Selected Ion Monitoring
TPH	Total Petroleum Hydrocarbons
tr	Trace level is the concentration of an analyte that is less than the PQL but greater than or equal to the MDL.

ALS Group USA, Corp.
dba ALS Environmental

Analyst Summary report

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28

Service Request: J1806723

Sample Name: PC7
Lab Code: J1806723-001
Sample Matrix: Water

Date Collected: 09/19/18
Date Received: 09/19/18

Analysis Method
ASTM D6503-99
SM 9222 D

Extracted/Digested By

Analyzed By
HHERNANDEZ
HHERNANDEZ

Sample Name: PC6
Lab Code: J1806723-002
Sample Matrix: Water

Date Collected: 09/19/18
Date Received: 09/19/18

Analysis Method
ASTM D6503-99
SM 9222 D

Extracted/Digested By

Analyzed By
HHERNANDEZ
HHERNANDEZ

Sample Name: PC5
Lab Code: J1806723-003
Sample Matrix: Water

Date Collected: 09/19/18
Date Received: 09/19/18

Analysis Method
ASTM D6503-99
SM 9222 D

Extracted/Digested By

Analyzed By
HHERNANDEZ
HHERNANDEZ

Sample Name: PC4
Lab Code: J1806723-004
Sample Matrix: Water

Date Collected: 09/19/18
Date Received: 09/19/18

Analysis Method
ASTM D6503-99
SM 9222 D

Extracted/Digested By

Analyzed By
HHERNANDEZ
HHERNANDEZ

ALS Group USA, Corp.

dba ALS Environmental

Analyst Summary report

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28

Service Request: J1806723

Sample Name: CB
Lab Code: J1806723-005
Sample Matrix: Water

Date Collected: 09/19/18**Date Received:** 09/19/18

Analysis Method
ASTM D6503-99
SM 9222 D

Extracted/Digested By

Analyzed By
HHERNANDEZ
HHERNANDEZ

Sample Name: PC3
Lab Code: J1806723-006
Sample Matrix: Water

Date Collected: 09/19/18**Date Received:** 09/19/18

Analysis Method
ASTM D6503-99
SM 9222 D

Extracted/Digested By

Analyzed By
HHERNANDEZ
HHERNANDEZ

Sample Name: PC2
Lab Code: J1806723-007
Sample Matrix: Water

Date Collected: 09/19/18**Date Received:** 09/19/18

Analysis Method
ASTM D6503-99
SM 9222 D

Extracted/Digested By

Analyzed By
HHERNANDEZ
HHERNANDEZ

Sample Name: PC1
Lab Code: J1806723-008
Sample Matrix: Water

Date Collected: 09/19/18**Date Received:** 09/19/18

Analysis Method
ASTM D6503-99
SM 9222 D

Extracted/Digested By

Analyzed By
HHERNANDEZ
HHERNANDEZ

ALS Group USA, Corp.
dba ALS Environmental

Analyst Summary report

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28

Service Request: J1806723

Sample Name: SB
Lab Code: J1806723-009
Sample Matrix: Water

Date Collected: 09/19/18
Date Received: 09/19/18

Analysis Method
ASTM D6503-99
SM 9222 D

Extracted/Digested By

Analyzed By
HHERNANDEZ
HHERNANDEZ

Sample Name: DB
Lab Code: J1806723-010
Sample Matrix: Water

Date Collected: 09/19/18
Date Received: 09/19/18

Analysis Method
ASTM D6503-99
SM 9222 D

Extracted/Digested By

Analyzed By
HHERNANDEZ
HHERNANDEZ



Sample Results

ALS Environmental—Jacksonville Laboratory
9143 Philips Highway, Suite 200, Jacksonville, FL 32256
Phone (904) 739-2277 Fax (904) 739-2011
www.alsglobal.com



Microbiology

ALS Environmental—Jacksonville Laboratory
9143 Philips Highway, Suite 200, Jacksonville, FL 32256
Phone (904)739-2277 Fax (904)739-2011
www.alsglobal.com

ALS Group USA, Corp.
dba ALS Environmental

Analytical Report

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28
Sample Matrix: Water
Sample Name: PC7
Lab Code: J1806723-001

Service Request: J1806723
Date Collected: 09/19/18 09:31
Date Received: 09/19/18 14:00
Basis: NA

General Chemistry Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Coliform, Fecal	SM 9222 D	10 B	CFU/100mL	10	10	09/19/18 17:01	
Enterococcus	ASTM D6503-99	31	MPN/100mL	10	10	09/19/18 16:05	

ALS Group USA, Corp.
dba ALS Environmental

Analytical Report

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28
Sample Matrix: Water
Sample Name: PC6
Lab Code: J1806723-002

Service Request: J1806723
Date Collected: 09/19/18 10:10
Date Received: 09/19/18 14:00
Basis: NA

General Chemistry Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Coliform, Fecal	SM 9222 D	200	CFU/100mL	10	10	09/19/18 17:01	
Enterococcus	ASTM D6503-99	210	MPN/100mL	10	10	09/19/18 16:05	

ALS Group USA, Corp.
dba ALS Environmental

Analytical Report

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28
Sample Matrix: Water
Sample Name: PC5
Lab Code: J1806723-003

Service Request: J1806723
Date Collected: 09/19/18 10:43
Date Received: 09/19/18 14:00
Basis: NA

General Chemistry Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Coliform, Fecal	SM 9222 D	280	CFU/100mL	10	10	09/19/18 17:58	
Enterococcus	ASTM D6503-99	95	MPN/100mL	10	10	09/19/18 16:05	

ALS Group USA, Corp.
dba ALS Environmental

Analytical Report

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28
Sample Matrix: Water
Sample Name: PC4
Lab Code: J1806723-004

Service Request: J1806723
Date Collected: 09/19/18 11:13
Date Received: 09/19/18 14:00
Basis: NA

General Chemistry Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Coliform, Fecal	SM 9222 D	420	CFU/100mL	10	10	09/19/18 17:58	
Enterococcus	ASTM D6503-99	190	MPN/100mL	10	10	09/19/18 16:05	

ALS Group USA, Corp.
dba ALS Environmental

Analytical Report

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28
Sample Matrix: Water
Sample Name: CB
Lab Code: J1806723-005

Service Request: J1806723
Date Collected: 09/19/18 11:35
Date Received: 09/19/18 14:00
Basis: NA

General Chemistry Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Coliform, Fecal	SM 9222 D	240	CFU/100mL	10	10	09/19/18 17:58	
Enterococcus	ASTM D6503-99	310	MPN/100mL	10	10	09/19/18 16:05	

ALS Group USA, Corp.
dba ALS Environmental

Analytical Report

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28
Sample Matrix: Water
Sample Name: PC3
Lab Code: J1806723-006

Service Request: J1806723
Date Collected: 09/19/18 12:06
Date Received: 09/19/18 14:00
Basis: NA

General Chemistry Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Coliform, Fecal	SM 9222 D	430	CFU/100mL	10	10	09/19/18 17:58	
Enterococcus	ASTM D6503-99	490	MPN/100mL	10	10	09/19/18 16:05	

ALS Group USA, Corp.
dba ALS Environmental

Analytical Report

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28
Sample Matrix: Water
Sample Name: PC2
Lab Code: J1806723-007

Service Request: J1806723
Date Collected: 09/19/18 09:37
Date Received: 09/19/18 14:00
Basis: NA

General Chemistry Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Coliform, Fecal	SM 9222 D	310	CFU/100mL	10	10	09/19/18 17:01	
Enterococcus	ASTM D6503-99	420	MPN/100mL	10	10	09/19/18 16:05	

ALS Group USA, Corp.
dba ALS Environmental

Analytical Report

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28
Sample Matrix: Water
Sample Name: PC1
Lab Code: J1806723-008

Service Request: J1806723
Date Collected: 09/19/18 10:33
Date Received: 09/19/18 14:00
Basis: NA

General Chemistry Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Coliform, Fecal	SM 9222 D	340	CFU/100mL	10	10	09/19/18 17:01	
Enterococcus	ASTM D6503-99	330	MPN/100mL	10	10	09/19/18 16:05	

ALS Group USA, Corp.
dba ALS Environmental

Analytical Report

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28
Sample Matrix: Water
Sample Name: SB
Lab Code: J1806723-009

Service Request: J1806723
Date Collected: 09/19/18 09:40
Date Received: 09/19/18 14:00
Basis: NA

General Chemistry Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Coliform, Fecal	SM 9222 D	440	CFU/100mL	10	10	09/19/18 17:01	
Enterococcus	ASTM D6503-99	280	MPN/100mL	10	10	09/19/18 16:05	

ALS Group USA, Corp.
dba ALS Environmental

Analytical Report

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28
Sample Matrix: Water
Sample Name: DB
Lab Code: J1806723-010

Service Request: J1806723
Date Collected: 09/19/18 11:24
Date Received: 09/19/18 14:00
Basis: NA

General Chemistry Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Coliform, Fecal	SM 9222 D	90 B	CFU/100mL	10	10	09/19/18 17:58	
Enterococcus	ASTM D6503-99	63	MPN/100mL	10	10	09/19/18 16:05	



QC Summary Forms

ALS Environmental - Jacksonville Laboratory
9143 Philips Highway, Suite 200, Jacksonville, FL 32256
Phone (904) 739-2277 Fax (904) 739-2011
www.alsglobal.com



Microbiology

ALS Environmental—Jacksonville Laboratory
9143 Philips Highway, Suite 200, Jacksonville, FL 32256
Phone (904)739-2277 Fax (904)739-2011
www.alsglobal.com

ALS Group USA, Corp.
dba ALS Environmental

Analytical Report

Client: Florida DEP
Project: Pellicer WQ/RQ-2018-09-03-28
Sample Matrix: Water
Sample Name: Method Blank
Lab Code: J1806723-MB

Service Request: J1806723
Date Collected: NA
Date Received: NA
Basis: NA

General Chemistry Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Coliform, Fecal	SM 9222 D	1.0 U	CFU/100mL	1.0	1	09/19/18 17:01	
Enterococcus	ASTM D6503-99	1.0 U	MPN/100mL	1.0	1	09/19/18 14:24	