

Chemical Analysis Report

NW-DIST-2018-11-16-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Eglin Run East**
Request ID: **RQ-2018-11-12-23**
Customer: **NW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
160 W Government St
Suite 208A
Pensacola, FL 32502-5794
Attn: Michael King

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 13-DEC-2018 10:18

A handwritten signature in black ink, appearing to read "Colin Wright", is displayed on a light gray rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: TITI CREEK ABOVE SR 285

Collection Date/Time: 11/15/2018 11:00

Field ID: G4NW0365

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2042241	EPA 180.1 Rev. 2.0	Turbidity	1.9		NTU	P354882	
	EPA 300.0 Rev. 2.1	Chloride	2.1		mg Cl/L	P355287	
		Sulfate	0.69		mg SO4/L	P355290	
	SM 2120 B	Color (true)	120		PCU	P354840	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P355189	
	SM 2540 C-97	TDS	29		mg/L	P355137	
	SM 2540 D-97	TSS	2	I	mg/L	P354862	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P355192	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	I	mg N/L	P355075	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.25		mg N/L	P354979	
2042244	EPA 353.2 Rev. 2.0	NO2NO3-N	0.046		mg N/L	P355091	
	EPA 365.1 Rev. 2.0	Total-P	0.004	I	mg P/L	P354988	
	SM 5310 B-00	Organic Carbon	9.8		mg C/L	P354966	
	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P354843	
2042247	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P354843	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: DIRECT RUNOFF TO CHOCTAWHATCHEE BAY BELOW OREGON DRIVE **Collection Date/Time: 11/15/2018 12:00**

Field ID: G3NW0001

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2042242	EPA 180.1 Rev. 2.0	Turbidity	7.4		NTU	P354882	
	EPA 300.0 Rev. 2.1	Chloride	13		mg Cl/L	P355287	
		Sulfate	6.9		mg SO4/L	P355290	
		Color (true)	62		PCU	P354840	
	SM 2320 B-97	Total Alkalinity	58		mg CaCO3/L	P355189	
	SM 2540 C-97	TDS	100		mg/L	P355137	
	SM 2540 D-97	TSS	3	I	mg/L	P354862	
	SM 4500 F-C-97	Fluoride	0.073	I	mg F/L	P355192	
2042245	EPA 350.1 Rev. 2.0	Ammonia-N	0.033		mg N/L	P355075	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.30		mg N/L	P354979	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.16		mg N/L	P355091	
	EPA 365.1 Rev. 2.0	Total-P	0.028		mg P/L	P354988	
	SM 5310 B-00	Organic Carbon	6.8		mg C/L	P354964	
2042248	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.014		mg P/L	P354843	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: DEAN CREEK UPSTREAM OF SR 87

Collection Date/Time: 11/15/2018 13:00

Field ID: G4NW0362

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2042243	EPA 180.1 Rev. 2.0	Turbidity	2.0		NTU	P354882	
	EPA 300.0 Rev. 2.1	Chloride	5.7		mg Cl/L	P355287	
		Sulfate	1.4		mg SO4/L	P355290	
	SM 2120 B	Color (true)	110		PCU	P354840	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P355189	
	SM 2540 C-97	TDS	34		mg/L	P355137	
	SM 2540 D-97	TSS	3	I	mg/L	P354862	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P355192	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	I	mg N/L	P355075	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.20		mg N/L	P354979	
2042246	EPA 353.2 Rev. 2.0	NO2NO3-N	0.028		mg N/L	P355091	
	EPA 365.1 Rev. 2.0	Total-P	0.003	I	mg P/L	P354988	
	SM 5310 B-00	Organic Carbon	8.8		mg C/L	P354964	
	EPA 365.1 Rev. 2.0	O-Phosphate-P	0.004	U	mg P/L	P354844	
	dissolved						

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0 dissolved: O-Phosphate-P: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P354882

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355287

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355290

Component	Result	Code	Units
Sulfate	0.20	U	mg SO ₄ /L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P355075

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P354979

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P355091

Component	Result	Code	Units
NO ₂ NO ₃ -N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P354988

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P354843

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P354844

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 2120 B
Batch ID: P354840

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P355189

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-97
Batch ID: P355137

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P354862

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P355192

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P354964

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Reference Method: SM 5310 B-00
Batch ID: P354966

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355287

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	99.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355290

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	99.4		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P355075

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P354979

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	107		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P355091

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P354988

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	99.3		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P354843

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.4		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P354844

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	102		P	90 - 110

Reference Method: SM 2120 B
Batch ID: P354840

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	98.6		P	90 - 110

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 2320 B-97
Batch ID: P355189

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	100		P	98 - 107

Reference Method: SM 4500 F-C-97
Batch ID: P355192

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.8		P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P354964

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	93.2		P	85 - 115

Reference Method: SM 5310 B-00
Batch ID: P354966

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	91.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355287

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2042091	Chloride	103		P	90 - 110
2042180	Chloride	102		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355290

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2042091	Sulfate	102		P	90 - 110
2042180	Sulfate	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P355075

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2042244	Ammonia-N	95.0	96.0	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P354979

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2042182	Kjeldahl Nitrogen	93.2	97.9	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P355091

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2042244	NO2NO3-N	99.4	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P354988

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2042181	Total-P	100	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P354843

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2042110	O-Phosphate-P	95.3	97.0	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P355192

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2042491	Fluoride	102	102	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P354964

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2041793	Organic Carbon	97.0	92.2	P/P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-00
Batch ID: P354966

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2041973	Organic Carbon	96.4	95.6	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P354882

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2042125	Turbidity	0.816	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355287

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2042091	Chloride	1.72	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355290

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2042091	Sulfate	0.192	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P355075

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2042244	Ammonia-N	0.989	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P354979

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2042182	Kjeldahl Nitrogen	4.26	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P355091

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2042244	NO2NO3-N	2.02	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P354988

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2042181	Total-P	3.16	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P354843

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2042110	O-Phosphate-P	1.77	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P354844

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2042295	O-Phosphate-P	0.314	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P354840

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2042142	Color (true)	2.31	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P355189

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2042491	Total Alkalinity	0.155	Sample	P	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P355137

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2042142	TDS	2.68	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P355192

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2042491	Fluoride	0.463	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P354964

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2041793	Organic Carbon	2.82	Spike	P	0 - 20

Reference Method: SM 5310 B-00
Batch ID: P354966

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2041973	Organic Carbon	0.938	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B

Run ID: A87829

Included Lab Sample IDs: 2042241, 2042242, 2042243

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	99.7	101	P/P	85 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A87830

Included Lab Sample IDs: 2042247, 2042248, 2042249

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.6	98.8	P/P	90 - 110
O-Phosphate-P	99.7	99.6	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A87844

Included Lab Sample IDs: 2042241, 2042242, 2042243

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: SM 5310 B-00

Run ID: A87888

Included Lab Sample IDs: 2042244, 2042245, 2042246

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	93.8	94.0	P/P	90 - 110
Organic Carbon	94.1	93.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A87911

Included Lab Sample IDs: 2042244, 2042245, 2042246

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	101	96.2	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A87925

Included Lab Sample IDs: 2042244, 2042245, 2042246

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	99.5	99.0	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A87946

Included Lab Sample IDs: 2042244, 2042245, 2042246

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.3	97.8	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A87948

Included Lab Sample IDs: 2042244, 2042245, 2042246

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
-----------	---------	---------	------------	----------------

Quality Assurance Report Calibration Verification

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A87948

Included Lab Sample IDs: 2042244, 2042245, 2042246

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A87993

Included Lab Sample IDs: 2042241, 2042242, 2042243

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	104	104	P/P	90 - 110
Sulfate	103	103	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A87997

Included Lab Sample IDs: 2042241, 2042242, 2042243

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	98.8	101	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A87997

Included Lab Sample IDs: 2042241, 2042242, 2042243

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	103	104	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					0.816	
EPA 300.0 Rev. 2.1	Chloride	99.8	103	102		1.72	
	Sulfate	99.4	102	101		0.192	
EPA 350.1 Rev. 2.0	Ammonia-N	99.8	95.0	96.0			0.989
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	107	93.2	97.9			4.26
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	100	99.4	102			2.02
EPA 365.1 Rev. 2.0	Total-P	99.3	100	104			3.16
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.4	95.3	97.0			1.77
	O-Phosphate-P	102					0.314
SM 2120 B	Color (true)	98.6				2.31	
SM 2320 B-97	Total Alkalinity	100				0.155	
SM 2540 C-97	TDS					2.68	
SM 4500 F-C-97	Fluoride	99.8	102	102			0.463
SM 5310 B-00	Organic Carbon	93.2	97.0	92.2			2.82
	Organic Carbon	91.8	96.4	95.6			0.938

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2042241, 2042242, 2042243
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2042241, 2042242, 2042243
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2042241, 2042242, 2042243
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2042244, 2042245, 2042246
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2042244, 2042245, 2042246
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2042244, 2042245, 2042246
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2042244, 2042245, 2042246
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2042247, 2042248, 2042249
SM 2120 B / E31780	True Color measured at 450 nm	2042241, 2042242, 2042243
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2042241, 2042242, 2042243
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2042241, 2042242, 2042243
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2042241, 2042242, 2042243
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2042241, 2042242, 2042243
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2042244, 2042245, 2042246

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	11/16/2018			11/16/2018 16:18	Megan Treadwell	2042241, 2042242, 2042243
EPA 300.0 Rev. 2.1	11/16/2018			11/29/2018 08:40	Lipi Saha	2042241
	11/16/2018			11/29/2018 08:52	Lipi Saha	2042242
	11/16/2018			11/29/2018 09:05	Lipi Saha	2042243
EPA 350.1 Rev. 2.0	11/16/2018			11/19/2018 16:00	Ping Hua	2042244
	11/16/2018			11/19/2018 16:04	Ping Hua	2042245
	11/16/2018			11/19/2018 16:06	Ping Hua	2042246
EPA 351.2 Rev. 2.0	11/16/2018	11/20/2018 10:30	Adrian Shelby	11/21/2018 11:05	Joseph Suarez	2042244
	11/16/2018	11/20/2018 10:30	Adrian Shelby	11/21/2018 11:07	Joseph Suarez	2042245
	11/16/2018	11/20/2018 10:30	Adrian Shelby	11/21/2018 11:08	Joseph Suarez	2042246
EPA 353.2 Rev. 2.0	11/16/2018			11/26/2018 12:38	Lauren Bottorf	2042244
	11/16/2018			11/26/2018 12:46	Lauren Bottorf	2042245
	11/16/2018			11/26/2018 12:47	Lauren Bottorf	2042246
EPA 365.1 Rev. 2.0	11/16/2018	11/20/2018 12:35	Sima Feizi	11/21/2018 13:53	Beverly Y. Sanders	2042244
	11/16/2018	11/20/2018 12:35	Sima Feizi	11/21/2018 13:54	Beverly Y. Sanders	2042245
	11/16/2018	11/20/2018 12:35	Sima Feizi	11/21/2018 13:55	Beverly Y. Sanders	2042246
EPA 365.1 Rev. 2.0 dissolved	11/16/2018			11/16/2018 16:45	Jhamieka S. Greenwood	2042248
	11/16/2018			11/16/2018 16:51	Jhamieka S. Greenwood	2042247
	11/16/2018			11/16/2018 16:58	Jhamieka S. Greenwood	2042249
SM 2120 B	11/16/2018			11/16/2018 16:22	Chelsea Hernandez	2042241, 2042242
	11/16/2018			11/16/2018 16:23	Chelsea Hernandez	2042243
SM 2320 B-97	11/16/2018			11/27/2018 09:18	Dale L. Simmons	2042241
	11/16/2018			11/27/2018 09:28	Dale L. Simmons	2042242
	11/16/2018			11/27/2018 09:41	Dale L. Simmons	2042243
SM 2540 C-97	11/16/2018			11/16/2018 17:06	Yijie Li	2042241, 2042242, 2042243
SM 2540 D-97	11/16/2018			11/16/2018 17:06	Yijie Li	2042241, 2042242, 2042243
SM 4500 F-C-97	11/16/2018			11/27/2018 09:18	Dale L. Simmons	2042241
	11/16/2018			11/27/2018 09:28	Dale L. Simmons	2042242
	11/16/2018			11/27/2018 09:41	Dale L. Simmons	2042243
SM 5310 B-00	11/16/2018			11/19/2018 16:37	Megan Treadwell	2042245
	11/16/2018			11/19/2018 16:50	Megan Treadwell	2042246
	11/16/2018			11/20/2018 04:09	Megan Treadwell	2042244