

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter ID: Biology #2

RQ: 2018-03-05-26

Project: SMP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site _____

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification _____

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Ben Ralys	3/7/18	7:10	20.96	8.93	8.92	100.0	62.5	-	(P) X (F)	(L) / (F)
ICV	Ben Ralys	3/7/18	7:12	20.97	8.93	8.92	100.0	61.4	-	(P) / (F)	(L) / (F)
CCV	Ben Ralys	3/7/18	1:38	21.55	8.82	8.94	101.3	60.4	-	(P) / (F)	(L) / (F)
CCV										P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Ben Ralys	3/7/18	7:14	171005D	10/2018	1,000	999	(P) / (F)	(L) / (F)
ICV	Ben Ralys	3/7/18	7:14	171005D	10/2018	1,000	1,001	(P) / (F)	(L) / (F)
CCV	Ben Ralys	3/7/18	1:40	171005E	10/2018	100	99	(P) / (F)	(L) / (F)
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Ben Ralys	3/7/18	7:18	170614B	12/2018	7.0	6.98	2.3	(P) / (F)	(L) / (F)
Calibr.	Ben Ralys	3/7/18	7:20	171005A	10/2018	4.0	3.98	178.9	(P) / (F)	(L) / (F)
Calibr.	Ben Ralys	3/7/18	7:22	171005C	4/2019	10.0	9.81	-151.0	(P) / (F)	(L) / (F)
ICV	Ben Ralys	3/7/18	7:25	171005C	4/2019	10.0	9.81	-148.8	(P) / (F)	(L) / (F)
CCV	Ben Ralys	3/7/18	1:43	171005A	10/2018	4.0	3.94	178.9	(P) / (F)	(L) / (F)
CCV									P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

Florida Department of Environmental Protection
Central Laboratory Submittal Form

Customer: WAS-SECT

Project ID: SMP

Requester: Curtis Musson

Collected By: Ben Rabys, Jason Storns

Field Report Prepared By: Ben Rabys, Jason Storns

Sampling Agency: FDEP

Location		WIN ID/Field ID		Collection (comp begin or grab)		Composite end		Bottle Group(s)	
Field ID		Pettys Gulf		Date 3/7/18		Date		(See pg 2)	
WIN/STORET #		G1WA0022		Matrix (type, e.g., Salt, Fresh, etc)		mg/L % sat		Total Res. Chlorine (mg/L)	
Latitude		30.58915		pH		ft		Sp. Conductance (umho/cm)	
Longitude		-84.22637		Salinity (ppt)		m		58	
Location		Pine Hill		Collection (comp begin or grab)		Composite end		Bottle Group(s)	
Field ID		Lake Middle		Date 3/7/18		Date		(See pg 2)	
WIN/STORET #		G1WA0054		Matrix (type, e.g., Salt, Fresh, etc)		mg/L % sat		Total Res. Chlorine (mg/L)	
Latitude		30.58492		pH		ft		Sp. Conductance (umho/cm)	
Longitude		-84.2195		Salinity (ppt)		m		44	
Location		Lower Dianne		Collection (comp begin or grab)		Composite end		Bottle Group(s)	
Field ID		Lake Middle		Date 3/7/18		Date		(See pg 2)	
WIN/STORET #		G1WA0040		Matrix (type, e.g., Salt, Fresh, etc)		mg/L % sat		Total Res. Chlorine (mg/L)	
Latitude		30.59355		pH		ft		Sp. Conductance (umho/cm)	
Longitude		-84.23534		Salinity (ppt)		m		85	
Location		Upper Dianne		Collection (comp begin or grab)		Composite end		Bottle Group(s)	
Field ID		Lake Middle		Date 3/7/18		Date		(See pg 2)	
WIN/STORET #		G1WA0041		Matrix (type, e.g., Salt, Fresh, etc)		mg/L % sat		Total Res. Chlorine (mg/L)	
Latitude		30.59986		pH		ft		Sp. Conductance (umho/cm)	
Longitude		-84.23778		Salinity (ppt)		m		92	

Relinquished By:	Date/Time	Shipping Method	Number of Coolers Shipped:	Received By:	Date/Time
	3/7/18 1323	Drop-off			03-07-18 1323

Customer: WAS-SECT

Requester: Curtis Musson

Field Report Prepared By: Ben Ralys, Jason Storrs

Project ID: SMP

Collected By: Ben Ralys, Jason Storrs

Sampling Agency: FDEP

Location		WIN ID/Field ID		☐ Comp		Collection (comp begin or grab)		☒ Eastern		Composite end		Bottle Group(s)	
Field ID		Lake Monkey		☒ Grab		Date 3/7/18 Time 1105		Central		Date		(See pg 2)	
WIN/STORET #		Business Middle		Matrix (type, e.g., Salt, Fresh, etc)		Fresh		Diss. Oxygen		mg/L		Total Res. Chlorine	
Latitude		30.60779		Temp (C)		18.61		pH		Sample Depth		75	
Longitude		-84.23076		Salinity (ppt)		0.03		Comments		ft		Sp. Conductance (umho/cm)	
dd dms		☒ dd ☐ dms		☒ dd ☐ dms		☐ dd ☐ dms		☐ dd ☐ dms		☐ ft ☒ m		☐ mg/L ☒ % sat	
Location		WIN/STORET #		Matrix (type, e.g., Salt, Fresh, etc)		pH		Diss. Oxygen		mg/L		Total Res. Chlorine	
Field ID		Latitude		Longitude		Temp (C)		Salinity (ppt)		Sample Depth		Sp. Conductance (umho/cm)	
dd dms		☐ dd ☐ dms		☐ dd ☐ dms		☐ dd ☐ dms		☐ dd ☐ dms		☐ ft ☐ m		☐ mg/L ☐ % sat	
Location		WIN/STORET #		Matrix (type, e.g., Salt, Fresh, etc)		pH		Diss. Oxygen		mg/L		Total Res. Chlorine	
Field ID		Latitude		Longitude		Temp (C)		Salinity (ppt)		Sample Depth		Sp. Conductance (umho/cm)	
dd dms		☐ dd ☐ dms		☐ dd ☐ dms		☐ dd ☐ dms		☐ dd ☐ dms		☐ ft ☐ m		☐ mg/L ☐ % sat	
Location		WIN/STORET #		Matrix (type, e.g., Salt, Fresh, etc)		pH		Diss. Oxygen		mg/L		Total Res. Chlorine	
Field ID		Latitude		Longitude		Temp (C)		Salinity (ppt)		Sample Depth		Sp. Conductance (umho/cm)	
dd dms		☐ dd ☐ dms		☐ dd ☐ dms		☐ dd ☐ dms		☐ dd ☐ dms		☐ ft ☐ m		☐ mg/L ☐ % sat	
Location		WIN/STORET #		Matrix (type, e.g., Salt, Fresh, etc)		pH		Diss. Oxygen		mg/L		Total Res. Chlorine	
Field ID		Latitude		Longitude		Temp (C)		Salinity (ppt)		Sample Depth		Sp. Conductance (umho/cm)	
dd dms		☐ dd ☐ dms		☐ dd ☐ dms		☐ dd ☐ dms		☐ dd ☐ dms		☐ ft ☐ m		☐ mg/L ☐ % sat	

Relinquished By:	Date/Time	Shipping Method	Number of Coolers Shipped:	Received By:	Date/Time
	3/7/18 1323	Drop-off			03-07-18 13:23

Group: A	Bottle Type: ALKALINITY TURBIDITY W-CL-IC W-COLOR W-F W-SO4-IC W-TDS W-TSS	P-1L	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab
Group: A	Bottle Type: CHLSUITE-W	BP-1L	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab
Group: A	Bottle Type: ECOLI-18QT	P-250ML-WO-THI	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab
Group: A	Bottle Type: W-NH3 W-NO2NO3 W-S-A-TP W-TKN W-TOC	P-500ML	# of Bottles: 5	Preservative: ICE And H2SO4		Enter Number of Bottles Sent to Lab
Group: A	Bottle Type: W-PO4-F	P-125ML	# of Bottles: 5	Preservative: FILTER-ICE		Enter Number of Bottles Sent to Lab

Cooler Packing Worksheet for Request: RQ-2018-03-05-26

NE Tallahassee lakes

Ship Cooler On: 3/2/2018

Customer/Project: WAS-SECT/SMP

Assigned To: SHAIK_A

Requester: Curtis Musson

Priority: 3

850-245-8453
2600 Blair Stone Rd

Tallahassee, FL 32399

Attn: Curtis Musson

Comments:

No preservatives needed.

Group A: Freshwater Kit + Ecoli

Packing Notes

No FedEx labels

no trash bags

2 Submittal forms

Requested Analyses:

Bottle Group: A

of Sites: 5

Container ID: P-1L

Qty: 5 Preservation: ICE

Lot # 00069122

Description: Plastic Bottle 1L

Analysis

ALKALINITY

TURBIDITY

W-CL-IC

W-COLOR

W-F

W-SO4-IC

W-TDS

W-TSS

Description

Alkalinity in aqueous matrices as mg CaCO3/L

Turbidity in aqueous matrices

Chloride in aqueous matrices

True Color measured at 450 nm

Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)

Sulfate in aqueous matrices

Total Dissolved Solids in aqueous matrices

Total Suspended Solids in aqueous matrices

Container ID: BP-1L

Qty: 5 Preservation: ICE

Lot # 1214297

Description: Brown Plastic Bottle 1L

Analysis

CHLSUITE-W

Description

Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Container ID: P-250ML-WO-THI

Qty: 5 Preservation: ICE

Lot # 417167

Description:

Analysis

ECOLI-18QT

Description

Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems

Container ID: P-500ML

Qty: 5 Preservation: ICE And H2SO4

Lot # 5507 0302

Description: Plastic Bottle 500 mL

Affix YELLOW sulfuric acid dot to container.

Analysis

W-NH3

W-NO2NO3

W-S-A-TP

W-TKN

W-TOC

Description

Ammonia in aqueous matrices as mg N/L

Nitrite/Nitrate in aqueous matrices as mg N/L

Total Phosphorus in aqueous matrices as mg P/L

Total Kjeldahl Nitrogen in aqueous matrices

Nonpurgeable Organic Carbon in aqueous matrices

Container ID: P-125ML

Qty: 5 Preservation: FILTER-ICE

Lot # 5506 9732

Description: Plastic Bottle 125 mL

Analysis

W-PO4-F

Description

Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Cooler Packed By: ATG Number of Coolers: 2 Date: 2-28-18

Kit must also include:

- ☒ Field Sheets
- ☒ Temperature Control Bottle
- ☒ FedEx Bills, if applicable
- ☒ Plastic Bags
- ☒ Zip Ties

If Preservation Included:

ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____

PLEASE RETURN ALL COOLERS FOR RQ-2018-03-05-26