

Chemical Analysis Report

NE-DIST-2019-05-03-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Interlachen 2**
Request ID: **RQ-2019-04-29-23**
Customer: **NE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
8800 Baymeadows Way West
Suite 100
Jacksonville, FL 32256
Attn: Jeremy Parrish

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 17-MAY-2019 13:50

A handwritten signature in black ink, appearing to read "Liang Lin", on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: LAKE FANNY CENTER

Collection Date/Time: 05/02/2019 10:45

Field ID: 20020100

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2083258	EPA 180.1 Rev. 2.0	Turbidity	1.1		NTU	P363374	
	EPA 300.0 Rev. 2.1	Chloride	12		mg Cl/L	P363745	
		Sulfate	6.1		mg SO4/L	P363749	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P363807	
	SM 2540 C-97	TDS	46		mg/L	P363915	
	SM 2540 D-97	TSS	3	I	mg/L	P363892	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P363969	
2083261	EPA 350.1 Rev. 2.0	Ammonia-N	0.007		mg N/L	P363777	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.61		mg N/L	P363711	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.15		mg N/L	P363675	
	EPA 365.1 Rev. 2.0	Total-P	0.012		mg P/L	P363525	
	SM 5310 B-00	Organic Carbon	5.4		mg C/L	P363920	
2083264	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P363334	

Sample Location: GILLIS POND CENTER

Collection Date/Time: 05/02/2019 11:20

Field ID: G1NE0863

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2083259	EPA 180.1 Rev. 2.0	Turbidity	1.1		NTU	P363374	
	EPA 300.0 Rev. 2.1	Chloride	10		mg Cl/L	P363745	
		Sulfate	2.6		mg SO4/L	P363749	
		Total Alkalinity	0.65	U	mg CaCO3/L	P363807	
	SM 2320 B-97	TDS	44		mg/L	P363915	
	SM 2540 D-97	TSS	2	I	mg/L	P363892	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P363969	
2083262	EPA 350.1 Rev. 2.0	Ammonia-N	1.1		mg N/L	P363777	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	2.4		mg N/L	P363711	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.52		mg N/L	P363675	
	EPA 365.1 Rev. 2.0	Total-P	0.019		mg P/L	P363525	
	SM 5310 B-00	Organic Carbon	17		mg C/L	P363920	
2083265	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P363334	

Sample Location: HARDESTY LAKE CENTER

Collection Date/Time: 05/02/2019 12:20

Field ID: G1NE0869

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2083260	EPA 180.1 Rev. 2.0	Turbidity	1.6	A	NTU	P363375	
	EPA 300.0 Rev. 2.1	Chloride	7.4		mg Cl/L	P363745	
		Sulfate	1.2		mg SO4/L	P363749	
	SM 2320 B-97	Total Alkalinity	1.7	I	mg CaCO3/L	P363807	
	SM 2540 C-97	TDS	44		mg/L	P363915	
	SM 2540 D-97	TSS	4	I	mg/L	P363892	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P363969	
2083263	EPA 350.1 Rev. 2.0	Ammonia-N	0.004	I	mg N/L	P363777	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.84		mg N/L	P363711	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P363675	
	EPA 365.1 Rev. 2.0	Total-P	0.047		mg P/L	P363525	
	SM 5310 B-00	Organic Carbon	16		mg C/L	P363920	
2083266	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.006	I	mg P/L	P363334	

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P363374

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P363375

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P363745

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P363749

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P363777

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P363711

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P363675

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P363525

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P363334

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Quality Assurance Report Method Blank Results

Reference Method: SM 2320 B-97
Batch ID: P363807

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-97
Batch ID: P363915

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P363892

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P363969

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P363920

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P363745

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P363749

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	99.1		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P363777

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.5		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P363711

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	103		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P363675

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P363525

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	97.0		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P363334

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	96.9		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P363807

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	98 - 106

Reference Method: SM 4500 F-C-97
Batch ID: P363969

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.6		P	92 - 103

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
Batch ID: P363920

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	90.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P363745

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2083258	Chloride	102		P	90 - 110
2083385	Chloride	99.2		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P363749

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2083258	Sulfate	102		P	90 - 110
2083385	Sulfate	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P363777

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2083179	Ammonia-N	102	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P363711

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2083229	Kjeldahl Nitrogen	106	109	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P363675

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2083246	NO2NO3-N	100	99.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P363525

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2083289	Total-P	100	98.6	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P363334

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2083135	O-Phosphate-P	95.0	95.3	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P363969

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2083317	Fluoride	99.1	97.4	P/P	89 - 106

Reference Method: SM 5310 B-00
Batch ID: P363920

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2083261	Organic Carbon	94.7	91.5	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P363374

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2083226	Turbidity	5.77	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P363375

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2083260	Turbidity	1.88	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P363745

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2083385	Chloride	0.391	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P363749

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2083385	Sulfate	0.478	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P363777

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2083179	Ammonia-N	0.224	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P363711

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2083229	Kjeldahl Nitrogen	1.82	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P363675

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2083246	NO2NO3-N	1.50	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P363525

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2083289	Total-P	1.55	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P363334

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2083135	O-Phosphate-P	0.228	Spike	P	0 - 20

Reference Method: SM 2320 B-97
 Batch ID: P363807

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2083136	Total Alkalinity	0.0676	Sample	P	0 - 2.8

Reference Method: SM 2540 C-97
 Batch ID: P363915

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2083245	TDS	1.98	Sample	P	0 - 10

Reference Method: SM 2540 D-97
 Batch ID: P363892

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2083245	TSS	5.61	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
 Batch ID: P363969

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2083317	Fluoride	1.46	Spike	P	0 - 2.5

Reference Method: SM 5310 B-00
 Batch ID: P363920

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2083261	Organic Carbon	2.70	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
 Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A91195

Included Lab Sample IDs: 2083264, 2083265, 2083266

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	95.8	95.8	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A91209

Included Lab Sample IDs: 2083258, 2083259, 2083260

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105
Turbidity	100	99.5	P/P	95 - 105

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A91228

Included Lab Sample IDs: 2083258, 2083259, 2083260

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	102	P/P	90 - 110
Sulfate	102	103	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A91321

Included Lab Sample IDs: 2083261, 2083262, 2083263

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A91339

Included Lab Sample IDs: 2083261, 2083262, 2083263

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	99.8	98.5	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A91350

Included Lab Sample IDs: 2083261, 2083262, 2083263

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	102	102	P/P	90 - 110
Ammonia-N	103	104	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A91372

Included Lab Sample IDs: 2083258, 2083259, 2083260

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A91387

Included Lab Sample IDs: 2083261, 2083262, 2083263

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	106	102	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A91414

Included Lab Sample IDs: 2083261, 2083262, 2083263

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	91.8	92.2	P/P	90 - 110
Organic Carbon	92.2	92.4	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A91436

Included Lab Sample IDs: 2083258, 2083259, 2083260

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	98.1	97.6	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					5.77	
	Turbidity					1.88	
EPA 300.0 Rev. 2.1	Chloride	100	99.2	102		0.391	
	Sulfate	99.1	101	102		0.478	
EPA 350.1 Rev. 2.0	Ammonia-N	99.5	102	102			0.224
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	103	106	109			1.82
EPA 353.2 Rev. 2.0	NO2NO3-N	100	100	99.0			1.50
EPA 365.1 Rev. 2.0	Total-P	97.0	100	98.6			1.55
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	96.9	95.0	95.3			0.228
SM 2320 B-97	Total Alkalinity	102				0.0676	
SM 2540 C-97	TDS					1.98	
SM 2540 D-97	TSS					5.61	
SM 4500 F-C-97	Fluoride	96.6	99.1	97.4			1.46
SM 5310 B-00	Organic Carbon	90.8	94.7	91.5			2.70

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2083258, 2083259, 2083260
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2083258, 2083259, 2083260
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2083258, 2083259, 2083260
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2083261, 2083262, 2083263
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2083261, 2083262, 2083263
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2083261, 2083262, 2083263
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2083261, 2083262, 2083263
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2083264, 2083265, 2083266
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	2083258, 2083259, 2083260
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2083258, 2083259, 2083260
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2083258, 2083259, 2083260
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2083258, 2083259, 2083260
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2083261, 2083262, 2083263

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	05/03/2019			05/03/2019 16:06	Sima Feizi	2083258, 2083259, 2083260
EPA 300.0 Rev. 2.1	05/03/2019			05/10/2019 03:36	Lipi Saha	2083258
	05/03/2019			05/10/2019 04:00	Lipi Saha	2083259
	05/03/2019			05/10/2019 04:12	Lipi Saha	2083260
EPA 350.1 Rev. 2.0	05/03/2019			05/09/2019 11:43	Ping Hua	2083261
	05/03/2019			05/09/2019 14:03	Ping Hua	2083262
	05/03/2019			05/09/2019 14:04	Ping Hua	2083263
EPA 351.2 Rev. 2.0	05/03/2019	05/09/2019 13:00	Adrian Shelby	05/13/2019 13:51	Joseph Suarez	2083261
	05/03/2019	05/09/2019 13:00	Adrian Shelby	05/13/2019 13:52	Joseph Suarez	2083262
	05/03/2019	05/09/2019 13:00	Adrian Shelby	05/13/2019 13:54	Joseph Suarez	2083263
EPA 353.2 Rev. 2.0	05/03/2019			05/09/2019 09:39	Beverly Y. Sanders	2083261
	05/03/2019			05/09/2019 09:40	Beverly Y. Sanders	2083262
	05/03/2019			05/09/2019 09:41	Beverly Y. Sanders	2083263
EPA 365.1 Rev. 2.0	05/03/2019	05/07/2019 14:35	Sima Feizi	05/09/2019 17:49	Rosalyn Ballman	2083261
	05/03/2019	05/07/2019 14:35	Sima Feizi	05/09/2019 17:50	Rosalyn Ballman	2083262
	05/03/2019	05/07/2019 14:35	Sima Feizi	05/09/2019 17:51	Rosalyn Ballman	2083263
EPA 365.1 Rev. 2.0 dissolved	05/03/2019			05/03/2019 14:39	Lipi Saha	2083264
	05/03/2019			05/03/2019 14:40	Lipi Saha	2083265
	05/03/2019			05/03/2019 14:41	Lipi Saha	2083266
SM 2320 B-97	05/03/2019			05/11/2019 17:55	Dale L. Simmons	2083258
	05/03/2019			05/11/2019 18:04	Dale L. Simmons	2083259
	05/03/2019			05/11/2019 18:15	Dale L. Simmons	2083260
SM 2540 C-97	05/03/2019			05/08/2019 14:21	Yijie Li	2083258, 2083259, 2083260
SM 2540 D-97	05/03/2019			05/08/2019 14:21	Yijie Li	2083258, 2083259, 2083260
SM 4500 F-C-97	05/03/2019			05/15/2019 04:35	Dale L. Simmons	2083258
	05/03/2019			05/15/2019 04:42	Dale L. Simmons	2083259
	05/03/2019			05/15/2019 04:48	Dale L. Simmons	2083260
SM 5310 B-00	05/03/2019			05/13/2019 04:13	Julia Storbeck	2083261
	05/03/2019			05/13/2019 05:10	Julia Storbeck	2083262
	05/03/2019			05/13/2019 08:13	Julia Storbeck	2083263