



September 24, 2019

Service Request No:J1905597

Mr. Jeremy Parrish
Florida DEP
8800 Baymeadows Way West
Jacksonville, FL 32256-7577

Laboratory Results for: LSJR Downtown

Dear Mr.Parrish,

Enclosed are the results of the sample(s) submitted to our laboratory September 19, 2019
For your reference, these analyses have been assigned our service request number **J1905597**.

All analyses were performed according to our laboratory's quality assurance program. The test results meet requirements of the NELAP standards except as noted in the case narrative report. All results are intended to be considered in their entirety, and ALS Environmental is not responsible for use of less than the complete report. Results apply only to the items submitted to the laboratory for analysis and individual items (samples) analyzed, as listed in the report. In accordance to the NELAC 2003 Standard, a statement on the estimated uncertainty of measurement of any quantitative analysis will be supplied upon request.

Please contact me if you have any questions. My extension is 4406. You may also contact me via email at Mandy.Sullivan@alsglobal.com.

Respectfully submitted,

ALS Group USA, Corp. dba ALS Environmental

Mandy Sullivan
Project Manager

ADDRESS 9143 Philips Highway, Suite 200, Jacksonville, FL 32256
PHONE +1 904 739 2277 | FAX +1 904 739 2011
ALS Group USA, Corp.
dba ALS Environmental



Narrative Documents

ALS Environmental—Jacksonville Laboratory
9143 Philips Highway, Suite 200, Jacksonville, FL 32256
Phone (904) 739-2277 Fax (904) 739-2011
www.alsglobal.com



Client: Florida Department of Environmental Protection
(DEP)
Project: LSJR Downtown
Sample Matrix: Water

Service Request: J1905597

Date Received: 09/19/2019

CASE NARRATIVE

All analyses were performed consistent with the quality assurance program of ALS Environmental. This report contains analytical results for samples for the Tier I level requested by the client.

Sample Receipt:

Three water samples were received for analysis at ALS Environmental on 09/19/2019. Any discrepancies upon initial sample inspection are annotated on the sample receipt and preservation form included within this report. The samples were stored at minimum in accordance with the analytical method requirements.

Microbiology:

Method ASTM D6503-99: Sample(s) J1905597-003 was/were removed from the incubator and read on 9/20/19 at 14:39.

Method SM 9223 B: Sample(s) J1905597 (1, 2) were removed from the incubator and read on 9/20/19 at 14:25.

Approved by 

Date 09/24/2019

SAMPLE DETECTION SUMMARY

CLIENT ID: G2NE1192				Lab ID: J1905597-001		
---------------------	--	--	--	----------------------	--	--

Analyte	Results	Flag	MDL	MRL	Units	Method
Escherichia coli	45.0			1.0	MPN/100m L	SM 9223 B

CLIENT ID: G2NE1159				Lab ID: J1905597-002		
---------------------	--	--	--	----------------------	--	--

Analyte	Results	Flag	MDL	MRL	Units	Method
Escherichia coli	727			1.0	MPN/100m L	SM 9223 B

CLIENT ID: 20030792				Lab ID: J1905597-003		
---------------------	--	--	--	----------------------	--	--

Analyte	Results	Flag	MDL	MRL	Units	Method
Enterococcus	7300			10	MPN/100m L	ASTM D6503-99



Sample Receipt Information

ALS Environmental—Jacksonville Laboratory
9143 Philips Highway, Suite 200, Jacksonville, FL 32256
Phone (904) 739-2277 Fax (904) 739-2011
www.alsglobal.com

Client: Florida Department of Environmental Protection (DEP)
Project: LSJR Downtown/RQ-2019-09-16-17

Service Request:J1905597

SAMPLE CROSS-REFERENCE

<u>SAMPLE #</u>	<u>CLIENT SAMPLE ID</u>	<u>DATE</u>	<u>TIME</u>
J1905597-001	G2NE1192	9/19/2019	0940
J1905597-002	G2NE1159	9/19/2019	1050
J1905597-003	20030792	9/19/2019	1125

Cooler Receipt Form

Client: FDEP Service Request #: 11905597
 Project: LSJR Downturn Shipping paid by ALS? ☒ Yes ☐ No ☐ N/A
 Cooler received on 9/19/19 and opened on 9/19/19 by BB
 COURIER: ALS UPS FEDEX DHL ☒ Client Other _____ Airbill # _____


- | | | | | |
|----|---|--------------------------------------|---|--|
| 1 | Were custody seals on outside of cooler? | Yes | ☒ No | |
| | If yes, how many and where? | #: _____ on lid other _____ | | |
| 2 | Were seals intact and signature and date correct? | Yes | No | ☒ N/A |
| 3 | Were custody papers properly filled out? | ☒ Yes | No | N/A |
| 4 | Temperature of cooler(s) upon receipt (Should be 0°C and ≤ 6°C) | <u>LSJR</u> | | |
| 5 | Thermometer ID | <u>7124</u> | | |
| 6 | Temperature Blank Present? | Yes | ☒ No | |
| 7 | Were Ice or Ice Packs present | ☒ Ice | Ice Packs | No |
| 8 | Did all bottles arrive in good condition (unbroken, etc....)? | ☒ Yes | No | N/A |
| 9 | Type of packing material present | Netting | Vial Holder | Bubble Wrap |
| | | Paper | Styrofoam | Other ☒ N/A |
| 10 | Were all bottle labels complete (sample ID, preservation, etc....)? | ☒ Yes | No | N/A |
| 11 | Did all bottle labels and tags agree with custody papers? | ☒ Yes | No | N/A |
| 12 | Were the correct bottles used for the tests indicated? | ☒ Yes | No | N/A |
| 13 | Were all of the preserved bottles received with the appropriate preservative? | Yes | No | ☒ N/A |
| | HNO3 pH<2 H2SO4 pH<2 ZnAc2/NaOH pH>9 NaOH pH>12 HCl pH<2 | Preservative additions noted below | | |
| 14 | Were all samples received within analysis holding times? | ☒ Yes | No | N/A |
| 15 | Were VOA vials free of air bubbles greater than 6mm? If present, note below | Yes | No | ☒ N/A |
| 16 | Where did the bottles originate? | ALS | ☒ Client | |

Sample ID	Reagent	Lot #	ml added	Initials Date/Time

Additional comments and/or explanation of all discrepancies noted above:

Client approval to run samples if discrepancies noted:

Date:



CHAIN OF CUSTODY / LABORATORY ANALYSIS REQUEST FORM

J1905597

5

Florida DEP
LSJR Downtown



9143 Philips Highway, Suite 200. Jacksonville, FL 32256
Phone (904) 739-2277 / 800-695-7222 x06 / FAX (904) 739-2011

www.alsglobal.com

Page 1 of 1

Project Name LSJR Downtown		Project Number RQ-2019-09-16-17		ANALYSIS REQUESTED (Include Method Number and Container Preservative)																											
Report To J. Parrish		Report CC DEP/FLA/MSD/depstat.FL.US		Preservative																Preservative Key 0. None 1. HCL 2. HNO3 3. H2SO4 4. NaOH 5. Zn. Acetate 6. MeOH 7. NaHSO4 8. Other REMARKS											
Company/Address FDEP DEAR NE ROC 8800 Baymeadows way w Jacksonville, FL		Phone # 		FAX # 		NUMBER OF CONTAINERS Ecoli Enterococci																									
Sampler's Signature Chyanne R		Sampler's Printed Name Chyanne Riben																													
CLIENT SAMPLE ID	LAB ID	SAMPLING DATE TIME		Matrix																											
G2NE1192		9.19.19 0940		SW	1	X															Fresh										
G2NE1159		9.19.19 1050		SW	1	X															Fresh										
20030792		9.19.19 1125		SW	1		X															Salt									
Special Instructions/Comments: LSO																				TURNAROUND REQUIREMENTS RUSH (SURCHARGES APPLY) ☒ STANDARD REQUESTED FAX DATE REQUESTED REPORT DATE				REPORT REQUIREMENTS I. Results Only II. Results + QC Summaries (LCS, DUP, MS/MSD as required) III. Results + QC and Calibration Summaries IV. Data Validation Report with Raw Data Edata Yes No				INVOICE INFORMATION P.O. # Bill to:			
SAMPLE RECEIPT: CONDITION/COOLER TEMP:																				CUSTODY SEALS: Y N											
Relinquished By Signature Chyanne R Printed Name Chyanne Riben Firm FDEP DEAR NE ROC Date/Time 9/19/19 1151					Received By Signature [Signature] Printed Name [Name] Firm ALS Date/Time 9/19/19 1151					Relinquished By Signature Printed Name Firm Date/Time 					Received By Signature Printed Name Firm Date/Time 					Relinquished By Signature Printed Name Firm Date/Time 					Received By Signature Printed Name Firm Date/Time 						



Miscellaneous Forms

ALS Environmental—Jacksonville Laboratory
9143 Philips Highway, Suite 200, Jacksonville, FL 32256
Phone (904) 739-2277 Fax (904) 739-2011
www.alsglobal.com



FLORIDA DEP DATA QUALIFIERS

- B Results based upon colony counts outside the acceptable range.
- D Measurement was made in the field.
- H Value based on field kit determination; results may not be accurate.
- I The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J Estimated value (one of the following reasons is discussed in the project case narrative).
1. The result may be inaccurate because the surrogate recovery limits have been exceeded.
 2. No known quality control criteria exists for the component.
 3. The reported value failed to meet the established quality control criteria for either precision or accuracy.
 4. The sample matrix interfered with the ability to make any accurate determination (e.g., primary and confirmation results show greater than 40% RPD).
 5. The data is questionable because of improper laboratory or field protocols (e.g., GC/MS Tune did not meet method criteria).
- K Off scale low. The value is less than the lowest calibration standard but greater than the method reporting limit (MRL).
- L Off scale high. The analyte is above the upper limit of the linear calibration range.
- M The MDL/MRL has been elevated because the analyte could not be accurately quantified due to matrix interference.
- N Presumptive evidence of the analyte. Confirmation was not performed.
- Q Sample held beyond the accepted holding time.
- T Value reported is less than the laboratory method detection limit. The value is reported for informational purposes only.
- U Indicates that the compound was analyzed for but not detected.
- V Indicates that the analyte was detected in both the sample and the associated method blank.
- Y The laboratory analysis was from an improperly preserved sample.
- Z Too many colonies were present (TNTC). The numeric value represents the filtration volume.



Jacksonville Lab ID # for State Certifications¹

Agency	Number	Expiration Date
Department of Defense	66206	6/30/2020
Florida Department of Health	E82502	6/30/2020
Georgia Department of Natural Resources	958	6/30/2020
Kentucky Division of Waste Management	123042	6/30/2020
Louisiana Department of Environmental Quality	02086	6/30/2020
Maine Department of Health and Human Services	2019004	2/3/2021
New York State Department of Health	12113	4/1/2020
North Carolina Department of Environment and Natural Resources	527	12/31/2019
South Carolina Department of Health and Environmental Control	96021001	6/30/2019
Texas Commission on Environmental Quality	T104704197-19-12	5/31/2020
Virginia Environmental Accreditation Program	460191	12/14/2019

¹ Analyses were performed according to our laboratory's NELAP-approved quality assurance program and any applicable state or agency requirements. The test results meet requirements of the current NELAP/TNI standards or state or agency requirements, where applicable, except as noted in the laboratory case narrative provided. For a specific list of accredited analytes, refer to <http://www.alsglobal.com/en/Our-Services/Life-Sciences/Environmental/Downloads/North-America-Downloads>



ACRONYMS

ASTM	American Society for Testing and Materials
A2LA	American Association for Laboratory Accreditation
CARB	California Air Resources Board
CAS Number	Chemical Abstract Service registry Number
CFC	Chlorofluorocarbon
CFU	Colony-Forming Unit
DEC	Department of Environmental Conservation
DEQ	Department of Environmental Quality
DHS	Department of Health Services
DOE	Department of Ecology
DOH	Department of Health
EPA	U. S. Environmental Protection Agency
ELAP	Environmental Laboratory Accreditation Program
GC	Gas Chromatography
GC/MS	Gas Chromatography/Mass Spectrometry
LUFT	Leaking Underground Fuel Tank
M	Modified
MCL	Maximum Contaminant Level is the highest permissible concentration of a substance allowed in drinking water as established by the USEPA.
MDL	Method Detection Limit
MPN	Most Probable Number
MRL	Method Reporting Limit
NA	Not Applicable
NC	Not Calculated
NCASI	National Council of the Paper Industry for Air and Stream Improvement
ND	Not Detected
NIOSH	National Institute for Occupational Safety and Health
PQL	Practical Quantitation Limit
RCRA	Resource Conservation and Recovery Act
SIM	Selected Ion Monitoring
TPH	Total Petroleum Hydrocarbons
tr	Trace level is the concentration of an analyte that is less than the PQL but greater than or equal to the MDL.

ALS Group USA, Corp.

dba ALS Environmental

Analyst Summary report

Client: Florida Department of Environmental Protection (DEP)
Project: LSJR Downtown/RQ-2019-09-16-17

Service Request: J1905597

Sample Name: G2NE1192
Lab Code: J1905597-001
Sample Matrix: Water

Date Collected: 09/19/19**Date Received:** 09/19/19

Analysis Method
SM 9223 B

Extracted/Digested By

Analyzed By
HHERNANDEZ

Sample Name: G2NE1159
Lab Code: J1905597-002
Sample Matrix: Water

Date Collected: 09/19/19**Date Received:** 09/19/19

Analysis Method
SM 9223 B

Extracted/Digested By

Analyzed By
HHERNANDEZ

Sample Name: 20030792
Lab Code: J1905597-003
Sample Matrix: Water

Date Collected: 09/19/19**Date Received:** 09/19/19

Analysis Method
ASTM D6503-99

Extracted/Digested By

Analyzed By
HHERNANDEZ



Sample Results

ALS Environmental—Jacksonville Laboratory
9143 Philips Highway, Suite 200, Jacksonville, FL 32256
Phone (904) 739-2277 Fax (904) 739-2011
www.alsglobal.com



Microbiology

ALS Environmental—Jacksonville Laboratory
9143 Philips Highway, Suite 200, Jacksonville, FL 32256
Phone (904)739-2277 Fax (904)739-2011
www.alsglobal.com

ALS Group USA, Corp.
dba ALS Environmental

Analytical Report

Client: Florida Department of Environmental Protection (DEP)
Project: LSJR Downtown/RQ-2019-09-16-17
Sample Matrix: Water
Sample Name: 20030792
Lab Code: J1905597-003

Service Request: J1905597
Date Collected: 09/19/19 11:25
Date Received: 09/19/19 11:51
Basis: NA

General Chemistry Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Enterococcus	ASTM D6503-99	7300	MPN/100mL	10	10	09/19/19 13:27	

ALS Group USA, Corp.
dba ALS Environmental

Analytical Report

Client: Florida Department of Environmental Protection (DEP)
Project: LSJR Downtown/RQ-2019-09-16-17
Sample Matrix: Water
Sample Name: G2NE1192
Lab Code: J1905597-001

Service Request: J1905597
Date Collected: 09/19/19 09:40
Date Received: 09/19/19 11:51
Basis: NA

General Chemistry Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Escherichia coli	SM 9223 B	45.0	MPN/100mL	1.0	1	09/19/19 13:51	

ALS Group USA, Corp.
dba ALS Environmental

Analytical Report

Client: Florida Department of Environmental Protection (DEP)
Project: LSJR Downtown/RQ-2019-09-16-17
Sample Matrix: Water
Sample Name: G2NE1159
Lab Code: J1905597-002

Service Request: J1905597
Date Collected: 09/19/19 10:50
Date Received: 09/19/19 11:51
Basis: NA

General Chemistry Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Escherichia coli	SM 9223 B	727	MPN/100mL	1.0	1	09/19/19 13:51	



QC Summary Forms

ALS Environmental - Jacksonville Laboratory
9143 Philips Highway, Suite 200, Jacksonville, FL 32256
Phone (904) 739-2277 Fax (904) 739-2011
www.alsglobal.com



Microbiology

ALS Environmental—Jacksonville Laboratory
9143 Philips Highway, Suite 200, Jacksonville, FL 32256
Phone (904)739-2277 Fax (904)739-2011
www.alsglobal.com

ALS Group USA, Corp.
dba ALS Environmental

Analytical Report

Client: Florida Department of Environmental Protection (DEP)
Project: LSJR Downtown/RQ-2019-09-16-17
Sample Matrix: Water
Sample Name: Method Blank
Lab Code: J1905597-MB

Service Request: J1905597
Date Collected: NA
Date Received: NA
Basis: NA

General Chemistry Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Enterococcus	ASTM D6503-99	1.0 U	MPN/100mL	1.0	1	09/19/19 13:27	
Escherichia coli	SM 9223 B	1.0 U	MPN/100mL	1.0	1	09/19/19 13:51	