

Chemical Analysis Report

NW-DIST-2019-08-07-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **CHOCTAWHATCHEE NORTH RUN**
Request ID: **RQ-2019-08-05-88**
Customer: **NW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
160 Government Center
Suite 208
Pensacola, FL 32502-5794
Attn: Cheryl Bunch

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 03-SEP-2019 14:43

A handwritten signature in black ink, appearing to read "Liang Lin", on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals and Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Choctawhatchee River South of Carlisle Lake Collection Date/Time: 08/06/2019 10:45

Field ID: G3NW0130

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2109302	EPA 180.1 Rev. 2.0	Turbidity	21	A	NTU	P368633	
	EPA 300.0 Rev. 2.1	Chloride	5.1		mg Cl/L	P369056	
		Sulfate	3.4		mg SO4/L	P369058	
		Color (true)	9.9	A	PCU	P368615	
	SM 2320 B-97	Total Alkalinity	44		mg CaCO3/L	P369141	
	SM 2540 C-97	TDS	71		mg/L	P369207	
	SM 2540 D-97	TSS	24		mg/L	P369315	
	SM 4500 F-C-97	Fluoride	0.052	I	mg F/L	P369145	
2109305	EPA 350.1 Rev. 2.0	Ammonia-N	0.018		mg N/L	P369086	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.35		mg N/L	P369196	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.68		mg N/L	P368667	
	EPA 365.1 Rev. 2.0	Total-P	0.073		mg P/L	P368783	
	SM 5310 B-00	Organic Carbon	2.5		mg C/L	P368743	
2109308	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.023		mg P/L	P368628	
2109317	EPA 200.7 Rev. 4.4	Barium	25.9		ug/L	P368724	
		Calcium	15.4		mg/L	P368724	
		Iron	1.57E+03		ug/L	P368724	
		Magnesium	2.94		mg/L	P368724	
		Potassium	1.5		mg/L	P368724	
		Sodium	4.9		mg/L	P368724	
		Zinc	5.0	U	ug/L	P368724	
	EPA 200.8 Rev. 5.4	Aluminum	878		ug/L	P369671	
		Antimony	0.054	I	ug/L	P369671	
		Arsenic	0.86		ug/L	P369671	
		Boron**	15.7		ug/L	P369671	
		Cadmium	0.020	U	ug/L	P369671	
		Chromium	1.9		ug/L	P369671	
		Copper	0.58	I	ug/L	P369671	
		Lead	0.69		ug/L	P369671	
		Nickel	0.67	I	ug/L	P369671	
		Selenium	0.20	U	ug/L	P369671	
		Silver	0.010	U	ug/L	P369671	
		Thallium	0.10	U	ug/L	P369671	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Choctawhatchee River upstream of Holmes Creek

Collection Date/Time: 08/06/2019 13:15

Field ID: G3NW0129

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2109303	EPA 180.1 Rev. 2.0	Turbidity	13		NTU	P368633	
	EPA 300.0 Rev. 2.1	Chloride	5.6		mg Cl/L	P369056	
		Sulfate	3.0		mg SO4/L	P369058	
		Color (true)	25	A	PCU	P368614	
	SM 2320 B-97	Total Alkalinity	38		mg CaCO3/L	P369141	
	SM 2540 C-97	TDS	62		mg/L	P369207	
	SM 2540 D-97	TSS	13		mg/L	P369315	
	SM 4500 F-C-97	Fluoride	0.039	I	mg F/L	P369145	
2109306	EPA 350.1 Rev. 2.0	Ammonia-N	0.024		mg N/L	P369086	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.31		mg N/L	P369196	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.34		mg N/L	P368849	
	EPA 365.1 Rev. 2.0	Total-P	0.038		mg P/L	P368783	
	SM 5310 B-00	Organic Carbon	2.4		mg C/L	P368743	
2109309	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.006	I	mg P/L	P368628	
2109318	EPA 200.7 Rev. 4.4	Barium	21.7		ug/L	P368724	
		Calcium	12.7		mg/L	P368724	
		Iron	620		ug/L	P368724	
		Magnesium	2.43		mg/L	P368724	
		Potassium	0.93	I	mg/L	P368724	
		Sodium	4.8		mg/L	P368724	
		Zinc	5.0	U	ug/L	P368724	
	EPA 200.8 Rev. 5.4	Aluminum	475		ug/L	P369671	
		Antimony	0.050	U	ug/L	P369671	
		Arsenic	0.51		ug/L	P369671	
		Boron**	12.5		ug/L	P369671	
		Cadmium	0.020	U	ug/L	P369671	
		Chromium	0.99	I	ug/L	P369671	
		Copper	0.40	U	ug/L	P369671	
		Lead	0.33	I	ug/L	P369671	
		Nickel	0.50	U	ug/L	P369671	
		Selenium	0.20	U	ug/L	P369671	
		Silver	0.010	U	ug/L	P369671	
		Thallium	0.10	U	ug/L	P369671	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: SISTER RIVER MOUTH

Collection Date/Time: 08/06/2019 12:00

Field ID: G3NW0154

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2109304	EPA 180.1 Rev. 2.0	Turbidity	3.4		NTU	P368633	
	EPA 300.0 Rev. 2.1	Chloride	180		mg Cl/L	P369056	
		Sulfate	27		mg SO4/L	P369058	
		Color (true)	31		PCU	P368614	
	SM 2320 B-97	Total Alkalinity	52		mg CaCO3/L	P369141	
	SM 2540 C-97	TDS	389		mg/L	P369207	
	SM 2540 D-97	TSS	3	I	mg/L	P369315	
	SM 4500 F-C-97	Fluoride	0.049	I	mg F/L	P369145	
2109307	EPA 350.1 Rev. 2.0	Ammonia-N	0.008		mg N/L	P369086	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.33		mg N/L	P369196	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.16		mg N/L	P368849	
	EPA 365.1 Rev. 2.0	Total-P	0.016		mg P/L	P368783	
	SM 5310 B-00	Organic Carbon	4.2		mg C/L	P368743	
2109310	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P368628	
2109319	EPA 200.7 Rev. 4.4	Barium	22.7		ug/L	P368724	
		Calcium	21.2		mg/L	P368724	
		Iron	460		ug/L	P368724	
		Magnesium	14.5		mg/L	P368724	
		Potassium	4.4		mg/L	P368724	
		Sodium	107		mg/L	P368724	
		Zinc	5.0	U	ug/L	P368724	
	EPA 200.8 Rev. 5.4	Aluminum	62		ug/L	P369671	
		Antimony	0.050	U	ug/L	P369671	
		Arsenic	0.48		ug/L	P369671	
		Boron**	52.3		ug/L	P369671	
		Cadmium	0.020	U	ug/L	P369671	
		Chromium	0.40	U	ug/L	P369671	
		Copper	0.40	U	ug/L	P369671	
		Lead	0.20	U	ug/L	P369671	
		Nickel	0.50	U	ug/L	P369671	
		Selenium	0.20	U	ug/L	P369671	
		Silver	0.010	U	ug/L	P369671	
		Thallium	0.10	U	ug/L	P369671	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P368633

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P368724

Component	Result	Code	Units
Barium	1.0	U	ug/L
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P369671

Component	Result	Code	Units
Aluminum	5.0	U	ug/L
Antimony	0.050	U	ug/L
Arsenic	0.050	U	ug/L
Boron	2.0	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.40	U	ug/L
Lead	0.20	U	ug/L
Nickel	0.50	U	ug/L
Selenium	0.20	U	ug/L
Silver	0.010	U	ug/L
Thallium	0.10	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P369056

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P369058

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P369086

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P369196

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P368667

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P368849

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P368783

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P368628

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P368614

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2120 B
Batch ID: P368615

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P369141

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P369207

Component	Result	Code	Units
TDS	15	U	mg/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 2540 D-97
Batch ID: P369315

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P369145

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P368743

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P368724

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Barium	100		P	85 - 115
Calcium	100		P	85 - 115
Iron	100		P	85 - 115
Magnesium	99.9		P	85 - 115
Potassium	106		P	85 - 115
Sodium	102		P	85 - 115
Zinc	98.5		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P369671

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aluminum	99.6		P	85 - 115
Antimony	99.0		P	85 - 115
Arsenic	104		P	85 - 115
Boron	98.4		P	85 - 115
Cadmium	99.4		P	85 - 115
Chromium	101		P	85 - 115
Copper	100		P	85 - 115
Lead	101		P	85 - 115
Nickel	98.6		P	85 - 115
Selenium	101		P	85 - 115
Silver	105		P	85 - 115
Thallium	104		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P369056

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	96.0		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P369058

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P369086

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.9		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P369196

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	101		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P368667

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P368849

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	99.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P368783

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	95.6		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P368628

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.0		P	90 - 110

Reference Method: SM 2120 B
 Batch ID: P368614

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	102		P	85 - 115

Reference Method: SM 2120 B
 Batch ID: P368615

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	103		P	85 - 115

Reference Method: SM 2320 B-97
 Batch ID: P369141

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	104		P	98 - 106

Reference Method: SM 4500 F-C-97
 Batch ID: P369145

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.0		P	92 - 103

Reference Method: SM 5310 B-00
 Batch ID: P368743

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	97.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P368724

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2108886	Barium	99.3		P	80 - 120
2108886	Calcium	95.1		P	80 - 120
2108886	Iron	101		P	80 - 120
2108886	Magnesium	101		P	80 - 120
2108886	Potassium	108		P	80 - 120
2108886	Sodium	115		P	80 - 120
2108886	Zinc	98.1		P	80 - 120
2109519	Barium	99.9	99.9	P/P	80 - 120
2109519	Calcium	98.7		P	80 - 120
2109519	Iron	99.8	102	P/P	80 - 120
2109519	Magnesium	97.0		P	80 - 120
2109519	Potassium	105	107	P/P	80 - 120
2109519	Sodium	111	110	P/P	80 - 120
2109519	Zinc	97.9	97.3	P/P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P369671

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2108885	Aluminum	98.0	100	P/P	80 - 120
2108885	Antimony	99.8	100	P/P	80 - 120
2108885	Arsenic	99.9	101	P/P	80 - 120
2108885	Boron	94.5	99.8	P/P	80 - 120
2108885	Cadmium	99.8	103	P/P	80 - 120
2108885	Chromium	102	103	P/P	80 - 120
2108885	Copper	98.3	99.9	P/P	80 - 120
2108885	Lead	99.9	101	P/P	80 - 120
2108885	Nickel	89.4	93.8	P/P	80 - 120
2108885	Selenium	97.7	98.4	P/P	80 - 120
2108885	Silver	104	105	P/P	80 - 120
2108885	Thallium	103	105	P/P	80 - 120
2109520	Aluminum	97.5		P	80 - 120
2109520	Antimony	102		P	80 - 120
2109520	Arsenic	101		P	80 - 120
2109520	Boron	99.1		P	80 - 120
2109520	Cadmium	101		P	80 - 120
2109520	Chromium	99.7		P	80 - 120
2109520	Copper	94.6		P	80 - 120
2109520	Lead	99.3		P	80 - 120
2109520	Nickel	93.2		P	80 - 120
2109520	Selenium	98.2		P	80 - 120
2109520	Silver	105		P	80 - 120
2109520	Thallium	102		P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P369056

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109138	Chloride	97.7		P	90 - 110
2109413	Chloride	96.9		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P369058

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109413	Sulfate	102		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P369086

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109145	Ammonia-N	98.4	99.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P369196

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109568	Kjeldahl Nitrogen	103	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P368667

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109145	NO2NO3-N	102	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P368849

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109306	NO2NO3-N	96.5	96.2	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P368783

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109307	Total-P	95.1	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P368628

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109309	O-Phosphate-P	96.0	96.5	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P369145

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109510	Fluoride	98.3	97.8	P/P	89 - 106

Reference Method: SM 5310 B-00
Batch ID: P368743

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2109307	Organic Carbon	99.3	99.3	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
 Batch ID: P368633

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109302	Turbidity	3.77	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
 Batch ID: P368724

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109519	Barium	0.0479	Spike	P	0 - 20
2109519	Calcium	0.0511	Spike	P	0 - 20
2109519	Iron	2.04	Spike	P	0 - 20
2109519	Magnesium	1.48	Spike	P	0 - 20
2109519	Potassium	1.71	Spike	P	0 - 20
2109519	Sodium	0.276	Spike	P	0 - 20
2109519	Zinc	0.597	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
 Batch ID: P369671

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2108885	Aluminum	1.78	Spike	P	0 - 20
2108885	Antimony	0.454	Spike	P	0 - 20
2108885	Arsenic	1.44	Spike	P	0 - 20
2108885	Boron	5.05	Spike	P	0 - 20
2108885	Cadmium	3.00	Spike	P	0 - 20
2108885	Chromium	0.782	Spike	P	0 - 20
2108885	Copper	1.68	Spike	P	0 - 20
2108885	Lead	0.913	Spike	P	0 - 20
2108885	Nickel	4.47	Spike	P	0 - 20
2108885	Selenium	0.654	Spike	P	0 - 20
2108885	Silver	1.18	Spike	P	0 - 20
2108885	Thallium	2.46	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P369056

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109138	Chloride	0.233	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P369058

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109138	Sulfate	0.253	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P369086

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109145	Ammonia-N	0.833	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P369196

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109568	Kjeldahl Nitrogen	1.59	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P368667

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109145	NO2NO3-N	0.491	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P368849

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109306	NO2NO3-N	0.137	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P368783

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109307	Total-P	5.56	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P368628

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109309	O-Phosphate-P	0.491	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P368614

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109303	Color (true)	3.16	Sample	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P368615

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109302	Color (true)	18.5	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-97
Batch ID: P369141

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109510	Total Alkalinity	0.757	Sample	P	0 - 2.8

Reference Method: SM 2540 C-97
Batch ID: P369207

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109138	TDS	7.54	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P369145

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109510	Fluoride	0.479	Spike	P	0 - 2.5

Reference Method: SM 5310 B-00
Batch ID: P368743

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2109307	Organic Carbon	0.0	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B

Run ID: A93335

Included Lab Sample IDs: 2109302, 2109303, 2109304

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	104	P/P	90 - 115
Color (true)	103	102	P/P	90 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A93339

Included Lab Sample IDs: 2109308, 2109309, 2109310

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	99.0	P/P	90 - 110
O-Phosphate-P	98.3	99.0	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A93340

Included Lab Sample IDs: 2109302, 2109303, 2109304

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	102	P/P	95 - 105

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A93356

Included Lab Sample IDs: 2109305

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A93373

Included Lab Sample IDs: 2109305, 2109306, 2109307

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	101	101	P/P	90 - 110
Organic Carbon	101	103	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A93423

Included Lab Sample IDs: 2109306, 2109307

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	99.5	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A93438

Included Lab Sample IDs: 2109317, 2109318, 2109319

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Barium	102	102	P/P	90 - 110
Calcium	97.4	98.2	P/P	90 - 110
Iron	98.3	99.0	P/P	90 - 110
Magnesium	97.4	98.5	P/P	90 - 110
Potassium	99.7	99.1	P/P	90 - 110
Sodium	107	104	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A93438

Included Lab Sample IDs: 2109317, 2109318, 2109319

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Zinc	98.9	99.7	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A93467

Included Lab Sample IDs: 2109305

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	97.7	98.4	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A93469

Included Lab Sample IDs: 2109306, 2109307

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	98.5	99.0	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A93486

Included Lab Sample IDs: 2109302, 2109303, 2109304

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	102	P/P	90 - 110
Chloride	102	102	P/P	90 - 110
Sulfate	107	107	P/P	90 - 110
Sulfate	107	108	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A93532

Included Lab Sample IDs: 2109305, 2109306, 2109307

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	103	103	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A93545

Included Lab Sample IDs: 2109302, 2109303, 2109304

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	103	104	P/P	90 - 110
Total Alkalinity	104	103	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A93545

Included Lab Sample IDs: 2109302, 2109303, 2109304

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	100	P/P	90 - 110
Fluoride	100	99.5	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A93600

Included Lab Sample IDs: 2109305, 2109306, 2109307

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	103	104	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A93823

Included Lab Sample IDs: 2109317, 2109318, 2109319

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aluminum	97.5	97.7	P/P	90 - 110
Aluminum	97.7	96.8	P/P	90 - 110
Antimony	96.4	97.0	P/P	90 - 110
Antimony	97.0	96.3	P/P	90 - 110
Arsenic	98.1	98.1	P/P	90 - 110
Arsenic	99.5	98.1	P/P	90 - 110
Cadmium	98.4	97.6	P/P	90 - 110
Cadmium	98.7	98.4	P/P	90 - 110
Chromium	100	99.6	P/P	90 - 110
Chromium	101	100	P/P	90 - 110
Copper	97.1	97.6	P/P	90 - 110
Copper	99.0	97.1	P/P	90 - 110
Lead	100	99.6	P/P	90 - 110
Lead	99.6	99.8	P/P	90 - 110
Selenium	97.0	96.5	P/P	90 - 110
Selenium	97.8	97.0	P/P	90 - 110
Silver	106	106	P/P	90 - 110
Silver	106	105	P/P	90 - 110
Thallium	96.9	97.0	P/P	90 - 110
Thallium	98.7	96.9	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A93935

Included Lab Sample IDs: 2109317, 2109318, 2109319

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Boron	96.1	97.9	P/P	90 - 110
Boron	97.9	97.3	P/P	90 - 110
Nickel	95.1	94.6	P/P	90 - 110
Nickel	96.2	95.1	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			Precision SMP	MS
			LCS				
EPA 180.1 Rev. 2.0	Turbidity					3.77	
EPA 200.7 Rev. 4.4	Barium	100	99.3	99.9	99.9		0.0479
	Calcium	100	95.1	98.7			0.0511
	Iron	100	101	99.8	102		2.04
	Magnesium	99.9	101	97.0			1.48
	Potassium	106	108	105	107		1.71
	Sodium	102	115	111	110		0.276
	Zinc	98.5	98.1	97.9	97.3		0.597
EPA 200.8 Rev. 5.4	Aluminum	99.6	98.0	100	97.5		1.78
	Antimony	99.0	99.8	100	102		0.454
	Arsenic	104	99.9	101	101		1.44
	Boron	98.4	94.5	99.8	99.1		5.05
	Cadmium	99.4	99.8	103	101		3.00
	Chromium	101	102	103	99.7		0.782
	Copper	100	98.3	99.9	94.6		1.68
	Lead	101	99.9	101	99.3		0.913
	Nickel	98.6	89.4	93.8	93.2		4.47
	Selenium	101	97.7	98.4	98.2		0.654
	Silver	105	104	105	105		1.18
	Thallium	104	103	105	102		2.46
EPA 300.0 Rev. 2.1	Chloride	96.0	97.7	96.9		0.233	
	Sulfate	101	102			0.253	
EPA 350.1 Rev. 2.0	Ammonia-N	99.9	98.4	99.4			0.833
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	101	103	100			1.59
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	100	102	102			0.491
	NO ₂ NO ₃ -N	99.5	96.5	96.2			0.137
EPA 365.1 Rev. 2.0	Total-P	95.6	95.1	101			5.56
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.0	96.0	96.5			0.491
SM 2120 B	Color (true)	102				3.16	
	Color (true)	103				18.5	
SM 2320 B-97	Total Alkalinity	104				0.757	
SM 2540 C-97	TDS					7.54	
SM 4500 F-C-97	Fluoride	97.0	98.3	97.8			0.479
SM 5310 B-00	Organic Carbon	97.8	99.3	99.3			0.0

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2109302, 2109303, 2109304
EPA 200.7 Rev. 4.4 / E31780	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2109317, 2109318, 2109319
EPA 200.8 Rev. 5.4 / E31780	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2109317, 2109318, 2109319
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2109302, 2109303, 2109304
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2109302, 2109303, 2109304
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2109305, 2109306, 2109307
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2109305, 2109306, 2109307
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2109305, 2109306, 2109307

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2109305, 2109306, 2109307
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2109308, 2109309, 2109310
SM 2120 B / E31780	True Color measured at 450 nm	2109302, 2109303, 2109304
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2109302, 2109303, 2109304
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2109302, 2109303, 2109304
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2109302, 2109303, 2109304
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2109302, 2109303, 2109304
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2109305, 2109306, 2109307

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/07/2019			08/07/2019 17:10	Rosalyn Ballman	2109302, 2109303, 2109304
EPA 200.7 Rev. 4.4	08/07/2019	08/08/2019 17:10	Elliott D. Healy	08/12/2019 18:57	Betina Topolski	2109317
	08/07/2019	08/08/2019 17:10	Elliott D. Healy	08/12/2019 19:03	Betina Topolski	2109318
	08/07/2019	08/08/2019 17:10	Elliott D. Healy	08/12/2019 19:11	Betina Topolski	2109319
EPA 200.8 Rev. 5.4	08/07/2019	08/26/2019 15:30	Elliott D. Healy	08/27/2019 21:40	Justin Cutchin	2109317
	08/07/2019	08/26/2019 15:30	Elliott D. Healy	08/27/2019 22:19	Justin Cutchin	2109318
	08/07/2019	08/26/2019 15:30	Elliott D. Healy	08/27/2019 22:28	Justin Cutchin	2109319
	08/07/2019	08/26/2019 15:30	Elliott D. Healy	08/30/2019 14:54	Robert K Palmer	2109317
	08/07/2019	08/26/2019 15:30	Elliott D. Healy	08/30/2019 15:00	Robert K Palmer	2109318
	08/07/2019	08/26/2019 15:30	Elliott D. Healy	08/30/2019 15:18	Robert K Palmer	2109319
EPA 300.0 Rev. 2.1	08/07/2019			08/15/2019 01:43	Jhamieka S. Greenwood	2109302
	08/07/2019			08/15/2019 02:19	Jhamieka S. Greenwood	2109303
	08/07/2019			08/15/2019 02:31	Jhamieka S. Greenwood	2109304
EPA 350.1 Rev. 2.0	08/07/2019			08/14/2019 13:05	Ping Hua	2109305
	08/07/2019			08/14/2019 13:07	Ping Hua	2109306
	08/07/2019			08/14/2019 13:08	Ping Hua	2109307
EPA 351.2 Rev. 2.0	08/07/2019	08/16/2019 14:55	Sima Feizi	08/19/2019 12:01	Joseph Suarez	2109305
	08/07/2019	08/16/2019 14:55	Sima Feizi	08/19/2019 12:03	Joseph Suarez	2109306
	08/07/2019	08/16/2019 14:55	Sima Feizi	08/19/2019 12:04	Joseph Suarez	2109307
EPA 353.2 Rev. 2.0	08/07/2019			08/08/2019 09:52	Beverly Y. Sanders	2109305
	08/07/2019			08/12/2019 12:54	Beverly Y. Sanders	2109306
	08/07/2019			08/12/2019 13:01	Beverly Y. Sanders	2109307
EPA 365.1 Rev. 2.0	08/07/2019	08/09/2019 13:46	Vijayalakshmi Reddy	08/13/2019 13:33	Lipi Saha	2109305
	08/07/2019	08/09/2019 13:46	Vijayalakshmi Reddy	08/13/2019 14:14	Lipi Saha	2109307
	08/07/2019	08/09/2019 13:46	Vijayalakshmi Reddy	08/13/2019 14:17	Lipi Saha	2109306
EPA 365.1 Rev. 2.0 dissolved	08/07/2019			08/07/2019 16:12	Jhamieka S. Greenwood	2109309
	08/07/2019			08/07/2019 16:38	Jhamieka S. Greenwood	2109310
	08/07/2019			08/07/2019 16:43	Jhamieka S. Greenwood	2109308
SM 2120 B	08/07/2019			08/07/2019 16:15	Madeline Funaro	2109303, 2109304
	08/07/2019			08/07/2019 16:24	Madeline Funaro	2109302
SM 2320 B-97	08/07/2019			08/15/2019 04:48	Dale L. Simmons	2109302
	08/07/2019			08/15/2019 05:00	Dale L. Simmons	2109303
	08/07/2019			08/15/2019 05:56	Dale L. Simmons	2109304
SM 2540 C-97	08/07/2019			08/12/2019 16:08	Yijie Li	2109302, 2109303, 2109304
SM 2540 D-97	08/07/2019			08/12/2019 16:08	Yijie Li	2109302, 2109303, 2109304
SM 4500 F-C-97	08/07/2019			08/15/2019 04:48	Dale L. Simmons	2109302
	08/07/2019			08/15/2019 05:00	Dale L. Simmons	2109303
	08/07/2019			08/15/2019 05:56	Dale L. Simmons	2109304
SM 5310 B-00	08/07/2019			08/08/2019 14:51	Madeline Funaro	2109307
	08/07/2019			08/08/2019 15:40	Madeline Funaro	2109305

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
SM 5310 B-00	08/07/2019			08/08/2019 20:24	Madeline Funaro	2109306