

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Hodor

RQ: 2019-03-11-39

Project: SMP - North

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 01/30/19

DO FDEP SOP FT 1500	Name	Date	Time GT-ET	Temp °C	Bath meter	D.O. Chem	Meter D.O.	% DO	Probe charge	Probe gain	Pass Fail	Lab Field
Calibr.	Kristene B.	03/13/19	8:20	23.6	29.694	8.48	8.47	99.9	✓	1.038	P/F	L/F
ICV	"	"	8:22	23.5	29.89	8.49	8.49	99.8	✓	1.035	P/F	L/F
CCV	Kristene B.	03/13/19	15:40	23.3	29.869	8.53	8.51	99.9	✓	1.035	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec Cond. FT 200	Name	Date	Time GT-ET	Lot#	Expir. Date	Standard umhos/cm	Meter Reading umhos/cm	Pass Fail	Lab Field
Calibr.	Kristene B.	03/13/19	8:25	281E59	11/20	1000	1001	P/F	L/F
ICV	"	"	8:27	280884	07/20	100	101	P/F	L/F
CCV	Kristene B.	03/13/19	15:42	2709C7	09/19	5000	4950	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH FDEP SOP FT 1100	Name	Date	Time GT-ET	Lot#	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass Fail	Lab Field
Calibr.	Kristene B.	03/13/19	8:36	186926	10/20	7.01	23.3	7.01	-33.2	P/F	L/F
Calibr.	"	"	8:32	185909	09/20	4.00	23.1	4.00	143.0	P/F	L/F
Calibr.	"	"	8:34	183381	05/20	10.02	22.9	10.02	-216.4	P/F	L/F
ICV	"	"	8:36	183381	05/20	10	22.9	10.03	-215.6	P/F	L/F
CCV	Kristene B.	03/13/19	15:44	"	"	4	21.0	4.09	139.5	P/F	L/F
CCV								4.07		P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 183.4, slope from 4 to 7 176.2 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 01/30/19

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time GT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Kristene B.	03/13/19	8:46	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: