

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Hodon

RQ: 2019-03-11-34

Project: SMP-OKeechoke

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 01/30/19

DO	Name	Date	Time	Lot#	Expir. Date	pH Buffer	Temp	Meter Reading	mV	Pass / Fail	Lab / Field
Calibr.	Drew Lidick	03/27/19	0750	21.8	756.0	8.77	8.77	100	—	1.04	(P)F (L)F
ICV	"	"	0752	21.9	756.0	8.76	8.76	99.5	—	1.04	(P)F (L)F
CCV	"	"	1530	22.4	756.0	8.67	8.63	100.1	—	—	(P)F (L)F
CCV											P/F L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

DO	Name	Date	Time	Lot#	Expir. Date	pH Buffer	Temp	Meter Reading	mV	Pass / Fail	Lab / Field
Calibr.	Drew Lidick	03/27/19	0754	2901833	12/20	1000	1000	1000	101.2	(P)F (L)F	
ICV	"	"	0756	280465	12/20	100	101.2	101.2	101.2	(P)F (L)F	
CCV	"	"	1532	2901833	12/20	1000	1000	1000	1000	(P)F (L)F	
CCV										P/F L/F	

Conductivity Acceptance criteria $\pm 5\%$

pH	Name	Date	Time	Lot#	Expir. Date	pH Buffer	Temp	Meter Reading	mV	Pass / Fail	Lab / Field
Calibr.	Drew Lidick	03/27/19	0758	186926	10/20	7.05	22.6	7.00	-34.3	(P)F (L)F	
Calibr.	"	"	0800	187188	11/20	4.00	23.1	4.00	143.1	(P)F (L)F	
Calibr.	"	"	0802	187680	12/20	10.00	22.6	10.00	-214.0	(P)F (L)F	
ICV	"	"	0804	187680	12/20	10.06	22.6	10.00	-213.5	(P)F (L)F	
CCV	"	"	1534	187188	11/20	4.00	23.6	4.10	139.3	(P)F (L)F	
CCV										P/F L/F	

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 179.2, slope from 4 to 7 171.8 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 01/30/19

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Drew Lidick	03/27/19	0806	0.00	0.00	(P)F	(L)F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: