

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: Arya

RQ: 2019-07-29-60

Project: Trend

Boldly "X" box if then qualified do this page

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 07/10/19

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro meter mmHg	D.O. meter mg/L	Meter DO mg/L	2400	Probe Cal/age	Probe Gain	Pass / Fail
Calibr.	Kristene Taylor	08/05/19	8:15	24.4	756	8.35	8.35	99.5		1.03	P/F
ICV	"	"	8:17	24.2	756	8.38	8.37	99.8		1.03	P/F
CCV	Kristene Taylor	"	16:48	25.1	755	8.24	8.22	99.6		1.03	P/F
CCV											P/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75. Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A

Spec. Cond. FT 200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard Amhos/cm	Meter Reading Amhos/cm	Pass / Fail	Lab
Calibr.	Kristene Taylor	08/05/19	8:22	2901833	12/20	1000	1000	P/F	L/F
ICV	"	"	8:25	2812A65	12/20	100	100.2	P/F	L/F
CCV	Kristene Taylor	11	16:49	"	"	1000	1000.2	P/F	L/F
CCV							1000.2	P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1400	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab
Calibr.	Kristene Taylor	08/05/19	8:18	190073	02/21	7.01	24.1	7.01	38.1	P/F	L/F
Calibr.	"	"	8:31	191303	03/21	4.00	24.1	4.00	128.5	P/F	L/F
Calibr.	"	"	8:37	190071	01/21	10.01	24.1	10.01	209	P/F	L/F
ICV	"	"	8:40	"	"	10	24.1	10.02	-200.6	P/F	L/F
CCV	Kristene Taylor	"	"	"	"	4	24.3	4.00	140	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 172.9;

slope from 4 to 7 166.6

(both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 07/10/19

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____

Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Kristene Taylor	08/05/19	8:41	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: