RQ-2021-01-18-16

Login Checklist

Login Station ID: #2

Shipping Method: FedEx | UPS | HD | Greyhound

Date/Time of Receipt: 1/22/2021 9:35

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
					Present?		*Intact?		
	YES	HNO_3	Formalin		Yes	No	Yes	No	
1.2				10		✓			929380094080

COOLER CHECK

***All samples received on ice unless otherwise noted. HNO3 and Formalin preserved samples do NOT have a temperature requirement.**

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. **NCR#** _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. **NCR#** _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. **NCR#** _____
- If coolers or samples are received late; identify the affected samples in an NCR. **NCR#** _____

Chain of Custody/ Field Sheet(s) Included? Yes ✓ No _____

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included? Yes ✓ No _____

CONTAINER CHECK

Evidence Tape on Bottles? Yes ____ No ____ NA ✓ Evidence Tape intact? Yes ____ No ____

If Criminal, photographs taken? Yes ____ No ____ Containers intact? Yes ✓ No ____

Caps on tight? Yes ✓ No ____

Sufficient Sample Volume: Yes ✓ No ____

If NO is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? *(Blue dot on container)* Yes ____ No ✓ Init: IM

Chlorine detected? *(One container checked per Field ID)* Yes ____ No ____ Init: ____

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK *(EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")*

**Acid Preserved Samples: pH< 2.0? Yes ✓ No ____ NA ____ Init: IM

**W-1-4-DIOX *(Green dot on container)* preserved to pH<4.0? Yes ____ No ____ NA ✓ Init: IM

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: IM **Date:** 1/22/2021 **Event contains NCRs (Y/N)?** N

Event ID: LSJR South West Boat 2021
NE-DIST-2021-01-22-01 (SMP)
Date Received: 22-JAN-2021 09:35

Login Checklist

Login Preservation Equipment	Lot # & Exp: Date
Infrared Thermometer ID:	IR TEMP GUN #2 exp:10/28/21
pH Test Strips 0 – 6	
pH Paper 0 – 3	
pH Paper 0 – 13	231019 exp: 11/1/22
Chlorine Test Strips	0261 exp: 05/2023

Additional Comments:

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes_____ Init:_____

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes_____ No_____

If no, were samples shipped on dry ice? Yes_____ No_____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check (W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes_____ No_____

Date and time samples preserved: _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 1/21/2021
Customer: NE-DIST
RQ: RQ-2021-01-18-16

Collected By: Kelly McDonald, Kimberly Appleby
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Black Cr @ Rr Bridge Below Milllog Cr	20030204
Swimming Pen Cr @ Sr 220	20030191
Swimming Pen Creek At Eagles Creek Dr.	G2NE1188

RELINQUISHED BY(SIGNATURE): Kelly A McDonald **DATE/TIME:** 1/21/2021 1600

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By: IM **Rec'd/Inspected Date:** 1/22/21

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "x" this box if there are qualified data on this page.

Meter ID: ProDSS Stitch RQ- 2021-01-18-16 Project: LSJR SW Boat

Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400 Date of Last Temperature Verification 10/14/2020

DO DEP SOP FT 1500	Name	Date	Time CT-(ET)	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	K McDonald	1/21	1823	20.2	764.9	9.1	-	100.6	-	1.048	(P) F	(L) F
ICV			0825	20.3	765.0	9.1	9.10	100.7	-	1.048	(P) F	(L) F
CCV			1319	19.4	761.5	9.2	9.40	102.2	-	1.048	(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L. **Rapid-Pulse Sensors:** DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.
Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50), DO charge N/A. **Steady-state & Galvanic Sensors: DO Gain** & Charge N/A

Spec. Cond. FT 1200	Name	Date	Time CT-(ET)	Lot #	Expir. Date	Standard µmhos/cm	Meter Reading µmhos/cm	Pass / Fail	Lab / Field
Calibr.	K McDonald	1/21	0827	2012228	01/22	1000	1000	(P) F	(L) F
ICV			0830	200902A	09/21	100	103.2	(P) F	(L) F
CCV			1321	200902C	09/21	1413	1395	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria ± 5%

pH DEP SOP FT 1100	Name	Date	Time CT-(ET)	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	K McDonald	1/21	0831	2006F52	06/22	7.01	21.7	7.02	-15.2	(P) F	(L) F
Calibr.			0836	2912B71	12/21	4.00	21.8	4.00	167.9	(P) F	(L) F
Calibr.			0838	2006E28	12/21	10.04	21.7	10.04	-183.2	(P) F	(L) F
ICV			0840	2912B71	12/21	4.00	21.6	4.016	162.4	(P) F	(L) F
CCV			1322	2006E28	12/21	10.03	22.0	10.10	-189.2	(P) F	(L) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50; mV pH 4 Range +180 ± 50; mV pH 10 Range -180 ± 50;
If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)
If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____
If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-(ET)	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	K McDonald	1/21	0845	0.000	0.000	(P) F	(L) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria ± 5 % or ± 0.05m, whichever is greater.

COMMENTS:

RQ-2021-01-18-16
Customer: WQAP
Project ID: SMP

Collected By: Kelly McDonald, Kimberly Applebee
Field Report Prepared By: Kelly McDonald
Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID: 20030204

Location: Black Cr @ Rr Bridge Below Milllog Cr

Comments: Microbiology sample(s) submitted to contract laboratory
Acid preserved back in lab

Matrix: Fresh

(Check one) ☒ GRAB ☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPTh)
1/21/2021	1030 EST	7.64	72.5	13	6.54	0.3	0.7 ☐ VOB (S)	6	106.1	0.05
	1032	7.62	72.1	12.9	6.6	2.0	Central Lab please use top row for login purposes		106	0.05
	1034 EST	7.59	71.8	12.9	6.7	5.5			106.1	0.05

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☒ $H2SO_4$*	☒ <2	1	B
	Lot # SA0293040 Exp: 10.19.2021				
Metals (P-500ML)	☐ W-ICPMS ☐ W-ICP ☐ W-HARD	☐ Ice ☐ HNO_3*	☐ <2		
	Lot # Exp:				
Filtered Nutrients (P125ML)	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter	☒ Ice		1	B
	Filter Lot #				
	Syringe Lot # 9240133				
Chlorophyll (BP-1L)	☒ CHLSUITE-W	☒ Ice		1	B
Physical/Aggregate (Fresh) (P-1L)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice		1	B
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☒ E Coli ☐ F Coli ☐ Enterococci ☒ Contract Lab	☒ Ice		1	B
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN	☐ Ice			
Phytoplankton/Periphyton	☐ DTM-QL-C ☐ PEN-QL-C ☐ DTY-QN-C ☐ PKD-QN-C ☐ DTM-QN-C ☐ PRN-QN-C ☐ PKD-QN-BV	☐ Ice			

Name: Kelly McDonald, Kimberly Applebee

Signature: Kelly A McDonald

Date: 1/21/21

*2.0 ml of Preservative used unless noted otherwise.
**MCAA: Monochloroacetic acid buffer solution

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End of Day CCV
Yes / No

WIN Station Name: Black Cr @ Rr Bridge Below Milllog Cr **Date:** 1/21/2021

WIN/Field ID: 20030204 **WBID:** 2415A **ENR:** _____ **Latitude:** 30.0806

County: CLAY **Org ID:** 21FLA **Longitude:** -81.7611

Receiving Body of Water: Lower St. Johns **RQ-** 2021-01-18-16

Sampling Team Members	WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
[x] Kelly McDonald	x		x		x
[x] Kimberly Applebee		x		x	x
[ ]					
[ ]					

Sampling Equipment

- [x] Sample Container
- [ ] Dipper
- [ ] Syringe
- [ ] Carboy
- [ ] Van Dorn
- [ ] Pole

Depth measured with: Sonar

Secchi Equipment ID:

Collection Method

- [ ] Wading
- [ ] From Shore
- [ ] Other _____
- [ ] Canoe/Kayak
- [ ] Electric Motor
- [x] Gasoline Motor

Water Level: [X] Normal [ ] Low [ ] High [ ] Dry [ ] Pooled [ ] Flooded

Meter ID# STITCH

Photos: No **Flow:** FLOW **Tide:**

Tide Times: High: Low:

Biological Samples Collected: [X] None [ ] HA [ ] SCI [ ] RPS [ ] Algae ID [ ] LVS [ ] LVI Other:

SBIO ID Numbers:

Notes:

Duplicate Sample

Time Zone: EST **Time:** _____ **Field ID:** N/A (WIN ID_DUP)

WIN DMT Activity ID: _____

Blank

Time Zone: _____ **Time:** _____ **Field ID:** _____ (Blank Type_DDMMYYYY)

DI Source: _____ **Location:** _____

DI Type: _____ **Equipment ID:** _____

DI Container (Name/ID): **DI Container Type:**

LIMS EVENT ID: _____ **WIN Import ID:** _____

Enter Field Parameters, Verify Station ID In LIMS DMT: _____ (Initials/Date) **QC Station ID, Field Parameters In LIMS DMT:** _____ (Initials/Date)

Scan/Save Field Sheets _____ (Initials/Date) **Migrate Data to WIN:** _____ (Initials/Date) **Verify Data In WIN:** _____ (Initials/Date)

RQ-2021-01-18-16
Customer: WQAP
Project ID: SMP

Collected By: Kelly McDonald, Kimberly Applebee
Field Report Prepared By: Kelly McDonald
Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID: 20030191

Location: Swimming Pen Cr @ Sr 220

Comments: Microbiology sample(s) submitted to contract laboratory
Acid preserved back in lab

Matrix: Fresh

(Check one) ☒ GRAB ☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPTh)
1/21/2021	1135 EST	6.45	50.4	14.1	7.29	0.3	0.711 ☒ VOB (S)	0.711	1085	0.54
							Central Lab please use top row for login purposes			
	EST									

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☒ H_2SO_4*	☒ <2	1	A
	Lot # SA0293040 Exp:	I10.19.21			
Metals (P-500ML)	☒ W-ICPMS ☒ W-ICP ☒ W-HARD	☒ Ice ☒ HNO_3*	☒ <2	1	A
	Lot # NA0038050 Exp:	2.7.2021			
Filtered Nutrients (P125ML)	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter	☒ Ice		1	A
	Filter Lot #				
	Syringe Lot # 9240133				
Chlorophyll (BP-1L)	☒ CHLSUITE-W	☒ Ice		1	A
Physical/Aggregate (Fresh) (P-1L)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice		1	A
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☒ E Coli ☐ F Coli ☐ Enterococci ☒ Contract Lab	☒ Ice		1	A
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	W-E8321-DI / WE8321-MS / W-PSNP-TQ / W-CT-CL-R / W-PCL-SQ-R / W-CARB-AA (5 ml of MCAA** Buffer Solution) / W-DIELDRIN	☐ Ice			
Phytoplankton/Periphyton	DTM-QL-C / PEN-QL-C / DTY-QN-C / PKD-QN-C / DTM-QN-C / PRN-QN-C / PKD-QN-BV	☐ Ice			

Name: Kelly McDonald, Kimberly Applebee

Signature: Kelly A McDonald

Date: 1/21/2021

*2.0 ml of Preservative used unless noted otherwise.
**MCAA: Monochloroacetic acid buffer solution

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End of Day CCV
[Yes] / No

WIN Station Name: Swimming Pen Cr @ Sr 220 **Date:** 1/21/2021

WIN/Field ID: 20030191 **WBID:** 2410 **ENR:** _____ **Latitude:** 30.1

County: CLAY **Org ID:** 21FLA **Longitude:** -81.7463

Receiving Body of Water: Lower St. Johns **RQ-** 2021-01-18-16

Sampling Team Members	WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
[x] Kelly McDonald	x		x		x
[x] Kimberly Applebee		x		x	x
[ ]					
[ ]					

Sampling Equipment

- [x] Sample Container
- [ ] Dipper
- [ ] Syringe
- [ ] Carboy
- [ ] Van Dorn
- [x] Pole

Depth measured with: Meter
Secchi Equipment ID:

Collection Method

- [ ] Wading
- [ ] From Shore
- [x] Other dock
- [ ] Canoe/Kayak
- [ ] Electric Motor
- [ ] Gasoline Motor

Water Level: [X] Normal [ ] Low [ ] High [ ] Dry [ ] Pooled [ ] Flooded

Meter ID# STITCH

Photos: No **Flow:** FLOW **Tide:**

Tide Times: High: Low:

Biological Samples Collected: [X] None [ ] HA [ ] SCI [ ] RPS [ ] Algae ID [ ] LVS [ ] LVI Other:

SBIO ID Numbers:

Notes:

Duplicate Sample

Time Zone: EST **Time:** _____ **Field ID:** N/A (WIN ID_DUP)

WIN DMT Activity ID: _____

Blank

Time Zone: _____ **Time:** _____ **Field ID:** _____ (Blank Type_DDMMYYYY)

DI Source: _____ **Location:** _____

DI Type: _____ **Equipment ID:** _____

DI Container (Name/ID): **DI Container Type:**

LIMS EVENT ID: _____ **WIN Import ID:** _____

Enter Field Parameters, Verify Station ID In LIMS DMT: _____ (Initials/Date) **QC Station ID, Field Parameters In LIMS DMT:** _____ (Initials/Date)

Scan/Save Field Sheets _____ (Initials/Date) **Migrate Data to WIN:** _____ (Initials/Date) **Verify Data In WIN:** _____ (Initials/Date)

RQ-2021-01-18-16
Customer: WQAP
Project ID: SMP

Collected By: Kelly McDonald, Kimberly Applebee
Field Report Prepared By: Kelly McDonald
Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID: G2NE1188

Location: Swimming Pen Creek At Eagles Creek Dr.

Comments: Microbiology sample(s) submitted to contract laboratory
Acid preserved back in lab

Matrix: Fresh

(Check one) ☒ GRAB ☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPTh)
1/21/2021	1210 EST	8.75	87.9	15.4	7.59	0.09	0.182 ☒ VOB (S)	0.182	262.2	0.13
							Central Lab please use top row for login purposes			
	EST									

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☐ Ice ☐ H_2SO_4*	☐ <2		
	Lot # Exp:				
Metals (P-500ML)	☒ W-ICPMS ☒ W-ICP ☒ W-HARD	☒ Ice ☒ HNO_3*	☒ <2	1	C
	Lot # NA0038050 Exp:	2.7.21			
Filtered Nutrients (P125ML)	☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter	☐ Ice			
	Filter Lot #				
	Syringe Lot #				
Chlorophyll (BP-1L)	☐ CHLSUITE-W	☐ Ice			
Physical/Aggregate (Fresh) (P-1L)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☒ E Coli ☐ F Coli ☐ Enterococci ☒ Contract Lab	☒ Ice		1	C
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	W-E8321-DI / WE8321-MS / W-PSNP-TQ / W-CT-CL-R / W-PCL-SQ-R / W-CARB-AA (5 ml of MCAA** Buffer Solution) / W-DIELDRIN	☐ Ice			
Phytoplankton/Periphyton	DTM-QL-C / PEN-QL-C / DTY-QN-C / PKD-QN-C / DTM-QN-C / PRN-QN-C / PKD-QN-BV	☐ Ice			

Name:
Kelly McDonald
Kimberly Applebee

Signature: Kelly A McDonald

Date: 1/21/2021

*2.0 ml of Preservative used unless noted otherwise.
**MCAA: Monochloroacetic acid buffer solution

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End of Day CCV: [Yes] / No

WIN Station Name: Swimming Pen Creek At Eagles Creek Dr. **Date:** 1/21/2021

WIN/Field ID: G2NE1188 **WBID:** 2410 **ENR:** _____ **Latitude:** 30.089

County: CLAY **Org ID:** 21FLA **Longitude:** -81.7267

Receiving Body of Water: Lower St. Johns **RQ-** 2021-01-18-16

Sampling Team Members	WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
[x] Kelly McDonald	x		x		x
[x] Kimberly Applebee		x		x	x
[ ]					
[ ]					

Sampling Equipment

- [x] Sample Container
- [ ] Dipper
- [ ] Syringe
- [ ] Carboy
- [ ] Van Dorn
- [x] Pole

Depth measured with: Meter

Secchi Equipment ID:

Collection Method

- [ ] Wading
- [x] From Shore
- [ ] Other _____
- [ ] Canoe/Kayak
- [ ] Electric Motor
- [ ] Gasoline Motor

Water Level: [X] Normal [ ] Low [ ] High [ ] Dry [ ] Pooled [ ] Flooded

Meter ID# STITCH

Photos: NO **Flow:** FLOW **Tide:**

Tide Times: High: Low:

Biological Samples Collected: [X] None [ ] HA [ ] SCI [ ] RPS [ ] Algae ID [ ] LVS [ ] LVI Other:

SBIO ID Numbers:

Notes:

MUSHY BANKS

Duplicate Sample

Time Zone: EST **Time:** _____ **Field ID:** N/A (WIN ID_DUP)

WIN DMT Activity ID: _____

Blank

Time Zone: _____ **Time:** _____ **Field ID:** _____ (Blank Type_DDMMYYYY)

DI Source: _____ **Location:** _____

DI Type: _____ **Equipment ID:** _____

DI Container (Name/ID): **DI Container Type:**

LIMS EVENT ID: _____ **WIN Import ID:** _____

Enter Field Parameters, Verify Station ID In LIMS DMT: _____ (Initials/Date) **QC Station ID, Field Parameters In LIMS DMT:** _____ (Initials/Date)

Scan/Save Field Sheets _____ (Initials/Date) **Migrate Data to WIN:** _____ (Initials/Date) **Verify Data In WIN:** _____ (Initials/Date)

CHAIN OF CUSTODY / LABORATORY ANALYSIS REQUEST FORM

ALS Environmental

9143 Philips Highway, Suite 200 Jacksonville, FL 32256 / Ph(904) 739-2277 / FAX (904) 739-2011

SR #

ALS Contact: Mandy Sullivan

Page 1 of 1

Client Name	Project Number
FDEP-NE ROC	RQ-~~2020~~ 2021-01-18-16
Report To	**Report CC**
Dep-Overflowmicro@dep.state.fl.us	
Company/Address	
8800 Baymeadows Way W Jacksonville, Fl 32256	
Phone #	**FAX #**
904-256-1700	
Sampler's Signature	**Sampler's Printed Name**
	Kim Appleby

CLIENT SAMPLE ID	LAB ID	SAMPLING DATE	SAMPLING TIME	Matrix	NUMBER OF CONTAINERS	E.Coli	Enterococci	REMARKS
20030204		1/21/21	1030	SW	1	X		Fresh
20030191		1/21/21	1135	SW	1	X		↓
G2NE1188		1/21/21	1210	SW	1	X		

ANALYSIS REQUESTED (Include Method Number and Container Preservative)

Preservative Key: 0. None; 1. HCL; 2. HNO3; 3. H2SO4; 4. NaOH; 5. Zn. Acetate; 6. MeOH; 7. NaHSO4; 8. ICE

Special Instructions/Comments: Email Report to: Dep-Overflowmicro@dep.state.fl.us

0.0

TURNAROUND REQUIREMENTS

___ RUSH (SURCHARGES APPLY)

X STANDARD

REQUESTED FAX DATE: N/A

REQUESTED REPORT DATE: N/A

REPORT REQUIREMENTS

___ I. Results Only

___ II. Results + QC Summaries (LCS, DUP, MS/MSD as required)

___ III. Results + QC and Calibration Summaries

___ IV. Data Validation Report with Raw Data

Edata ___ Yes ___ No

INVOICE INFORMATION

P.O. # ~~2020~~

Bill to:

Relinquished By	Received By	Relinquished By	Received By	Relinquished By	Received By
Signature:	Signature:	Signature:	Signature:	Signature:	Signature:
Printed Name: Kim Appleby	Printed Name: Mandie Spears	Printed Name:	Printed Name:	Printed Name:	Printed Name:
Firm: FDEP NE ROC	Firm: ALS	Firm:	Firm:	Firm:	Firm:
Date/Time: 1/21/21 1252	Date/Time: 1/21/21 1252	Date/Time:	Date/Time:	Date/Time:	Date/Time:

RQ-2021-01-18-16

Date of Request: 22-DEC-2020
Created By: PARRISH_J On: 22-DEC-2020
Modified By: JASROTIA_P On: 23-DEC-2020
Customer: NE-DIST
Project: SMP
Division: Division of Environmental Assessment and Restoration
District: Northeast District
Event Description: LSJR South West Boat 2021

Send Coolers To:

Phone: 904-256-1700
8800 Baymeadows Way West
Suite 100
Jacksonville, FL 32256
Attn: Jeremy Parrish
Cooler Ship EMail: Jeremy.M.Parrish@dep.state.fl.us

Priority: 3
Event Status: S
Criminal Investigation: NO
Chemistry Request Reviewed By: WATTS_J On: 23-DEC-2020
Biology Request Reviewed By: JASROTIA_P On: 23-DEC-2020
Sampling Kit Required: YES Ship on: 08-JAN-2021
Sampling Kit Shipped: YES On: 07-JAN-2021
Sampling Kit Packed By: SHAIK_A
Preservatives Needed: NO
Date to Receive Samples: 18-JAN-2021
Logged In By:

Send Final Report To:

8800 Baymeadows Way West
Suite 100
Jacksonville, FL 32256
Attn: Jeremy Parrish
Report Notification EMail: Jeremy.M.Parrish@dep.state.fl.us

Comments: Group A: (Surf.Fresh)Nutrients,Metals,Bacteria
Group B:(Surf.Fresh) Nutrients, Bacteria
Group C:(Surf.Fresh)Metals,Bacteria

NE-ALS-ECQ

Packing Notes:

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-1L Number of Bottles: 1 Preserved With: ICE

ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO3/L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR Color (true)	Template: DEFAULT	(SM 2120 B-2011) True Color measured at 450 nm (not valid for NPDES monitoring)
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Bottle Type: BP-1L Number of Bottles: 1 Preserved With: ICE

CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Bottle Type: P-250ML-WI-THI Number of Bottles: 1 Preserved With: ICE

NE-ALS-ECQ	Template: DEFAULT	(SM 9223 B Quanti-Tray) E. coli by Colilert/QT by ALS in Jacksonville

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: HNO3

W-ICP	Template:	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects

Barium
Calcium
Iron
Magnesium

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: HNO3

W-ICP	Template:	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Manganese		
Potassium		
Sodium		
Zinc		
W-ICPMS	Template:	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
*Boron		
Aluminum		
Antimony		
Arsenic		
Beryllium		
Cadmium		
Chromium		
Copper		
Lead		
Nickel		
Selenium		
Silver		
Thallium		

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: ICE And H2SO4

W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Type: P-125ML Number of Bottles: 1 Preserved With: FILTER-ICE

W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Bottle Group B (Water) With 1 Samples:

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: ICE And H2SO4

W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Type: P-125ML Number of Bottles: 1 Preserved With: FILTER-ICE

W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Bottle Type: P-1L Number of Bottles: 1 Preserved With: ICE

ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO3/L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(SM 2120 B-2011) True Color measured at 450 nm (not valid for NPDES monitoring)
Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Bottle Type: BP-1L Number of Bottles: 1 Preserved With: ICE

CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Bottle Type: P-250ML-WI-THI Number of Bottles: 1 Preserved With: ICE

NE-ALS-ECQ	Template: DEFAULT	(SM 9223 B Quanti-Tray) E. coli by Colilert/QT by ALS in Jacksonville

Bottle Group C (Water) With 1 Samples:

Bottle Type: P-250ML-WI-THI Number of Bottles: 1 Preserved With: ICE

NE-ALS-ECQ	Template: DEFAULT	(SM 9223 B Quanti-Tray) E. coli by Colilert/QT by ALS in Jacksonville

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: HNO3

Bottle Group C (Water) With 1 Samples:

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: HNO3

W-ICP Template: (EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects

- Barium
- Calcium
- Iron
- Magnesium
- Manganese
- Potassium
- Sodium
- Zinc

W-ICPMS Template: (EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

- *Boron
- Aluminum
- Antimony
- Arsenic
- Beryllium
- Cadmium
- Chromium
- Copper
- Lead
- Nickel
- Selenium
- Silver
- Thallium

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.