RQ-2021-08-23-05(4)

RQ-2021-08-23-04 (17)

Login Checklist

Login Station ID: #4

Shipping Method: FedEx | UPS | HD | Greyhound

Date/Time of Receipt: 08 | 31 | 2021 @ 11:20

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
					Present?		*Intact?		
	YES	HNO_3	Formalin		Yes	No	Yes	No	
0.8°C				10		X			N/A
1.9°C				11		X			N/A

COOLER CHECK

***All samples received on ice unless otherwise noted. HNO3 and Formalin preserved samples do NOT have a temperature requirement.**

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. **NCR#** _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. **NCR#** _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. **NCR#** _____
- If coolers or samples are received late; identify the affected samples in an NCR. **NCR#** _____

Chain of Custody/ Field Sheet(s) Included? Yes ✓ No _____

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included? Yes ✓ No _____

CONTAINER CHECK

Evidence Tape on Bottles? Yes ____ No ____ NA X

Evidence Tape intact? Yes ____ No ____

If Criminal, photographs taken? Yes ____ No ____

Containers intact? Yes ✓ No ____

Caps on tight? Yes ✓ No ____

Sufficient Sample Volume: Yes ✓ No ____

If NO is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? *(Blue dot on container)* Yes ✓ (TR) No X Init: Tyler

Chlorine detected? *(One container checked per Field ID)* Yes _____ No X (TR) Init: Tyler

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK *(EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")*

**Acid Preserved Samples: pH< 2.0? Yes ✓ No _____ NA _____ Init: Tyler

**W-1-4-DIOX *(Green dot on container)* preserved to pH<4.0? Yes _____ No _____ NA X Init: Tyler

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: Tyler **Date:** 08/31/2021 **Event contains NCRs (Y/N)?** N

Event ID: Nassau Lakes and makeup
NE-DIST-2021-08-31-01 (SMP)
Date Received: 31-AUG-2021 11:20

Login Checklist

Login Preservation Equipment	Lot # & Exp: Date
Infrared Thermometer ID:	#4 Temp Gun EXP:10/28/2021
pH Test Strips 0 – 6	N/A
pH Paper 0 – 3	N/A
pH Paper 0 – 13	210620 EXP:04/15/2023
Chlorine Test Strips	0261 EXP:05/2023

Additional Comments:

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes_____ Init:_____

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes_____ No_____

If no, were samples shipped on dry ice? Yes_____ No_____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check(W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes_____ No_____

Date and time samples preserved: _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 8/30/2021
Customer: NE-DIST
RQ: RQ-2021-08-23-05

Collected By: Jeremy Parrish, Brian Davis
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Deep Creek @ Hamp Hicks Rd	19010084

RELINQUISHED BY(SIGNATURE):

DATE/TIME: 8/30/2021 1600EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By: Tyler N.

Rec'd/Inspected Date: 08/31/2021

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: YSI DEST RQ- 2021-08-23-04 Project: Summit Control

2021-08-23-05 Nassau Manned

Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400 Date of Last Temperature Verification 8/16/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	JPamsh	8/30/21	731	22.5	761.9	8.7	8.7	100.2	—	1.06	(P)/F	(L) F
ICV	"	8/30/21	733	22.5	761.8	8.7	8.7	100.3	—	1.06	(P) F	(L)/ F
CCV	Brian D.	↓	1446	24.1	759.2	8.4	8.41	100.1	/	1.06	(P)/F	(L) F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L. Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.
Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard µmhos/cm	Meter Reading µmhos/cm	Pass / Fail	Lab / Field
Calibr.	JPamsh	8/30/21	734	210426B	5/22	1000	1000	(P)/F	(L) F
ICV	"	↓	735	210426A	5/22	100	99.5	(P) F	(L)/ F
CCV	Brian D.	↓	1449	210426B	5/22	1000	1004	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria ± 5%

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	JPamsh	8/30/21	737	2104E81	4/23	7.	22.8	7.01	-21.8	(P) F	(L)/ F
Calibr.		↓	738	2103654	3/23	4.	22.8	4.00	160.6	(P)/F	(L) F
Calibr.			739	2107A04	12/22	10.	22.8	10.08	-184.2	(P)/F	(L) F
ICV			740	2105654	3/23	4.	22.8	4.02	156.8	(P)/F	(L) F
CCV	Brian D.	↓	1452	21036	4/23	7.01	22.7	7.09	-21.2	P/F	L/F
CCV				2104E81						P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50; mV pH 4 Range +180 ± 50; mV pH 10 Range -180 ± 50;
If mV are recorded: slope from 7 to 10 168.2, slope from 4 to 7 181.7 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? (YES) / NO / NA (not surf. water project)
If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 8/18/21
If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	JPam	8/30/21	750	0.00	0.02	(P) F	(L)/ F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria ± 5 % or ± 0.05m, whichever is greater.

COMMENTS:

RQ-2021-08-23-05
Customer: WQAP
Project ID: SMP

Collected By: Jeremy Parrish, Brian Davis
Field Report Prepared By: Jeremy Parrish
Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID: 19010084

Location: Deep Creek @ Hamp Hicks Rd

Comments: Microbiology sample(s) submitted to contract laboratory

Matrix: Fresh

(Check one) ☒ GRAB ☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPTh)
	1305 EST	3.8	47.4	26.6	4.05	0.1	0.2 ☒ VOB (S)	0.2	52.3	0.02
8/30/2021										
	EST									

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☒ H_2SO_4*	☒ <2	1	C
	Lot # SA1105060 Exp:	4/30/22			
Metals (P-500ML)	☐ W-ICPMS ☐ W-ICP ☐ W-HARD	☐ Ice ☐ HNO_3*	☐ <2		
	Lot # Exp:				
Filtered Nutrients (P125ML)	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter	☒ Ice		1	C
	Filter Lot # SA1J92H5				
	Syringe Lot #				
Chlorophyll (BP-1L)	☒ CHLSUITE-W	☒ Ice		1	C
Physical/Aggregate (Fresh) (P-1L)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice		1	C
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC Formalin Lot #	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☒ E Coli ☐ F Coli ☐ Enterococci ☒ Contract Lab	☒ Ice		1	C
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	W-E8321-DI WE8321-MS W-PSNP-TQ W-CT-CL-R W-PCL-TQ-R W-CARB-AA (5 ml of MCAA** Buffer Solution)	☐ Ice			
Phytoplankton/Periphyton	DTY-QN-C PKD-QN-BV	☐ Ice			

Name: Jeremy Parrish, Brian Davis

Signature: B. D.

Date: 08/30/2021

*2.0 ml of Preservative used unless noted otherwise.
**MCAA: Monochloroacetic acid buffer solution

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
STRATEGIC MONITORING PLAN FIELD SHEET

January 2020

Meter Passed End of Day CCV
Yes / No

WIN Station Name: Deep Creek @ Hamp Hicks Rd **Date:** 8/30/2021

WIN/Field ID: 19010084 **WBID:** 2196 **ENR:** **Latitude:** 30.4617305

County: Nassau **Org ID:** 21FLA **Longitude:** -81.98987

Receiving Body of Water: St Marys River **RQ-** 2021-08-23-05

Sampling Team Members	WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
[x] Jeremy Parrish		x		x	x
[x] Brian Davis	x		x		
[ ]					
[ ]					

Sampling Equipment

- [x] Sample Container
- [ ] Dipper
- [ ] Syringe
- [ ] Carboy
- [ ] Van Dorn
- [x] Pole

Depth measured with: Meter

Secchi Equipment ID:

Collection Method

- [x] Wading
- [ ] From Shore
- [ ] Other
- [ ] Canoe/Kayak
- [ ] Electric Motor
- [ ] Gasoline Motor

Water Level: [x] Normal [ ] Low [ ] High [ ] Dry [ ] Pooled [ ] Flooded

Meter ID# DESI

Photos: No **Flow:** NA **Tide:** **Tide Times:** High: Low:

Biological Samples Collected: [x] None [ ] HA [ ] SCI [ ] RPS [ ] Algae ID [ ] LVS [ ] LVI Other:

SBIO ID Numbers:

Notes:

NEARLY FLOODED, COULD NOT GET TO MAIN CHANNEL.

Duplicate Sample

Time Zone: EST **Time:** **Field ID:** (WIN ID_DUP)

WIN DMT Activity ID:

Blank

Time Zone: **Time:** **Field ID:** (Blank Type_DDMMYYYY)

DI Source: **Location:**

DI Type: **Equipment ID:**

DI Container (Name/ID): **DI Container Type:**

LIMS EVENT ID: **WIN Import ID:**

Enter Field Parameters, Verify Station ID In LIMS DMT: (Initials/Date) **QC Station ID, Field Parameters In LIMS DMT:** (Initials/Date)

Scan/Save Field Sheets BD 08/30/2021 (Initials/Date) **Migrate Data to WIN:** (Initials/Date) **Verify Data In WIN:** (Initials/Date)

ALS Environmental

CHAIN OF CUSTODY / LABORATORY ANALYSIS REQUEST FORM

SR #

ALS Contact

9143 Philips Highway, Suite 120, Jacksonville, FL 32256
Phone (904) 739-2277
www.alsglobal.com

Page _____ of _____

Project Name: SMP	Project Number: 2021-08-23 05
Report To: VEPOverflowNotice@fldep.gov	Report CC:
Phone #	Email Address
Company/Address: 5500 Meadows Way W Suite 100 Jax FL 32256	
Sampler's Signature: B-V-	Sampler's Printed Name: Brian Vann

ANALYSIS REQUESTED (Include Method Number and Container Preservative)

CLIENT SAMPLE ID	LAB ID	SAMPLING DATE	SAMPLING TIME	Matrix	NUMBER OF CONTAINERS	Preservative	REMARKS
19010084		8/10/21	1300	SW	1	1	Fecal

Preservative Key: 0. None; 1. HCL; 2. HNO3; 3. H2SO4; 4. NaOH; 5. Zn. Acetate; 6. MeOH; 7. NaHSO4; 8. Other

Special Instructions/Comments: -0.8

SAMPLE RECEIPT - Cooler Temp(s): _____ Custody Seals: Y N

TURNAROUND REQUIREMENTS
- RUSH (SURCHARGES APPLY)
- ✓ STANDARD

REQUESTED REPORT DATE

REPORT REQUIREMENTS
- I. Results Only
- II. Results + QC Summaries (LCS, DUP, MS/MSD as required)
- III. Results + QC and Calibration Summaries
- IV. Data Validation Report with Raw Data

Edata ___ Yes ___ No

INVOICE INFORMATION
P.O. # _____
Bill to _____

Relinquished By	Received By	Relinquished By	Received By	Relinquished By	Received By
Signature: B-V-	Signature: Valene Spears	Signature	Signature	Signature	Signature
Printed Name: Brian Vann	Printed Name: Valene Spears	Printed Name	Printed Name	Printed Name	Printed Name
Firm: FDEP	Firm: ALS	Firm	Firm	Firm	Firm
Date/Time: 8/10/21 1415	Date/Time: 8/30/21 1415	Date/Time	Date/Time	Date/Time	Date/Time

RQ-2021-08-23-05

Date of Request: 19-JUL-2021
Created By: PARRISH_J On: 19-JUL-2021
Modified By: AYRES_J On: 24-AUG-2021
Customer: NE-DIST
Project: SMP
Division: Division of Environmental Assessment and Restoration
District: Northeast District
Event Description: Nassau Lakes and makeup

Send Coolers To:

Phone: 904-256-1700
8800 Baymeadows Way West
Suite 100
Jacksonville, FL 32256
Attn: Jeremy Parrish
Cooler Ship EMail: Jeremy.M.Parrish@dep.state.fl.us

Priority: 3
Event Status: R
Criminal Investigation: NO
Chemistry Request Reviewed By: AYRES_J On: 24-AUG-2021
Biology Request Reviewed By: JASROTIA_P On: 24-AUG-2021
Sampling Kit Required: NO Ship on:
Sampling Kit Shipped: YES On: 09-AUG-2021
Sampling Kit Packed By: SHAIK_A
Preservatives Needed: NO
Date to Receive Samples: 23-AUG-2021
Logged In By: SHAIK_A On: 27-AUG-2021

Send Final Report To:

8800 Baymeadows Way West
Suite 100
Jacksonville, FL 32256
Attn: Jeremy Parrish
Report Notification EMail: Jeremy.M.Parrish@dep.state.fl.us

Comments: Group A: (Surf.Fresh)Nutrients, Metals, Tracers
Group B:(Surf.Fresh)Nutrients, Metals
Group C;(Surf.Fresh)Nutrients, Ecoli

Packing Notes:

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-1L Number of Bottles: 1 Preserved With: ICE

ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO3/L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(SM 2120 B-2011) True Color measured at 450 nm (not valid for NPDES monitoring)
Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Bottle Type: BP-1L Number of Bottles: 1 Preserved With: ICE

CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Bottle Type: BG-500ML Number of Bottles: 1 Preserved With: ICE

W-E8321-DI	Template: DEFAULT	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
W-E8321-MS	Template: DEFAULT	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: HNO3

W-HARD	Template: DEFAULT	(SM 2340B) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness		
W-ICP	Template:	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Barium		
Calcium		

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: HNO3

W-ICP	Template:	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Iron		
Magnesium		
Manganese		
Potassium		
Sodium		
Zinc		
W-ICPMS	Template:	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
*Boron		
Aluminum		
Antimony		
Arsenic		
Beryllium		
Cadmium		
Chromium		
Copper		
Lead		
Nickel		
Selenium		
Silver		
Thallium		

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: ICE And H2SO4

W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Type: P-125ML Number of Bottles: 1 Preserved With: FILTER-ICE

W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Bottle Group B (Water) With 1 Samples:

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: ICE And H2SO4

W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Type: P-125ML Number of Bottles: 1 Preserved With: FILTER-ICE

W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Bottle Type: P-1L Number of Bottles: 1 Preserved With: ICE

ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO3/L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(SM 2120 B-2011) True Color measured at 450 nm (not valid for NPDES monitoring)
Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Bottle Type: BP-1L Number of Bottles: 1 Preserved With: ICE

CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: HNO3

W-HARD	Template: DEFAULT	(SM 2340B) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness		
W-ICP	Template:	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects

Bottle Group B (Water) With 1 Samples:

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: HNO3

W-ICP	Template:	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Barium		
Calcium		
Iron		
Magnesium		
Manganese		
Potassium		
Sodium		
Zinc		
W-ICPMS	Template:	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
*Boron		
Aluminum		
Antimony		
Arsenic		
Beryllium		
Cadmium		
Chromium		
Copper		
Lead		
Nickel		
Selenium		
Silver		
Thallium		

Bottle Group C (Water) With 1 Samples:

Bottle Type: P-1L Number of Bottles: 1 Preserved With: ICE

ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO3/L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(SM 2120 B-2011) True Color measured at 450 nm (not valid for NPDES monitoring)
Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Bottle Type: BP-1L Number of Bottles: 1 Preserved With: ICE

CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Bottle Type: P-250ML-WI-THI Number of Bottles: 1 Preserved With: ICE

NE-ALS-ECQ	Template: DEFAULT	(SM 9223 B Quanti-Tray) E. coli by Colilert/QT by ALS in Jacksonville

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: ICE And H2SO4

W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Type: P-125ML Number of Bottles: 1 Preserved With: FILTER-ICE

W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.