Login Checklist

RQ- **2021 - 03 - 01 - 15** Login Station ID: **# 3**

Shipping Method: FedEx | UPS | HD | Greyhound Date/Time of Receipt: **03-04-2021 / 10:20**

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
					Present?		*Intact?		
	YES	HNO_3	Formalin		Yes	No	Yes	No	
0.9C				2		x			**9293 8010 4166**

COOLER CHECK

***All samples received on ice unless otherwise noted. HNO3 and Formalin preserved samples do NOT have a temperature requirement.**

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. **NCR#** _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. **NCR#** _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. **NCR#** _____
- If coolers or samples are received late; identify the affected samples in an NCR. **NCR#** _____

Chain of Custody/ Field Sheet(s) Included? Yes **x** No _____

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included? Yes _____ No **X**

CONTAINER CHECK

Evidence Tape on Bottles? Yes ____ No ____ NA **X** Evidence Tape intact? Yes ____ No ____

If Criminal, photographs taken? Yes ____ No ____ Containers intact? Yes **x** No ____

Caps on tight? Yes **x** No ____

Sufficient Sample Volume: Yes **x** No ____

If NO is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? *(Blue dot on container)* Yes____ No **X** Init: **ABS/**

Chlorine detected? *(One container checked per Field ID)* Yes____ No____ Init: **ABS/**

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK *(EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")*

Acid Preserved Samples: pH< 2.0? Yes **X No____ NA ____ Init: **ABS/**

W-1-4-DIOX *(Green dot on container)* preserved to pH<4.0? Yes____ No____ NA **x Init: **ABS/**

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: **ABS/** **Date:** **03-04-2021** **Event contains NCRs (Y/N)?** N

EventID: Little Haw Creek / Lake Dias -CEN-DIST-2021-03-04-02 (SMP)

Login Checklist

Login Preservation Equipment	Lot # & Exp: Date
Infrared Thermometer ID:	IRTG # 3 EXP:10/28/2021
pH Test Strips 0 – 6	213316 BV EXP:05/15/21
pH Paper 0 – 3	N/A
pH Paper 0 – 13	231019 EXP:11/01/2021
Chlorine Test Strips	0261 2023 / 05

Additional Comments:

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes_____ Init: ABS

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes_____ No_____

If no, were samples shipped on dry ice? Yes_____ No_____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check (W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes_____ No_____

Date and time samples preserved: _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 3/3/2021
Customer: CEN-DIST
RQ: RQ-2021-03-01-15

Collected By: Leif Boman, Victoria Schwartz
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
LITTLE HAW CREEK @ 25M DS OF SR40	G2CE0128

RELINQUISHED BY(SIGNATURE): Victoria Schwartz **DATE/TIME:** 3/3/21 16:00

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By: A.SHAIK **Rec'd/Inspected Date:** 03-04-2021

RQ-2021-03-01-15
Customer: WQAP
Project ID: SMP

Collected By: Leif Boman, Victoria Schwartz
Field Report Prepared By: Leif Boman
Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID: G2CE0128

Location: LITTLE HAW CREEK @ 25M DS OF SR40

Comments:

Matrix: Fresh

(Check one) ☒ GRAB ☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPTh)
3/3/2021	1152 EST	2.32	25.9	20.2	4.53	0.3	0.5 ☐ VOB (S)	1.3	91	0.04
							Central Lab please use top row for login purposes			
	EST									

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☐ Ice ☐ H_2SO_4*	☐ <2		
	Lot #	**Exp:**			
Metals (P-500ML)	☒ W-ICPMS ☒ W-ICP ☒ W-HARD	☒ Ice ☒ HNO_3*	☒ <2	1	A
	Lot # NA0344060	**Exp:** 01/04/22			
Filtered Nutrients (P125ML)	☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter	☐ Ice			
	Filter Lot #				
	Syringe Lot #				
Chlorophyll (BP-1L)	☐ CHLSUITE-W	☐ Ice			
Physical/Aggregate (Fresh) (P-1L)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☒ E Coli ☐ F Coli ☐ Enterococci ☐ Contract Lab	☒ Ice		1	A
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-CAR8-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN	☐ Ice			
Phytoplankton/Periphyton	☐ DTM-QL-C ☐ PEN-QL-C ☐ DTY-QN-C ☐ PKD-QN-C ☐ DTM-QN-C ☐ PRN-QN-C ☐ PKD-QN-BV	☐ Ice			

Name: Leif Boman, Victoria Schwartz

Signature: Victoria Schwartz

Date: 3/3/21

*2.0 ml of Preservative used unless noted otherwise.
**MCAA: Monochloroacetic acid buffer solution

RQ-2021-03-01-15

Date of Request: 02-FEB-2021
Created By: BOMAN_L On: 02-FEB-2021
Modified By: JASROTIA_P On: 08-FEB-2021
Customer: CEN-DIST
Project: SMP
Division:
District: Central District
Event Description: Little Haw Creek / Lake Dias

Send Coolers To:

Phone: 407-897-4178
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Natalie Ayala
Cooler Ship EMail: natalie.ayala@dep.state.fl.us

Priority: 3
Event Status: R
Criminal Investigation: NO
Chemistry Request Reviewed By: WATTS_J On: 08-FEB-2021
Biology Request Reviewed By: JASROTIA_P On: 08-FEB-2021
Sampling Kit Required: YES Ship on: 19-FEB-2021
Sampling Kit Shipped: YES On: 16-FEB-2021
Sampling Kit Packed By: SHAIK_A
Preservatives Needed: NO
Date to Receive Samples: 01-MAR-2021
Logged In By: SHAIK_A On: 04-MAR-2021

Send Final Report To:

FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Kalina Warren
Report Notification EMail: natalie.ayala@dep.state.fl.us

Comments: Bottle Group A: Little Haw (WBID 2630C)
Bottle Group B: Lake Dias

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-250ML-WO-THI Number of Bottles: 1 Preserved With: ICE

ECOLI-18QT Template: DEFAULT (SM 9223 Quanti-Tray) Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: HNO3

W-HARD Template: DEFAULT (SM 2340B) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
- Hardness

W-ICP Template: (EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
- Barium
- Calcium
- Iron
- Magnesium
- Potassium
- Sodium
- Zinc

W-ICPMS Template: (EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
- *Boron
- Aluminum
- Antimony
- Arsenic
- Beryllium
- Cadmium
- Chromium
- Copper
- Lead
- Nickel
- Selenium
- Silver

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: HNO3

W-ICPMS	Template:	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
Thallium		

Bottle Group B (Water) With 1 Samples:

Bottle Type: P-1L Number of Bottles: 1 Preserved With: ICE

ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO3/L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(SM 2120 B-2011) True Color measured at 450 nm (not valid for NPDES monitoring)
Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Bottle Type: BP-1L Number of Bottles: 1 Preserved With: ICE

CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: ICE And H2SO4

W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Type: P-125ML Number of Bottles: 1 Preserved With: FILTER-ICE

W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.