

Chemical Analysis Report

CEN-DIST-2021-03-23-03

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Packingham/Blanket**
Request ID: **RQ-2021-03-15-64**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Kalina Warren

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 14-APR-2021 17:51



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Livingston Creek @ Old Avon Park Rd

Collection Date/Time: 03/22/2021 11:17

Field ID: G4CE0210

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2233114	EPA 180.1 Rev. 2.0	Turbidity	1.0		NTU	P395386	
	EPA 300.0 Rev. 2.1	Chloride	17		mg Cl/L	P395584	
		Sulfate	7.4		mg SO4/L	P395586	
	SM 2120 B-2011	Color (true)	200		PCU	P395382	
	SM 2320 B-2011	Total Alkalinity	13		mg CaCO3/L	P395512	
	SM 2540 C-2011	TDS	110		mg/L	P395815	
	SM 2540 D-2011	TSS	2	U	mg/L	P395852	
	SM 4500 F-C-2011	Fluoride	0.084	I	mg F/L	P395517	
2233115	EPA 350.1 Rev. 2.0	Ammonia-N	0.009		mg N/L	P395904	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.94		mg N/L	P395719	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P395411	
	EPA 365.1 Rev. 2.0	Total-P	0.067		mg P/L	P395487	
	SM 5310 B-2011	Organic Carbon	23		mg C/L	P395668	
2233116	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.057		mg P/L	P395363	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P395386

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395584

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395586

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P395904

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P395719

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P395411

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P395487

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P395363

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P395382

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-2011
Batch ID: P395512

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P395815

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P395852

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P395517

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P395668

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P395584

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.9		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P395586

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	97.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P395904

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.0		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P395719

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	106		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P395411

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	105		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P395487

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	106		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P395363

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.9		P	90 - 110

Reference Method: SM 2120 B-2011
 Batch ID: P395382

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	100		P	85 - 115

Reference Method: SM 2320 B-2011
 Batch ID: P395512

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	101		P	93 - 104

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-2011
Batch ID: P395517

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.9		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P395668

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.4		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395584

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233010	Chloride	100		P	90 - 110
2233235	Chloride	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P395586

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233010	Sulfate	99.0		P	90 - 110
2233235	Sulfate	98.1		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P395904

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233236	Ammonia-N	95.8	95.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P395719

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233098	Kjeldahl Nitrogen	101	90.1	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P395411

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233134	NO2NO3-N	103	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P395487

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233085	Total-P	104	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P395363

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233007	O-Phosphate-P	100	99.1	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P395517

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233059	Fluoride	98.7	96.9	P/P	94 - 107

Reference Method: SM 5310 B-2011
Batch ID: P395668

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2233281	Organic Carbon	99.6		P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P395386

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233087	Turbidity	1.94	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P395584

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233235	Chloride	0.570	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P395586

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233235	Sulfate	0.441	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P395904

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233236	Ammonia-N	0.326	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P395719

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233098	Kjeldahl Nitrogen	5.80	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P395411

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233134	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P395487

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233085	Total-P	0.157	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P395363

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233007	O-Phosphate-P	0.715	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B-2011
Batch ID: P395382

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233128	Color (true)	6.40	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P395512

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233059	Total Alkalinity	0.279	Sample	P	0 - 8.1

Reference Method: SM 2540 C-2011
Batch ID: P395815

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233100	TDS	5.83	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P395517

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233059	Fluoride	0.946	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
Batch ID: P395668

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2233281	Organic Carbon	0.273	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A104216

Included Lab Sample IDs: 2233114

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	98.8	99.0	P/P	90 - 110
Sulfate	98.5	98.7	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A104225

Included Lab Sample IDs: 2233116

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.0	98.3	P/P	90 - 110

Reference Method: SM 2120 B-2011

Run ID: A104228

Included Lab Sample IDs: 2233114

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	98.6	101	P/P	90 - 115

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A104229

Included Lab Sample IDs: 2233114

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	101	P/P	95 - 105

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A104242

Included Lab Sample IDs: 2233115

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	104	105	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A104288

Included Lab Sample IDs: 2233114

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	98.5	98.8	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A104288

Included Lab Sample IDs: 2233114

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A104296

Included Lab Sample IDs: 2233115

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 5310 B-2011

Run ID: A104362

Included Lab Sample IDs: 2233115

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	95.8	96.6	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A104425

Included Lab Sample IDs: 2233115

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.2	94.4	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A104463

Included Lab Sample IDs: 2233115

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.9	98.6	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					1.94	
EPA 300.0 Rev. 2.1	Chloride	97.9	100	100		0.570	
	Sulfate	97.3	98.1	99.0		0.441	
EPA 350.1 Rev. 2.0	Ammonia-N	99.0	95.8	95.4			0.326
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	106	101	90.1			5.80
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	105	103	103			0.0
EPA 365.1 Rev. 2.0	Total-P	106	104	104			0.157
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.9	100	99.1			0.715
SM 2120 B-2011	Color (true)	100				6.40	
SM 2320 B-2011	Total Alkalinity	101				0.279	
SM 2540 C-2011	TDS					5.83	
SM 4500 F-C-2011	Fluoride	98.9	98.7	96.9			0.946
SM 5310 B-2011	Organic Carbon	95.4	99.6				0.273

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2233114
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2233114
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2233114
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2233115
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2233115
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2233115
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2233115
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2233116
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2233114
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2233114
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2233114
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2233114
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2233114
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2233115

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	03/23/2021			03/23/2021 15:57	Yen Ta	2233114
EPA 300.0 Rev. 2.1	03/23/2021			03/25/2021 17:02	Lipi Saha	2233114
EPA 350.1 Rev. 2.0	03/23/2021			04/02/2021 18:43	Ping Hua	2233115
EPA 351.2 Rev. 2.0	03/23/2021	03/31/2021 13:16	Benjamin T. Nordmann	04/01/2021 13:10	Virginia P. Brown	2233115
EPA 353.2 Rev. 2.0	03/23/2021			03/24/2021 10:27	Beverly Y. Sanders	2233115
EPA 365.1 Rev. 2.0	03/23/2021	03/25/2021 12:45	Yen Ta	03/26/2021 11:38	Shengyu Ding	2233115
EPA 365.1 Rev. 2.0 dissolved	03/23/2021			03/23/2021 14:08	Lipi Saha	2233116
SM 2120 B-2011	03/23/2021			03/23/2021 16:16	Benjamin T. Nordmann	2233114
SM 2320 B-2011	03/23/2021			03/24/2021 20:04	Dale L. Simmons	2233114
SM 2540 C-2011	03/23/2021			03/26/2021 15:08	Yijie Li	2233114
SM 2540 D-2011	03/23/2021			03/26/2021 15:08	Yijie Li	2233114
SM 4500 F-C-2011	03/23/2021			03/24/2021 20:04	Dale L. Simmons	2233114
SM 5310 B-2011	03/23/2021			03/30/2021 05:40	Madeline Gruver	2233115