From: katie sanderslabs.net
To: DEP-Overflowmicro
Subject: Re: Analytical Report: DEP Fecals,2106243,6/7/2021
Date: Friday, June 18, 2021 11:20:31 AM

Well we did run incorrect dilutions. Would you like us add note or refund invoice. She ran sample one as her duplicate that's why DF 2 because she ran 50ml and 50 ml but rest were ran at 100. When I asked her about it I think she did not have paperwork and ran off bottle and ran wrong. That is our fault. Fecal was ran 1/100 1/10 1/2 all absent so ND 2 is correct.

Let me know what you want for incorrect DF being ran? I can delete note g

Thanks
Katie Strothman
Laboratory Director/ QCM
Sanders Laboratories, Inc.
10090 Bavaria Rd
Ft Myers, FL 33913
941-915-7447 C
239-590-0337 O
Monday - Thursday : 8:30 am - 4:00 pm
Friday: 8:30 am - 1:00 pm
Saturday / Sunday: By appointment only
Closed Holidays

COVID-19 update: In order to do our part to reduce the spread we will have Bay/Suite 2 open and main office will be closed. Please leave your samples in the black fridge and COC in the bin on the cart. Make a copy of the COC for your records at your facility, get copies from client portal, or take picture with your phone we will not be making on site until the CDC space guidelines are removed. Call 239-590-0337 if you need one of our team to assist you with a mask. Call kit pickups in ahead and we will put your bottles in Bay 2 on the cart. Staff will remain at least 6 feet apart at all times in order to do our part to reduce the spread and protect you and your staff. Thank you for helping us meet CDC guidelines. As always our hearts with all of you.

Sanders will always match or beat any quoted price.
The information contained in this email is privileged and confidential information, intended only for the use of the individual (s) distributed to. If the reader of this message is not the intended recipient, you are hereby notified that any distribution or copying of this is strictly prohibited. If you have received this in error, please notify promptly.

From: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Sent: Friday, June 18, 2021 10:50 AM
To: katie sanderslabs.net <katie@sanderslabs.net>
Subject: FW: Analytical Report: DEP Fecals,2106243,6/7/2021

Hi Katie,

Good morning, and thanks for sending these reports. I just have a couple of questions about 2106243 and 2106052.

It looks like in the EDD for 2106243, the incubation start times and end dates/times are the same. Could you take a look and let me know the correct incubation end times? You can either send a revised EDD or just let me know the times and I can fill them in.

Also, I am wondering about the dilutions on report 2106052.

I noticed the Enterolert samples G1SD0104 and G1SD0057 have a dilution factor of 1 even though they are marine, and G2SD0047 has a dilution factor of 2 for both the Enterolert and Fecal coliform tests. I just wanted to check if that was correct because you guys are usually really good about diluting marine samples 1:10. Was that a typo or was 50mL actually used for G2SD0047 and 100mL used for G1SD0104 and G1SD0057? Also, G2SD0047 has a note that just says "g", does that need to be deleted?

Thank you for the help,

Abigail

Abigail L. Baker
Environmental Specialist II
FDEP Laboratories – Biology
2600 Blair Stone Road
Tallahassee, Florida 32399-24000
Abigail.L.Baker@FloridaDEP.gov

-----Original Message-----
From: Sanders Laboratories <reporting@sanderslabs.net>
Sent: Thursday, June 17, 2021 2:50 PM
To: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Subject: Analytical Report: DEP Fecals,2106243,6/7/2021

Please note this a NO reply outgoing emailer only. If you have questions in response to this email. Send your email to Katie@sanderslabs.net or Jeff@sanderslabs.net.

Your sample(s) has (have) been tested in accordance with NELAP requirements unless stated otherwise.

Your report has been emailed by electronic format. A hard copy of the report can be obtained by contacting our office for a fee of $25 per initial report.

Best Regards,

Tami Bright
Office Manager
Tami@sanderslabs.net
10090 Bavaria Rd.
Fort Myers, FL 33913
www.sanderslabs.net

Sanders is dedicated to always putting our clients first. We pride ourselves on having quality data, great customer service, and competitive pricing. Sanders always strives to match or beat any local competitor's pricing. As always thanks for choosing Sanders Laboratories.

The information contained in this email is privileged and confidential information, intended only for the use of the individual (s) distributed to. If the reader of this message is not the intended recipient, you are hereby notified that any distribution or copying of this is strictly prohibited. If you have received this in error, please notify promptly.
[Dep Customer Survey]<http://survey.dep.state.fl.us/?refemail=DEP-Overflowmicro@dep.state.fl.us>