

Chemical Analysis Report

NW-DIST-2021-04-29-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Western Lakes**
Request ID: **RQ-2021-04-26-18**
Customer: **NW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
160 Government Center
Suite 208
Pensacola, FL 32502-5794
Attn: Cheryl Bunch

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 24-MAY-2021 11:31

A handwritten signature in black ink, appearing to read "Liang Lin", is shown on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals and Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Locklin Lake Center

Collection Date/Time: 04/28/2021 13:35

Field ID: G4NW0437

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2243962	EPA 180.1 Rev. 2.0	Turbidity	0.60		NTU	P397432	
	EPA 300.0 Rev. 2.1	Chloride	5.7		mg Cl/L	P398427	
		Sulfate	1.5		mg SO4/L	P398428	
	SM 2120 B-2011	Color (true)	7.9		PCU	P397442	
	SM 2320 B-2011	Total Alkalinity	11		mg CaCO3/L	P398016	
	SM 2540 C-2011	TDS	22	I	mg/L	P397913	
	SM 2540 D-2011	TSS	2	U	mg/L	P397792	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P397577	
2243963	EPA 350.1 Rev. 2.0	Ammonia-N	0.005		mg N/L	P398150	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.16	I	mg N/L	P397767	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.33		mg N/L	P397738	
	EPA 365.1 Rev. 2.0	Total-P	0.006		mg P/L	P397678	
	SM 5310 B-2011	Organic Carbon	1.8		mg C/L	P397804	
2243964	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P397423	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Lake Stone Center

Collection Date/Time: 04/28/2021 11:35

Field ID: G4NW0426

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2243959	EPA 180.1 Rev. 2.0	Turbidity	2.4		NTU	P397432	
	EPA 300.0 Rev. 2.1	Chloride	3.2		mg Cl/L	P398427	
		Sulfate	1.9		mg SO4/L	P398428	
		Color (true)	27	A	PCU	P397442	
	SM 2320 B-2011	Total Alkalinity	3.9		mg CaCO3/L	P398016	
	SM 2540 C-2011	TDS	19	I	mg/L	P397913	
	SM 2540 D-2011	TSS	3	I	mg/L	P397792	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P397577	
2243960	EPA 350.1 Rev. 2.0	Ammonia-N	0.004	I	mg N/L	P398150	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.31		mg N/L	P397767	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P397738	
	EPA 365.1 Rev. 2.0	Total-P	0.013		mg P/L	P397678	
	SM 5310 B-2011	Organic Carbon	5.1		mg C/L	P397804	
2243961	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P397422	
2243979	EPA 200.7 Rev. 4.4	Barium	8.4		ug/L	P398008	
		Iron	240		ug/L	P398008	
		Magnesium	0.83		mg/L	P398008	
		Manganese	13.4		ug/L	P398008	
		Potassium	0.30	I	mg/L	P398008	
		Sodium	1.9	I	mg/L	P398008	
		Zinc	5.0	U	ug/L	P398008	
	EPA 200.8 Rev. 5.4	Aluminum	84		ug/L	P398008	
		Antimony	0.050	U	ug/L	P398254	
		Arsenic	0.30		ug/L	P398254	
		Beryllium	0.030	I	ug/L	P398008	
		Boron**	11.0		ug/L	P398254	
		Cadmium	0.020	U	ug/L	P398254	
		Chromium	0.40	U	ug/L	P398254	
		Copper	0.40	U	ug/L	P398254	
		Lead	0.20	U	ug/L	P398254	
		Nickel	0.50	U	ug/L	P398254	
		Selenium	0.40	U	ug/L	P398254	
		Silver	0.010	U	ug/L	P398008	
		Thallium	0.10	U	ug/L	P398254	
		Hardness	8.8		mg/L	P398008	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: FB

Collection Date/Time: 04/28/2021 12:00

Field ID: FB_04282021

Matrix: W-FIELD-BK

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2243965	EPA 180.1 Rev. 2.0	Turbidity	0.10	U	NTU	P397432	
	EPA 300.0 Rev. 2.1	Chloride	0.20	U	mg Cl/L	P398427	
		Sulfate	0.20	U	mg SO4/L	P398428	
		Color (true)	2.5	U	PCU	P397442	
	SM 2320 B-2011	Total Alkalinity	0.65	U	mg CaCO3/L	P398016	
	SM 2540 C-2011	TDS	15	U	mg/L	P397911	
	SM 2540 D-2011	TSS	2	U	mg/L	P397790	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P397577	
2243966	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P398150	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P397767	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P397738	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P397678	
	SM 5310 B-2011	Organic Carbon	0.50	U	mg C/L	P397881	
2243967	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P397422	
2243980	EPA 200.7 Rev. 4.4	Barium	1.0	U	ug/L	P398008	
		Iron	30	U	ug/L	P398008	
		Magnesium	0.040	U	mg/L	P398008	
		Manganese	1.0	U	ug/L	P398008	
		Potassium	0.30	U	mg/L	P398008	
		Sodium	0.50	U	mg/L	P398008	
		Zinc	5.0	U	ug/L	P398008	
	EPA 200.8 Rev. 5.4	Aluminum	5.0	U	ug/L	P398008	
		Antimony	0.050	U	ug/L	P398254	
		Arsenic	0.050	U	ug/L	P398254	
		Beryllium	0.010	U	ug/L	P398008	
		Boron**	2.0	U	ug/L	P398254	
		Cadmium	0.020	U	ug/L	P398254	
		Chromium	0.40	U	ug/L	P398254	
		Copper	0.40	U	ug/L	P398254	
		Lead	0.20	U	ug/L	P398254	
		Nickel	0.50	U	ug/L	P398254	
		Selenium	0.40	U	ug/L	P398254	
		Silver	0.010	U	ug/L	P398008	
		Thallium	0.10	U	ug/L	P398254	
		Hardness	0.35	U	mg/L	P398008	
	SM 2340B	Hardness	0.35	U	mg/L	P398008	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P397432

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P398008

Component	Result	Code	Units
Barium	1.0	U	ug/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Manganese	1.0	U	ug/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P398008

Component	Result	Code	Units
Aluminum	5.0	U	ug/L
Beryllium	0.010	U	ug/L
Silver	0.010	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P398254

Component	Result	Code	Units
Antimony	0.050	U	ug/L
Arsenic	0.050	U	ug/L
Boron	2.0	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.40	U	ug/L
Lead	0.20	U	ug/L
Nickel	0.50	U	ug/L
Selenium	0.40	U	ug/L
Thallium	0.10	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P398427

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P398428

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P398150

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P397767

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P397738

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P397678

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P397422

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P397423

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B-2011
Batch ID: P397442

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-2011
Batch ID: P398016

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-2011
Batch ID: P397911

Component	Result	Code	Units
TDS	15	U	mg/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 2540 C-2011
Batch ID: P397913

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P397790

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P397792

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P397577

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P397804

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Reference Method: SM 5310 B-2011
Batch ID: P397881

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P398008

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Barium	103		P	85 - 115
Iron	106		P	85 - 115
Magnesium	106		P	85 - 115
Manganese	105		P	85 - 115
Potassium	103		P	85 - 115
Sodium	108		P	85 - 115
Zinc	104		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4

Batch ID: P398008

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aluminum	99.7		P	85 - 115
Beryllium	94.6		P	85 - 115
Silver	104		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4

Batch ID: P398254

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Antimony	94.1		P	85 - 115
Arsenic	98.8		P	85 - 115
Boron	97.0		P	85 - 115
Cadmium	99.2		P	85 - 115
Chromium	95.3		P	85 - 115
Copper	95.0		P	85 - 115
Lead	90.5		P	85 - 115
Nickel	95.2		P	85 - 115
Selenium	96.7		P	85 - 115
Thallium	97.5		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P398427

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	99.9		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P398428

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P398150

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	98.8		P	90 - 110

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P397767

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	99.8		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P397738

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P397678

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P397422

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.2		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P397423

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.8		P	90 - 110

Reference Method: SM 2120 B-2011

Batch ID: P397442

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	85 - 115

Reference Method: SM 2320 B-2011

Batch ID: P398016

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	98.4		P	93 - 104

Reference Method: SM 4500 F-C-2011

Batch ID: P397577

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.7		P	90 - 106

Reference Method: SM 5310 B-2011

Batch ID: P397804

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.4		P	85 - 115

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-2011
Batch ID: P397881

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	97.2		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P398008

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2243833	Barium	101		P	80 - 120
2243833	Iron	107		P	80 - 120
2243833	Magnesium	102		P	80 - 120
2243833	Manganese	104		P	80 - 120
2243833	Potassium	104		P	80 - 120
2243833	Sodium	103		P	80 - 120
2243833	Zinc	105		P	80 - 120
2243979	Barium	102	103	P/P	80 - 120
2243979	Iron	107	108	P/P	80 - 120
2243979	Magnesium	107	108	P/P	80 - 120
2243979	Manganese	104	105	P/P	80 - 120
2243979	Potassium	102	103	P/P	80 - 120
2243979	Sodium	105	104	P/P	80 - 120
2243979	Zinc	101	101	P/P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4

Batch ID: P398008

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2243833	Aluminum	99.0		P	80 - 120
2243833	Beryllium	99.0		P	80 - 120
2243833	Silver	103		P	80 - 120
2243979	Aluminum	117	108	P/P	80 - 120
2243979	Beryllium	110	93.3	P/P	80 - 120
2243979	Silver	106	105	P/P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4

Batch ID: P398254

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2243340	Antimony	93.0	92.4	P/P	80 - 120
2243340	Arsenic	98.9	98.3	P/P	80 - 120
2243340	Boron	100	99.7	P/P	80 - 120
2243340	Cadmium	101	100	P/P	80 - 120
2243340	Chromium	95.2	95.3	P/P	80 - 120
2243340	Copper	98.2	97.4	P/P	80 - 120
2243340	Lead	92.5	92.6	P/P	80 - 120
2243340	Nickel	96.0	95.5	P/P	80 - 120
2243340	Selenium	95.6	95.7	P/P	80 - 120
2243340	Thallium	101	102	P/P	80 - 120
2243647	Antimony	95.9		P	80 - 120
2243647	Arsenic	103		P	80 - 120
2243647	Boron	101		P	80 - 120
2243647	Cadmium	103		P	80 - 120
2243647	Chromium	96.5		P	80 - 120
2243647	Copper	100		P	80 - 120
2243647	Lead	93.3		P	80 - 120
2243647	Nickel	97.6		P	80 - 120
2243647	Selenium	99.3		P	80 - 120
2243647	Thallium	102		P	80 - 120

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P398427

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2243654	Chloride	99.5		P	90 - 110
2244094	Chloride	99.5		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P398428

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2243654	Sulfate	100		P	90 - 110
2244094	Sulfate	99.0		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P398150

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2243949	Ammonia-N	95.4	97.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P397767

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2243963	Kjeldahl Nitrogen	105	106	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P397738

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2243960	NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P397678

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2243960	Total-P	108	107	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P397422

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2243939	O-Phosphate-P	97.7	98.4	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P397423

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2243964	O-Phosphate-P	96.0	98.0	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P397577

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2244187	Fluoride	102	103	P/P	94 - 107

Quality Assurance Report

Matrix Spike Accuracy

Reference Method: SM 5310 B-2011
Batch ID: P397804

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2243830	Organic Carbon	95.3	96.4	P/P	85 - 115

Reference Method: SM 5310 B-2011
Batch ID: P397881

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2244011	Organic Carbon	100	104	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P397432

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2243944	Turbidity	0.255	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P398008

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2243979	Barium	0.361	Spike	P	0 - 20
2243979	Iron	0.350	Spike	P	0 - 20
2243979	Magnesium	0.441	Spike	P	0 - 20
2243979	Manganese	0.273	Spike	P	0 - 20
2243979	Potassium	1.04	Spike	P	0 - 20
2243979	Sodium	0.221	Spike	P	0 - 20
2243979	Zinc	0.0603	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P398008

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2243979	Aluminum	7.42	Spike	P	0 - 20
2243979	Beryllium	15.6	Spike	P	0 - 20
2243979	Silver	1.14	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P398254

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2243340	Antimony	0.650	Spike	P	0 - 20
2243340	Arsenic	0.638	Spike	P	0 - 20
2243340	Boron	0.597	Spike	P	0 - 20
2243340	Cadmium	1.05	Spike	P	0 - 20
2243340	Chromium	0.0842	Spike	P	0 - 20
2243340	Copper	0.831	Spike	P	0 - 20
2243340	Lead	0.0389	Spike	P	0 - 20
2243340	Nickel	0.551	Spike	P	0 - 20
2243340	Selenium	0.114	Spike	P	0 - 20
2243340	Thallium	0.498	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P398427

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2244094	Chloride	2.57	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P398428

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2244094	Sulfate	5.54	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P398150

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2243949	Ammonia-N	1.79	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P397767

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2243963	Kjeldahl Nitrogen	1.10	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P397738

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2243960	NO2NO3-N	0.499	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P397678

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2243960	Total-P	1.40	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P397422

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2243939	O-Phosphate-P	0.714	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P397423

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2243964	O-Phosphate-P	2.06	Spike	P	0 - 20

Reference Method: SM 2120 B-2011
Batch ID: P397442

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2243959	Color (true)	0.479	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-2011
Batch ID: P398016

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2244187	Total Alkalinity	1.31	Sample	P	0 - 8.1

Reference Method: SM 2540 C-2011
Batch ID: P397911

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2243771	TDS	2.55	Sample	P	0 - 10

Reference Method: SM 2540 C-2011
Batch ID: P397913

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2244094	TDS	8.51	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P397577

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2244187	Fluoride	0.929	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
Batch ID: P397804

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2243830	Organic Carbon	0.689	Spike	P	0 - 20

Reference Method: SM 5310 B-2011
Batch ID: P397881

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2244011	Organic Carbon	2.93	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A105087

Included Lab Sample IDs: 2243961, 2243964, 2243967

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	102	P/P	90 - 110
O-Phosphate-P	101	100	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A105090

Included Lab Sample IDs: 2243959, 2243962, 2243965

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	98.5	98.0	P/P	95 - 105

Reference Method: SM 2120 B-2011

Run ID: A105091

Included Lab Sample IDs: 2243959, 2243962, 2243965

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	103	P/P	90 - 115

Reference Method: SM 4500 F-C-2011

Run ID: A105140

Included Lab Sample IDs: 2243959, 2243962, 2243965

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.5	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A105214

Included Lab Sample IDs: 2243960, 2243963, 2243966

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	99.8	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A105217

Included Lab Sample IDs: 2243960, 2243963, 2243966

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	100	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A105235

Included Lab Sample IDs: 2243960, 2243963

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	96.2	96.0	P/P	90 - 110
Organic Carbon	96.9	96.2	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A105257

Included Lab Sample IDs: 2243960, 2243963, 2243966

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
-----------	---------	---------	------------	----------------

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A105257

Included Lab Sample IDs: 2243960, 2243963, 2243966

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	97.0	98.7	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A105278

Included Lab Sample IDs: 2243966

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	96.8	99.0	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A105344

Included Lab Sample IDs: 2243959, 2243962, 2243965

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	100	97.8	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A105387

Included Lab Sample IDs: 2243979, 2243980

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Barium	102	104	P/P	90 - 110
Barium	102	103	P/P	90 - 110
Iron	103	103	P/P	90 - 110
Iron	98.9	101	P/P	90 - 110
Magnesium	103	103	P/P	90 - 110
Magnesium	98.7	101	P/P	90 - 110
Manganese	103	105	P/P	90 - 110
Manganese	104	104	P/P	90 - 110
Potassium	100	103	P/P	90 - 110
Potassium	98.0	101	P/P	90 - 110
Sodium	97.5	100	P/P	90 - 110
Sodium	99.9	101	P/P	90 - 110
Zinc	103	103	P/P	90 - 110
Zinc	99.7	97.4	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A105391

Included Lab Sample IDs: 2243979, 2243980

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Beryllium	101	103	P/P	90 - 110
Beryllium	103	103	P/P	90 - 110
Silver	103	103	P/P	90 - 110
Silver	103	103	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A105414

Included Lab Sample IDs: 2243960, 2243963, 2243966

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
-----------	---------	---------	------------	----------------

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A105414

Included Lab Sample IDs: 2243960, 2243963, 2243966

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.8	98.4	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A105428

Included Lab Sample IDs: 2243979, 2243980

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aluminum	103	101	P/P	90 - 110
Aluminum	105	103	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A105480

Included Lab Sample IDs: 2243979, 2243980

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Antimony	94.3	95.9	P/P	90 - 110
Antimony	96.2	95.3	P/P	90 - 110
Arsenic	101	100	P/P	90 - 110
Arsenic	101	101	P/P	90 - 110
Boron	98.8	98.8	P/P	90 - 110
Boron	99.5	97.8	P/P	90 - 110
Cadmium	101	100	P/P	90 - 110
Cadmium	98.3	99.1	P/P	90 - 110
Chromium	95.7	97.3	P/P	90 - 110
Chromium	96.2	95.4	P/P	90 - 110
Copper	95.7	95.9	P/P	90 - 110
Copper	97.8	97.5	P/P	90 - 110
Lead	95.0	95.5	P/P	90 - 110
Lead	96.5	95.8	P/P	90 - 110
Nickel	95.7	96.1	P/P	90 - 110
Nickel	98.1	97.8	P/P	90 - 110
Selenium	101	102	P/P	90 - 110
Selenium	101	99.9	P/P	90 - 110
Thallium	91.9	91.2	P/P	90 - 110
Thallium	93.7	92.6	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A105509

Included Lab Sample IDs: 2243959, 2243962, 2243965

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	102	P/P	90 - 110
Sulfate	102	102	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			Precision SMP	MS
			LCS				
EPA 180.1 Rev. 2.0	Turbidity					0.255	
EPA 200.7 Rev. 4.4	Barium	103	101	102	103		0.361
	Iron	106	107	107	108		0.350
	Magnesium	106	102	107	108		0.441
	Manganese	105	104	104	105		0.273
	Potassium	103	104	102	103		1.04
	Sodium	108	103	105	104		0.221
	Zinc	104	105	101	101		0.0603
EPA 200.8 Rev. 5.4	Aluminum	99.7	117	108	99.0		7.42
	Antimony	94.1	93.0	92.4	95.9		0.650
	Arsenic	98.8	98.9	98.3	103		0.638
	Beryllium	94.6	110	93.3	99.0		15.6
	Boron	97.0	100	99.7	101		0.597
	Cadmium	99.2	101	100	103		1.05
	Chromium	95.3	95.2	95.3	96.5		0.0842
	Copper	95.0	98.2	97.4	100		0.831
	Lead	90.5	92.5	92.6	93.3		0.0389
	Nickel	95.2	96.0	95.5	97.6		0.551
	Selenium	96.7	95.6	95.7	99.3		0.114
	Silver	104	106	105	103		1.14
	Thallium	97.5	101	102	102		0.498
EPA 300.0 Rev. 2.1	Chloride	99.9	99.5	99.5		2.57	
	Sulfate	100	99.0	100		5.54	
EPA 350.1 Rev. 2.0	Ammonia-N	98.8	95.4	97.4			1.79
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	99.8	105	106			1.10
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	101	100	100			0.499
EPA 365.1 Rev. 2.0	Total-P	104	108	107			1.40
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.2	97.7	98.4			0.714
	O-Phosphate-P	97.8	96.0	98.0			2.06
SM 2120 B-2011	Color (true)	101				0.479	
SM 2320 B-2011	Total Alkalinity	98.4				1.31	
SM 2540 C-2011	TDS					2.55	
	TDS					8.51	
SM 4500 F-C-2011	Fluoride	98.7	102	103			0.929
SM 5310 B-2011	Organic Carbon	95.4	95.3	96.4			0.689
	Organic Carbon	97.2	100	104			2.93

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2243959, 2243962, 2243965
EPA 200.7 Rev. 4.4	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2243979, 2243980
EPA 200.8 Rev. 5.4	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2243979, 2243980
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2243959, 2243962, 2243965
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2243959, 2243962, 2243965
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2243960, 2243963, 2243966
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2243960, 2243963, 2243966
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2243960, 2243963, 2243966
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2243960, 2243963, 2243966
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2243961, 2243964, 2243967

Reference Method Descriptions

Method	Description	Associated Samples
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2243959, 2243962, 2243965
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2243959, 2243962, 2243965
SM 2340B	Hardness by calculation, in aqueous matrices, reported as mg/L CaCO ₃	2243979, 2243980
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2243959, 2243962, 2243965
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2243959, 2243962, 2243965
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2243959, 2243962, 2243965
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2243960, 2243963, 2243966

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	04/29/2021			04/29/2021 17:28	Yen Ta	2243959, 2243962, 2243965
EPA 200.7 Rev. 4.4	04/29/2021	05/11/2021 17:00	Elliott D. Healy	05/12/2021 19:26	Betina Topolski	2243980
	04/29/2021	05/11/2021 17:00	Elliott D. Healy	05/12/2021 22:11	Betina Topolski	2243979
EPA 200.8 Rev. 5.4	04/29/2021	05/11/2021 17:00	Elliott D. Healy	05/12/2021 19:56	McKinley C Carter	2243980
	04/29/2021	05/11/2021 17:00	Elliott D. Healy	05/12/2021 21:46	McKinley C Carter	2243979
	04/29/2021	05/11/2021 17:00	Elliott D. Healy	05/13/2021 23:12	Alexander Thompson	2243980
	04/29/2021	05/11/2021 17:00	Elliott D. Healy	05/14/2021 01:16	Alexander Thompson	2243979
	04/29/2021	05/17/2021 10:05	Alexander Thompson	05/17/2021 17:13	Alexander Thompson	2243980
	04/29/2021	05/17/2021 10:05	Alexander Thompson	05/17/2021 20:24	Alexander Thompson	2243979
EPA 300.0 Rev. 2.1	04/29/2021			05/19/2021 01:06	Lipi Saha	2243959
	04/29/2021			05/19/2021 01:25	Lipi Saha	2243962
	04/29/2021			05/19/2021 01:44	Lipi Saha	2243965
EPA 350.1 Rev. 2.0	04/29/2021			05/13/2021 12:13	Ping Hua	2243960
	04/29/2021			05/13/2021 12:14	Ping Hua	2243963
	04/29/2021			05/13/2021 12:21	Ping Hua	2243966
EPA 351.2 Rev. 2.0	04/29/2021	05/06/2021 13:53	Benjamin T. Nordmann	05/07/2021 15:16	Virginia P. Brown	2243960
	04/29/2021	05/06/2021 13:53	Benjamin T. Nordmann	05/07/2021 15:18	Virginia P. Brown	2243963
	04/29/2021	05/06/2021 13:53	Benjamin T. Nordmann	05/07/2021 15:23	Virginia P. Brown	2243966
EPA 353.2 Rev. 2.0	04/29/2021			05/06/2021 09:58	Beverly Y. Sanders	2243960
	04/29/2021			05/06/2021 09:59	Beverly Y. Sanders	2243963
	04/29/2021			05/06/2021 10:00	Beverly Y. Sanders	2243966
EPA 365.1 Rev. 2.0	04/29/2021	05/05/2021 14:05	Yen Ta	05/06/2021 11:17	Shengyu Ding	2243966
	04/29/2021	05/05/2021 14:05	Yen Ta	05/06/2021 11:19	Shengyu Ding	2243960
	04/29/2021	05/05/2021 14:05	Yen Ta	05/06/2021 11:20	Shengyu Ding	2243963
EPA 365.1 Rev. 2.0 dissolved	04/29/2021			04/29/2021 14:51	Blanca Fach	2243967
	04/29/2021			04/29/2021 14:57	Blanca Fach	2243961
	04/29/2021			04/29/2021 15:05	Blanca Fach	2243964
SM 2120 B-2011	04/29/2021			04/29/2021 16:59	Sarah A Nixon	2243959
	04/29/2021			04/29/2021 17:02	Sarah A Nixon	2243962
	04/29/2021			04/29/2021 17:03	Sarah A Nixon	2243965
SM 2320 B-2011	04/29/2021			05/11/2021 21:32	Dale L. Simmons	2243959
	04/29/2021			05/11/2021 21:40	Dale L. Simmons	2243962
	04/29/2021			05/11/2021 21:46	Dale L. Simmons	2243965
SM 2340B	04/29/2021			05/12/2021 19:26	Betina Topolski	2243980
	04/29/2021			05/12/2021 22:11	Betina Topolski	2243979
SM 2540 C-2011	04/29/2021			05/04/2021 15:06	Yijie Li	2243959, 2243962, 2243965
SM 2540 D-2011	04/29/2021			05/04/2021 15:06	Yijie Li	2243959, 2243962, 2243965
SM 4500 F-C-2011	04/29/2021			05/03/2021 23:44	Dale L. Simmons	2243959
	04/29/2021			05/03/2021 23:55	Dale L. Simmons	2243962
	04/29/2021			05/04/2021 00:04	Dale L. Simmons	2243965

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
SM 5310 B-2011	04/29/2021			05/06/2021 10:33	Madeline Gruver	2243960
	04/29/2021			05/06/2021 11:07	Madeline Gruver	2243963
	04/29/2021			05/07/2021 20:28	Madeline Gruver	2243966