

Chemical Analysis Report

NW-DIST-2021-08-11-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **North Escambia County Run**

Request ID: **RQ-2021-08-09-31**

Customer: **NW-DIST**

Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
160 Government Center
Suite 208
Pensacola, FL 32502-5794
Attn: Myranda Butler

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Environmental Consultant

Date Certified: 02-SEP-2021 11:01



NON-CONFORMANCE REPORT INCLUDED

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals, Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Big Escambia Creek downstream of Fannie Rd Collection Date/Time: 08/09/2021 10:05

Field ID: G4NW0421

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2268464	EPA 180.1 Rev. 2.0	Turbidity	8.3	A	NTU	P402343	
	EPA 300.0 Rev. 2.1	Chloride	4.3		mg Cl/L	P402550	
		Sulfate	0.94		mg SO4/L	P402552	
	SM 2120 B-2011	Color (true)	45	Q	PCU	P402302	
	SM 2320 B-2011	Total Alkalinity	4.6		mg CaCO3/L	P402593	
	SM 2540 C-2011	TDS	42		mg/L	P402858	
	SM 2540 D-2011	TSS	7	I	mg/L	P402749	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P402596	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.010		mg N/L	P402459	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.23		mg N/L	P402966	MS
2268466	EPA 353.2 Rev. 2.0	NO2NO3-N	0.51		mg N/L	P402439	
	EPA 365.1 Rev. 2.0	Total-P	0.016		mg P/L	P402464	MS
	SM 5310 B-2011	Organic Carbon	3.6		mg C/L	P402919	
	EPA 365.1 Rev. 2.0	O-Phosphate-P	0.004	I Q	mg P/L	P402339	
	dissolved						

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2120 B-2011: The sample expired upon receipt.

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0 dissolved: O-Phosphate-P: The sample expired upon receipt.

Sample Location: McDavid Creek 100 m upstream SR99

Collection Date/Time: 08/09/2021 11:35

Field ID: G5NW0150

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2268465	EPA 180.1 Rev. 2.0	Turbidity	5.7		NTU	P402343	
	EPA 300.0 Rev. 2.1	Chloride	3.5		mg Cl/L	P402551	
		Sulfate	0.48	I	mg SO4/L	P402553	
	SM 2120 B-2011	Color (true)	69	A	PCU	P402302	
	SM 2320 B-2011	Total Alkalinity	0.78	I	mg CaCO3/L	P402593	
	SM 2540 C-2011	TDS	26		mg/L	P402859	
	SM 2540 D-2011	TSS	6	I	mg/L	P402750	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P402596	
2268467	EPA 350.1 Rev. 2.0	Ammonia-N	0.008		mg N/L	P402459	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.34		mg N/L	P402966	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.20		mg N/L	P402439	
	EPA 365.1 Rev. 2.0	Total-P	0.013		mg P/L	P402464	MS
	SM 5310 B-2011	Organic Carbon	7.2		mg C/L	P402919	
2268469	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	UQ	mg P/L	P402339	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0 dissolved: O-Phosphate-P: The sample expired upon receipt.

Sample Location: Little Pine Barren Creek below CR99A

Collection Date/Time: 08/09/2021 11:00

Field ID: G4NW0192

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2268477	EPA 8321B	Aldicarb	0.020	U	ug/L	P402170	
		Aldicarb Sulfone	0.0020	U	ug/L	P402170	
		Aldicarb Sulfoxide	0.0020	U	ug/L	P402170	
		Carbaryl	0.0020	U	ug/L	P402170	
		Carbofuran	0.0020	U	ug/L	P402170	
		3-Hydroxycarbofuran	0.0020	U	ug/L	P402170	
		Methiocarb	0.0020	U	ug/L	P402170	
		Methomyl	0.0020	U	ug/L	P402170	
		Oxamyl	0.0020	U	ug/L	P402170	
		Propoxur	0.0020	U	ug/L	P402170	
2268478		2,4-D	0.0020	U	ug/L	P402303	
		Acesulfame K	0.10	U	ug/L	P402297	
		2,4,5-T	0.0040	U	ug/L	P402303	
		AMPA	0.10	U	ug/L	P402297	
		Acetaminophen	0.078		ug/L	P402303	
		Acetamiprid	0.0020	U	ug/L	P402303	
		Endothall	0.25	U	ug/L	P402297	
		Glufosinate	0.10	U	ug/L	P402297	
		Glyphosate	0.10	U	ug/L	P402297	
		Afidopyropen	0.0020	U	ug/L	P402303	
		Bentazon	8.0E-04	U	ug/L	P402303	
		Sucralose	0.077		ug/L	P402297	
		Benzovindiflupyr	0.0020	U	ug/L	P402303	
		Carbamazepine	8.0E-04	U	ug/L	P402303	
		Clothianidin	0.0050	I	ug/L	P402303	
		Dinotefuran	0.0020	U	ug/L	P402303	
		Diuron	0.0020	U	ug/L	P402303	
		Fenuron	0.016	U	ug/L	P402303	
		Fluridone	8.0E-04	U	ug/L	P402303	
		Imazapyr	0.0040	U	ug/L	P402303	
		Hydrocodone	0.0020	U	ug/L	P402303	
		Ibuprofen	0.020	U	ug/L	P402303	
		Imidacloprid	0.018		ug/L	P402303	
		Linuron	0.0040	U	ug/L	P402303	
		Mandestrobin	0.0020	U	ug/L	P402303	
		MCPP	0.0020	U	ug/L	P402303	
		Naproxen	0.0080	U	ug/L	P402303	
		Primidone	0.0040	U	ug/L	P402303	
		Pyraclostrobin	0.0020	U	ug/L	P402303	
		Thiamethoxam	0.0020	U	ug/L	P402303	
		Tolfenpyrad	0.0020	U	ug/L	P402303	
		Triclopyr	0.0040	U	ug/L	P402303	
2268479	EPA 8270E	Aldrin	0.66	U	ng/L	P402230	
		Alpha-BHC	0.10	U	ng/L	P402230	

Field ID: G4NW0192

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2268479	EPA 8270E	Alpha-Chlordane	0.20	U	ng/L	P402230	
		Beta-BHC	0.10	U	ng/L	P402230	
		Beta-Cyfluthrin**	1.3	UJ	ng/L	P402230	LCS
		Bifenthrin**	0.51	U	ng/L	P402230	
		Chlorothalonil	0.51	U	ng/L	P402230	
		Cis-Nonachlor**	0.20	U	ng/L	P402230	
		Cypermethrin	1.5	U	ng/L	P402230	
		Dacthal**	0.15	U	ng/L	P402230	
		DDD-p,p'	0.26	U	ng/L	P402230	
		DDE-p,p'	0.20	U	ng/L	P402230	
		DDT-p,p'	0.26	U	ng/L	P402230	
		Delta-BHC	0.41	U	ng/L	P402230	
		Deltamethrin**	1.3	U	ng/L	P402230	
		Dichlobenil**	0.10	U	ng/L	P402230	
		Dicofol	1.3	U	ng/L	P402230	
		Dieldrin	0.41	U	ng/L	P402230	
		Endosulfan I	0.41	U	ng/L	P402230	
		Endosulfan II	0.41	U	ng/L	P402230	
		Endosulfan Sulfate	0.31	U	ng/L	P402230	
		Endrin	0.41	U	ng/L	P402230	
		Endrin Aldehyde	0.77	U	ng/L	P402230	
		Endrin Ketone	0.41	U	ng/L	P402230	
		Esfenvalerate**	0.77	U	ng/L	P402230	
		Ethalfuralin**	0.15	U	ng/L	P402230	
		Fenpropathrin**	0.26	U	ng/L	P402230	
		Gamma-BHC	0.13	U	ng/L	P402230	
		Gamma-Chlordane	0.23	I	ng/L	P402230	
		Heptachlor	0.20	U	ng/L	P402230	
		Heptachlor Epoxide	0.26	U	ng/L	P402230	MS
		Hexachlorobenzene	1.0	U	ng/L	P402230	
		Kepone**	5.1	U	ng/L	P402230	MS
		Lambda-Cyhalothrin**	0.77	U	ng/L	P402230	
		Methoxychlor	0.31	U	ng/L	P402230	
		MGK-264**	0.20	U	ng/L	P402230	
		Mirex	0.71	U	ng/L	P402230	
		Permethrin	1.6	U	ng/L	P402230	
		Phenothrin**	0.92	U	ng/L	P402230	
		Piperonyl Butoxide**	0.15	U	ng/L	P402230	
		Prallethrin**	2.0	U	ng/L	P402230	
		Tau-Fluvalinate**	0.26	U	ng/L	P402230	
		Tetramethrin**	0.77	U	ng/L	P402230	
		Trans-Nonachlor**	0.20	U	ng/L	P402230	
		Triclosan Methyl**	0.15	U	ng/L	P402230	
		Trifluralin	0.20	U	ng/L	P402230	
	EPA 8276	Chlordane	5.1	U	ng/L	P402230	

Field ID: G4NW0192

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2268479	EPA 8276	Toxaphene	31	UJ	ng/L	P402230	CCV
2268480	EPA 8270E	Acetochlor	0.24	U	ng/L	P402306	
		Alachlor	0.24	U	ng/L	P402306	
		Ametryn	0.71	I	ng/L	P402306	
		Atrazine	4.4		ng/L	P402306	
		Atrazine Desethyl	4.8	I	ng/L	P402306	
		Azinphos Methyl	1.9	U	ng/L	P402306	
		Azoxystrobin**	7.4	J	ng/L	P402306	
		Bromacil	0.48	U	ng/L	P402306	
		Butylate	0.38	U	ng/L	P402306	
		Caffeine**	8.8	I	ng/L	P402306	
		Carbophenothion	0.48	U	ng/L	P402306	
		Carfentrazone Ethyl**	0.096	U	ng/L	P402306	
		Chlorpyrifos Ethyl	0.24	U	ng/L	P402306	
		Chlorpyrifos Methyl	0.048	U	ng/L	P402306	
		Cyanazine	3.1	U	ng/L	P402306	
		Demeton	1.4	UJ	ng/L	P402306	CCV
		Diazinon	0.12	U	ng/L	P402306	
		Difenoconazole**	0.58	U	ng/L	P402306	
		Dimethenamid**	0.096	U	ng/L	P402306	
		Disulfoton	0.48	U	ng/L	P402306	
		Dithiopyr**	0.17	U	ng/L	P402306	
		EPTC	1.2	U	ng/L	P402306	
		Ethion	0.14	U	ng/L	P402306	RPD
		Ethoprop	0.096	U	ng/L	P402306	
		Fenamiphos	0.48	U	ng/L	P402306	
		Fenbuconazole**	0.12	U	ng/L	P402306	
		Fipronil	0.24	U	ng/L	P402306	
		Fipronil Desulfinyl**	0.12	U	ng/L	P402306	
		Fipronil Sulfide	0.14	U	ng/L	P402306	
		Fipronil Sulfone	0.14	U	ng/L	P402306	
		Fluazifop-P-butyl**	0.096	U	ng/L	P402306	
		Fludioxonil**	0.12	U	ng/L	P402306	
		Flumioxazin**	1.7	U	ng/L	P402306	
		Flutolanil**	2.6		ng/L	P402306	
		Fluxapyroxad**	4.2		ng/L	P402306	
		Fonofos	0.19	U	ng/L	P402306	
		Hexazinone	0.48	U	ng/L	P402306	
		Imazalil**	0.72	U	ng/L	P402306	
		Iprodione**	0.58	U	ng/L	P402306	
		Malathion	0.34	U	ng/L	P402306	
		Metalaxyl	0.72	U	ng/L	P402306	
		Metolachlor	11		ng/L	P402306	
		Metribuzin	3.1	U	ng/L	P402306	
		Mevinphos	1.0	U	ng/L	P402306	

Field ID: G4NW0192

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2268480	EPA 8270E	Molinate	0.29	U	ng/L	P402306	
		Myclobutanil**	0.24	U	ng/L	P402306	
		Naled**	1.4	U	ng/L	P402306	
		Norflurazon	0.49	I	ng/L	P402306	
		Oxadiazon	0.29	U	ng/L	P402306	
		Oxyfluorfen**	0.14	U	ng/L	P402306	
		Parathion Ethyl	0.24	U	ng/L	P402306	
		Parathion Methyl	0.53	U	ng/L	P402306	
		Pendimethalin	0.96	U	ng/L	P402306	
		Phorate	0.50	U	ng/L	P402306	
		Prodiamine**	0.17	U	ng/L	P402306	
		Prometon	0.86	U	ng/L	P402306	
		Prometryn	0.19	U	ng/L	P402306	
		Pronamide**	0.14	U	ng/L	P402306	
		Propiconazole**	3.2		ng/L	P402306	
		Pyriproxyfen**	0.12	U	ng/L	P402306	
		Simazine	0.48	U	ng/L	P402306	
		Sulfentrazone**	0.48	U	ng/L	P402306	
		Tebuconazole**	25		ng/L	P402306	
		Terbufos	0.096	U	ng/L	P402306	
		Terbuthylazine	0.41	U	ng/L	P402306	
2268482	EPA 200.7 Rev. 4.4	Barium	61.7		ug/L	P402516	
		Calcium	2.67		mg/L	P402516	
		Iron	2.73E+03		ug/L	P402516	
		Magnesium	1.99		mg/L	P402516	
		Manganese	118		ug/L	P402516	
		Potassium	1.4		mg/L	P402516	
		Sodium	2.6		mg/L	P402516	
		Zinc	5.5	I	ug/L	P402516	
	EPA 200.8 Rev. 5.4	Aluminum	54		ug/L	P402516	
		Antimony	0.050	U	ug/L	P402516	
		Arsenic	0.79		ug/L	P402516	
		Beryllium	0.020	I	ug/L	P402516	
		Boron**	12.2		ug/L	P402516	
		Cadmium	0.020	U	ug/L	P402516	
		Chromium	0.40	U	ug/L	P402516	
		Copper	0.40	U	ug/L	P402516	
		Lead	0.20	U	ug/L	P402516	
		Nickel	0.50	U	ug/L	P402516	
		Selenium	0.20	U	ug/L	P402516	
		Silver	0.010	U	ug/L	P402516	
		Thallium	0.10	U	ug/L	P402516	
		Hardness	14.9		mg/L	P402516	
	SM 2340B						

Field ID: G4NW0192

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
-----------	-------------	-----------	--------	------	-------	----------	----------

Ref. Method and Comment:
EPA 8270E: Refer to the Lab Analysis Report for an explanation of QC Codes.
EPA 8276: Refer to the Lab Analysis Report for an explanation of QC Codes.
EPA 8270E: MS accuracy for Atrazine and Caffeine not assessed due to high content of these parameters in the sample spiked. Estimated value for azoxystrobin due to internal standard recovery outside control limits. Refer to the Lab Analysis Report for an explanation of QC Codes.

Sample Location: Wiggins Branch below Highway 29

Collection Date/Time: 08/09/2021 09:40

Field ID: G4NW0163

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2268481	EPA 200.7 Rev. 4.4	Barium	26.0		ug/L	P402516	
		Calcium	2.51		mg/L	P402516	
		Iron	2.38E+03		ug/L	P402516	
		Magnesium	0.93		mg/L	P402516	
		Manganese	59.9		ug/L	P402516	
		Potassium	0.35	I	mg/L	P402516	
		Sodium	1.9	I	mg/L	P402516	
	EPA 200.8 Rev. 5.4	Zinc	5.0	U	ug/L	P402516	
		Aluminum	92		ug/L	P402516	
		Antimony	0.050	U	ug/L	P402516	
		Arsenic	0.53		ug/L	P402516	
		Beryllium	0.043		ug/L	P402516	
		Boron**	10.8		ug/L	P402516	
		Cadmium	0.020	U	ug/L	P402516	
		Chromium	0.40	U	ug/L	P402516	
		Copper	0.40	U	ug/L	P402516	
		Lead	0.23	I	ug/L	P402516	
		Nickel	0.57	I	ug/L	P402516	
		Selenium	0.20	U	ug/L	P402516	
		Silver	0.010	U	ug/L	P402516	
		Thallium	0.10	U	ug/L	P402516	
	SM 2340B	Hardness	10.1		mg/L	P402516	

Non-Conformance Report

NCR ID: 9016

Event(s)

NW-DIST-2021-08-11-01

Job(s)

Sample(s)

Test(s)

NCR Type: SHIPPING/RECEIVING

NCR Category: Samples arrived late

Observation: Samples arrived late, samples received within the acceptable temperature range.

Resolution: Samples will be qualified as appropriate.

Authorized by/Date: Joshua Ayres 9/2/2021

The Non-Conformance Report details exceptions or problems encountered with the events/jobs/samples/test.
Please address questions to:

Chemistry	Colin Wright	(850) 245-8085
Biology	Cheryl Swanson	(850) 245-8177

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P402343

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P402516

Component	Result	Code	Units
Barium	1.0	U	ug/L
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Manganese	1.0	U	ug/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P402516

Component	Result	Code	Units
Aluminum	5.0	U	ug/L
Antimony	0.050	U	ug/L
Arsenic	0.050	U	ug/L
Beryllium	0.010	U	ug/L
Boron	2.0	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.40	U	ug/L
Lead	0.20	U	ug/L
Nickel	0.50	U	ug/L
Selenium	0.20	U	ug/L
Silver	0.010	U	ug/L
Thallium	0.10	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402550

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402551

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402552

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402553

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P402459

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P402966

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P402439

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P402464

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P402339

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8270E
Batch ID: P402230

Component	Result	Code	Units
Aldrin	0.65	U	ng/L
Aldrin	0.65	U	ng/L
Alpha-BHC	0.10	U	ng/L
Alpha-BHC	0.10	U	ng/L
Alpha-Chlordane	0.20	U	ng/L
Alpha-Chlordane	0.20	U	ng/L
Beta-BHC	0.10	U	ng/L
Beta-BHC	0.10	U	ng/L
Beta-Cyfluthrin	1.2	U	ng/L
Beta-Cyfluthrin	1.2	U	ng/L
Bifenthrin	0.50	U	ng/L
Bifenthrin	0.50	U	ng/L
Chlorothalonil	0.50	U	ng/L
Chlorothalonil	0.50	U	ng/L
Cis-Nonachlor	0.20	U	ng/L
Cis-Nonachlor	0.20	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E

Batch ID: P402230

Component	Result	Code	Units
Cypermethrin	1.5	U	ng/L
Cypermethrin	1.5	U	ng/L
Dacthal	0.15	U	ng/L
Dacthal	0.15	U	ng/L
DDD-p,p'	0.25	U	ng/L
DDD-p,p'	0.25	U	ng/L
DDE-p,p'	0.20	U	ng/L
DDE-p,p'	0.20	U	ng/L
DDT-p,p'	0.25	U	ng/L
DDT-p,p'	0.25	U	ng/L
Delta-BHC	0.40	U	ng/L
Delta-BHC	0.40	U	ng/L
Deltamethrin	1.2	U	ng/L
Deltamethrin	1.2	U	ng/L
Dichlobenil	0.10	U	ng/L
Dichlobenil	0.10	U	ng/L
Dicofol	1.2	U	ng/L
Dicofol	1.2	U	ng/L
Dieldrin	0.40	U	ng/L
Dieldrin	0.40	U	ng/L
Endosulfan I	0.40	U	ng/L
Endosulfan I	0.40	U	ng/L
Endosulfan II	0.40	U	ng/L
Endosulfan II	0.40	U	ng/L
Endosulfan Sulfate	0.30	U	ng/L
Endosulfan Sulfate	0.30	U	ng/L
Endrin	0.40	U	ng/L
Endrin	0.40	U	ng/L
Endrin Aldehyde	0.75	U	ng/L
Endrin Aldehyde	0.75	U	ng/L
Endrin Ketone	0.40	U	ng/L
Endrin Ketone	0.40	U	ng/L
Esfenvalerate	0.75	U	ng/L
Esfenvalerate	0.75	U	ng/L
Ethalfuralin	0.15	U	ng/L
Ethalfuralin	0.15	U	ng/L
Fenpropathrin	0.25	U	ng/L
Fenpropathrin	0.25	U	ng/L
Gamma-BHC	0.12	U	ng/L
Gamma-BHC	0.12	U	ng/L
Gamma-Chlordane	0.20	U	ng/L
Gamma-Chlordane	0.20	U	ng/L
Heptachlor	0.20	U	ng/L
Heptachlor	0.20	U	ng/L
Heptachlor Epoxide	0.25	U	ng/L
Heptachlor Epoxide	0.25	U	ng/L
Hexachlorobenzene	1.0	U	ng/L
Hexachlorobenzene	1.0	U	ng/L
Kepone	5.0	U	ng/L
Kepone	5.0	U	ng/L
Lambda-Cyhalothrin	0.75	U	ng/L
Lambda-Cyhalothrin	0.75	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E

Batch ID: P402230

Component	Result	Code	Units
Methoxychlor	0.30	U	ng/L
Methoxychlor	0.30	U	ng/L
MGK-264	0.20	U	ng/L
MGK-264	0.20	U	ng/L
Mirex	0.70	U	ng/L
Mirex	0.70	U	ng/L
Permethrin	1.6	U	ng/L
Permethrin	1.6	U	ng/L
Phenothrin	0.90	U	ng/L
Phenothrin	0.90	U	ng/L
Piperonyl Butoxide	0.15	U	ng/L
Piperonyl Butoxide	0.15	U	ng/L
Prallethrin	2.0	U	ng/L
Prallethrin	2.0	U	ng/L
Tau-Fluvalinate	0.25	U	ng/L
Tau-Fluvalinate	0.25	U	ng/L
Tetramethrin	0.75	U	ng/L
Tetramethrin	0.75	U	ng/L
Trans-Nonachlor	0.20	U	ng/L
Trans-Nonachlor	0.20	U	ng/L
Triclosan Methyl	0.15	U	ng/L
Triclosan Methyl	0.15	U	ng/L
Trifluralin	0.20	U	ng/L
Trifluralin	0.20	U	ng/L

Reference Method: EPA 8270E

Batch ID: P402306

Component	Result	Code	Units
Acetochlor	0.25	U	ng/L
Acetochlor	0.25	U	ng/L
Alachlor	0.25	U	ng/L
Alachlor	0.25	U	ng/L
Ametryn	0.60	U	ng/L
Ametryn	0.60	U	ng/L
Atrazine	0.25	U	ng/L
Atrazine	0.25	U	ng/L
Atrazine Desethyl	1.5	U	ng/L
Atrazine Desethyl	1.5	U	ng/L
Azinphos Methyl	2.0	U	ng/L
Azinphos Methyl	2.0	U	ng/L
Azoxystrobin	0.50	U	ng/L
Azoxystrobin	0.50	U	ng/L
Bromacil	0.50	U	ng/L
Bromacil	0.50	U	ng/L
Butylate	0.40	U	ng/L
Butylate	0.40	U	ng/L
Caffeine	7.5	U	ng/L
Caffeine	7.5	U	ng/L
Carbophenothion	0.50	U	ng/L
Carbophenothion	0.50	U	ng/L
Carfentrazone Ethyl	0.10	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E

Batch ID: P402306

Component	Result	Code	Units
Carfentrazone Ethyl	0.10	U	ng/L
Chlorpyrifos Ethyl	0.25	U	ng/L
Chlorpyrifos Ethyl	0.25	U	ng/L
Chlorpyrifos Methyl	0.050	U	ng/L
Chlorpyrifos Methyl	0.050	U	ng/L
Cyanazine	3.2	U	ng/L
Cyanazine	3.2	U	ng/L
Demeton	1.5	U	ng/L
Demeton	1.5	U	ng/L
Diazinon	0.12	U	ng/L
Diazinon	0.12	U	ng/L
Difenoconazole	0.60	U	ng/L
Difenoconazole	0.60	U	ng/L
Dimethenamid	0.10	U	ng/L
Dimethenamid	0.10	U	ng/L
Disulfoton	0.50	U	ng/L
Disulfoton	0.50	U	ng/L
Dithiopyr	0.18	U	ng/L
Dithiopyr	0.18	U	ng/L
EPTC	1.2	U	ng/L
EPTC	1.2	U	ng/L
Ethion	0.15	U	ng/L
Ethion	0.15	U	ng/L
Ethoprop	0.10	U	ng/L
Ethoprop	0.10	U	ng/L
Fenamiphos	0.50	U	ng/L
Fenamiphos	0.50	U	ng/L
Fenbuconazole	0.12	U	ng/L
Fenbuconazole	0.12	U	ng/L
Fipronil	0.25	U	ng/L
Fipronil	0.25	U	ng/L
Fipronil Desulfinyl	0.12	U	ng/L
Fipronil Desulfinyl	0.12	U	ng/L
Fipronil Sulfide	0.15	U	ng/L
Fipronil Sulfide	0.15	U	ng/L
Fipronil Sulfone	0.15	U	ng/L
Fipronil Sulfone	0.15	U	ng/L
Fluazifop-P-butyl	0.10	U	ng/L
Fluazifop-P-butyl	0.10	U	ng/L
Fludioxonil	0.12	U	ng/L
Fludioxonil	0.12	U	ng/L
Flumioxazin	1.8	U	ng/L
Flumioxazin	1.8	U	ng/L
Flutolanil	0.10	U	ng/L
Flutolanil	0.10	U	ng/L
Fluxapyroxad	0.10	U	ng/L
Fluxapyroxad	0.10	U	ng/L
Fonofos	0.20	U	ng/L
Fonofos	0.20	U	ng/L
Hexazinone	0.50	U	ng/L
Hexazinone	0.50	U	ng/L
Imazalil	0.75	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E

Batch ID: P402306

Component	Result	Code	Units
Imazalil	0.75	U	ng/L
Iprodione	0.60	U	ng/L
Iprodione	0.60	U	ng/L
Malathion	0.35	U	ng/L
Malathion	0.35	U	ng/L
Metalaxyl	0.75	U	ng/L
Metalaxyl	0.75	U	ng/L
Metolachlor	0.35	U	ng/L
Metolachlor	0.35	U	ng/L
Metribuzin	3.2	U	ng/L
Metribuzin	3.2	U	ng/L
Mevinphos	1.0	U	ng/L
Mevinphos	1.0	U	ng/L
Molinate	0.30	U	ng/L
Molinate	0.30	U	ng/L
Myclobutanil	0.25	U	ng/L
Myclobutanil	0.25	U	ng/L
Naled	1.5	U	ng/L
Naled	1.5	U	ng/L
Norflurazon	0.50	U	ng/L
Norflurazon	0.50	U	ng/L
Oxadiazon	0.30	U	ng/L
Oxadiazon	0.30	U	ng/L
Oxyfluorfen	0.15	U	ng/L
Oxyfluorfen	0.15	U	ng/L
Parathion Ethyl	0.25	U	ng/L
Parathion Ethyl	0.25	U	ng/L
Parathion Methyl	0.55	U	ng/L
Parathion Methyl	0.55	U	ng/L
Pendimethalin	1.0	U	ng/L
Pendimethalin	1.0	U	ng/L
Phorate	0.52	U	ng/L
Phorate	0.52	U	ng/L
Prodiamine	0.18	U	ng/L
Prodiamine	0.18	U	ng/L
Prometon	0.90	U	ng/L
Prometon	0.90	U	ng/L
Prometryn	0.20	U	ng/L
Prometryn	0.20	U	ng/L
Pronamide	0.15	U	ng/L
Pronamide	0.15	U	ng/L
Propiconazole	0.50	U	ng/L
Propiconazole	0.50	U	ng/L
Pyriproxyfen	0.12	U	ng/L
Pyriproxyfen	0.12	U	ng/L
Simazine	0.50	U	ng/L
Simazine	0.50	U	ng/L
Sulfentrazone	0.50	U	ng/L
Sulfentrazone	0.50	U	ng/L
Tebuconazole	0.50	U	ng/L
Tebuconazole	0.50	U	ng/L
Terbufos	0.10	U	ng/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 8270E
Batch ID: P402306

Component	Result	Code	Units
Terbufos	0.10	U	ng/L
Terbutylazine	0.42	U	ng/L
Terbutylazine	0.42	U	ng/L

Reference Method: EPA 8276
Batch ID: P402230

Component	Result	Code	Units
Chlordane	5.0	U	ng/L
Chlordane	5.0	U	ng/L
Toxaphene	30	U	ng/L
Toxaphene	30	U	ng/L

Reference Method: EPA 8321B
Batch ID: P402170

Component	Result	Code	Units
3-Hydroxycarbofuran	0.0020	U	ug/L
Aldicarb	0.020	U	ug/L
Aldicarb Sulfone	0.0020	U	ug/L
Aldicarb Sulfoxide	0.0020	U	ug/L
Carbaryl	0.0020	U	ug/L
Carbofuran	0.0020	U	ug/L
Methiocarb	0.0020	U	ug/L
Methomyl	0.0020	U	ug/L
Oxamyl	0.0020	U	ug/L
Propoxur	0.0020	U	ug/L

Reference Method: EPA 8321B
Batch ID: P402297

Component	Result	Code	Units
Acesulfame K	0.10	U	ug/L
AMPA	0.10	U	ug/L
Endothall	0.25	U	ug/L
Glufosinate	0.10	U	ug/L
Glyphosate	0.10	U	ug/L
Sucralose	0.010	U	ug/L

Reference Method: EPA 8321B
Batch ID: P402303

Component	Result	Code	Units
2,4-D	0.0020	U	ug/L
2,4,5-T	0.0040	U	ug/L
Acetaminophen	0.0080	U	ug/L
Acetamiprid	0.0020	U	ug/L
Afidopyropen	0.0020	U	ug/L
Bentazon	8.0E-04	U	ug/L
Benzovindiflupyr	0.0020	U	ug/L
Carbamazepine	8.0E-04	U	ug/L
Clothianidin	0.0020	U	ug/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8321B

Batch ID: P402303

Component	Result	Code	Units
Dinotefuran	0.0020	U	ug/L
Diuron	0.0020	U	ug/L
Fenuron	0.016	U	ug/L
Fluridone	8.0E-04	U	ug/L
Hydrocodone	0.0020	U	ug/L
Ibuprofen	0.020	U	ug/L
Imazapyr	0.0040	U	ug/L
Imidacloprid	0.0020	U	ug/L
Linuron	0.0040	U	ug/L
Mandestrobin	0.0020	U	ug/L
MCPP	0.0020	U	ug/L
Naproxen	0.0080	U	ug/L
Primidone	0.0040	U	ug/L
Pyraclostrobin	0.0020	U	ug/L
Thiamethoxam	0.0020	U	ug/L
Tolfenpyrad	0.0020	U	ug/L
Triclopyr	0.0040	U	ug/L

Reference Method: SM 2120 B-2011

Batch ID: P402302

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-2011

Batch ID: P402593

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-2011

Batch ID: P402858

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 C-2011

Batch ID: P402859

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011

Batch ID: P402749

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 2540 D-2011

Batch ID: P402750

Component	Result	Code	Units
-----------	--------	------	-------

Quality Assurance Report

Method Blank Results

Reference Method: SM 2540 D-2011
Batch ID: P402750

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P402596

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P402919

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P402516

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Barium	110		P	85 - 115
Calcium	107		P	85 - 115
Iron	112		P	85 - 115
Magnesium	111		P	85 - 115
Manganese	105		P	85 - 115
Potassium	107		P	85 - 115
Sodium	104		P	85 - 115
Zinc	106		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P402516

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aluminum	103		P	85 - 115
Antimony	102		P	85 - 115
Arsenic	104		P	85 - 115
Beryllium	95.9		P	85 - 115
Boron	103		P	85 - 115
Cadmium	102		P	85 - 115
Chromium	102		P	85 - 115
Copper	105		P	85 - 115
Lead	106		P	85 - 115
Nickel	105		P	85 - 115
Selenium	105		P	85 - 115
Silver	111		P	85 - 115
Thallium	106		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402550

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402551

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	99.1		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402552

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402553

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	101		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P402459

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	98.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P402966

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	105		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P402439

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P402464

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	109		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P402339

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	102		P	90 - 110

Reference Method: EPA 8270E
Batch ID: P402230

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aldrin	69.1	70.0	P/P	20 - 150
Alpha-BHC	107	115	P/P	50 - 150
Alpha-Chlordane	109	115	P/P	50 - 150
Beta-BHC	123	142	P/P	50 - 150
Beta-Cyfluthrin	134	150	P/F	50 - 150
Bifenthrin	108	116	P/P	40 - 150
Chlorothalonil	119	140	P/P	20 - 150
Cis-Nonachlor	93.9	103	P/P	50 - 150
Cypermethrin	135	150	P/P	50 - 150
Dacthal	119	122	P/P	50 - 150
DDD-p,p'	101	109	P/P	50 - 150
DDE-p,p'	108	113	P/P	50 - 150
DDT-p,p'	109	116	P/P	50 - 150
Delta-BHC	120	139	P/P	50 - 150
Deltamethrin	124	136	P/P	50 - 150
Dichlobenil	81.5	90.0	P/P	40 - 150
Dicofol	98.3	106	P/P	50 - 150
Dieldrin	107	114	P/P	50 - 150
Endosulfan I	101	106	P/P	50 - 150
Endosulfan II	96.5	105	P/P	50 - 150
Endosulfan Sulfate	112	121	P/P	50 - 150
Endrin	89.9	94.8	P/P	50 - 150

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8270E
 Batch ID: P402230

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Endrin Aldehyde	102	102	P/P	50 - 150
Endrin Ketone	104	116	P/P	50 - 150
Esfenvalerate	129	143	P/P	50 - 180
Ethalfuralin	91.6	98.1	P/P	50 - 150
Fenpropathrin	102	113	P/P	50 - 150
Gamma-BHC	113	123	P/P	50 - 150
Gamma-Chlordane	103	109	P/P	50 - 150
Heptachlor	86.9	90.8	P/P	50 - 150
Heptachlor Epoxide	102	108	P/P	50 - 150
Hexachlorobenzene	89.0	93.8	P/P	50 - 150
Kepone	122	144	P/P	30 - 180
Lambda-Cyhalothrin	129	149	P/P	50 - 200
Methoxychlor	105	111	P/P	50 - 150
MGK-264	109	119	P/P	30 - 150
Mirex	83.8	89.7	P/P	50 - 180
Permethrin	131	144	P/P	50 - 150
Phenothrin	82.0	96.6	P/P	20 - 150
Piperonyl Butoxide	111	119	P/P	50 - 150
Prallethrin	100	101	P/P	30 - 150
Tau-Fluvalinate	108	123	P/P	40 - 150
Tetramethrin	91.2	105	P/P	30 - 150
Trans-Nonachlor	99.6	112	P/P	50 - 150
Triclosan Methyl	114	118	P/P	50 - 150
Trifluralin	91.3	101	P/P	50 - 150

Reference Method: EPA 8270E
 Batch ID: P402306

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Acetochlor	100	102	P/P	70 - 121
Alachlor	93.0	94.1	P/P	70 - 119
Ametryn	102	102	P/P	61 - 135
Atrazine	95.2	96.2	P/P	76 - 123
Atrazine Desethyl	89.2	93.8	P/P	35 - 140
Azinphos Methyl	131	129	P/P	57 - 163
Azoxystrobin	98.2	106	P/P	43 - 231
Bromacil	87.0	90.0	P/P	42 - 123
Butylate	42.9	44.3	P/P	34 - 92.0
Caffeine	70.3	71.4	P/P	70 - 130
Carbophenothion	105	103	P/P	61 - 121
Carfentrazone Ethyl	117	117	P/P	62 - 139
Chlorpyrifos Ethyl	88.8	87.5	P/P	67 - 121
Chlorpyrifos Methyl	96.3	93.9	P/P	67 - 120
Cyanazine	135	141	P/P	20 - 226
Demeton	48.5	46.2	P/P	35 - 125
Diazinon	97.6	95.9	P/P	70 - 122
Difenoconazole	183	184	P/P	56 - 186
Dimethenamid	79.7	80.8	P/P	67 - 119
Disulfoton	91.1	94.0	P/P	47 - 120
Dithiopyr	90.6	94.2	P/P	67 - 118
EPTC	44.9	45.4	P/P	33 - 90.0
Ethion	76.9	74.7	P/P	40 - 133

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8270E
Batch ID: P402306

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ethoprop	87.4	87.0	P/P	61 - 121
Fenamiphos	94.1	97.9	P/P	31 - 139
Fenbuconazole	115	119	P/P	66 - 141
Fipronil	106	107	P/P	62 - 129
Fipronil Desulfinyl	106	117	P/P	59 - 126
Fipronil Sulfide	87.8	90.6	P/P	63 - 117
Fipronil Sulfone	81.2	78.3	P/P	57 - 117
Fluazifop-P-butyl	105	104	P/P	63 - 124
Fludioxonil	121	121	P/P	63 - 123
Flumioxazin	189	191	P/P	34 - 245
Flutolanil	93.1	91.4	P/P	56 - 131
Fluxapyroxad	94.1	95.2	P/P	57 - 117
Fonofos	86.6	84.8	P/P	61 - 116
Hexazinone	96.1	100	P/P	55 - 138
Imazalil	93.7	96.7	P/P	33 - 143
Iprodione	109	102	P/P	59 - 118
Malathion	93.3	91.0	P/P	62 - 138
Metalaxyl	92.8	95.7	P/P	24 - 147
Metolachlor	89.6	90.4	P/P	68 - 118
Metribuzin	90.6	92.6	P/P	20 - 239
Mevinphos	75.7	84.4	P/P	46 - 124
Molinate	46.8	50.0	P/P	46 - 100
Myclobutanil	92.2	95.7	P/P	32 - 149
Naled	47.7	46.8	P/P	32 - 95.0
Norflurazon	110	112	P/P	57 - 127
Oxadiazon	82.0	83.9	P/P	63 - 118
Oxyfluorfen	102	97.1	P/P	69 - 137
Parathion Ethyl	87.9	93.1	P/P	70 - 113
Parathion Methyl	112	110	P/P	71 - 130
Pendimethalin	65.2	66.7	P/P	55 - 118
Phorate	74.0	67.8	P/P	46 - 125
Prodiamine	28.8	25.6	P/P	20 - 141
Prometon	87.6	91.4	P/P	54 - 122
Prometryn	95.4	95.7	P/P	66 - 124
Pronamide	93.4	91.5	P/P	79 - 122
Propiconazole	94.6	97.8	P/P	61 - 126
Pyriproxyfen	89.2	91.8	P/P	52 - 131
Simazine	98.1	103	P/P	75 - 128
Sulfentrazone	100	97.6	P/P	20 - 132
Tebuconazole	77.6	76.9	P/P	20 - 116
Terbufos	93.7	95.4	P/P	61 - 117
Terbuthylazine	83.0	89.0	P/P	74 - 116

Reference Method: EPA 8276
Batch ID: P402230

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chlordane	72.5	64.4	P/P	56 - 117
Toxaphene	60.4	57.5	P/P	45 - 116

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8321B

Batch ID: P402170

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
3-Hydroxycarbofuran	87.4		P	40 - 150
Aldicarb	62.8		P	30 - 150
Aldicarb Sulfone	90.8		P	40 - 150
Aldicarb Sulfoxide	107		P	40 - 150
Carbaryl	80.5		P	40 - 150
Carbofuran	81.6		P	40 - 150
Methiocarb	73.0		P	40 - 150
Methomyl	87.6		P	40 - 150
Oxamyl	87.6		P	40 - 150
Propoxur	82.2		P	40 - 150

Reference Method: EPA 8321B

Batch ID: P402297

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Acesulfame K	106		P	40 - 150
AMPA	88.6		P	40 - 150
Endothall	87.2		P	40 - 150
Glufosinate	92.7		P	40 - 150
Glyphosate	98.0		P	40 - 150
Sucralose	99.0		P	40 - 150

Reference Method: EPA 8321B

Batch ID: P402303

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2,4-D	135		P	40 - 150
2,4,5-T	111		P	40 - 150
Acetaminophen	100		P	30 - 150
Acetamiprid	93.7		P	40 - 150
Afidopyropen	75.0		P	40 - 150
Bentazon	113		P	30 - 150
Benzovindiflupyr	94.7		P	40 - 150
Carbamazepine	103		P	40 - 150
Clothianidin	107		P	40 - 150
Dinotefuran	104		P	40 - 150
Diuron	94.0		P	40 - 150
Fenuron	104		P	40 - 150
Fluridone	95.3		P	40 - 150
Hydrocodone	98.6		P	40 - 150
Ibuprofen	84.8		P	40 - 150
Imazapyr	108		P	40 - 150
Imidacloprid	100		P	40 - 150
Linuron	83.4		P	40 - 150
Mandestrobin	100		P	40 - 150
MCPP	117		P	40 - 150
Naproxen	89.9		P	40 - 150
Primidone	103		P	40 - 150
Pyraclostrobin	92.6		P	40 - 150
Thiamethoxam	105		P	40 - 150
Tolfenpyrad	60.2		P	30 - 150
Triclopyr	114		P	40 - 150

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 2120 B-2011
Batch ID: P402302

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	94.7		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P402593

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	94 - 106

Reference Method: SM 4500 F-C-2011
Batch ID: P402596

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.0		P	93 - 104

Reference Method: SM 5310 B-2011
Batch ID: P402919

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.5		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P402516

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2267978	Barium	108	112	P/P	80 - 120
2267978	Calcium	106		P	80 - 120
2267978	Iron	108	118	P/P	80 - 120
2267978	Magnesium	108	115	P/P	80 - 120
2267978	Manganese	103	109	P/P	80 - 120
2267978	Potassium	104	109	P/P	80 - 120
2267978	Sodium	103		P	80 - 120
2267978	Zinc	106	110	P/P	80 - 120
2268448	Barium	108		P	80 - 120
2268448	Calcium	108		P	80 - 120
2268448	Iron	110		P	80 - 120
2268448	Magnesium	109		P	80 - 120
2268448	Manganese	104		P	80 - 120
2268448	Potassium	111		P	80 - 120
2268448	Sodium	106		P	80 - 120
2268448	Zinc	103		P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4

Batch ID: P402516

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2267978	Aluminum	98.2	102	P/P	80 - 120
2267978	Antimony	97.2	99.4	P/P	80 - 120
2267978	Arsenic	98.0	100	P/P	80 - 120
2267978	Beryllium	98.4	100	P/P	80 - 120
2267978	Boron	99.9	104	P/P	80 - 120
2267978	Cadmium	96.5	100	P/P	80 - 120
2267978	Chromium	97.0	98.3	P/P	80 - 120
2267978	Copper	95.8	99.4	P/P	80 - 120
2267978	Lead	103	106	P/P	80 - 120
2267978	Nickel	97.5	98.8	P/P	80 - 120
2267978	Selenium	98.3	99.8	P/P	80 - 120
2267978	Silver	107	108	P/P	80 - 120
2267978	Thallium	104	106	P/P	80 - 120
2268448	Aluminum	99.8		P	80 - 120
2268448	Antimony	98.5		P	80 - 120
2268448	Arsenic	99.8		P	80 - 120
2268448	Beryllium	96.7		P	80 - 120
2268448	Boron	98.7		P	80 - 120
2268448	Cadmium	99.5		P	80 - 120
2268448	Chromium	99.5		P	80 - 120
2268448	Copper	101		P	80 - 120
2268448	Lead	105		P	80 - 120
2268448	Nickel	101		P	80 - 120
2268448	Selenium	99.8		P	80 - 120
2268448	Silver	108		P	80 - 120
2268448	Thallium	106		P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P402550

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
---------------	-----------	---------	---------	-----------	----------------

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402550

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268349	Chloride	96.3		P	90 - 110
2268421	Chloride	99.1		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402551

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268465	Chloride	101		P	90 - 110
2268944	Chloride	99.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402552

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268421	Sulfate	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402553

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268465	Sulfate	101		P	90 - 110
2268944	Sulfate	98.9		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P402459

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268424	Ammonia-N	99.2	99.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P402966

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268511	Kjeldahl Nitrogen	112	102	F/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P402439

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268424	NO2NO3-N	101	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P402464

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268624	Total-P	113	106	F/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P402339

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268430	O-Phosphate-P	99.8	100	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270E
Batch ID: P402230

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268682	Aldrin	68.5	61.7	P/P	20 - 170
2268682	Alpha-BHC	110	103	P/P	50 - 170
2268682	Alpha-Chlordane	120	107	P/P	50 - 170
2268682	Beta-BHC	153	159	P/P	50 - 170
2268682	Beta-Cyfluthrin	118	112	P/P	50 - 170
2268682	Bifenthrin	103	91.4	P/P	40 - 170
2268682	Chlorothalonil	63.5	56.9	P/P	20 - 170
2268682	Cis-Nonachlor	98.6	86.8	P/P	50 - 170
2268682	Cypermethrin	117	107	P/P	50 - 170
2268682	Dacthal	119	116	P/P	50 - 170
2268682	DDD-p,p'	100	92.8	P/P	50 - 170
2268682	DDE-p,p'	112	104	P/P	50 - 170
2268682	DDT-p,p'	112	104	P/P	50 - 170
2268682	Delta-BHC	158	160	P/P	50 - 170
2268682	Deltamethrin	127	131	P/P	50 - 170
2268682	Dichlobenil	64.2	79.1	P/P	40 - 170
2268682	Dicofol	99.3	91.6	P/P	50 - 170
2268682	Dieldrin	131	118	P/P	50 - 170
2268682	Endosulfan I	103	98.0	P/P	50 - 170
2268682	Endosulfan II	96.3	91.4	P/P	50 - 170
2268682	Endosulfan Sulfate	120	130	P/P	50 - 170
2268682	Endrin	91.6	84.6	P/P	50 - 170
2268682	Endrin Aldehyde	76.0	76.8	P/P	50 - 170
2268682	Endrin Ketone	111	110	P/P	50 - 170
2268682	Esfenvalerate	114	103	P/P	50 - 180
2268682	Ethalfuralin	99.8	88.1	P/P	50 - 170
2268682	Fenpropathrin	101	84.9	P/P	50 - 170
2268682	Gamma-BHC	131	127	P/P	50 - 170
2268682	Gamma-Chlordane	98.0	88.2	P/P	50 - 170
2268682	Heptachlor	79.4	71.0	P/P	50 - 170
2268682	Heptachlor Epoxide	184	124	F/P	50 - 170
2268682	Hexachlorobenzene	89.4	85.3	P/P	50 - 170
2268682	Kepone	140	237	P/F	30 - 180
2268682	Lambda-Cyhalothrin	112	107	P/P	50 - 200
2268682	Methoxychlor	101	94.8	P/P	50 - 170
2268682	MGK-264	113	112	P/P	30 - 170
2268682	Mirex	88.6	77.7	P/P	50 - 180
2268682	Permethrin	120	112	P/P	50 - 170
2268682	Phenothrin	99.2	86.1	P/P	20 - 170
2268682	Piperonyl Butoxide	113	103	P/P	50 - 170
2268682	Prallethrin	103	103	P/P	30 - 170
2268682	Tau-Fluvalinate	111	101	P/P	40 - 170
2268682	Tetramethrin	104	92.7	P/P	30 - 170
2268682	Trans-Nonachlor	102	95.3	P/P	50 - 170
2268682	Triclosan Methyl	116	111	P/P	50 - 170
2268682	Trifluralin	98.5	86.5	P/P	50 - 170

Reference Method: EPA 8270E
Batch ID: P402306

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268339	Acetochlor	92.2	102	P/P	66 - 122

Quality Assurance Report

Matrix Spike Accuracy

Reference Method: EPA 8270E

Batch ID: P402306

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268339	Alachlor	83.8	96.5	P/P	65 - 122
2268339	Ametryn	75.4	87.2	P/P	45 - 173
2268339	Atrazine Desethyl	54.8	64.8	P/P	25 - 150
2268339	Azinphos Methyl	120	142	P/P	44 - 174
2268339	Azoxystrobin	108	98.8	P/P	10 - 268
2268339	Bromacil	84.2	95.7	P/P	31 - 145
2268339	Butylate	45.5	38.6	P/P	15 - 97.0
2268339	Carbophenothion	91.0	101	P/P	42 - 129
2268339	Carfentrazone Ethyl	106	89.0	P/P	62 - 145
2268339	Chlorpyrifos Ethyl	74.7	82.5	P/P	60 - 124
2268339	Chlorpyrifos Methyl	90.1	101	P/P	66 - 119
2268339	Cyanazine	138	162	P/P	17 - 225
2268339	Demeton	113	115	P/P	36 - 142
2268339	Diazinon	96.4	115	P/P	70 - 128
2268339	Difenoconazole	177	215	P/P	33 - 222
2268339	Dimethenamid	69.9	83.0	P/P	54 - 125
2268339	Disulfoton	89.8	99.4	P/P	49 - 133
2268339	Dithiopyr	78.7	83.8	P/P	55 - 123
2268339	EPTC	47.7	36.9	P/P	19 - 91.0
2268339	Ethion	57.3	82.5	P/P	13 - 143
2268339	Ethoprop	110	113	P/P	49 - 144
2268339	Fenamiphos	87.5	76.4	P/P	32 - 136
2268339	Fenbuconazole	121	128	P/P	73 - 148
2268339	Fipronil	104	111	P/P	57 - 129
2268339	Fipronil Desulfinyl	97.4	116	P/P	49 - 127
2268339	Fipronil Sulfide	89.5	101	P/P	44 - 127
2268339	Fipronil Sulfone	71.6	87.0	P/P	35 - 126
2268339	Fluazifop-P-butyl	96.7	114	P/P	67 - 129
2268339	Fludioxonil	104	125	P/P	61 - 126
2268339	Flumioxazin	182	235	P/P	38 - 279
2268339	Flutolanil	74.7	93.1	P/P	55 - 136
2268339	Fluxapyroxad	80.6	95.4	P/P	33 - 140
2268339	Fonofos	91.3	105	P/P	58 - 125
2268339	Hexazinone	92.0	106	P/P	57 - 148
2268339	Imazalil	77.7	84.3	P/P	10 - 221
2268339	Iprodione	110	123	P/P	32 - 140
2268339	Malathion	88.2	99.8	P/P	52 - 138
2268339	Metalaxyl	84.0	96.8	P/P	46 - 134
2268339	Metolachlor	82.4	96.5	P/P	58 - 122
2268339	Metribuzin	102	105	P/P	16 - 272
2268339	Mevinphos	105	111	P/P	45 - 140
2268339	Molinate	57.7	56.2	P/P	41 - 103
2268339	Myclobutanil	82.5	98.4	P/P	60 - 134
2268339	Naled	49.1	53.8	P/P	20 - 113
2268339	Norflurazon	97.7	112	P/P	51 - 123
2268339	Oxadiazon	76.2	85.7	P/P	44 - 125
2268339	Oxyfluorfen	104	123	P/P	66 - 176
2268339	Parathion Ethyl	90.9	105	P/P	57 - 132
2268339	Parathion Methyl	117	125	P/P	59 - 156
2268339	Pendimethalin	81.9	87.8	P/P	59 - 129
2268339	Phorate	87.3	84.8	P/P	47 - 136
2268339	Prodiamine	47.3	52.3	P/P	10 - 159

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270E

Batch ID: P402306

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268339	Prometon	82.8	108	P/P	59 - 127
2268339	Prometryn	67.9	72.5	P/P	59 - 129
2268339	Pronamide	81.9	95.0	P/P	65 - 145
2268339	Propiconazole	82.7	86.5	P/P	28 - 165
2268339	Pyriproxyfen	69.7	83.7	P/P	21 - 180
2268339	Simazine	101	110	P/P	63 - 147
2268339	Sulfentrazone	85.0	102	P/P	32 - 136
2268339	Tebuconazole	50.0	73.7	P/P	10 - 117
2268339	Terbufos	109	119	P/P	61 - 131
2268339	Terbuthylazine	87.1	99.2	P/P	62 - 126

Reference Method: EPA 8276

Batch ID: P402230

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268338	Chlordane	62.9	63.3	P/P	47 - 128
2268338	Toxaphene	59.4	59.3	P/P	11 - 163

Reference Method: EPA 8321B

Batch ID: P402170

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2267464	3-Hydroxycarbofuran	72.2	90.9	P/P	40 - 150
2267464	Aldicarb	78.8	79.6	P/P	30 - 150
2267464	Aldicarb Sulfone	110	104	P/P	40 - 150
2267464	Aldicarb Sulfoxide	54.7	47.0	P/P	40 - 150
2267464	Carbaryl	76.8	76.3	P/P	40 - 150
2267464	Carbofuran	78.1	79.8	P/P	40 - 150
2267464	Methiocarb	74.9	73.7	P/P	40 - 150
2267464	Methomyl	85.7	81.0	P/P	40 - 150
2267464	Oxamyl	95.2	88.9	P/P	40 - 150
2267464	Propoxur	84.0	82.7	P/P	40 - 150

Reference Method: EPA 8321B

Batch ID: P402297

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268337	Acesulfame K	123	133	P/P	40 - 150
2268337	AMPA	76.0	77.6	P/P	40 - 150
2268337	Endothall	105	106	P/P	40 - 150
2268337	Glufosinate	80.0	77.7	P/P	40 - 150
2268337	Glyphosate	82.0	77.0	P/P	40 - 150
2268337	Sucralose	102	102	P/P	40 - 150

Reference Method: EPA 8321B

Batch ID: P402303

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268133	2,4-D	114	104	P/P	40 - 150
2268133	2,4,5-T	99.4	107	P/P	40 - 150
2268133	Acetaminophen	88.2	82.1	P/P	30 - 150
2268133	Acetamiprid	81.9	68.5	P/P	40 - 150

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8321B
Batch ID: P402303

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268133	Afidopyropen	63.4	55.2	P/P	40 - 150
2268133	Bentazon	91.2	94.7	P/P	30 - 150
2268133	Benzovindiflupyr	86.6	87.0	P/P	40 - 150
2268133	Carbamazepine	89.0	91.5	P/P	40 - 150
2268133	Clothianidin	98.6	93.2	P/P	40 - 150
2268133	Dinotefuran	94.9	91.2	P/P	40 - 150
2268133	Diuron	81.0	84.9	P/P	40 - 150
2268133	Fenuron	92.9	86.6	P/P	40 - 150
2268133	Fluridone	84.9	91.7	P/P	40 - 150
2268133	Hydrocodone	86.0	89.8	P/P	40 - 150
2268133	Ibuprofen	77.5	82.1	P/P	40 - 150
2268133	Imazapyr	104	99.5	P/P	40 - 150
2268133	Imidacloprid	99.4	97.8	P/P	40 - 150
2268133	Linuron	75.9	80.5	P/P	40 - 150
2268133	Mandestrobin	87.2	82.5	P/P	40 - 150
2268133	MCP	110	117	P/P	40 - 150
2268133	Naproxen	84.1	85.6	P/P	40 - 150
2268133	Primidone	89.5	93.5	P/P	40 - 150
2268133	Pyraclostrobin	83.6	82.3	P/P	40 - 150
2268133	Thiamethoxam	99.6	95.0	P/P	40 - 150
2268133	Tolfenpyrad	54.7	57.3	P/P	30 - 150
2268133	Triclopyr	101	106	P/P	40 - 150

Reference Method: SM 4500 F-C-2011
Batch ID: P402596

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268419	Fluoride	103	103	P/P	92 - 107

Reference Method: SM 5310 B-2011
Batch ID: P402919

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268467	Organic Carbon	98.6	96.6	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P402343

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268464	Turbidity	2.06	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P402516

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267978	Barium	4.27	Spike	P	0 - 20
2267978	Calcium	5.55	Spike	P	0 - 20
2267978	Iron	7.99	Spike	P	0 - 20
2267978	Magnesium	5.41	Spike	P	0 - 20
2267978	Manganese	5.14	Spike	P	0 - 20
2267978	Potassium	4.43	Spike	P	0 - 20
2267978	Sodium	5.49	Spike	P	0 - 20
2267978	Zinc	4.42	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4

Batch ID: P402516

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267978	Aluminum	2.64	Spike	P	0 - 20
2267978	Antimony	2.29	Spike	P	0 - 20
2267978	Arsenic	2.24	Spike	P	0 - 20
2267978	Beryllium	2.11	Spike	P	0 - 20
2267978	Boron	3.22	Spike	P	0 - 20
2267978	Cadmium	3.63	Spike	P	0 - 20
2267978	Chromium	1.33	Spike	P	0 - 20
2267978	Copper	3.63	Spike	P	0 - 20
2267978	Lead	3.11	Spike	P	0 - 20
2267978	Nickel	1.32	Spike	P	0 - 20
2267978	Selenium	1.50	Spike	P	0 - 20
2267978	Silver	1.22	Spike	P	0 - 20
2267978	Thallium	1.79	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P402550

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268349	Chloride	0.488	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P402551

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268944	Chloride	1.97	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402552

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268349	Sulfate	0.910	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P402553

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268944	Sulfate	1.98	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P402459

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268424	Ammonia-N	0.201	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P402966

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268511	Kjeldahl Nitrogen	6.39	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P402439

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268424	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P402464

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268624	Total-P	5.15	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P402339

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268430	O-Phosphate-P	0.152	Spike	P	0 - 20

Reference Method: EPA 8270E
Batch ID: P402230

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268682	Aldrin	10.6	Spike	P	0 - 30
2268682	Alpha-BHC	6.44	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8270E

Batch ID: P402230

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268682	Alpha-Chlordane	6.93	Spike	P	0 - 30
2268682	Beta-BHC	4.15	Spike	P	0 - 30
2268682	Beta-Cyfluthrin	5.67	Spike	P	0 - 30
2268682	Bifenthrin	12.1	Spike	P	0 - 30
2268682	Chlorothalonil	11.1	Spike	P	0 - 50
2268682	Cis-Nonachlor	12.7	Spike	P	0 - 30
2268682	Cypermethrin	9.43	Spike	P	0 - 30
2268682	Dacthal	2.21	Spike	P	0 - 30
2268682	DDD-p,p'	7.48	Spike	P	0 - 30
2268682	DDE-p,p'	8.06	Spike	P	0 - 30
2268682	DDT-p,p'	7.73	Spike	P	0 - 30
2268682	Delta-BHC	1.52	Spike	P	0 - 30
2268682	Deltamethrin	3.04	Spike	P	0 - 30
2268682	Dichlobenil	20.8	Spike	P	0 - 30
2268682	Dicofol	8.10	Spike	P	0 - 30
2268682	Dieldrin	4.48	Spike	P	0 - 30
2268682	Endosulfan I	5.05	Spike	P	0 - 30
2268682	Endosulfan II	5.25	Spike	P	0 - 30
2268682	Endosulfan Sulfate	8.50	Spike	P	0 - 30
2268682	Endrin	7.87	Spike	P	0 - 30
2268682	Endrin Aldehyde	1.10	Spike	P	0 - 30
2268682	Endrin Ketone	0.828	Spike	P	0 - 30
2268682	Esfenvalerate	10.3	Spike	P	0 - 30
2268682	Ethalfuralin	12.4	Spike	P	0 - 30
2268682	Fenpropathrin	17.0	Spike	P	0 - 30
2268682	Gamma-BHC	3.19	Spike	P	0 - 30
2268682	Gamma-Chlordane	7.97	Spike	P	0 - 30
2268682	Heptachlor	11.2	Spike	P	0 - 30
2268682	Heptachlor Epoxide	5.66	Spike	P	0 - 30
2268682	Hexachlorobenzene	4.64	Spike	P	0 - 30
2268682	Kepone	8.36	Spike	P	0 - 30
2268682	Lambda-Cyhalothrin	3.97	Spike	P	0 - 30
2268682	Methoxychlor	6.83	Spike	P	0 - 30
2268682	MGK-264	1.45	Spike	P	0 - 30
2268682	Mirex	13.1	Spike	P	0 - 30
2268682	Permethrin	6.80	Spike	P	0 - 30
2268682	Phenothrin	14.1	Spike	P	0 - 30
2268682	Piperonyl Butoxide	8.82	Spike	P	0 - 30
2268682	Prallethrin	0.243	Spike	P	0 - 30
2268682	Tau-Fluvalinate	9.83	Spike	P	0 - 30
2268682	Tetramethrin	11.4	Spike	P	0 - 30
2268682	Trans-Nonachlor	5.90	Spike	P	0 - 30
2268682	Triclosan Methyl	4.29	Spike	P	0 - 30
2268682	Trifluralin	13.0	Spike	P	0 - 30
LFB	Aldrin	1.38	LCS	P	0 - 30
LFB	Alpha-BHC	7.19	LCS	P	0 - 30
LFB	Alpha-Chlordane	5.51	LCS	P	0 - 30
LFB	Beta-BHC	14.1	LCS	P	0 - 30
LFB	Beta-Cyfluthrin	11.6	LCS	P	0 - 30
LFB	Bifenthrin	6.82	LCS	P	0 - 30
LFB	Chlorothalonil	16.4	LCS	P	0 - 50

Quality Assurance Report Precision

Reference Method: EPA 8270E
Batch ID: P402230

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
LFB	Cis-Nonachlor	9.71	LCS	P	0 - 30
LFB	Cypermethrin	10.2	LCS	P	0 - 30
LFB	Dacthal	2.19	LCS	P	0 - 30
LFB	DDD-p,p'	7.79	LCS	P	0 - 30
LFB	DDE-p,p'	4.79	LCS	P	0 - 30
LFB	DDT-p,p'	6.25	LCS	P	0 - 30
LFB	Delta-BHC	14.4	LCS	P	0 - 30
LFB	Deltamethrin	8.75	LCS	P	0 - 30
LFB	Dichlobenil	9.93	LCS	P	0 - 30
LFB	Dicofol	7.55	LCS	P	0 - 30
LFB	Dieldrin	6.71	LCS	P	0 - 30
LFB	Endosulfan I	5.14	LCS	P	0 - 30
LFB	Endosulfan II	8.36	LCS	P	0 - 30
LFB	Endosulfan Sulfate	7.75	LCS	P	0 - 30
LFB	Endrin	5.38	LCS	P	0 - 30
LFB	Endrin Aldehyde	0.546	LCS	P	0 - 30
LFB	Endrin Ketone	10.7	LCS	P	0 - 30
LFB	Esfenvalerate	10.7	LCS	P	0 - 30
LFB	Ethalfuralin	6.85	LCS	P	0 - 30
LFB	Fenpropathrin	9.87	LCS	P	0 - 30
LFB	Gamma-BHC	8.52	LCS	P	0 - 30
LFB	Gamma-Chlordane	5.28	LCS	P	0 - 30
LFB	Heptachlor	4.37	LCS	P	0 - 30
LFB	Heptachlor Epoxide	5.38	LCS	P	0 - 30
LFB	Hexachlorobenzene	5.28	LCS	P	0 - 30
LFB	Kepone	16.0	LCS	P	0 - 30
LFB	Lambda-Cyhalothrin	14.6	LCS	P	0 - 30
LFB	Methoxychlor	5.65	LCS	P	0 - 30
LFB	MGK-264	9.30	LCS	P	0 - 30
LFB	Mirex	6.86	LCS	P	0 - 30
LFB	Permethrin	9.21	LCS	P	0 - 30
LFB	Phenothrin	16.4	LCS	P	0 - 30
LFB	Piperonyl Butoxide	6.36	LCS	P	0 - 30
LFB	Prallethrin	0.444	LCS	P	0 - 30
LFB	Tau-Fluvalinate	12.3	LCS	P	0 - 30
LFB	Tetramethrin	14.3	LCS	P	0 - 30
LFB	Trans-Nonachlor	11.9	LCS	P	0 - 30
LFB	Triclosan Methyl	3.81	LCS	P	0 - 30
LFB	Trifluralin	9.91	LCS	P	0 - 30

Reference Method: EPA 8270E
Batch ID: P402306

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268339	Acetochlor	9.98	Spike	P	0 - 30
2268339	Alachlor	14.0	Spike	P	0 - 30
2268339	Ametryn	10.6	Spike	P	0 - 30
2268339	Atrazine	7.25	Spike	P	0 - 30
2268339	Atrazine Desethyl	16.8	Spike	P	0 - 30
2268339	Azinphos Methyl	16.8	Spike	P	0 - 30

Quality Assurance Report

Precision

Reference Method: EPA 8270E

Batch ID: P402306

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268339	Azoxystrobin	7.70	Spike	P	0 - 30
2268339	Bromacil	11.9	Spike	P	0 - 30
2268339	Butylate	16.2	Spike	P	0 - 30
2268339	Caffeine	0.0433	Spike	P	0 - 30
2268339	Carbophenothion	10.3	Spike	P	0 - 30
2268339	Carfentrazone Ethyl	17.9	Spike	P	0 - 30
2268339	Chlorpyrifos Ethyl	9.91	Spike	P	0 - 30
2268339	Chlorpyrifos Methyl	11.6	Spike	P	0 - 30
2268339	Cyanazine	15.5	Spike	P	0 - 30
2268339	Demeton	2.03	Spike	P	0 - 30
2268339	Diazinon	17.7	Spike	P	0 - 30
2268339	Difenoconazole	19.7	Spike	P	0 - 30
2268339	Dimethenamid	17.2	Spike	P	0 - 30
2268339	Disulfoton	10.1	Spike	P	0 - 30
2268339	Dithiopyr	6.34	Spike	P	0 - 30
2268339	EPTC	25.6	Spike	P	0 - 30
2268339	Ethion	36.1	Spike	F	0 - 30
2268339	Ethoprop	3.09	Spike	P	0 - 30
2268339	Fenamiphos	13.5	Spike	P	0 - 30
2268339	Fenbuconazole	6.22	Spike	P	0 - 30
2268339	Fipronil	6.86	Spike	P	0 - 30
2268339	Fipronil Desulfinyf	16.2	Spike	P	0 - 30
2268339	Fipronil Sulfide	10.2	Spike	P	0 - 30
2268339	Fipronil Sulfone	14.8	Spike	P	0 - 30
2268339	Fluazifop-P-butyl	16.7	Spike	P	0 - 30
2268339	Fludioxonil	18.5	Spike	P	0 - 30
2268339	Flumioxazin	25.0	Spike	P	0 - 30
2268339	Flutolanil	21.9	Spike	P	0 - 30
2268339	Fluxapyroxad	16.8	Spike	P	0 - 30
2268339	Fonofos	13.7	Spike	P	0 - 30
2268339	Hexazinone	8.66	Spike	P	0 - 30
2268339	Imazalil	6.59	Spike	P	0 - 30
2268339	Iprodione	11.7	Spike	P	0 - 30
2268339	Malathion	12.4	Spike	P	0 - 30
2268339	Metalaxyl	14.1	Spike	P	0 - 30
2268339	Metolachlor	14.4	Spike	P	0 - 30
2268339	Metribuzin	3.35	Spike	P	0 - 30
2268339	Mevinphos	5.06	Spike	P	0 - 30
2268339	Molinate	2.53	Spike	P	0 - 30
2268339	Myclobutanil	15.9	Spike	P	0 - 30
2268339	Naled	9.03	Spike	P	0 - 30
2268339	Norflurazon	13.3	Spike	P	0 - 30
2268339	Oxadiazon	11.7	Spike	P	0 - 30
2268339	Oxyfluorfen	16.4	Spike	P	0 - 30
2268339	Parathion Ethyl	14.4	Spike	P	0 - 30
2268339	Parathion Methyl	6.41	Spike	P	0 - 30
2268339	Pendimethalin	6.97	Spike	P	0 - 30
2268339	Phorate	2.96	Spike	P	0 - 30
2268339	Prodiamine	10.1	Spike	P	0 - 30
2268339	Prometon	26.5	Spike	P	0 - 30
2268339	Prometryn	6.51	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8270E
Batch ID: P402306

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268339	Pronamide	14.9	Spike	P	0 - 30
2268339	Propiconazole	2.96	Spike	P	0 - 30
2268339	Pyriproxyfen	18.2	Spike	P	0 - 30
2268339	Simazine	9.23	Spike	P	0 - 30
2268339	Sulfentrazone	14.0	Spike	P	0 - 30
2268339	Tebuconazole	28.7	Spike	P	0 - 30
2268339	Terbufos	8.60	Spike	P	0 - 30
2268339	Terbutylazine	13.0	Spike	P	0 - 30
LFB	Acetochlor	1.12	LCS	P	0 - 30
LFB	Alachlor	1.19	LCS	P	0 - 30
LFB	Ametryn	0.471	LCS	P	0 - 30
LFB	Atrazine	1.03	LCS	P	0 - 30
LFB	Atrazine Desethyl	4.98	LCS	P	0 - 30
LFB	Azinphos Methyl	1.84	LCS	P	0 - 30
LFB	Azoxystrobin	7.60	LCS	P	0 - 30
LFB	Bromacil	3.45	LCS	P	0 - 30
LFB	Butylate	3.22	LCS	P	0 - 30
LFB	Caffeine	1.57	LCS	P	0 - 30
LFB	Carbophenothion	1.86	LCS	P	0 - 30
LFB	Carfentrazone Ethyl	0.223	LCS	P	0 - 30
LFB	Chlorpyrifos Ethyl	1.54	LCS	P	0 - 30
LFB	Chlorpyrifos Methyl	2.43	LCS	P	0 - 30
LFB	Cyanazine	4.14	LCS	P	0 - 30
LFB	Demeton	4.83	LCS	P	0 - 30
LFB	Diazinon	1.84	LCS	P	0 - 30
LFB	Difenoconazole	0.655	LCS	P	0 - 30
LFB	Dimethenamid	1.46	LCS	P	0 - 30
LFB	Disulfoton	3.09	LCS	P	0 - 30
LFB	Dithiopyr	3.92	LCS	P	0 - 30
LFB	EPTC	0.967	LCS	P	0 - 30
LFB	Ethion	2.82	LCS	P	0 - 30
LFB	Ethoprop	0.466	LCS	P	0 - 30
LFB	Fenamiphos	3.96	LCS	P	0 - 30
LFB	Fenbuconazole	3.17	LCS	P	0 - 30
LFB	Fipronil	0.397	LCS	P	0 - 30
LFB	Fipronil Desulfinyl	9.79	LCS	P	0 - 30
LFB	Fipronil Sulfide	3.10	LCS	P	0 - 30
LFB	Fipronil Sulfone	3.59	LCS	P	0 - 30
LFB	Fluazifop-P-butyl	1.44	LCS	P	0 - 30
LFB	Fludioxonil	0.268	LCS	P	0 - 30
LFB	Flumioxazin	0.990	LCS	P	0 - 30
LFB	Flutolanil	1.88	LCS	P	0 - 30
LFB	Fluxapyroxad	1.08	LCS	P	0 - 30
LFB	Fonofos	2.12	LCS	P	0 - 30
LFB	Hexazinone	4.10	LCS	P	0 - 30
LFB	Imazalil	3.09	LCS	P	0 - 30
LFB	Iprodione	6.70	LCS	P	0 - 30
LFB	Malathion	2.52	LCS	P	0 - 30
LFB	Metalaxyl	3.05	LCS	P	0 - 30
LFB	Metolachlor	0.882	LCS	P	0 - 30
LFB	Metribuzin	2.20	LCS	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8270E
Batch ID: P402306

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
LFB	Mevinphos	10.8	LCS	P	0 - 30
LFB	Molinate	6.59	LCS	P	0 - 30
LFB	Myclobutanil	3.69	LCS	P	0 - 30
LFB	Naled	1.89	LCS	P	0 - 30
LFB	Norflurazon	1.47	LCS	P	0 - 30
LFB	Oxadiazon	2.36	LCS	P	0 - 30
LFB	Oxyfluorfen	4.78	LCS	P	0 - 30
LFB	Parathion Ethyl	5.83	LCS	P	0 - 30
LFB	Parathion Methyl	2.11	LCS	P	0 - 30
LFB	Pendimethalin	2.38	LCS	P	0 - 30
LFB	Phorate	8.68	LCS	P	0 - 30
LFB	Prodiamine	11.9	LCS	P	0 - 30
LFB	Prometon	4.21	LCS	P	0 - 30
LFB	Prometryn	0.268	LCS	P	0 - 30
LFB	Pronamide	2.08	LCS	P	0 - 30
LFB	Propiconazole	3.33	LCS	P	0 - 30
LFB	Pyriproxyfen	2.91	LCS	P	0 - 30
LFB	Simazine	4.94	LCS	P	0 - 30
LFB	Sulfentrazone	2.89	LCS	P	0 - 30
LFB	Tebuconazole	0.883	LCS	P	0 - 30
LFB	Terbufos	1.84	LCS	P	0 - 30
LFB	Terbutylazine	7.07	LCS	P	0 - 30

Reference Method: EPA 8276
Batch ID: P402230

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268338	Chlordane	0.577	Spike	P	0 - 30
2268338	Toxaphene	0.155	Spike	P	0 - 30
LFB	Chlordane	11.8	LCS	P	0 - 30
LFB	Toxaphene	4.94	LCS	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P402170

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2267464	3-Hydroxycarbofuran	23.0	Spike	P	0 - 30
2267464	Aldicarb	0.979	Spike	P	0 - 30
2267464	Aldicarb Sulfone	5.32	Spike	P	0 - 30
2267464	Aldicarb Sulfoxide	15.0	Spike	P	0 - 30
2267464	Carbaryl	0.624	Spike	P	0 - 30
2267464	Carbofuran	2.11	Spike	P	0 - 30
2267464	Methiocarb	1.55	Spike	P	0 - 30
2267464	Methomyl	5.72	Spike	P	0 - 30
2267464	Oxamyl	6.86	Spike	P	0 - 30
2267464	Propoxur	1.53	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P402297

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268337	Acesulfame K	7.71	Spike	P	0 - 30
2268337	AMPA	1.94	Spike	P	0 - 30
2268337	Endothall	1.56	Spike	P	0 - 30
2268337	Glufosinate	2.85	Spike	P	0 - 30
2268337	Glyphosate	5.95	Spike	P	0 - 30
2268337	Sucralose	0.147	Spike	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P402303

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268133	2,4-D	8.89	Spike	P	0 - 30
2268133	2,4,5-T	7.50	Spike	P	0 - 30
2268133	Acetaminophen	7.25	Spike	P	0 - 30
2268133	Acetamiprid	17.8	Spike	P	0 - 30
2268133	Afidopyropen	13.9	Spike	P	0 - 30
2268133	Bentazon	3.76	Spike	P	0 - 30
2268133	Benzovindiflupyr	0.421	Spike	P	0 - 30
2268133	Carbamazepine	2.79	Spike	P	0 - 30
2268133	Clothianidin	5.57	Spike	P	0 - 30
2268133	Dinotefuran	4.03	Spike	P	0 - 30
2268133	Diuron	4.75	Spike	P	0 - 30
2268133	Fenuron	7.02	Spike	P	0 - 30
2268133	Fluridone	7.75	Spike	P	0 - 30
2268133	Hydrocodone	4.39	Spike	P	0 - 30
2268133	Ibuprofen	5.73	Spike	P	0 - 30
2268133	Imazapyr	4.86	Spike	P	0 - 30
2268133	Imidacloprid	1.64	Spike	P	0 - 30
2268133	Linuron	5.91	Spike	P	0 - 30
2268133	Mandestrobin	5.57	Spike	P	0 - 30
2268133	MCPP	6.04	Spike	P	0 - 30
2268133	Naproxen	1.76	Spike	P	0 - 30
2268133	Primidone	4.37	Spike	P	0 - 30
2268133	Pyraclostrobin	1.48	Spike	P	0 - 30
2268133	Thiamethoxam	4.70	Spike	P	0 - 30
2268133	Tolfenpyrad	4.62	Spike	P	0 - 30
2268133	Triclopyr	4.93	Spike	P	0 - 30

Reference Method: SM 2120 B-2011
Batch ID: P402302

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268465	Color (true)	0.684	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P402593

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
--------------------------	-----------	-----------	-------------------	-----------	----------------

Quality Assurance Report Precision

Reference Method: SM 2320 B-2011
Batch ID: P402593

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268419	Total Alkalinity	0.923	Sample	P	0 - 5.2

Reference Method: SM 2540 C-2011
Batch ID: P402858

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268423	TDS	6.32	Sample	P	0 - 10

Reference Method: SM 2540 C-2011
Batch ID: P402859

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268626	TDS	2.87	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P402596

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268419	Fluoride	0.457	Spike	P	0 - 2.1

Reference Method: SM 5310 B-2011
Batch ID: P402919

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268467	Organic Carbon	1.46	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Surrogates

Lab Sample ID: 2268477
Field Sample ID: G4NW0192

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Carbaryl-d7	74.9	P	30 - 160

Lab Sample ID: 2268478
Field Sample ID: G4NW0192

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	2,4-D-d3	120	P	30 - 160
EPA 8321B	Acetaminophen-d4	94.8	P	30 - 160
EPA 8321B	Carbamazepine-d10	100	P	30 - 160
EPA 8321B	Diuron-d6	84.9	P	30 - 160
EPA 8321B	Endothall-d6	95.6	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	69.4	P	30 - 160
EPA 8321B	Ibuprofen-d3	84.6	P	30 - 160
EPA 8321B	Primidone-d5	99.6	P	30 - 160
EPA 8321B	Sucralose-d6	100	P	30 - 160

Lab Sample ID: 2268479
Field Sample ID: G4NW0192

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270E	Decachlorobiphenyl	38.5	P	35 - 120
EPA 8270E	Tetrachloro-m-xylene	37.9	P	30 - 110
EPA 8276	Decachlorobiphenyl	36.9	P	29 - 92.0
EPA 8276	PCNB	47.8	P	43 - 134

Lab Sample ID: 2268480
Field Sample ID: G4NW0192

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270E	Caffeine-d9	86.3	P	20 - 200
EPA 8270E	Simazine-d10	127	P	73 - 135
EPA 8270E	Terbufos-d10	105	P	58 - 130

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B-2011

Run ID: A107174

Included Lab Sample IDs: 2268464, 2268465

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	97.9	96.6	P/P	85 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A107186

Included Lab Sample IDs: 2268468, 2268469

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.6	99.0	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A107187

Included Lab Sample IDs: 2268464, 2268465

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	98.5	99.0	P/P	95 - 105

Reference Method: EPA 8321B

Run ID: A107195

Included Lab Sample IDs: 2268477

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
3-Hydroxycarbofuran	104	104	P/P	60 - 150
Aldicarb	104	103	P/P	60 - 150
Aldicarb Sulfone	103	103	P/P	50 - 150
Aldicarb Sulfoxide	102	105	P/P	50 - 150
Carbaryl	105	106	P/P	60 - 150
Carbofuran	101	104	P/P	50 - 150
Methiocarb	103	97.8	P/P	50 - 150
Methomyl	104	106	P/P	50 - 150
Oxamyl	104	102	P/P	50 - 150
Propoxur	106	105	P/P	50 - 150

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A107236

Included Lab Sample IDs: 2268466, 2268467

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A107239

Included Lab Sample IDs: 2268478

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
AMPA	89.0	87.5	P/P	60 - 160
Glufosinate	88.1	91.9	P/P	60 - 160
Glyphosate	97.3	104	P/P	60 - 160

Quality Assurance Report Calibration Verification

Reference Method: EPA 8321B
Run ID: A107242
Included Lab Sample IDs: 2268478

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acesulfame K	123	129	P/P	60 - 160
Endothall	98.3	83.5	P/P	60 - 160
Sucralose	108	96.9	P/P	60 - 160

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A107243
Included Lab Sample IDs: 2268466, 2268467

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.0	99.4	P/P	90 - 110

Reference Method: EPA 8321B
Run ID: A107269
Included Lab Sample IDs: 2268478

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
2,4-D	96.1	97.8	P/P	50 - 150
2,4,5-T	96.9	91.6	P/P	50 - 150
Acetaminophen	97.3	90.4	P/P	60 - 160
Acetamiprid	98.5	96.1	P/P	50 - 150
Afidopyropen	92.7	85.6	P/P	50 - 150
Bentazon	109	106	P/P	50 - 160
Benzovindiflupyr	95.5	96.3	P/P	50 - 150
Carbamazepine	101	102	P/P	60 - 150
Clothianidin	101	93.3	P/P	60 - 150
Dinotefuran	98.2	95.1	P/P	50 - 150
Diuron	109	106	P/P	60 - 160
Fenuron	96.8	92.3	P/P	60 - 150
Fluridone	103	92.0	P/P	60 - 160
Hydrocodone	106	103	P/P	60 - 150
Ibuprofen	91.7	91.4	P/P	50 - 160
Imazapyr	94.3	92.3	P/P	50 - 150
Imidacloprid	94.6	93.5	P/P	60 - 150
Linuron	95.4	91.4	P/P	60 - 150
Mandestrobin	95.0	94.6	P/P	50 - 150
MCPP	99.8	92.4	P/P	50 - 150
Naproxen	85.0	92.0	P/P	50 - 160
Primidone	104	99.3	P/P	60 - 150
Pyraclostrobin	96.5	97.0	P/P	60 - 150
Thiamethoxam	102	99.0	P/P	50 - 150
Tolfenpyrad	93.3	93.8	P/P	60 - 150
Triclopyr	94.6	91.5	P/P	50 - 150

Reference Method: EPA 8270E
Run ID: A107278
Included Lab Sample IDs: 2268479

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aldrin	92.6		P	80 - 120
Alpha-BHC	96.4		P	80 - 120
Alpha-Chlordane	98.7		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270E
Run ID: A107278
Included Lab Sample IDs: 2268479

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Beta-BHC	98.7		P	80 - 120
Beta-Cyfluthrin	93.1		P	80 - 120
Bifenthrin	98.4		P	80 - 120
Chlorothalonil	92.5		P	80 - 120
Cis-Nonachlor	99.2		P	80 - 120
Cypermethrin	101		P	80 - 120
Dacthal	96.0		P	80 - 120
DDD-p,p'	98.6		P	80 - 120
DDE-p,p'	94.2		P	80 - 120
DDT-p,p'	91.0		P	80 - 120
Delta-BHC	96.2		P	80 - 120
Deltamethrin	90.1		P	80 - 120
Dichlobenil	98.6		P	80 - 120
Dicofol	101		P	80 - 120
Dieldrin	97.3		P	80 - 120
Endosulfan I	92.1		P	80 - 120
Endosulfan II	98.6		P	80 - 120
Endosulfan Sulfate	105		P	80 - 120
Endrin	96.1		P	80 - 120
Endrin Aldehyde	101		P	80 - 120
Endrin Ketone	101		P	80 - 120
Esfenvalerate	93.5		P	80 - 120
Ethalfuralin	98.0		P	80 - 120
Fenpropathrin	99.8		P	80 - 120
Gamma-BHC	92.0		P	80 - 120
Gamma-Chlordane	96.9		P	80 - 120
Heptachlor	96.0		P	80 - 120
Heptachlor Epoxide	95.6		P	80 - 120
Hexachlorobenzene	102		P	80 - 120
Kepone	88.7		P	80 - 120
Lambda-Cyhalothrin	93.4		P	80 - 120
Methoxychlor	101		P	80 - 120
MGK-264	101		P	80 - 120
Mirex	95.2		P	80 - 120
Permethrin	97.4		P	80 - 120
Phenothrin	97.8		P	80 - 120
Piperonyl Butoxide	97.8		P	80 - 120
Prallethrin	92.7		P	80 - 120
Tau-Fluvalinate	96.3		P	80 - 120
Tetramethrin	99.9		P	80 - 120
Trans-Nonachlor	96.9		P	80 - 120
Triclosan Methyl	94.8		P	80 - 120
Trifluralin	98.9		P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A107288
Included Lab Sample IDs: 2268464, 2268465

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	100	99.5	P/P	90 - 110
Chloride	99.9	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A107288

Included Lab Sample IDs: 2268464, 2268465

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sulfate	100	100	P/P	90 - 110
Sulfate	100	99.6	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A107300

Included Lab Sample IDs: 2268464, 2268465

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	99.5	102	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A107300

Included Lab Sample IDs: 2268464, 2268465

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	102	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A107304

Included Lab Sample IDs: 2268481, 2268482

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Barium	102	101	P/P	90 - 110
Calcium	102	104	P/P	90 - 110
Iron	101	104	P/P	90 - 110
Magnesium	100	102	P/P	90 - 110
Manganese	102	102	P/P	90 - 110
Potassium	101	102	P/P	90 - 110
Sodium	99.6	101	P/P	90 - 110
Zinc	97.9	95.1	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A107317

Included Lab Sample IDs: 2268466, 2268467

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	101	P/P	90 - 110

Reference Method: EPA 8276

Run ID: A107328

Included Lab Sample IDs: 2268479

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chlordane	94.2		P	80 - 120
Toxaphene	61.2*		F	80 - 120

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A107337

Included Lab Sample IDs: 2268481, 2268482

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
-----------	---------	---------	------------	----------------

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A107337

Included Lab Sample IDs: 2268481, 2268482

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aluminum	99.4	99.8	P/P	90 - 110
Antimony	98.5	98.5	P/P	90 - 110
Arsenic	97.9	98.4	P/P	90 - 110
Beryllium	94.9	94.9	P/P	90 - 110
Boron	98.0	99.3	P/P	90 - 110
Cadmium	98.8	98.7	P/P	90 - 110
Chromium	96.9	97.9	P/P	90 - 110
Copper	94.8	95.2	P/P	90 - 110
Lead	102	102	P/P	90 - 110
Nickel	96.7	97.4	P/P	90 - 110
Selenium	96.6	98.0	P/P	90 - 110
Silver	102	103	P/P	90 - 110
Thallium	102	102	P/P	90 - 110

Reference Method: EPA 8270E

Run ID: A107387

Included Lab Sample IDs: 2268480

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acetochlor	103		P	80 - 120
Alachlor	101		P	80 - 120
Ametryn	94.5		P	80 - 120
Atrazine	104		P	80 - 120
Atrazine Desethyl	105		P	80 - 120
Azinphos Methyl	87.2		P	80 - 120
Azoxystrobin	92.5		P	80 - 120
Bromacil	100		P	80 - 120
Butylate	99.9		P	80 - 120
Caffeine	96.0		P	80 - 120
Carbophenothion	100		P	80 - 120
Carfentrazone Ethyl	101		P	80 - 120
Chlorpyrifos Ethyl	103		P	80 - 120
Chlorpyrifos Methyl	108		P	80 - 120
Cyanazine	102		P	80 - 120
Demeton	49.6*		F	80 - 120
Diazinon	103		P	80 - 120
Difenoconazole	89.5		P	80 - 120
Dimethenamid	103		P	80 - 120
Disulfoton	98.2		P	80 - 120
Dithiopyr	103		P	80 - 120
EPTC	100		P	80 - 120
Ethion	97.1		P	80 - 120
Ethoprop	99.6		P	80 - 120
Fenamiphos	95.7		P	80 - 120
Fenbuconazole	93.0		P	80 - 120
Fipronil	101		P	80 - 120
Fipronil Desulfinyl	110		P	80 - 120
Fipronil Sulfide	103		P	80 - 120
Fipronil Sulfone	106		P	80 - 120
Fluazifop-P-butyl	113		P	80 - 120
Fludioxonil	101		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270E
Run ID: A107387
Included Lab Sample IDs: 2268480

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Flumioxazin	92.9		P	80 - 120
Flutolanil	102		P	80 - 120
Fluxapyroxad	102		P	80 - 120
Fonofos	100		P	80 - 120
Hexazinone	110		P	80 - 120
Imazalil	92.6		P	80 - 120
Iprodione	107		P	80 - 120
Malathion	95.0		P	80 - 120
Metaxyl	106		P	80 - 120
Metolachlor	105		P	80 - 120
Metribuzin	102		P	80 - 120
Mevinphos	101		P	80 - 120
Molinate	91.1		P	80 - 120
Myclobutanil	97.8		P	80 - 120
Naled	103		P	80 - 120
Norflurazon	90.2		P	80 - 120
Oxadiazon	104		P	80 - 120
Oxyfluorfen	92.3		P	80 - 120
Parathion Ethyl	98.6		P	80 - 120
Parathion Methyl	100		P	80 - 120
Pendimethalin	98.2		P	80 - 120
Phorate	100		P	80 - 120
Prodiamine	103		P	80 - 120
Prometon	99.4		P	80 - 120
Prometryn	98.6		P	80 - 120
Pronamide	94.5		P	80 - 120
Propiconazole	101		P	80 - 120
Pyriproxyfen	100		P	80 - 120
Simazine	108		P	80 - 120
Sulfentrazone	104		P	80 - 120
Tebuconazole	101		P	80 - 120
Terbufos	103		P	80 - 120
Terbutylazine	103		P	80 - 120

Reference Method: SM 5310 B-2011
Run ID: A107443
Included Lab Sample IDs: 2268466, 2268467

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.0	96.0	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A107494
Included Lab Sample IDs: 2268466, 2268467

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	98.1	104	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery			LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity							2.06	
EPA 200.7 Rev. 4.4	Barium	110		108	112	108			4.27
	Calcium	107		106	108				5.55
	Iron	112		108	118	110			7.99
	Magnesium	111		108	115	109			5.41
	Manganese	105		103	109	104			5.14
	Potassium	107		104	109	111			4.43
	Sodium	104		103	106				5.49
	Zinc	106		106	110	103			4.42
EPA 200.8 Rev. 5.4	Aluminum	103		98.2	102	99.8			2.64
	Antimony	102		97.2	99.4	98.5			2.29
	Arsenic	104		98.0	100	99.8			2.24
	Beryllium	95.9		98.4	100	96.7			2.11
	Boron	103		99.9	104	98.7			3.22
	Cadmium	102		96.5	100	99.5			3.63
	Chromium	102		97.0	98.3	99.5			1.33
	Copper	105		95.8	99.4	101			3.63
	Lead	106		103	106	105			3.11
	Nickel	105		97.5	98.8	101			1.32
	Selenium	105		98.3	99.8	99.8			1.50
	Silver	111		107	108	108			1.22
	Thallium	106		104	106	106			1.79
EPA 300.0 Rev. 2.1	Chloride	100		96.3	99.1			0.488	
	Chloride	99.1		99.8	101			1.97	
	Sulfate	101		101				0.910	
	Sulfate	101		98.9	101			1.98	
EPA 350.1 Rev. 2.0	Ammonia-N	98.8		99.2	99.4				0.201
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	105		112	102				6.39
EPA 353.2 Rev. 2.0	NO2NO3-N	102		101	101				0.0
EPA 365.1 Rev. 2.0	Total-P	109		113	106				5.15
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	102		99.8	100				0.152
EPA 8270E	Acetochlor	100	102	92.2	102		1.12		9.98
	Alachlor	93.0	94.1	83.8	96.5		1.19		14.0
	Aldrin	69.1	70.0	68.5	61.7		1.38		10.6
	Alpha-BHC	107	115	110	103		7.19		6.44
	Alpha-Chlordane	109	115	120	107		5.51		6.93
	Ametryn	102	102	75.4	87.2		0.471		10.6
	Atrazine	95.2	96.2				1.03		7.25
	Atrazine Desethyl	89.2	93.8	54.8	64.8		4.98		16.8
	Azinphos Methyl	131	129	120	142		1.84		16.8
	Azoxystrobin	98.2	106	108	98.8		7.60		7.70
	Beta-BHC	123	142	153	159		14.1		4.15
	Beta-Cyfluthrin	134	150	118	112		11.6		5.67
	Bifenthrin	108	116	103	91.4		6.82		12.1
	Bromacil	87.0	90.0	84.2	95.7		3.45		11.9
	Butylate	42.9	44.3	45.5	38.6		3.22		16.2
	Caffeine	70.3	71.4				1.57		0.0433
	Carbophenothion	105	103	91.0	101		1.86		10.3
	Carfentrazone Ethyl	117	117	106	89.0		0.223		17.9
	Chlorothalonil	119	140	63.5	56.9		16.4		11.1
	Chlorpyrifos Ethyl	88.8	87.5	82.5	74.7		1.54		9.91
	Chlorpyrifos Methyl	96.3	93.9	90.1	101		2.43		11.6
	Cis-Nonachlor	93.9	103	98.6	86.8		9.71		12.7
	Cyanazine	135	141	138	162		4.14		15.5

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery		Precision	
						LCS	MS
						SMP	
EPA 8270E	Cypermethrin	135	150	117	107	10.2	9.43
	Dacthal	119	122	119	116	2.19	2.21
	DDD-p,p'	101	109	100	92.8	7.79	7.48
	DDE-p,p'	108	113	112	104	4.79	8.06
	DDT-p,p'	109	116	112	104	6.25	7.73
	Delta-BHC	120	139	158	160	14.4	1.52
	Deltamethrin	124	136	127	131	8.75	3.04
	Demeton	48.5	46.2	113	115	4.83	2.03
	Diazinon	97.6	95.9	96.4	115	1.84	17.7
	Dichlobenil	81.5	90.0	64.2	79.1	9.93	20.8
	Dicofol	98.3	106	99.3	91.6	7.55	8.10
	Dieldrin	107	114	131	118	6.71	4.48
	Difenoconazole	183	184	177	215	0.655	19.7
	Dimethenamid	79.7	80.8	69.9	83.0	1.46	17.2
	Disulfoton	91.1	94.0	89.8	99.4	3.09	10.1
	Dithiopyr	90.6	94.2	78.7	83.8	3.92	6.34
	Endosulfan I	101	106	103	98.0	5.14	5.05
	Endosulfan II	96.5	105	96.3	91.4	8.36	5.25
	Endosulfan Sulfate	112	121	120	130	7.75	8.50
	Endrin	89.9	94.8	91.6	84.6	5.38	7.87
	Endrin Aldehyde	102	102	76.0	76.8	0.546	1.10
	Endrin Ketone	104	116	111	110	10.7	0.828
	EPTC	44.9	45.4	47.7	36.9	0.967	25.6
	Esfenvalerate	129	143	114	103	10.7	10.3
	Ethalfuralin	91.6	98.1	99.8	88.1	6.85	12.4
	Ethion	76.9	74.7	57.3	82.5	2.82	36.1
	Ethoprop	87.4	87.0	110	113	0.466	3.09
	Fenamiphos	94.1	97.9	87.5	76.4	3.96	13.5
	Fenbuconazole	115	119	121	128	3.17	6.22
	Fenpropathrin	102	113	101	84.9	9.87	17.0
	Fipronil	106	107	104	111	0.397	6.86
	Fipronil Desulfinyl	106	117	97.4	116	9.79	16.2
	Fipronil Sulfide	87.8	90.6	89.5	101	3.10	10.2
	Fipronil Sulfone	81.2	78.3	71.6	87.0	3.59	14.8
	Fluazifop-P-butyl	105	104	96.7	114	1.44	16.7
	Fludioxonil	121	121	104	125	0.268	18.5
	Flumioxazin	189	191	182	235	0.990	25.0
	Flutolanil	93.1	91.4	74.7	93.1	1.88	21.9
	Fluxapyroxad	94.1	95.2	80.6	95.4	1.08	16.8
	Fonofos	86.6	84.8	91.3	105	2.12	13.7
	Gamma-BHC	113	123	131	127	8.52	3.19
	Gamma-Chlordane	103	109	98.0	88.2	5.28	7.97
	Heptachlor	86.9	90.8	79.4	71.0	4.37	11.2
	Heptachlor Epoxide	102	108	184	124	5.38	5.66
	Hexachlorobenzene	89.0	93.8	89.4	85.3	5.28	4.64
	Hexazinone	96.1	100	92.0	106	4.10	8.66
	Imazalil	93.7	96.7	77.7	84.3	3.09	6.59
	Iprodione	109	102	110	123	6.70	11.7
	Kepone	122	144	140	237	16.0	8.36
	Lambda-Cyhalothrin	129	149	112	107	14.6	3.97
	Malathion	93.3	91.0	88.2	99.8	2.52	12.4
	Metaxyl	92.8	95.7	84.0	96.8	3.05	14.1
	Methoxychlor	105	111	101	94.8	5.65	6.83
	Metolachlor	89.6	90.4	82.4	96.5	0.882	14.4

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery		Precision	
						LCS	MS
						SMP	
EPA 8270E	Metribuzin	90.6	92.6	102	105	2.20	3.35
	Mevinphos	75.7	84.4	105	111	10.8	5.06
	MGK-264	109	119	113	112	9.30	1.45
	Mirex	83.8	89.7	88.6	77.7	6.86	13.1
	Molinate	46.8	50.0	57.7	56.2	6.59	2.53
	Myclobutanil	92.2	95.7	82.5	98.4	3.69	15.9
	Naled	47.7	46.8	49.1	53.8	1.89	9.03
	Norflurazon	110	112	97.7	112	1.47	13.3
	Oxadiazon	82.0	83.9	76.2	85.7	2.36	11.7
	Oxyfluorfen	102	97.1	104	123	4.78	16.4
	Parathion Ethyl	87.9	93.1	90.9	105	5.83	14.4
	Parathion Methyl	112	110	117	125	2.11	6.41
	Pendimethalin	65.2	66.7	81.9	87.8	2.38	6.97
	Permethrin	144	131	120	112	9.21	6.80
	Phenothrin	96.6	82.0	99.2	86.1	16.4	14.1
	Phorate	74.0	67.8	87.3	84.8	8.68	2.96
	Piperonyl Butoxide	119	111	113	103	6.36	8.82
	Prallethrin	101	100	103	103	0.444	0.243
	Prodiamine	28.8	25.6	47.3	52.3	11.9	10.1
	Prometon	87.6	91.4	82.8	108	4.21	26.5
	Prometryn	95.4	95.7	67.9	72.5	0.268	6.51
	Pronamide	93.4	91.5	81.9	95.0	2.08	14.9
	Propiconazole	94.6	97.8	82.7	86.5	3.33	2.96
	Pyriproxyfen	89.2	91.8	69.7	83.7	2.91	18.2
	Simazine	98.1	103	101	110	4.94	9.23
	Sulfentrazon	100	97.6	85.0	102	2.89	14.0
	Tau-Fluvalinate	123	108	111	101	12.3	9.83
	Tebuconazole	77.6	76.9	50.0	73.7	0.883	28.7
	Terbufos	93.7	95.4	109	119	1.84	8.60
	Terbutylazine	83.0	89.0	87.1	99.2	7.07	13.0
	Tetramethrin	105	91.2	104	92.7	14.3	11.4
	Trans-Nonachlor	112	99.6	102	95.3	11.9	5.90
	Triclosan Methyl	118	114	116	111	3.81	4.29
	Trifluralin	101	91.3	98.5	86.5	9.91	13.0
EPA 8276	Chlordane	72.5	64.4	62.9	63.3	11.8	0.577
	Toxaphene	60.4	57.5	59.4	59.3	4.94	0.155
EPA 8321B	2,4-D	135		104	114		8.89
	2,4,5-T	111		107	99.4		7.50
	3-Hydroxycarbofuran	87.4		72.2	90.9		23.0
	Acesulfame K	106		123	133		7.71
	Acetaminophen	100		82.1	88.2		7.25
	Acetamiprid	93.7		68.5	81.9		17.8
	Afidopyropen	75.0		55.2	63.4		13.9
	Aldicarb	62.8		78.8	79.6		0.979
	Aldicarb Sulfone	90.8		110	104		5.32
	Aldicarb Sulfoxide	107		54.7	47.0		15.0
	AMPA	88.6		76.0	77.6		1.94
	Bentazon	113		94.7	91.2		3.76
	Benzovindiflupyr	94.7		87.0	86.6		0.421
	Carbamazepine	103		91.5	89.0		2.79
	Carbaryl	80.5		76.8	76.3		0.624
	Carbofuran	81.6		78.1	79.8		2.11
	Clothianidin	107		93.2	98.6		5.57
	Dinotefuran	104		91.2	94.9		4.03

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision LCS	SMP	MS
EPA 8321B	Diuron	94.0	84.9	81.0			4.75
	Endothall	87.2	105	106			1.56
	Fenuron	104	86.6	92.9			7.02
	Fluridone	95.3	91.7	84.9			7.75
	Glufosinate	92.7	80.0	77.7			2.85
	Glyphosate	98.0	82.0	77.0			5.95
	Hydrocodone	98.6	89.8	86.0			4.39
	Ibuprofen	84.8	82.1	77.5			5.73
	Imazapyr	108	99.5	104			4.86
	Imidacloprid	100	97.8	99.4			1.64
	Linuron	83.4	80.5	75.9			5.91
	Mandestrobin	100	82.5	87.2			5.57
	MCPP	117	117	110			6.04
	Methiocarb	73.0	74.9	73.7			1.55
	Methomyl	87.6	85.7	81.0			5.72
	Naproxen	89.9	85.6	84.1			1.76
	Oxamyl	87.6	95.2	88.9			6.86
	Primidone	103	93.5	89.5			4.37
	Propoxur	82.2	84.0	82.7			1.53
	Pyraclostrobin	92.6	82.3	83.6			1.48
	Sucralose	99.0	102	102			0.147
	Thiamethoxam	105	95.0	99.6			4.70
	Tolfenpyrad	60.2	57.3	54.7			4.62
	Triclopyr	114	106	101			4.93
SM 2120 B-2011	Color (true)	94.7				0.684	
SM 2320 B-2011	Total Alkalinity	102				0.923	
SM 2540 C-2011	TDS					6.32	
	TDS					2.87	
SM 4500 F-C-2011	Fluoride	99.0	103	103			0.457
SM 5310 B-2011	Organic Carbon	95.5	98.6	96.6			1.46

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2268464, 2268465
EPA 200.7 Rev. 4.4	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2268481, 2268482
EPA 200.8 Rev. 5.4	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2268481, 2268482
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2268464, 2268465
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2268464, 2268465
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2268466, 2268467
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2268466, 2268467
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2268466, 2268467
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2268466, 2268467
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2268468, 2268469
EPA 8270E	Organochlorine pesticides, pyrethroids and tracers in water matrices by GC/MS/MS	2268479
EPA 8270E	Organonitrogen and phosphorus pesticides in water matrices by GC/MS/MS	2268480
EPA 8276	Organochlorine pesticides in water matrices by GC/MS with Negative Chemical Ionization	2268479
EPA 8321B	Carbamates in water matrices by HPLC/MS/MS	2268477
EPA 8321B	Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS	2268478
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2268464, 2268465

Reference Method Descriptions

Method	Description	<u>Associated Samples</u>
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2268464, 2268465
SM 2340B	Hardness by calculation, in aqueous matrices, reported as mg/L CaCO ₃	2268481, 2268482
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2268464, 2268465
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2268464, 2268465
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2268464, 2268465
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2268466, 2268467

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/11/2021			08/11/2021 10:01	Sarah A Nixon	2268464
	08/11/2021			08/11/2021 10:03	Sarah A Nixon	2268465
EPA 200.7 Rev. 4.4	08/11/2021	08/16/2021 16:10	Elliott D. Healy	08/17/2021 21:15	Josef B Mick	2268481
	08/11/2021	08/16/2021 16:10	Elliott D. Healy	08/17/2021 21:21	Josef B Mick	2268482
EPA 200.8 Rev. 5.4	08/11/2021	08/16/2021 16:10	Elliott D. Healy	08/18/2021 22:15	McKinley C Carter	2268481
	08/11/2021	08/16/2021 16:10	Elliott D. Healy	08/18/2021 22:24	McKinley C Carter	2268482
EPA 300.0 Rev. 2.1	08/11/2021			08/17/2021 04:45	Lipi Saha	2268464
	08/11/2021			08/17/2021 06:58	Lipi Saha	2268465
EPA 350.1 Rev. 2.0	08/11/2021			08/13/2021 13:20	Ping Hua	2268466
	08/11/2021			08/13/2021 13:22	Ping Hua	2268467
EPA 351.2 Rev. 2.0	08/11/2021	08/25/2021 13:54	Benjamin T. Nordmann	08/26/2021 12:05	Virginia P. Brown	2268466
	08/11/2021	08/25/2021 13:54	Benjamin T. Nordmann	08/26/2021 12:07	Virginia P. Brown	2268467
EPA 353.2 Rev. 2.0	08/11/2021			08/13/2021 10:16	Beverly Y. Sanders	2268466
	08/11/2021			08/13/2021 10:17	Beverly Y. Sanders	2268467
EPA 365.1 Rev. 2.0	08/11/2021	08/13/2021 15:00	Simonne.V. Calhoun	08/16/2021 11:34	Shengyu Ding	2268466
	08/11/2021	08/13/2021 15:00	Simonne.V. Calhoun	08/16/2021 11:35	Shengyu Ding	2268467
EPA 365.1 Rev. 2.0 dissolved	08/11/2021			08/11/2021 15:34	Patrick M. Roche	2268468
	08/11/2021			08/11/2021 15:35	Patrick M. Roche	2268469
EPA 8270E	08/11/2021	08/12/2021 08:00	Fe-Val Siddell	08/17/2021 01:58	Arthur Maknenko	2268479
	08/11/2021	08/13/2021 08:00	Rasheda Ghaffari	08/17/2021 21:42	Michael Poppell	2268480
EPA 8276	08/11/2021	08/12/2021 08:00	Fe-Val Siddell	08/17/2021 11:33	Michael Poppell	2268479
EPA 8321B	08/11/2021	08/11/2021 13:00	Hoor Shaik	08/12/2021 04:28	Juyoung Kim	2268477
	08/11/2021	08/13/2021 09:00	Hoor Shaik	08/14/2021 10:15	Juyoung Kim	2268478
	08/11/2021	08/13/2021 09:00	Juyoung Kim	08/13/2021 10:30	Juyoung Kim	2268478
	08/11/2021	08/13/2021 09:00	Juyoung Kim	08/13/2021 11:32	Pramila Ghimire	2268478
	08/11/2021			08/11/2021 11:23	Sarah A Nixon	2268465
SM 2120 B-2011	08/11/2021			08/11/2021 11:24	Sarah A Nixon	2268464
	08/11/2021			08/12/2021 18:52	Dale L. Simmons	2268464
SM 2320 B-2011	08/11/2021			08/12/2021 19:00	Dale L. Simmons	2268465
	08/11/2021			08/17/2021 21:15	Josef B Mick	2268481
SM 2340B	08/11/2021			08/17/2021 21:21	Josef B Mick	2268482
	08/11/2021			08/16/2021 09:10	Yijie Li	2268464
SM 2540 C-2011	08/11/2021			08/16/2021 09:15	Yijie Li	2268465
	08/11/2021			08/16/2021 09:10	Yijie Li	2268464
SM 2540 D-2011	08/11/2021			08/16/2021 09:15	Yijie Li	2268465
	08/11/2021			08/12/2021 18:52	Dale L. Simmons	2268464
SM 4500 F-C-2011	08/11/2021			08/12/2021 19:00	Dale L. Simmons	2268465
	08/11/2021			08/23/2021 16:11	Madeline Gruver	2268467
SM 5310 B-2011	08/11/2021			08/23/2021 17:04	Madeline Gruver	2268466