From:	Tales, Carolyn R
To:	DEP-Overflowmicro
Cc:	Soni, Seemakumari; Scilabro, Gail P.
Subject:	Re: non-potable 4.20.21
Date:	Thursday, April 22, 2021 1:37:26 PM
Attachments:	image001.png non-potable 4.20.21 updated.pdf

Hello Abigail
You are right it should be 290 not 390. It was a mistake on our part I attached an updated form.
Thanks
Carolyn

From: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Sent: Thursday, April 22, 2021 11:45 AM
To: Tales, Carolyn R <Carolyn.Tales@flhealth.gov>
Cc: Soni, Seemakumari <Seemakumari.Soni@flhealth.gov>; Scilabro, Gail P. <Gail.Scilabro@flhealth.gov>
Subject: RE: non-potable 4.20.21

Hi Carolyn,

Could you take a look at the entero result for G5SW0170 (#22212)? It looks like it was 29 colonies on a 10mL dilution so I would think it would be 290 CFU/100mL instead of 390.

Please check it out and let me know what you think.

Thank you,

Abigail

Abigail L. Baker
Environmental Specialist I
FDEP Laboratories – Biology
2600 Blair Stone Road
Tallahassee, Florida 32399-24000
Abigail.L.Baker@FloridaDEP.gov

From: Tales, Carolyn R <Carolyn.Tales@flhealth.gov>
Sent: Wednesday, April 21, 2021 3:25 PM
To: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>; Sprague, Kara <Kara.Sprague@FloridaDEP.gov>

Cc: Soni, Seemakumari <Seemakumari.Soni@flhealth.gov>; Scilabro, Gail P. <Gail.Scilabro@flhealth.gov>
Subject: non-potable 4.20.21