



Bureau of Laboratories
Jacksonville, Miami, Pensacola,
Tampa, West Palm Beach

I.D. NO. E14157
DATE/TIME ANALYZED

2021 APR 20 PM 3:20

N006486

I.D. NO. E14157
REC'D TAMPA
For Lab Use Only

2021 APR 20 PM 1:40

3602 SPECTRUM BLVD.
TAMPA, FL 33612

Sample Acceptance Criteria:

Preservation ☐ on ice ☐ not on ice ☒ 5.8°C
☐ This sample does not meet the following NELAC requirements:

SAMPLE COLLECTION AND REPORT FORM FOR
NON-POTABLE WATER BACTERIOLOGICAL ANALYSIS

Program or Site Name: RO-2021-04-19-20

County: Hillsborough

Collector: FDEP SMP

Collector Phone #: _____

Collection Address: 13051 N Telecom Parkway

Collection City: Temple Terrace

Collection State: FL

Collection Zip: 33637

Date Reported: 04/20/2021

4/21/21 Comments: PO#BTDEC3

Circle the analysis requested

Station No.	Date & Time Collected	Total Coliform/100ml MPN (SM9221B) MF (SM9222B)	Fecal Coliform/100ml MPN (SM9221E) MF (SM9222D)	Enterococci/100ml MF (EPA 1600)	Other	Lab Number
G1SW0099 Enter ☒ Fecal	04/20/2021 1005	CT 4/21/21 26(B)	26(B)	50		22210
G5SW0134 Enter	04/20/2021 1040			2200		22211
G5SW0170 Enter	04/20/2021 1115			CT 4/22/21 390 290		22212
G5SW0169 Enter	04/20/2021 1140			200		22213

Name and Mailing Address of Person to Receive Report

DEP-cie:flowmicro@dep.state.fl.us

benjamin.paswater@dep.state.fl.us

All tests are performed in accordance with NELAC standards. Qualifier codes: (U) = not detected; (B) = results based on colony counts outside the acceptable range; (Z) = TNTC at highest dilution performed; (Q) = sample holding time exceeded.

For enforcement samples, EPA recommends a maximum of 6 hours time lapse from sample collection to analysis of samples in the lab. Samples must be scheduled with the lab in advance to arrive by 3:30PM to allow sufficient time for analysis. Lab procedures conform to EPA recommendations. Results of samples not received within this time frame may not be reliable.

ENVIRONMENTAL MICROBIOLOGY: POTENTIAL CREATIONAL WATER WORKSHEET

COUNTY Hillsborough

COLLECTOR FDEP SMP

DATE RECEIVED 4/20/21

MEDIA LOT NUMBER 9337200 / 9337124 SURVEY NAME BA-2021-04-19-20

DATE/TIME SET UP/ TEMP 4/20/21 / 44.5°C 41.0°C / 3:10pm

MEDIA EXPIRATION 6/14/21 / 5/21/21

Tech intl CT TRA: MFC21 MAR 17-01 / TRA: MEI 21 FEB 24-01

DATE/TIME READ/ TEMP 4/21/21 / 44.5°C 41.0°C

Waterbath ID # 2001883 Waterbath ID # 601011232

Chlorine Strip lot # A0030

LAB #	COLL #	FECAL COLIFORM (SM 9222D)		ENTERO-COCCI (EPA 1600)		FECAL COLI.	ENTERO COCCI
22210	GIW0099 4/20/21 10:05	50ml	13	50ml	25		
		10ml	3/2	10ml	2	26(B)	50
		1ml	0	1ml	0	cl res'd	cl res'd
22211	G5SW0134 4/20/21 10:40	50ml		50ml	TNTC 60+		
		10ml		10ml	TNTC 60+		2200
		1ml		1ml	22	cl res'd	cl res'd
22212	G5SW0170 4/20/21 11:15	50ml		50ml	TNTC 60+		4/22/21 CT
		10ml		10ml	29		390
		1ml		1ml	5		290
22213	G5SW0169 4/20/21 11:40	50ml		50ml	79		
		10ml		10ml	20		200
		1ml		1ml	0	cl res'd	cl res'd

Controls Setup #			
Pre Rinse		A	Ent
		A	FC
Mid Rinse			Ent
			FC
Post Rinse		A	Ent
		A	FC

Controls	
E Coli	60+
E. Faecalis	60+
Filter	# FG1656
Rinse water	# 20210218-09 20210218-12

20210218-12

Key: FC- Fecal Coliform P-Coliforms present
Ent-Enterococci A- Coliforms absent

Reviewed By: ln Date: 4/21/21

Bol- T EM002a (10/17) Rev(9/10/20)