



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: Unnamed Creek At Palmview Rd **Date:** 1/25/2021
WIN/Field ID: G3WA0013 **WBID:** 233A **ENR:** **Latitude:** 30.8122
County: JACKSON **Org ID:** 21FLWQA **Longitude:** -85.4542
Receiving Body of Water: Choctawhatchee - St. Andrew **RQ:** 2021-01-25-05

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Blake Spencer	x		x	x	x
☒	Avril Wood-McGrath		x			
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☒ Syringe	☐ Pole
Depth measured with: MeterStick	
Secchi Equipment ID:	
Collection Method	
☒ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other _____	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# <u>EXO #1</u>					
Photos: <u>NO</u>		Flow: <u>FLOW</u>		Tide: _____		Tide Times: High: _____ Low: _____					
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other: _____											
SBIO ID Numbers: _____											
Notes: 											

Duplicate Sample

Time Zone: <u>EST</u>	Time: _____	Field ID: <u>N/A</u> (WIN ID_DUP)
WIN DMT Activity ID: _____		

Blank

Time Zone: _____	Time: _____	Field ID: _____ (Blank Type_DDDMMYYY)
DI Source: _____	Location: _____	
DI Type: _____	Equipment ID: _____	
DI Container (Name/ID): _____		DI Container Type: _____

LIMS EVENT ID: WAS-SECT-2021-01-25-01 **WIN Import ID:** 21FLWQA_WAS-SECT-2021-01-25-01_03-05-2021.TXT
Enter Field Parameters, Verify Station ID In LIMS DMT: SBS 2/26/21 **QC Station ID, Field Parameters In LIMS DMT:** ESB 3/3/21
(Initials/Date) (Initials/Date)
Scan/Save Field Sheets SBS 2/26/21 **Migrate Data to WIN:** DT 3/5/21 **Verify Data In WIN:** _____
(Initials/Date) (Initials/Date) (Initials/Date)



RQ-2021-01-25-05

Customer: WQAP

Project ID: NWQI

Collected By: Blake Spencer, Avril Wood-McGrath

Field Report Prepared By: Blake Spencer

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



G3WA0013

Location: Unnamed Creek At Palmview Rd

Comments:

Matrix: Fresh

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
1/25/2021	1303 EST	9.2	96.7	17.5	5.72	0.1	0.2 ☒ VOB (S)	0.2	36.1	0.02
	EST						Central Lab please use top row for login purposes			

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # SA0293040 Exp: 10-19-21	☒ Ice ☒ H2SO4*	☒ <2	1	A
Metals (P-500ML)	☒ W-ICPMS ☒ W-ICP ☒ W-HARD Lot # NA0038050 Exp: 02-07-21	☒ Ice ☒ HNO3*	☒ <2	1	A
Filtered Nutrients (P125ML)	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter Filter Lot # 16930646 Syringe Lot # 7086795	☒ Ice		1	A
Chlorophyll (BP-1L)	☒ CHLSUITE-W	☒ Ice		1	A
Physical/Aggregate (Fresh) (P-1L)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice		1	A
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates (PJ-2L)	☐ MI-FW-QLDC	☐ Formalin			
Periphyton (PT-50ML)	☐ ALGAE-ID	☐ Ice			
Microbiology (Contract Lab Bottle)	☒ E Coli ☐ F Coli ☐ Enterococci ☐ Contract Lab	☒ Ice		1	A
Molecular (QPCR-P-500ML)	☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice			
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-SQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-DIELDRIN	☐ Ice			
Phytoplankton/Periphyton	DTM-QL-C ☐ PEN-QL-C ☐ DTY-QN-C ☐ PKD-QN-C DTM-QN-C ☐ PRN-QN-C ☐ PKD-QN-BV	☐ Ice			

Name:
Blake Spencer
Avril Wood-McGrath

Signature:

Date:

1/25/21

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: <u>1/25/2021</u> Customer: <u>WAS</u> RQ: <u>RQ-2021-01-25-05</u>	Collected By: <u>Blake Spencer, Avril Wood-McGrath</u> Sampling Agency: <u>FDEP</u> Project ID: <u>NWQI</u>
--	--

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 2

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Alligator Creek Upstream Of Sr 273	 G3WA0005
Scurlock Creek At Pilgrim Rd	 G3WA0009
Pine Barn Creek At Woodrest Rd	 G3WA0011
Unnamed Creek At Woodrest Rd	 G3WA0010
Unnamed Creek At Palmview Rd	 G3WA0013
Minnow Creek At Lovewood Rd	 G3WA0008

RELINQUISHED BY(SIGNATURE): *Blake Spencer*

DATE/TIME: 1/25/21
15:15

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

Rec'd/Inspected Date:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Eto #1

RQ- 2021-01-25-05

Project: AWQI

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification _____

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	1/25/21	0653	20.0	752.0	9.03	9.06	99.8	—	1.133	P/F	L/F
ICV	Blake Spencer	1/25/21	0654	20.0	757.0	9.03	9.10	100.3	—	—	P/F	L/F
CCV	Amil Wood-McGath	4/25/21	16:19	20.8	755.8	8.95	9.02	100.8	—	—	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	1/25/20	0657	1911201	3/2021	1000	1000	P/F	L/F
ICV	Blake Spencer	1/25/20	0658	1911201	3/2021	1000	1000	P/F	L/F
CCV	Amil Wood-McGath	1/25/20	16:12	2911C93	11/2021	100	104.8	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	1/25/20	0700	200407B	10/2021	7.02	20.1	7.02	-12.7	P/F	L/F
Calibr.	Blake Spencer	1/25/20	0702	200221A	09/2021	4.00	20.1	4.00	155.8	P/F	L/F
Calibr.	Blake Spencer	1/25/20	0705	200407A	10/2021	10.06	20.0	10.66	176.1	P/F	L/F
ICV	Blake Spencer	1/25/20	0706	200407A	10/2021	10.06	20.0	10.07	177.8	P/F	L/F
CCV	Amil Wood-McGath	1/25/20	16:17	200221A	09/2021	4.00	20.3	4.03	153.3	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

Cooler Packing Worksheet for Request: RQ-2021-01-25-05

NWQI - Alligator Creek

Ship Cooler On: 1/21/2021

Customer/Project: WAS-SECT/SMP

Assigned To: KLUG_K

Requester: Curtis Musson

Priority: 3

850-245-8453

2600 Blair Stone Rd

Tallahassee, FL 32399

Attn: Curtis Musson

Comments:

No preservatives needed.

Group A: Nutrients, Metals, E. coli X 6

Packing Notes:

No FedEx labels

No Trash Bags

No Bubblewrap

Requested Analyses:

Bottle Group: A

of Sites: 6

Container ID: P-1L

Qty: 6

Preservation: ICE

Lot # 00078451

Description: Plastic Bottle 1L

Analysis

ALKALINITY

TURBIDITY

W-CL-IC

W-COLOR

W-F

W-SO4-IC

W-TDS

W-TSS

Description

Alkalinity in aqueous matrices as mg CaCO3/L

Turbidity in aqueous matrices

Chloride in aqueous matrices

True Color measured at 450 nm (not valid for NPDES monitoring)

Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)

Sulfate in aqueous matrices

Total Dissolved Solids in aqueous matrices

Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Container ID: BP-1L

Qty: 6

Preservation: ICE

Lot # 1299839

Description: Brown Plastic Bottle 1L

Analysis

CHLSUITE-W

Description

Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Container ID: P-250ML-WO-THI

Qty: 6 Preservation: ICE

Lot #

~~000304~~

Description: 250 ml Sterilized no tablet

L2C275

Analysis

ECOLI-18QT

Description

Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems

Container ID: P-500ML

Qty: 6 Preservation: HNO3

Lot #

00078385

Description: Plastic Bottle 500 mL

Affix RED nitric acid dot to container.

Analysis

W-HARD

W-ICP

W-ICPMS

Description

Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3

Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects

Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

Container ID: P-500ML

Qty: 6 Preservation: ICE And H2SO4

Lot #

00078385

Description: Plastic Bottle 500 mL

Affix YELLOW sulfuric acid dot to container.

Analysis

W-NH3

W-NO2NO3

W-S-A-TP

W-TKN

W-TOC

Description

Ammonia in aqueous matrices as mg N/L

Nitrite/Nitrate in aqueous matrices as mg N/L

Total Phosphorus in aqueous matrices as mg P/L

Total Kjeldahl Nitrogen in aqueous matrices

Nonpurgeable Organic Carbon in aqueous matrices

Container ID: P-125ML

Qty: 6 Preservation: FILTER-ICE

Lot #

00079279

Description: Plastic Bottle 125 mL

Analysis

W-PO4-F

Description

Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Cooler Packed By: ihh

Number of Coolers: 2

Date: 1.19.21

Kit must also include:

- ☒ Field Sheets
- ☒ Temperature Control Bottle
- ☐ FedEx Bills, if applicable
- ☐ Plastic Bags/Zip Ties
- ☐ Acid Labels (sets of 24)

If Preservation Included:

ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____

PLEASE RETURN ALL COOLERS FOR RQ-2021-01-25-05