

Chemical Analysis Report

SO-DIST-2022-04-12-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **2022 SMP: Hatchett Creek**
Request ID: **RQ-2022-04-11-32**
Customer: **SO-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FLDEP South District Office
2295 Victoria Ave. Suite 364
Fort Myers, FL 33901-3381
Attn: William Mullen

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 29-APR-2022 17:57



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Hatchett Creek Tidal 2

Collection Date/Time: 04/11/2022 11:20

Field ID: G3SD0192

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2318297	EPA 180.1 Rev. 2.0	Turbidity	1.8		NTU	P412262	
	SM 2320 B-2011	Total Alkalinity	134		mg CaCO3/L	P412808	
	SM 2540 D-2011	TSS	3	IA	mg/L	P412561	
	SM 4500-F- C-2011	Fluoride	1.2		mg F/L	P412811	
2318298	EPA 350.1 Rev. 2.0	Ammonia-N	0.008		mg N/L	P412302	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.36	I	mg N/L	P412510	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P412398	
	EPA 365.1 Rev. 2.0	Total-P	0.063		mg P/L	P412599	
	SM 5310 B-2011	Organic Carbon	4.6		mg C/L	P412477	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P412262

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P412302

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P412510

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P412398

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P412599

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P412808

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 D-2011
Batch ID: P412561

Component	Result	Code	Units
TSS	3	U	mg/L

Reference Method: SM 4500-F- C-2011
Batch ID: P412811

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P412477

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P412262

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Turbidity	96.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P412302

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.4		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P412510

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	95.0		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P412398

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P412599

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	102		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P412808

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	97.7		P	94 - 106

Reference Method: SM 4500-F- C-2011
Batch ID: P412811

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	101		P	93 - 104

Reference Method: SM 5310 B-2011
Batch ID: P412477

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	97.6		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P412302

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2318298	Ammonia-N	100	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P412510

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2317821	Kjeldahl Nitrogen	91.3	90.9	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P412398

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2317701	NO2NO3-N	98.5	97.5	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P412599

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2318318	Total-P	96.6	95.8	P/P	90 - 110

Reference Method: SM 4500-F- C-2011
Batch ID: P412811

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2319759	Fluoride	99.1	97.7	P/P	92 - 107

Reference Method: SM 5310 B-2011
Batch ID: P412477

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2318660	Organic Carbon	87.1	88.4	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P412262

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2318385	Turbidity	7.81	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P412302

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2318298	Ammonia-N	0.343	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P412510

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2317821	Kjeldahl Nitrogen	0.349	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P412398

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2317701	NO2NO3-N	1.02	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P412599

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2318318	Total-P	0.794	Spike	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P412808

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2319759	Total Alkalinity	3.36	Sample	P	0 - 5.2

Reference Method: SM 4500-F- C-2011
Batch ID: P412811

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2319759	Fluoride	0.953	Spike	P	0 - 2.1

Reference Method: SM 5310 B-2011
Batch ID: P412477

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2318660	Organic Carbon	1.12	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A111434

Included Lab Sample IDs: 2318297

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.0	99.0	P/P	95 - 105

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A111449

Included Lab Sample IDs: 2318298

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.5	99.3	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A111492

Included Lab Sample IDs: 2318298

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A111518

Included Lab Sample IDs: 2318298

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	98.0	94.6	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A111610

Included Lab Sample IDs: 2318298

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	98.2	102	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A111656

Included Lab Sample IDs: 2318297

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	97.6	98.2	P/P	90 - 110

Reference Method: SM 4500-F- C-2011

Run ID: A111656

Included Lab Sample IDs: 2318297

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A111662

Included Lab Sample IDs: 2318298

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	102	P/P	90 - 110

Quality Assurance Report

Calibration Verification

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity	96.3				7.81	
EPA 350.1 Rev. 2.0	Ammonia-N	97.4	100	100			0.343
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	95.0	91.3	90.9			0.349
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	100	98.5	97.5			1.02
EPA 365.1 Rev. 2.0	Total-P	102	96.6	95.8			0.794
SM 2320 B-2011	Total Alkalinity	97.7				3.36	
SM 4500-F- C-2011	Fluoride	101	99.1	97.7			0.953
SM 5310 B-2011	Organic Carbon	97.6	87.1	88.4			1.12

Reference Method Descriptions

Method	Description	<u>Associated Samples</u>
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2318297
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2318298
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2318298
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2318298
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2318298
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2318297
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2318297
SM 4500-F- C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2318297
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2318298

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	04/12/2022			04/12/2022 14:37	Khloe J Berg	2318297
EPA 350.1 Rev. 2.0	04/12/2022			04/13/2022 12:34	Ping Hua	2318298
EPA 351.2 Rev. 2.0	04/12/2022	04/20/2022 14:20	Samantha L Bates	04/21/2022 13:08	Samantha L Bates	2318298
EPA 353.2 Rev. 2.0	04/12/2022			04/14/2022 16:12	Nathaniel J Jones	2318298
EPA 365.1 Rev. 2.0	04/12/2022	04/21/2022 17:55	Simonne.V. Calhoun	04/22/2022 15:59	Sarah A Nixon	2318298
SM 2320 B-2011	04/12/2022			04/22/2022 08:54	Sarah A Nixon	2318297
SM 2540 D-2011	04/12/2022			04/13/2022 14:26	Yijie Li	2318297
SM 4500-F- C-2011	04/12/2022			04/22/2022 12:40	Sarah A Nixon	2318297
SM 5310 B-2011	04/12/2022			04/16/2022 00:38	Judith K. Clark	2318298