

Chemical Analysis Report

NW-DIST-2022-01-26-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Pace Run**
Request ID: **RQ-2022-01-24-40**
Customer: **NW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
160 Government Center
Suite 208
Pensacola, FL 32502-5794
Attn: Myranda Butler

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Environmental Consultant

Date Certified: 17-FEB-2022 09:41



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals, Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Coldwater Creek West Fork on CR 178

Collection Date/Time: 01/25/2022 13:00

Field ID: G4NW0016

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2301748	DEP SOP: NU-094-1	Color (true)**	13		PCU	P409004	
	EPA 180.1 Rev. 2.0	Turbidity	3.8		NTU	P409008	
	EPA 300.0 Rev. 2.1	Chloride	5.9		mg Cl/L	P409417	
		Sulfate	1.8		mg SO4/L	P409420	
	SM 2320 B-2011	Total Alkalinity	2.9		mg CaCO3/L	P409140	
	SM 2540 C-2011	TDS	41		mg/L	P409502	
	SM 2540 D-2011	TSS	5	I	mg/L	P409494	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P409145	
2301749	EPA 350.1 Rev. 2.0	Ammonia-N	0.018		mg N/L	P409349	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.20	I	mg N/L	P409231	
	EPA 353.2 Rev. 2.0	NO2NO3-N	1.8		mg N/L	P409208	
	EPA 365.1 Rev. 2.0	Total-P	0.005		mg P/L	P409035	
	SM 5310 B-2011	Organic Carbon	1.7		mg C/L	P409765	
2301750	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P409001	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Coldwater Creek West Fork downstream of HWY 87

Collection Date/Time: 01/25/2022 12:30

Field ID: G4NW0017

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2301736	DEP SOP: NU-094-1	Color (true)**	10		PCU	P409004	
	EPA 180.1 Rev. 2.0	Turbidity	2.3		NTU	P409008	
	EPA 300.0 Rev. 2.1	Chloride	5.8		mg Cl/L	P409417	
		Sulfate	1.2		mg SO4/L	P409420	
	SM 2320 B-2011	Total Alkalinity	2.4	I	mg CaCO3/L	P409140	
	SM 2540 C-2011	TDS	38		mg/L	P409502	
	SM 2540 D-2011	TSS	2	I	mg/L	P409494	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P409145	
2301737	EPA 350.1 Rev. 2.0	Ammonia-N	0.012		mg N/L	P409349	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.16	I	mg N/L	P409231	
	EPA 353.2 Rev. 2.0	NO2NO3-N	2.1		mg N/L	P409208	
	EPA 365.1 Rev. 2.0	Total-P	0.004	I	mg P/L	P409035	
	SM 5310 B-2011	Organic Carbon	1.3		mg C/L	P409765	
2301738	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P409001	
2301770	EPA 8321B	Aldicarb	0.020	U	ug/L	P409047	
		Aldicarb Sulfone	0.0020	U	ug/L	P409047	
		Aldicarb Sulfoxide	0.0020	U	ug/L	P409047	
		Carbaryl	0.0020	U	ug/L	P409047	
		Carbofuran	0.0020	U	ug/L	P409047	
		3-Hydroxycarbofuran	0.0020	U	ug/L	P409047	
		Methiocarb	0.0020	U	ug/L	P409047	
		Methomyl	0.0020	U	ug/L	P409047	
		Oxamyl	0.0020	U	ug/L	P409047	
		Propoxur	0.0020	U	ug/L	P409047	
		2,4-D	0.0020	U	ug/L	P409048	
		Acesulfame K	0.10	U	ug/L	P409043	
2301771		2,4,5-T	0.0040	U	ug/L	P409048	
		AMPA	0.10	U	ug/L	P409043	
		Acetaminophen	0.0080	U	ug/L	P409048	
		Acetamiprid	0.0020	U	ug/L	P409048	
		Endothall	0.25	U	ug/L	P409043	MS
		Glufosinate	0.10	U	ug/L	P409043	
		Glyphosate	0.10	U	ug/L	P409043	
		Afidopyropen	0.0020	U	ug/L	P409048	
		Bentazon	0.0033		ug/L	P409048	
		Sucralose	0.039	I	ug/L	P409043	
		Benzovindiflupyr	0.0020	U	ug/L	P409048	
		Carbamazepine	8.0E-04	U	ug/L	P409048	
		Clothianidin	0.0027	I	ug/L	P409048	
		Dinotefuran	0.0046	I	ug/L	P409048	
		Diuron	0.0020	U	ug/L	P409048	
		Fenuron	0.016	U	ug/L	P409048	

Field ID: G4NW0017

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2301771	EPA 8321B	Fluridone	8.0E-04	U	ug/L	P409048	
		Imazapyr	0.0045	I	ug/L	P409048	
		Hydrocodone	0.0020	U	ug/L	P409048	
		Ibuprofen	0.020	U	ug/L	P409048	
		Imidacloprid	0.0020	U	ug/L	P409048	
		Linuron	0.0040	U	ug/L	P409048	
		Mandestrobin	0.0020	U	ug/L	P409048	
		MCPP	0.0020	U	ug/L	P409048	
		Naproxen	0.0080	U	ug/L	P409048	
		Primidone	0.0040	U	ug/L	P409048	
		Pyraclostrobin	0.0020	U	ug/L	P409048	
		Silvex	0.0040	U	ug/L	P409048	
		Thiamethoxam	0.0020	U	ug/L	P409048	
		Tolfenpyrad	0.0020	U	ug/L	P409048	
		Triclopyr	0.0040	U	ug/L	P409048	
2301774	EPA 8270E	Aldrin	0.67	U	ng/L	P408985	RPD
		Alpha-BHC	0.10	U	ng/L	P408985	MS, RPD
		Alpha-Chlordane	0.20	U	ng/L	P408985	MS, RPD
		Beta-BHC	0.10	U	ng/L	P408985	
		Beta-Cyfluthrin**	1.3	U	ng/L	P408985	RPD
		Bifenthrin**	0.51	U	ng/L	P408985	RPD
		Chlorothalonil	1.3	U	ng/L	P408985	RPD
		Cis-Nonachlor**	0.20	U	ng/L	P408985	
		Cypermethrin	1.5	U	ng/L	P408985	RPD
		Dacthal**	0.15	U	ng/L	P408985	MS, RPD
		DDD-p,p'	0.26	U	ng/L	P408985	RPD
		DDE-p,p'	0.20	U	ng/L	P408985	RPD
		DDT-p,p'	0.26	U	ng/L	P408985	MS, RPD
		Delta-BHC	0.41	U	ng/L	P408985	RPD
		Deltamethrin**	1.3	U	ng/L	P408985	
		Dichlobenil**	0.26	U	ng/L	P408985	RPD
		Dicofol	1.3	U	ng/L	P408985	RPD
		Dieldrin	0.66	I	ng/L	P408985	RPD
		Endosulfan I	0.41	U	ng/L	P408985	MS, RPD
		Endosulfan II	0.41	U	ng/L	P408985	RPD
		Endosulfan Sulfate	0.31	U	ng/L	P408985	
		Endrin	0.41	U	ng/L	P408985	
		Endrin Aldehyde	0.77	U	ng/L	P408985	
		Endrin Ketone	0.67	I	ng/L	P408985	RPD
		Esfenvalerate**	0.77	U	ng/L	P408985	
		Ethalfuralin**	0.15	U	ng/L	P408985	MS, RPD
		Fenpropathrin**	0.26	U	ng/L	P408985	RPD
		Gamma-BHC	0.13	U	ng/L	P408985	RPD
		Gamma-Chlordane	0.20	U	ng/L	P408985	MS, RPD
		Heptachlor	0.20	U	ng/L	P408985	MS, RPD

Field ID: G4NW0017

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2301774	EPA 8270E	Heptachlor Epoxide	0.26	U	ng/L	P408985	RPD
		Hexachlorobenzene	1.0	U	ng/L	P408985	MS, RPD
		Kepone**	5.1	U	ng/L	P408985	
		Lambda-Cyhalothrin**	0.77	U	ng/L	P408985	
		Methoprene**	0.77	U	ng/L	P408985	RPD
		Methoxychlor	0.31	U	ng/L	P408985	
		MGK-264**	0.20	U	ng/L	P408985	
		Mirex	0.72	U	ng/L	P408985	RPD
		Oxychlorthane**	0.31	U	ng/L	P408985	RPD
		Permethrin	1.6	U	ng/L	P408985	RPD
		Phenothrin**	0.92	U	ng/L	P408985	RPD
		Piperonyl Butoxide**	0.15	U	ng/L	P408985	
		Prallethrin**	2.0	U	ng/L	P408985	RPD
		Tau-Fluvalinate**	0.26	U	ng/L	P408985	RPD
		Tetramethrin**	0.77	U	ng/L	P408985	RPD
		Trans-Nonachlor**	0.20	U	ng/L	P408985	MS, RPD
		Triclosan Methyl**	0.15	U	ng/L	P408985	MS, RPD
		Trifluralin	0.20	U	ng/L	P408985	MS, RPD
	EPA 8276	Chlordane	2.6	U	ng/L	P408985	
		Toxaphene	15	U	ng/L	P408985	
2301775	EPA 8270E	Acetochlor	0.24	U	ng/L	P409108	
		Alachlor	0.24	U	ng/L	P409108	
		Ametryn	0.59	U	ng/L	P409108	
		Atrazine	0.37	I	ng/L	P409108	
		Atrazine Desethyl	1.5	U	ng/L	P409108	
		Azinphos Methyl	2.0	U	ng/L	P409108	
		Azoxystrobin**	0.49	U	ng/L	P409108	
		Bromacil	0.97	I	ng/L	P409108	
		Butylate	0.39	U	ng/L	P409108	
		Caffeine**	7.3	U	ng/L	P409108	
		Carbophenothion	0.20	U	ng/L	P409108	
		Carfentrazone Ethyl**	0.098	U	ng/L	P409108	
		Chlorfenapyr**	0.17	U	ng/L	P409108	
		Chlorpyrifos Ethyl	0.24	U	ng/L	P409108	
		Chlorpyrifos Methyl	0.049	U	ng/L	P409108	
		Coumaphos**	0.49	U	ng/L	P409108	
		Cyanazine	3.2	U	ng/L	P409108	
		Cyprodinil**	0.098	U	ng/L	P409108	
		Demeton	1.5	U	ng/L	P409108	
		Diazinon	0.12	U	ng/L	P409108	
		Difenoconazole**	0.59	U	ng/L	P409108	
		Dimethenamid**	0.098	U	ng/L	P409108	
		Dimethomorph**	0.17	U	ng/L	P409108	
		Disulfoton	0.49	U	ng/L	P409108	
		Dithiopyr**	0.17	U	ng/L	P409108	

Field ID: G4NW0017

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2301775	EPA 8270E	EPTC	1.2	U	ng/L	P409108	
		Ethion	0.15	U	ng/L	P409108	
		Ethoprop	0.098	U	ng/L	P409108	
		Etridiazole**	0.15	U	ng/L	P409108	
		Fenamiphos	0.49	U	ng/L	P409108	
		Fenbuconazole**	0.12	U	ng/L	P409108	
		Fipronil	0.24	U	ng/L	P409108	
		Fipronil Desulfinyl**	0.12	U	ng/L	P409108	
		Fipronil Sulfide	0.15	U	ng/L	P409108	
		Fipronil Sulfone	0.15	U	ng/L	P409108	
		Fluazifop-P-butyl**	0.098	U	ng/L	P409108	
		Fludioxonil**	0.12	U	ng/L	P409108	
		Flumioxazin**	1.7	U	ng/L	P409108	
		Flutolanil**	0.32	I	ng/L	P409108	
		Fluxapyroxad**	0.13	I	ng/L	P409108	
		Fonofos	0.20	U	ng/L	P409108	
		Hexazinone	1.7	I	ng/L	P409108	
		Imazalil**	0.73	U	ng/L	P409108	
		Iprodione**	0.59	U	ng/L	P409108	
		Loratadine**	0.098	U	ng/L	P409108	
		Malathion	0.34	U	ng/L	P409108	
		Metalaxyl	0.73	U	ng/L	P409108	
		Metolachlor	18		ng/L	P409108	
		Metribuzin	2.4	U	ng/L	P409108	
		Mevinphos	1.0	U	ng/L	P409108	
		Molinate	0.29	U	ng/L	P409108	
		Myclobutanil**	0.24	U	ng/L	P409108	
		Naled**	1.5	UJ	ng/L	P409108	LCS, MS
		Napropamide**	0.39	U	ng/L	P409108	
		Norflurazon	0.93	I	ng/L	P409108	
		Oxadiazon	0.29	U	ng/L	P409108	
		Oxyfluorfen**	0.15	U	ng/L	P409108	
		Parathion Ethyl	0.24	U	ng/L	P409108	
		Parathion Methyl	0.54	U	ng/L	P409108	
		Pendimethalin	0.98	U	ng/L	P409108	
		Phenytol**	3.4	U	ng/L	P409108	
		Phorate	0.51	U	ng/L	P409108	
		Prodiamine**	0.17	U	ng/L	P409108	
		Prometon	0.88	U	ng/L	P409108	RPD
		Prometryn	0.20	U	ng/L	P409108	
		Pronamide**	0.15	U	ng/L	P409108	
		Propanil**	0.12	U	ng/L	P409108	
		Propargite**	1.5	U	ng/L	P409108	
		Propiconazole**	0.49	U	ng/L	P409108	
		Pyraflufen Ethyl**	0.24	U	ng/L	P409108	

Field ID: G4NW0017

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2301775	EPA 8270E	Pyridaben**	0.44	U	ng/L	P409108	
		Pyriproxyfen**	0.12	U	ng/L	P409108	
		Simazine	0.49	U	ng/L	P409108	
		Stirophos**	0.12	U	ng/L	P409108	
		Sulfentrazone**	0.49	U	ng/L	P409108	
		Tebuconazole**	0.49	U	ng/L	P409108	
		Terbufos	0.098	U	ng/L	P409108	
		Terbuthylazine	0.42	U	ng/L	P409108	
		Thiobencarb**	0.59	U	ng/L	P409108	
		Triadimefon**	0.24	U	ng/L	P409108	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

EPA 8321B: MS accuracy for some analytes could not be assessed due to high concentration of parameter in the spiked sample.

EPA 8270E: MS accuracy for Flutolanil not assessed due to high content of this parameter in the sample spiked. Refer to the Lab Analysis Report for an explanation of QC Codes.

Sample Location: Bucket Branch downstream of Pat Brown Road Collection Date/Time: 01/25/2022 10:40

Field ID: G4NW0532

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2301779	EPA 200.7 Rev. 4.4	Calcium	1.56		mg/L	P409121	
		Iron	2.82E+03		ug/L	P409121	
		Magnesium	1.28		mg/L	P409121	
		Potassium	0.60	I	mg/L	P409121	
		Sodium	1.8	I	mg/L	P409121	
		Strontium	10		ug/L	P409121	
		Tin	3.0	U	ug/L	P409121	
		Titanium	10.2		ug/L	P409121	
		Vanadium	2.4	I	ug/L	P409121	
		Zinc	6.8	I	ug/L	P409121	
	EPA 200.8 Rev. 5.4	Aluminum	1.23E+03		ug/L	P409121	
		Antimony	0.050	U	ug/L	P409121	
		Arsenic	0.64		ug/L	P409121	
		Barium	40.0		ug/L	P409121	
		Beryllium	0.080		ug/L	P409121	
		Boron**	12.1		ug/L	P409121	
		Cadmium	0.020	U	ug/L	P409121	
		Chromium	1.0	I	ug/L	P409121	
		Cobalt	3.46		ug/L	P409121	
		Copper	0.77	I	ug/L	P409121	
		Lead	1.93		ug/L	P409121	
		Manganese	448		ug/L	P409121	
		Molybdenum	0.15	U	ug/L	P409121	
		Nickel	0.70	I	ug/L	P409121	
		Selenium	0.20	U	ug/L	P409121	
		Silver	0.010	U	ug/L	P409121	
		Thallium	0.10	U	ug/L	P409121	
	SM 2340B	Hardness	9.2		mg/L	P409121	

Ref. Method and Comment:
EPA 200.8 Rev. 5.4: The lead result was confirmed in the undigested sample on 01/31/2022.

Sample Location: Clear Creek upstream of Hwy 87A

Collection Date/Time: 01/25/2022 11:55

Field ID: G4NW0510

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2301742	DEP SOP: NU-094-1	Color (true)**	2.5	U	PCU	P409004	
	EPA 180.1 Rev. 2.0	Turbidity	2.4		NTU	P409008	
	EPA 300.0 Rev. 2.1	Chloride	4.7		mg Cl/L	P409417	
		Sulfate	0.68		mg SO4/L	P409420	
	SM 2320 B-2011	Total Alkalinity	1.9	I	mg CaCO3/L	P409140	
	SM 2540 C-2011	TDS	34		mg/L	P409502	
	SM 2540 D-2011	TSS	2	I	mg/L	P409494	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P409145	
2301744	EPA 350.1 Rev. 2.0	Ammonia-N	0.006		mg N/L	P409406	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.097	I	mg N/L	P409191	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	1.8		mg N/L	P409208	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P409035	
	SM 5310 B-2011	Organic Carbon	0.61	I	mg C/L	P409765	
2301746	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P409001	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Big Coldwater Creek end of Steel Bridge Rd

Collection Date/Time: 01/25/2022 11:25

Field ID: G4NW0027

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2301743	DEP SOP: NU-094-1	Color (true)**	12		PCU	P409004	
	EPA 180.1 Rev. 2.0	Turbidity	2.4		NTU	P409008	
	EPA 300.0 Rev. 2.1	Chloride	4.4		mg Cl/L	P409417	
		Sulfate	0.99		mg SO4/L	P409420	
	SM 2320 B-2011	Total Alkalinity	1.7	I	mg CaCO3/L	P409140	
	SM 2540 C-2011	TDS	31		mg/L	P409502	
	SM 2540 D-2011	TSS	3	I	mg/L	P409494	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P409145	
2301745	EPA 350.1 Rev. 2.0	Ammonia-N	0.008		mg N/L	P409406	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.12	I	mg N/L	P409191	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	1.3		mg N/L	P409208	
	EPA 365.1 Rev. 2.0	Total-P	0.004	I	mg P/L	P409035	
	SM 5310 B-2011	Organic Carbon	1.4		mg C/L	P409765	
2301747	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P409001	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Bucket Branch end of Bucket Creek Rd

Collection Date/Time: 01/25/2022 11:00

Field ID: G4NW0059

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2301780	EPA 200.7 Rev. 4.4	Calcium	1.35		mg/L	P409121	
		Iron	220		ug/L	P409121	
		Magnesium	1.13		mg/L	P409121	
		Potassium	0.45	I	mg/L	P409121	
		Sodium	1.8	I	mg/L	P409121	
		Strontium	7.8	I	ug/L	P409121	
		Tin	3.0	U	ug/L	P409121	
		Titanium	0.93	I	ug/L	P409121	
		Vanadium	2.0	U	ug/L	P409121	
	EPA 200.8 Rev. 5.4	Zinc	5.0	U	ug/L	P409121	
		Aluminum	62		ug/L	P409121	
		Antimony	0.050	U	ug/L	P409121	
		Arsenic	0.099	I	ug/L	P409121	
		Barium	19.8		ug/L	P409121	
		Beryllium	0.010	U	ug/L	P409121	
		Boron**	11.3		ug/L	P409121	
		Cadmium	0.020	U	ug/L	P409121	
		Chromium	0.40	U	ug/L	P409121	
		Cobalt	0.065	I	ug/L	P409121	
		Copper	0.40	U	ug/L	P409121	
		Lead	0.20	U	ug/L	P409121	
		Manganese	11.1		ug/L	P409121	
		Molybdenum	0.15	U	ug/L	P409121	
		Nickel	0.50	U	ug/L	P409121	
		Selenium	0.20	U	ug/L	P409121	
		Silver	0.010	U	ug/L	P409121	
		Thallium	0.10	U	ug/L	P409121	
	SM 2340B	Hardness	8.0		mg/L	P409121	

Sample Location: Clear Creek upstream of Pat Brown Road

Collection Date/Time: 01/25/2022 10:00

Field ID: G4NW0360

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2301739	DEP SOP: NU-094-1	Color (true)**	6.1		PCU	P409004	
	EPA 180.1 Rev. 2.0	Turbidity	3.0		NTU	P409008	
	EPA 300.0 Rev. 2.1	Chloride	4.0		mg Cl/L	P409417	
		Sulfate	0.59		mg SO4/L	P409420	
	SM 2320 B-2011	Total Alkalinity	2.0	I	mg CaCO3/L	P409140	
	SM 2540 C-2011	TDS	25		mg/L	P409502	
	SM 2540 D-2011	TSS	3	I	mg/L	P409494	
2301740	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P409145	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.006		mg N/L	P409406	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.15	I	mg N/L	P409375	
	EPA 353.2 Rev. 2.0	NO2NO3-N	1.1		mg N/L	P409208	
	EPA 365.1 Rev. 2.0	Total-P	0.005	I	mg P/L	P409035	
2301741	SM 5310 B-2011	Organic Carbon	0.84	I	mg C/L	P409765	
2301741	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P409001	
2301778	EPA 200.7 Rev. 4.4	Calcium	1.32		mg/L	P409121	
		Iron	370		ug/L	P409121	
		Magnesium	1.20		mg/L	P409121	
		Potassium	0.35	I	mg/L	P409121	
		Sodium	2.0		mg/L	P409121	
		Strontium	11.0		ug/L	P409121	
		Tin	3.0	U	ug/L	P409121	
	EPA 200.8 Rev. 5.4	Titanium	1.5	I	ug/L	P409121	
		Vanadium	2.0	U	ug/L	P409121	
		Zinc	5.0	U	ug/L	P409121	
		Aluminum	104		ug/L	P409121	
		Antimony	0.050	U	ug/L	P409121	
		Arsenic	0.16	I	ug/L	P409121	
		Barium	26.8		ug/L	P409121	
		Beryllium	0.016	I	ug/L	P409121	
		Boron**	9.1		ug/L	P409121	
		Cadmium	0.020	U	ug/L	P409121	
		Chromium	0.40	U	ug/L	P409121	
		Cobalt	0.188		ug/L	P409121	
		Copper	0.40	U	ug/L	P409121	
		Lead	0.20	U	ug/L	P409367	
		Manganese	9.57		ug/L	P409121	
		Molybdenum	0.15	U	ug/L	P409121	
		Nickel	0.50	U	ug/L	P409121	
		Selenium	0.20	U	ug/L	P409121	
		Silver	0.010	U	ug/L	P409121	
		Thallium	0.10	U	ug/L	P409121	
	SM 2340B	Hardness	8.3		mg/L	P409121	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: FB

Collection Date/Time: 01/25/2022 10:45

Field ID: FB_01252022

Matrix: W-FIELD-BK

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2301751	DEP SOP: NU-094-1	Color (true)**	2.5	U	PCU	P409004	
	EPA 180.1 Rev. 2.0	Turbidity	0.10	U	NTU	P409008	
	EPA 300.0 Rev. 2.1	Chloride	0.20	U	mg Cl/L	P409417	
		Sulfate	0.20	U	mg SO4/L	P409420	
	SM 2320 B-2011	Total Alkalinity	0.65	U	mg CaCO3/L	P409140	
	SM 2540 C-2011	TDS	15	U	mg/L	P409499	
	SM 2540 D-2011	TSS	2	U	mg/L	P409491	
2301752	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P409145	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P409406	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P409191	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P409208	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P409035	
2301753	SM 5310 B-2011	Organic Carbon	0.50	U	mg C/L	P409765	
2301753	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P409001	
2301781	EPA 200.7 Rev. 4.4	Calcium	0.18	I	mg/L	P409121	
		Iron	30	U	ug/L	P409121	
		Magnesium	0.040	U	mg/L	P409121	
		Potassium	0.30	U	mg/L	P409121	
		Sodium	0.50	U	mg/L	P409121	
		Strontium	2.0	U	ug/L	P409121	
		Tin	3.0	U	ug/L	P409121	
	EPA 200.8 Rev. 5.4	Titanium	0.75	U	ug/L	P409121	
		Vanadium	2.0	U	ug/L	P409121	
		Zinc	5.0	U	ug/L	P409121	
		Aluminum	5.0	U	ug/L	P409121	
		Antimony	0.050	U	ug/L	P409121	
		Arsenic	0.050	U	ug/L	P409121	
		Barium	0.20	U	ug/L	P409121	
		Beryllium	0.010	U	ug/L	P409121	
		Boron**	2.0	U	ug/L	P409121	
		Cadmium	0.020	U	ug/L	P409121	
		Chromium	0.40	U	ug/L	P409121	
		Cobalt	0.020	U	ug/L	P409121	
		Copper	0.40	U	ug/L	P409121	
		Lead	0.20	U	ug/L	P409121	
		Manganese	0.10	U	ug/L	P409121	
		Molybdenum	0.15	U	ug/L	P409121	
		Nickel	0.50	U	ug/L	P409121	
		Selenium	0.20	U	ug/L	P409121	
		Silver	0.010	U	ug/L	P409121	
		Thallium	0.10	U	ug/L	P409121	
	SM 2340B	Hardness	0.48	I	mg/L	P409121	

Matrix: W-FIELD-BK

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
Ref. Method and Comment:							
SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.							
EPA 200.7 Rev. 4.4: The calcium result was confirmed in the undigested sample on 01/31/2022.							

Quality Assurance Report Method Blank Results

Reference Method: DEP SOP: NU-094-1
Batch ID: P409004

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P409008

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P409121

Component	Result	Code	Units
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Strontium	2.0	U	ug/L
Tin	3.0	U	ug/L
Titanium	0.75	U	ug/L
Vanadium	2.0	U	ug/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P409121

Component	Result	Code	Units
Aluminum	5.0	U	ug/L
Antimony	0.050	U	ug/L
Arsenic	0.050	U	ug/L
Barium	0.20	U	ug/L
Beryllium	0.010	U	ug/L
Boron	2.0	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Cobalt	0.020	U	ug/L
Copper	0.40	U	ug/L
Lead	0.20	U	ug/L
Manganese	0.10	U	ug/L
Molybdenum	0.15	U	ug/L
Nickel	0.50	U	ug/L
Selenium	0.20	U	ug/L
Silver	0.010	U	ug/L
Thallium	0.10	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P409367

Component	Result	Code	Units
Lead	0.20	U	ug/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P409417

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P409420

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P409349

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P409406

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409191

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409231

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409375

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P409208

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P409035

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P409001

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8270E
Batch ID: P408985

Component	Result	Code	Units
Aldrin	0.65	U	ng/L
Aldrin	0.65	U	ng/L
Alpha-BHC	0.10	U	ng/L
Alpha-BHC	0.10	U	ng/L
Alpha-Chlordane	0.20	U	ng/L
Alpha-Chlordane	0.20	U	ng/L
Beta-BHC	0.10	U	ng/L
Beta-BHC	0.10	U	ng/L
Beta-Cyfluthrin	1.2	U	ng/L
Beta-Cyfluthrin	1.2	U	ng/L
Bifenthrin	0.50	U	ng/L
Bifenthrin	0.50	U	ng/L
Chlorothalonil	1.2	U	ng/L
Chlorothalonil	1.2	U	ng/L
Cis-Nonachlor	0.20	U	ng/L
Cis-Nonachlor	0.20	U	ng/L
Cypermethrin	1.5	U	ng/L
Cypermethrin	1.5	U	ng/L
Dacthal	0.15	U	ng/L
Dacthal	0.15	U	ng/L
DDD-p,p'	0.25	U	ng/L
DDD-p,p'	0.25	U	ng/L
DDE-p,p'	0.20	U	ng/L
DDE-p,p'	0.20	U	ng/L
DDT-p,p'	0.25	U	ng/L
DDT-p,p'	0.25	U	ng/L
Delta-BHC	0.40	U	ng/L
Delta-BHC	0.40	U	ng/L
Deltamethrin	1.2	U	ng/L
Deltamethrin	1.2	U	ng/L
Dichlobenil	0.25	U	ng/L
Dichlobenil	0.25	U	ng/L
Dicofol	1.2	U	ng/L
Dicofol	1.2	U	ng/L
Dieldrin	0.40	U	ng/L
Dieldrin	0.40	U	ng/L
Endosulfan I	0.40	U	ng/L
Endosulfan I	0.40	U	ng/L
Endosulfan II	0.40	U	ng/L
Endosulfan II	0.40	U	ng/L
Endosulfan Sulfate	0.30	U	ng/L
Endosulfan Sulfate	0.30	U	ng/L
Endrin	0.40	U	ng/L
Endrin	0.40	U	ng/L
Endrin Aldehyde	0.75	U	ng/L
Endrin Aldehyde	0.75	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E

Batch ID: P408985

Component	Result	Code	Units
Endrin Ketone	0.40	U	ng/L
Endrin Ketone	0.40	U	ng/L
Esfenvalerate	0.75	U	ng/L
Esfenvalerate	0.75	U	ng/L
Ethalfuralin	0.15	U	ng/L
Ethalfuralin	0.15	U	ng/L
Fenpropathrin	0.25	U	ng/L
Fenpropathrin	0.25	U	ng/L
Gamma-BHC	0.12	U	ng/L
Gamma-BHC	0.12	U	ng/L
Gamma-Chlordane	0.20	U	ng/L
Gamma-Chlordane	0.20	U	ng/L
Heptachlor	0.20	U	ng/L
Heptachlor	0.20	U	ng/L
Heptachlor Epoxide	0.25	U	ng/L
Heptachlor Epoxide	0.25	U	ng/L
Hexachlorobenzene	1.0	U	ng/L
Hexachlorobenzene	1.0	U	ng/L
Kepone	5.0	U	ng/L
Kepone	5.0	U	ng/L
Lambda-Cyhalothrin	0.75	U	ng/L
Lambda-Cyhalothrin	0.75	U	ng/L
Methoprene	0.75	U	ng/L
Methoprene	0.75	U	ng/L
Methoxychlor	0.30	U	ng/L
Methoxychlor	0.30	U	ng/L
MGK-264	0.20	U	ng/L
MGK-264	0.20	U	ng/L
Mirex	0.70	U	ng/L
Mirex	0.70	U	ng/L
Oxychlordane	0.30	U	ng/L
Oxychlordane	0.30	U	ng/L
Permethrin	1.6	U	ng/L
Permethrin	1.6	U	ng/L
Phenothrin	0.90	U	ng/L
Phenothrin	0.90	U	ng/L
Piperonyl Butoxide	0.15	U	ng/L
Piperonyl Butoxide	0.15	U	ng/L
Prallethrin	2.0	U	ng/L
Prallethrin	2.0	U	ng/L
Tau-Fluvalinate	0.25	U	ng/L
Tau-Fluvalinate	0.25	U	ng/L
Tetramethrin	0.75	U	ng/L
Tetramethrin	0.75	U	ng/L
Trans-Nonachlor	0.20	U	ng/L
Trans-Nonachlor	0.20	U	ng/L
Triclosan Methyl	0.15	U	ng/L
Triclosan Methyl	0.15	U	ng/L
Trifluralin	0.20	U	ng/L
Trifluralin	0.20	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E

Batch ID: P409108

Component	Result	Code	Units
Acetochlor	0.25	U	ng/L
Acetochlor	0.25	U	ng/L
Alachlor	0.25	U	ng/L
Alachlor	0.25	U	ng/L
Ametryn	0.60	U	ng/L
Ametryn	0.60	U	ng/L
Atrazine	0.25	U	ng/L
Atrazine	0.25	U	ng/L
Atrazine Desethyl	1.5	U	ng/L
Atrazine Desethyl	1.5	U	ng/L
Azinphos Methyl	2.0	U	ng/L
Azinphos Methyl	2.0	U	ng/L
Azoxystrobin	0.50	U	ng/L
Azoxystrobin	0.50	U	ng/L
Bromacil	0.50	U	ng/L
Bromacil	0.50	U	ng/L
Butylate	0.40	U	ng/L
Butylate	0.40	U	ng/L
Caffeine	7.5	U	ng/L
Caffeine	7.5	U	ng/L
Carbophenothion	0.20	U	ng/L
Carbophenothion	0.20	U	ng/L
Carfentrazone Ethyl	0.10	U	ng/L
Carfentrazone Ethyl	0.10	U	ng/L
Chlorfenapyr	0.18	U	ng/L
Chlorfenapyr	0.18	U	ng/L
Chlorpyrifos Ethyl	0.25	U	ng/L
Chlorpyrifos Ethyl	0.25	U	ng/L
Chlorpyrifos Methyl	0.050	U	ng/L
Chlorpyrifos Methyl	0.050	U	ng/L
Coumaphos	0.50	U	ng/L
Coumaphos	0.50	U	ng/L
Cyanazine	3.2	U	ng/L
Cyanazine	3.2	U	ng/L
Cyprodinil	0.10	U	ng/L
Cyprodinil	0.10	U	ng/L
Demeton	1.5	U	ng/L
Demeton	1.5	U	ng/L
Diazinon	0.12	U	ng/L
Diazinon	0.12	U	ng/L
Difenoconazole	0.60	U	ng/L
Difenoconazole	0.60	U	ng/L
Dimethenamid	0.10	U	ng/L
Dimethenamid	0.10	U	ng/L
Dimethomorph	0.18	U	ng/L
Dimethomorph	0.18	U	ng/L
Disulfoton	0.50	U	ng/L
Disulfoton	0.50	U	ng/L
Dithiopyr	0.18	U	ng/L
Dithiopyr	0.18	U	ng/L
EPTC	1.2	U	ng/L
EPTC	1.2	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E

Batch ID: P409108

Component	Result	Code	Units
Ethion	0.15	U	ng/L
Ethion	0.15	U	ng/L
Ethoprop	0.10	U	ng/L
Ethoprop	0.10	U	ng/L
Etridiazole	0.15	U	ng/L
Etridiazole	0.15	U	ng/L
Fenamiphos	0.50	U	ng/L
Fenamiphos	0.50	U	ng/L
Fenbuconazole	0.12	U	ng/L
Fenbuconazole	0.12	U	ng/L
Fipronil	0.25	U	ng/L
Fipronil	0.25	U	ng/L
Fipronil Desulfinyl	0.12	U	ng/L
Fipronil Desulfinyl	0.12	U	ng/L
Fipronil Sulfide	0.15	U	ng/L
Fipronil Sulfide	0.15	U	ng/L
Fipronil Sulfone	0.15	U	ng/L
Fipronil Sulfone	0.15	U	ng/L
Fluazifop-P-butyl	0.10	U	ng/L
Fluazifop-P-butyl	0.10	U	ng/L
Fludioxonil	0.12	U	ng/L
Fludioxonil	0.12	U	ng/L
Flumioxazin	1.8	U	ng/L
Flumioxazin	1.8	U	ng/L
Flutolanil	0.10	U	ng/L
Flutolanil	0.10	U	ng/L
Fluxapyroxad	0.10	U	ng/L
Fluxapyroxad	0.10	U	ng/L
Fonofos	0.20	U	ng/L
Fonofos	0.20	U	ng/L
Hexazinone	0.50	U	ng/L
Hexazinone	0.50	U	ng/L
Imazalil	0.75	U	ng/L
Imazalil	0.75	U	ng/L
Iprodione	0.60	U	ng/L
Iprodione	0.60	U	ng/L
Loratadine	0.10	U	ng/L
Loratadine	0.10	U	ng/L
Malathion	0.35	U	ng/L
Malathion	0.35	U	ng/L
Metalaxyl	0.75	U	ng/L
Metalaxyl	0.75	U	ng/L
Metolachlor	0.35	U	ng/L
Metolachlor	0.35	U	ng/L
Metribuzin	2.5	U	ng/L
Metribuzin	2.5	U	ng/L
Mevinphos	1.0	U	ng/L
Mevinphos	1.0	U	ng/L
Molinate	0.30	U	ng/L
Molinate	0.30	U	ng/L
Myclobutanil	0.25	U	ng/L
Myclobutanil	0.25	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E

Batch ID: P409108

Component	Result	Code	Units
Naled	1.5	U	ng/L
Naled	1.5	U	ng/L
Napropamide	0.40	U	ng/L
Napropamide	0.40	U	ng/L
Norflurazon	0.50	U	ng/L
Norflurazon	0.50	U	ng/L
Oxadiazon	0.30	U	ng/L
Oxadiazon	0.30	U	ng/L
Oxyfluorfen	0.15	U	ng/L
Oxyfluorfen	0.15	U	ng/L
Parathion Ethyl	0.25	U	ng/L
Parathion Ethyl	0.25	U	ng/L
Parathion Methyl	0.55	U	ng/L
Parathion Methyl	0.55	U	ng/L
Pendimethalin	1.0	U	ng/L
Pendimethalin	1.0	U	ng/L
Phenytol	3.5	U	ng/L
Phenytol	3.5	U	ng/L
Phorate	0.52	U	ng/L
Phorate	0.52	U	ng/L
Prodiamine	0.18	U	ng/L
Prodiamine	0.18	U	ng/L
Prometon	0.90	U	ng/L
Prometon	0.90	U	ng/L
Prometryn	0.20	U	ng/L
Prometryn	0.20	U	ng/L
Pronamide	0.15	U	ng/L
Pronamide	0.15	U	ng/L
Propanil	0.12	U	ng/L
Propanil	0.12	U	ng/L
Propargite	1.5	U	ng/L
Propargite	1.5	U	ng/L
Propiconazole	0.50	U	ng/L
Propiconazole	0.50	U	ng/L
Pyraflufen Ethyl	0.25	U	ng/L
Pyraflufen Ethyl	0.25	U	ng/L
Pyridaben	0.45	U	ng/L
Pyridaben	0.45	U	ng/L
Pyriproxyfen	0.12	U	ng/L
Pyriproxyfen	0.12	U	ng/L
Simazine	0.50	U	ng/L
Simazine	0.50	U	ng/L
Stirophos	0.12	U	ng/L
Stirophos	0.12	U	ng/L
Sulfentrazon	0.50	U	ng/L
Sulfentrazon	0.50	U	ng/L
Tebuconazole	0.50	U	ng/L
Tebuconazole	0.50	U	ng/L
Terbufos	0.10	U	ng/L
Terbufos	0.10	U	ng/L
Terbuthylazine	0.42	U	ng/L
Terbuthylazine	0.42	U	ng/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 8270E
Batch ID: P409108

Component	Result	Code	Units
Thiobencarb	0.60	U	ng/L
Thiobencarb	0.60	U	ng/L
Triadimefon	0.25	U	ng/L
Triadimefon	0.25	U	ng/L

Reference Method: EPA 8276
Batch ID: P408985

Component	Result	Code	Units
Chlordane	2.5	U	ng/L
Chlordane	2.5	U	ng/L
Toxaphene	15	U	ng/L
Toxaphene	15	U	ng/L

Reference Method: EPA 8321B
Batch ID: P409043

Component	Result	Code	Units
Acesulfame K	0.10	U	ug/L
AMPA	0.10	U	ug/L
Endothall	0.25	U	ug/L
Glufosinate	0.10	U	ug/L
Glyphosate	0.10	U	ug/L
Sucralose	0.010	U	ug/L

Reference Method: EPA 8321B
Batch ID: P409047

Component	Result	Code	Units
3-Hydroxycarbofuran	0.0020	U	ug/L
Aldicarb	0.020	U	ug/L
Aldicarb Sulfone	0.0020	U	ug/L
Aldicarb Sulfoxide	0.0020	U	ug/L
Carbaryl	0.0020	U	ug/L
Carbofuran	0.0020	U	ug/L
Methiocarb	0.0020	U	ug/L
Methomyl	0.0020	U	ug/L
Oxamyl	0.0020	U	ug/L
Propoxur	0.0020	U	ug/L

Reference Method: EPA 8321B
Batch ID: P409048

Component	Result	Code	Units
2,4-D	0.0020	U	ug/L
2,4,5-T	0.0040	U	ug/L
Acetaminophen	0.0080	U	ug/L
Acetamiprid	0.0020	U	ug/L
Afidopyropen	0.0020	U	ug/L
Bentazon	8.0E-04	U	ug/L
Benzovindiflupyr	0.0020	U	ug/L
Carbamazepine	8.0E-04	U	ug/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8321B

Batch ID: P409048

Component	Result	Code	Units
Clothianidin	0.0020	U	ug/L
Dinotefuran	0.0020	U	ug/L
Diuron	0.0020	U	ug/L
Fenuron	0.016	U	ug/L
Fluridone	8.0E-04	U	ug/L
Hydrocodone	0.0020	U	ug/L
Ibuprofen	0.020	U	ug/L
Imazapyr	0.0040	U	ug/L
Imidacloprid	0.0020	U	ug/L
Linuron	0.0040	U	ug/L
Mandestrobin	0.0020	U	ug/L
MCPP	0.0020	U	ug/L
Naproxen	0.0080	U	ug/L
Primidone	0.0040	U	ug/L
Pyraclostrobin	0.0020	U	ug/L
Silvex	0.0040	U	ug/L
Thiamethoxam	0.0020	U	ug/L
Tolfenpyrad	0.0020	U	ug/L
Triclopyr	0.0040	U	ug/L

Reference Method: SM 2320 B-2011

Batch ID: P409140

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-2011

Batch ID: P409499

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 C-2011

Batch ID: P409502

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011

Batch ID: P409491

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 2540 D-2011

Batch ID: P409494

Component	Result	Code	Units
TSS	2	U	mg/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 4500 F-C-2011
Batch ID: P409145

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P409765

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: DEP SOP: NU-094-1
Batch ID: P409004

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	99.2		P	85 - 115

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P409121

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Calcium	102		P	85 - 115
Iron	107		P	85 - 115
Magnesium	102		P	85 - 115
Potassium	97.5		P	85 - 115
Sodium	97.7		P	85 - 115
Strontium	98.7		P	85 - 115
Tin	104		P	85 - 115
Titanium	102		P	85 - 115
Vanadium	102		P	85 - 115
Zinc	101		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P409121

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aluminum	99.7		P	85 - 115
Antimony	97.7		P	85 - 115
Arsenic	101		P	85 - 115
Barium	103		P	85 - 115
Beryllium	100		P	85 - 115
Boron	98.7		P	85 - 115
Cadmium	98.8		P	85 - 115
Chromium	99.0		P	85 - 115
Cobalt	100		P	85 - 115
Copper	101		P	85 - 115
Lead	100		P	85 - 115
Manganese	97.6		P	85 - 115
Molybdenum	100		P	85 - 115
Nickel	100		P	85 - 115
Selenium	99.8		P	85 - 115
Silver	102		P	85 - 115
Thallium	103		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P409367

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Lead	101		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P409417

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	101		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P409420

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	96.8		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P409349

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	98.4		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P409406

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.4		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409191

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	99.5		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409231

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	98.2		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409375

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	99.7		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P409208

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P409035

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	105		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P409001

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	100		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8270E
 Batch ID: P408985

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aldrin	66.0		P	10 - 93.0
Alpha-BHC	96.5		P	75 - 106
Alpha-Chlordane	90.4		P	50 - 109
Beta-BHC	109		P	65 - 135
Beta-Cyfluthrin	112		P	23 - 167
Bifenthrin	92.6		P	11 - 123
Chlorothalonil	113		P	26 - 182
Cis-Nonachlor	90.6		P	40 - 111
Cypermethrin	86.7		P	25 - 155
Dacthal	95.8		P	73 - 112
DDD-p,p'	90.3		P	47 - 108
DDE-p,p'	89.0		P	48 - 107
DDT-p,p'	88.4		P	49 - 111
Delta-BHC	112		P	68 - 143
Deltamethrin	132		P	16 - 176
Dichlobenil	85.9		P	32 - 113
Dicofol	83.4		P	50 - 108
Dieldrin	91.3		P	56 - 110
Endosulfan I	90.6		P	60 - 105
Endosulfan II	88.1		P	50 - 120
Endosulfan Sulfate	91.6		P	55 - 121
Endrin	89.2		P	35 - 117
Endrin Aldehyde	91.3		P	36 - 116
Endrin Ketone	101		P	56 - 130
Esfenvalerate	109		P	10 - 179
Ethalfuralin	87.6		P	49 - 99.0
Fenpropathrin	101		P	35 - 119
Gamma-BHC	87.0		P	72 - 115
Gamma-Chlordane	85.9		P	45 - 107
Heptachlor	80.1		P	41 - 100
Heptachlor Epoxide	90.7		P	58 - 106
Hexachlorobenzene	86.5		P	39 - 100
Kepone	156		P	10 - 191
Lambda-Cyhalothrin	111		P	10 - 186
Methoprene	76.6		P	10 - 129
Methoxychlor	103		P	61 - 111
MGK-264	86.9		P	20 - 123
Mirex	74.4		P	10 - 182
Oxychlordane	87.6		P	39 - 110
Permethrin	112		P	26 - 152
Phenothrin	87.5		P	10 - 109
Piperonyl Butoxide	103		P	27 - 142
Prallethrin	71.1		P	14 - 109
Tau-Fluvalinate	101		P	16 - 134
Tetramethrin	80.3		P	10 - 125
Trans-Nonachlor	85.2		P	45 - 107
Triclosan Methyl	92.0		P	60 - 110
Trifluralin	85.4		P	49 - 98.0

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8270E

Batch ID: P409108

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Acetochlor	114		P	70 - 121
Alachlor	103		P	68 - 119
Ametryn	112		P	64 - 139
Atrazine	106		P	76 - 120
Atrazine Desethyl	86.6		P	44 - 126
Azinphos Methyl	72.2		P	43 - 192
Azoxystrobin	109		P	36 - 224
Bromacil	109		P	52 - 121
Butylate	66.8		P	28 - 91.0
Caffeine	85.8		P	50 - 134
Carbophenothion	111		P	56 - 129
Carfentrazone Ethyl	125		P	60 - 141
Chlorfenapyr	103		P	56 - 115
Chlorpyrifos Ethyl	113		P	60 - 118
Chlorpyrifos Methyl	110		P	68 - 119
Coumaphos	171		P	18 - 250
Cyanazine	133		P	61 - 250
Cyprodinil	140		P	55 - 145
Demeton	96.8		P	30 - 159
Diazinon	106		P	70 - 123
Difenoconazole	108		P	49 - 204
Dimethenamid	91.0		P	64 - 115
Dimethomorph	111		P	12 - 219
Disulfoton	89.0		P	44 - 125
Dithiopyr	104		P	64 - 115
EPTC	61.1		P	28 - 86.0
Ethion	103		P	33 - 125
Ethoprop	85.4		P	64 - 116
Etridiazole	69.1		P	34 - 91.0
Fenamiphos	126		P	12 - 127
Fenbuconazole	145		P	59 - 151
Fipronil	110		P	67 - 129
Fipronil Desulfinyl	111		P	65 - 127
Fipronil Sulfide	95.7		P	61 - 116
Fipronil Sulfone	105		P	55 - 117
Fluazifop-P-butyl	97.3		P	62 - 127
Fludioxonil	111		P	62 - 128
Flumioxazin	248		P	33 - 250
Flutolanil	104		P	56 - 127
Fluxapyroxad	127		P	45 - 128
Fonofos	85.7		P	62 - 111
Hexazinone	114		P	46 - 139
Imazalil	88.8		P	38 - 142
Iprodione	70.6		P	47 - 135
Loratadine	87.4		P	10 - 244
Malathion	128		P	61 - 139
Metalaxyl	99.7		P	10 - 119
Metolachlor	111		P	65 - 118
Metribuzin	111		P	51 - 250
Mevinphos	89.0		P	53 - 121
Molinate	76.1		P	42 - 96.0
Myclobutanil	109		P	55 - 128

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8270E

Batch ID: P409108

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Naled	119		F	10 - 98.0
Napropamide	103		P	49 - 121
Norflurazon	119		P	61 - 126
Oxadiazon	92.9		P	59 - 116
Oxyfluorfen	117		P	64 - 137
Parathion Ethyl	106		P	70 - 112
Parathion Methyl	120		P	76 - 133
Pendimethalin	79.5		P	52 - 117
Phenytoin	113		P	10 - 199
Phorate	98.5		P	48 - 121
Prodiamine	98.2		P	26 - 142
Prometon	105		P	43 - 123
Prometryn	94.7		P	66 - 127
Pronamide	121		P	75 - 121
Propanil	118		P	45 - 150
Propargite	163		P	42 - 208
Propiconazole	99.8		P	60 - 125
Pyraflufen Ethyl	124		P	70 - 138
Pyridaben	159		P	35 - 245
Pyriproxyfen	61.7		P	30 - 149
Simazine	123		P	72 - 125
Stirophos	84.7		P	29 - 164
Sulfentrazone	136		P	21 - 139
Tebuconazole	78.7		P	10 - 115
Terbufos	102		P	60 - 123
Terbuthylazine	101		P	72 - 115
Thiobencarb	112		P	31 - 139
Triadimefon	98.3		P	65 - 116

Reference Method: EPA 8276

Batch ID: P408985

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chlordane	81.7		P	61 - 118
Toxaphene	76.4		P	51 - 122

Reference Method: EPA 8321B

Batch ID: P409043

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Acesulfame K	99.1		P	40 - 150
AMPA	91.4		P	40 - 150
Endothall	136		P	40 - 150
Glufosinate	97.9		P	40 - 150
Glyphosate	91.2		P	40 - 150
Sucralose	91.7		P	40 - 150

Reference Method: EPA 8321B

Batch ID: P409047

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
3-Hydroxycarbofuran	79.7		P	30 - 160

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8321B
Batch ID: P409047

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aldicarb	45.1		P	30 - 160
Aldicarb Sulfone	93.7		P	30 - 160
Aldicarb Sulfoxide	90.8		P	30 - 160
Carbaryl	81.3		P	30 - 160
Carbofuran	86.3		P	30 - 160
Methiocarb	70.4		P	30 - 160
Methomyl	88.6		P	30 - 160
Oxamyl	95.6		P	30 - 160
Propoxur	78.6		P	30 - 160

Reference Method: EPA 8321B
Batch ID: P409048

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2,4-D	89.2		P	30 - 160
2,4,5-T	75.5		P	30 - 160
Acetaminophen	96.0		P	30 - 160
Acetamiprid	98.2		P	30 - 160
Afidopyropen	71.9		P	30 - 160
Bentazon	105		P	30 - 160
Benzovindiflupyr	93.1		P	30 - 160
Carbamazepine	99.5		P	30 - 160
Clothianidin	106		P	30 - 160
Dinotefuran	105		P	30 - 160
Diuron	92.4		P	30 - 160
Fenuron	111		P	30 - 160
Fluridone	97.5		P	30 - 160
Hydrocodone	115		P	30 - 160
Ibuprofen	79.1		P	30 - 160
Imazapyr	149		P	30 - 160
Imidacloprid	91.8		P	30 - 160
Linuron	84.1		P	30 - 160
Mandestrobin	99.1		P	30 - 160
MCPP	108		P	30 - 160
Naproxen	98.5		P	30 - 160
Primidone	99.4		P	30 - 160
Pyraclostrobin	92.4		P	30 - 160
Silvex	69.8		P	30 - 160
Thiamethoxam	96.0		P	30 - 160
Tolfenpyrad	52.0		P	30 - 160
Triclopyr	75.5		P	30 - 160

Reference Method: SM 2320 B-2011
Batch ID: P409140

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	99.0		P	94 - 106

Reference Method: SM 4500 F-C-2011
Batch ID: P409145

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-2011
Batch ID: P409145

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.5		P	93 - 104

Reference Method: SM 5310 B-2011
Batch ID: P409765

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P409121

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301666	Calcium	102		P	80 - 120
2301666	Iron	110	107	P/P	80 - 120
2301666	Magnesium	104		P	80 - 120
2301666	Potassium	96.9		P	80 - 120
2301666	Sodium	97.4		P	80 - 120
2301666	Strontium	99.8	97.9	P/P	80 - 120
2301666	Tin	103	101	P/P	80 - 120
2301666	Titanium	104	101	P/P	80 - 120
2301666	Vanadium	105	102	P/P	80 - 120
2301666	Zinc	104	101	P/P	80 - 120
2301780	Calcium	106		P	80 - 120
2301780	Iron	110		P	80 - 120
2301780	Magnesium	110		P	80 - 120
2301780	Potassium	101		P	80 - 120
2301780	Sodium	103		P	80 - 120
2301780	Strontium	101		P	80 - 120
2301780	Tin	108		P	80 - 120
2301780	Titanium	103		P	80 - 120
2301780	Vanadium	104		P	80 - 120
2301780	Zinc	105		P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4

Batch ID: P409121

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301666	Aluminum	97.8	99.8	P/P	80 - 120
2301666	Antimony	99.7	101	P/P	80 - 120
2301666	Arsenic	102	101	P/P	80 - 120
2301666	Barium	102	104	P/P	80 - 120
2301666	Beryllium	101	100	P/P	80 - 120
2301666	Boron	96.0	101	P/P	80 - 120
2301666	Cadmium	98.5	100	P/P	80 - 120
2301666	Chromium	100	102	P/P	80 - 120
2301666	Cobalt	100	101	P/P	80 - 120
2301666	Copper	101	102	P/P	80 - 120
2301666	Lead	101	102	P/P	80 - 120
2301666	Manganese	98.6	99.5	P/P	80 - 120
2301666	Molybdenum	102	103	P/P	80 - 120
2301666	Nickel	100	101	P/P	80 - 120
2301666	Selenium	98.7	99.3	P/P	80 - 120
2301666	Silver	102	103	P/P	80 - 120
2301666	Thallium	109	111	P/P	80 - 120
2301780	Aluminum	101		P	80 - 120
2301780	Antimony	100		P	80 - 120
2301780	Arsenic	99.6		P	80 - 120
2301780	Barium	104		P	80 - 120
2301780	Beryllium	101		P	80 - 120
2301780	Boron	104		P	80 - 120
2301780	Cadmium	101		P	80 - 120
2301780	Chromium	101		P	80 - 120
2301780	Cobalt	99.2		P	80 - 120
2301780	Copper	102		P	80 - 120

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P409121

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301780	Lead	103		P	80 - 120
2301780	Manganese	100		P	80 - 120
2301780	Molybdenum	100		P	80 - 120
2301780	Nickel	101		P	80 - 120
2301780	Selenium	96.3		P	80 - 120
2301780	Silver	103		P	80 - 120
2301780	Thallium	111		P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P409367

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2302658	Lead	101	100	P/P	80 - 120
2302782	Lead	99.6		P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P409417

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301653	Chloride	95.0		P	90 - 110
2301742	Chloride	98.6		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P409420

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301653	Sulfate	93.2		P	90 - 110
2301742	Sulfate	96.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P409349

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301749	Ammonia-N	100	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P409406

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301740	Ammonia-N	95.0	96.2	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409191

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301639	Kjeldahl Nitrogen	102	111	P/F	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409375

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301510	Kjeldahl Nitrogen	101	96.2	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P409208

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301761	NO2NO3-N	102	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P409035

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301762	Total-P	102	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P409001

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301741	O-Phosphate-P	100	102	P/P	90 - 110

Reference Method: EPA 8270E

Batch ID: P408985

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301521	Aldrin	23.6	40.1	P/P	10 - 89.0
2301521	Alpha-BHC	60.7	81.9	F/P	71 - 105
2301521	Alpha-Chlordane	37.7	56.7	F/P	43 - 107
2301521	Beta-BHC	70.0	95.6	P/P	57 - 152
2301521	Beta-Cyfluthrin	48.5	75.1	P/P	29 - 186
2301521	Bifenthrin	36.7	62.5	P/P	19 - 119
2301521	Chlorothalonil	91.9	169	P/P	10 - 245
2301521	Cis-Nonachlor	44.2	54.7	P/P	36 - 110
2301521	Cypermethrin	47.1	70.5	P/P	24 - 169
2301521	Dacthal	53.6	73.7	F/P	62 - 115
2301521	DDD-p,p'	36.3	54.1	P/P	36 - 109
2301521	DDE-p,p'	32.7	48.2	P/P	30 - 114
2301521	DDT-p,p'	40.2	60.6	F/P	42 - 108
2301521	Delta-BHC	77.8	113	P/P	60 - 164
2301521	Deltamethrin	60.7	87.2	P/P	10 - 210
2301521	Dichlobenil	34.8	52.9	P/P	23 - 108
2301521	Dicofol	45.9	64.1	P/P	44 - 109
2301521	Dieldrin	40.2	58.7	P/P	37 - 119
2301521	Endosulfan I	43.1	62.4	F/P	52 - 105
2301521	Endosulfan II	38.8	54.6	P/P	35 - 127
2301521	Endosulfan Sulfate	52.6	69.3	P/P	39 - 130
2301521	Endrin	38.0	56.4	P/P	30 - 116
2301521	Endrin Aldehyde	32.2	50.5	P/P	20 - 110
2301521	Endrin Ketone	52.9	75.7	P/P	48 - 134
2301521	Esfenvalerate	53.3	77.5	P/P	10 - 199
2301521	Ethalfuralin	38.3	57.3	F/P	48 - 95.0
2301521	Fenpropathrin	39.6	54.6	P/P	37 - 121
2301521	Gamma-BHC	72.7	96.3	P/P	61 - 126
2301521	Gamma-Chlordane	36.8	55.4	F/P	39 - 105
2301521	Heptachlor	31.3	50.7	F/P	38 - 96.0
2301521	Heptachlor Epoxide	46.1	64.1	P/P	42 - 114
2301521	Hexachlorobenzene	33.2	54.5	F/P	39 - 93.0
2301521	Kepone	81.8	103	P/P	10 - 215
2301521	Lambda-Cyhalothrin	53.5	90.1	P/P	10 - 204

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270E
Batch ID: P408985

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301521	Methoprene	29.0	51.3	P/P	17 - 108
2301521	Methoxychlor	58.6	78.4	P/P	56 - 115
2301521	MGK-264	48.4	69.3	P/P	35 - 120
2301521	Mirex	25.8	41.0	P/P	10 - 178
2301521	Oxychlordane	52.4	86.7	P/P	47 - 91.0
2301521	Permethrin	53.0	79.0	P/P	27 - 170
2301521	Phenothrin	38.7	68.7	P/P	10 - 133
2301521	Piperonyl Butoxide	57.4	87.3	P/P	23 - 150
2301521	Prallethrin	36.8	60.2	P/P	21 - 113
2301521	Tau-Fluvalinate	41.1	69.0	P/P	16 - 158
2301521	Tetramethrin	42.1	70.6	P/P	22 - 132
2301521	Trans-Nonachlor	37.0	54.6	F/P	43 - 102
2301521	Triclosan Methyl	37.8	51.5	F/P	49 - 109
2301521	Trifluralin	39.8	56.9	F/P	48 - 93.0

Reference Method: EPA 8270E
Batch ID: P409108

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301326	Acetochlor	106	107	P/P	65 - 124
2301326	Alachlor	96.4	98.2	P/P	61 - 121
2301326	Ametryn	152	138	P/P	52 - 216
2301326	Atrazine	102	99.3	P/P	68 - 133
2301326	Atrazine Desethyl	58.1	62.5	P/P	15 - 160
2301326	Azinphos Methyl	72.1	78.8	P/P	40 - 292
2301326	Azoxystrobin	98.6	105	P/P	12 - 242
2301326	Bromacil	101	99.7	P/P	54 - 167
2301326	Butylate	39.5	47.7	P/P	14 - 94.0
2301326	Caffeine	115	126	P/P	10 - 226
2301326	Carbophenothion	101	106	P/P	49 - 122
2301326	Carfentrazone Ethyl	121	118	P/P	53 - 138
2301326	Chlorfenapyr	85.7	91.5	P/P	50 - 150
2301326	Chlorpyrifos Ethyl	105	109	P/P	52 - 122
2301326	Chlorpyrifos Methyl	106	109	P/P	66 - 117
2301326	Coumaphos	185	200	P/P	50 - 220
2301326	Cyanazine	120	128	P/P	43 - 245
2301326	Cyprodinil	121	128	P/P	50 - 150
2301326	Demeton	86.5	97.8	P/P	43 - 188
2301326	Diazinon	104	103	P/P	69 - 131
2301326	Difenoconazole	84.6	86.6	P/P	34 - 231
2301326	Dimethenamid	99.1	98.2	P/P	55 - 118
2301326	Dimethomorph	102	102	P/P	50 - 150
2301326	Disulfoton	93.8	104	P/P	38 - 141
2301326	Dithiopyr	101	96.5	P/P	42 - 116
2301326	EPTC	31.3	37.2	P/P	18 - 85.0
2301326	Ethion	93.8	85.9	P/P	10 - 131
2301326	Ethoprop	100	104	P/P	65 - 129
2301326	Etridiazole	52.0	56.1	P/P	50 - 150
2301326	Fenamiphos	89.1	80.6	P/P	31 - 137
2301326	Fenbuconazole	147	144	P/P	57 - 160
2301326	Fipronil	109	109	P/P	62 - 132
2301326	Fipronil Desulfinyl	106	101	P/P	57 - 131

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270E
Batch ID: P409108

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301326	Fipronil Sulfide	98.3	96.3	P/P	15 - 138
2301326	Fipronil Sulfone	101	100	P/P	21 - 134
2301326	Fluazifop-P-butyl	97.6	96.4	P/P	54 - 133
2301326	Fludioxonil	93.0	95.1	P/P	58 - 127
2301326	Flumioxazin	269	295	P/P	33 - 300
2301326	Fluxapyroxad	99.3	70.5	P/P	23 - 141
2301326	Fonofos	87.1	95.9	P/P	58 - 119
2301326	Hexazinone	106	106	P/P	68 - 172
2301326	Imazalil	104	89.3	P/P	48 - 283
2301326	Iprodione	61.7	60.6	P/P	38 - 140
2301326	Loratadine	75.7	76.6	P/P	50 - 150
2301326	Malathion	115	116	P/P	40 - 137
2301326	Metalaxyl	97.8	90.6	P/P	21 - 129
2301326	Metolachlor	102	98.6	P/P	55 - 122
2301326	Metribuzin	113	107	P/P	38 - 187
2301326	Mevinphos	99.9	108	P/P	54 - 134
2301326	Molinate	69.9	79.7	P/P	41 - 97.0
2301326	Myclobutanil	103	102	P/P	56 - 125
2301326	Naled	105	110	P/F	10 - 108
2301326	Napropamide	95.0	93.1	P/P	50 - 150
2301326	Norflurazon	122	116	P/P	32 - 122
2301326	Oxadiazon	83.8	80.1	P/P	40 - 119
2301326	Oxyfluorfen	126	121	P/P	61 - 153
2301326	Parathion Ethyl	92.2	97.1	P/P	59 - 130
2301326	Parathion Methyl	109	118	P/P	65 - 155
2301326	Pendimethalin	86.2	87.3	P/P	49 - 138
2301326	Phenytoin	114	108	P/P	50 - 200
2301326	Phorate	116	107	P/P	54 - 161
2301326	Prodiamine	129	126	P/P	33 - 161
2301326	Prometon	104	72.1	P/P	48 - 140
2301326	Prometryn	93.8	85.0	P/P	55 - 134
2301326	Pronamide	116	110	P/P	66 - 137
2301326	Propanil	112	115	P/P	50 - 150
2301326	Propargite	151	144	P/P	50 - 200
2301326	Propiconazole	131	128	P/P	36 - 150
2301326	Pyraflufen Ethyl	132	118	P/P	50 - 150
2301326	Pyridaben	174	194	P/P	50 - 200
2301326	Pyriproxyfen	55.1	47.2	P/P	10 - 165
2301326	Simazine	105	111	P/P	57 - 145
2301326	Stirophos	98.3	98.9	P/P	50 - 150
2301326	Sulfentrazone	103	89.4	P/P	34 - 140
2301326	Tebuconazole	75.1	46.2	P/P	10 - 127
2301326	Terbufos	118	121	P/P	59 - 138
2301326	Terbuthylazine	99.0	94.1	P/P	68 - 119
2301326	Thiobencarb	102	105	P/P	50 - 150
2301326	Triadimefon	80.7	78.1	P/P	50 - 150

Reference Method: EPA 8276
Batch ID: P408985

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2302073	Chlordane	64.4	56.9	P/P	55 - 128

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8276
Batch ID: P408985

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2302073	Toxaphene	83.5	69.7	P/P	57 - 211

Reference Method: EPA 8321B
Batch ID: P409043

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301661	Acesulfame K	100	84.0	P/P	40 - 150
2301661	AMPA	94.5	88.6	P/P	40 - 150
2301661	Endothall	159	164	F/F	40 - 150
2301661	Glufosinate	99.0	82.7	P/P	40 - 150
2301661	Glyphosate	90.4	80.7	P/P	40 - 150
2301661	Sucralose	91.2	96.4	P/P	40 - 150

Reference Method: EPA 8321B
Batch ID: P409047

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301321	3-Hydroxycarbofuran	106	86.7	P/P	30 - 160
2301321	Aldicarb	82.8	80.8	P/P	30 - 160
2301321	Aldicarb Sulfone	116	119	P/P	30 - 160
2301321	Aldicarb Sulfoxide	46.1	60.2	P/P	30 - 160
2301321	Carbaryl	73.7	76.8	P/P	30 - 160
2301321	Carbofuran	81.2	87.0	P/P	30 - 160
2301321	Methiocarb	68.6	72.8	P/P	30 - 160
2301321	Methomyl	86.6	90.7	P/P	30 - 160
2301321	Oxamyl	94.1	108	P/P	30 - 160
2301321	Propoxur	81.9	91.9	P/P	30 - 160

Reference Method: EPA 8321B
Batch ID: P409048

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301732	2,4-D	148	149	P/P	30 - 160
2301732	2,4,5-T	102	114	P/P	30 - 160
2301732	Acetaminophen	125	134	P/P	30 - 160
2301732	Acetamiprid	106	90.2	P/P	30 - 160
2301732	Afidopyropen	58.6	51.0	P/P	30 - 160
2301732	Benzovindiflupyr	92.0	94.0	P/P	30 - 160
2301732	Clothianidin	150	150	P/P	30 - 160
2301732	Dinotefuran	146	146	P/P	30 - 160
2301732	Diuron	88.9	84.5	P/P	30 - 160
2301732	Fenuron	89.8	92.5	P/P	30 - 160
2301732	Fluridone	106	102	P/P	30 - 160
2301732	Hydrocodone	120	120	P/P	30 - 160
2301732	Ibuprofen	71.8	77.6	P/P	30 - 160
2301732	Linuron	76.3	78.5	P/P	30 - 160
2301732	Mandestrobin	97.2	98.3	P/P	30 - 160
2301732	MCCP	138	142	P/P	30 - 160
2301732	Naproxen	89.4	106	P/P	30 - 160
2301732	Primidone	133	121	P/P	30 - 160
2301732	Pyraclostrobin	94.0	97.3	P/P	30 - 160
2301732	Silvex	102	111	P/P	30 - 160

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8321B
Batch ID: P409048

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301732	Thiamethoxam	159	159	P/P	30 - 160
2301732	Tolfenpyrad	54.7	58.9	P/P	30 - 160
2301732	Triclopyr	110	120	P/P	30 - 160

Reference Method: SM 4500 F-C-2011
Batch ID: P409145

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301654	Fluoride	99.1	97.8	P/P	92 - 107

Reference Method: SM 5310 B-2011
Batch ID: P409765

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2301745	Organic Carbon	93.0	92.2	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: DEP SOP: NU-094-1
Batch ID: P409004

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301673	Color (true)	1.68	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P409008

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301673	Turbidity	0.720	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P409121

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301666	Calcium	3.21	Spike	P	0 - 20
2301666	Iron	3.05	Spike	P	0 - 20
2301666	Magnesium	3.71	Spike	P	0 - 20
2301666	Potassium	3.24	Spike	P	0 - 20
2301666	Sodium	2.03	Spike	P	0 - 20
2301666	Strontium	1.86	Spike	P	0 - 20
2301666	Tin	1.87	Spike	P	0 - 20
2301666	Titanium	2.69	Spike	P	0 - 20
2301666	Vanadium	3.44	Spike	P	0 - 20
2301666	Zinc	3.04	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P409121

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301666	Aluminum	1.99	Spike	P	0 - 20
2301666	Antimony	0.873	Spike	P	0 - 20
2301666	Arsenic	0.130	Spike	P	0 - 20
2301666	Barium	1.92	Spike	P	0 - 20
2301666	Beryllium	0.962	Spike	P	0 - 20
2301666	Boron	3.57	Spike	P	0 - 20
2301666	Cadmium	1.57	Spike	P	0 - 20
2301666	Chromium	1.72	Spike	P	0 - 20
2301666	Cobalt	0.913	Spike	P	0 - 20
2301666	Copper	0.838	Spike	P	0 - 20
2301666	Lead	1.05	Spike	P	0 - 20
2301666	Manganese	0.796	Spike	P	0 - 20
2301666	Molybdenum	0.319	Spike	P	0 - 20
2301666	Nickel	0.523	Spike	P	0 - 20
2301666	Selenium	0.656	Spike	P	0 - 20
2301666	Silver	0.969	Spike	P	0 - 20
2301666	Thallium	1.74	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P409367

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2302658	Lead	0.322	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P409417

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301653	Chloride	0.275	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P409420

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301653	Sulfate	0.673	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P409349

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301749	Ammonia-N	0.729	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P409406

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301740	Ammonia-N	1.10	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409191

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301639	Kjeldahl Nitrogen	5.08	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409231

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301822	Kjeldahl Nitrogen	1.05	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409375

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301510	Kjeldahl Nitrogen	3.59	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P409208

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301761	NO2NO3-N	0.451	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P409035

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301762	Total-P	1.93	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P409001

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301741	O-Phosphate-P	1.87	Spike	P	0 - 20

Reference Method: EPA 8270E

Batch ID: P408985

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301521	Aldrin	51.7	Spike	F	0 - 37
2301521	Alpha-BHC	29.7	Spike	F	0 - 16
2301521	Alpha-Chlordane	40.3	Spike	F	0 - 28
2301521	Beta-BHC	30.9	Spike	P	0 - 32
2301521	Beta-Cyfluthrin	43.1	Spike	F	0 - 41
2301521	Bifenthrin	51.9	Spike	F	0 - 36
2301521	Chlorothalonil	59.2	Spike	F	0 - 54
2301521	Cis-Nonachlor	21.2	Spike	P	0 - 33
2301521	Cypermethrin	39.7	Spike	F	0 - 38
2301521	Dacthal	31.5	Spike	F	0 - 23
2301521	DDD-p,p'	39.3	Spike	F	0 - 35
2301521	DDE-p,p'	38.3	Spike	F	0 - 27
2301521	DDT-p,p'	40.4	Spike	F	0 - 31
2301521	Delta-BHC	37.1	Spike	F	0 - 30
2301521	Deltamethrin	35.9	Spike	P	0 - 76
2301521	Dichlobenil	41.1	Spike	F	0 - 33
2301521	Dicofol	33.0	Spike	F	0 - 24
2301521	Dieldrin	37.2	Spike	F	0 - 35
2301521	Endosulfan I	36.4	Spike	F	0 - 25
2301521	Endosulfan II	33.8	Spike	F	0 - 31
2301521	Endosulfan Sulfate	27.4	Spike	P	0 - 38
2301521	Endrin	39.0	Spike	P	0 - 53
2301521	Endrin Aldehyde	44.4	Spike	P	0 - 81
2301521	Endrin Ketone	35.4	Spike	F	0 - 33
2301521	Esfenvalerate	36.9	Spike	P	0 - 46
2301521	Ethalfuralin	39.8	Spike	F	0 - 22
2301521	Fenpropathrin	31.8	Spike	F	0 - 29
2301521	Gamma-BHC	27.9	Spike	F	0 - 17
2301521	Gamma-Chlordane	40.3	Spike	F	0 - 28
2301521	Heptachlor	47.3	Spike	F	0 - 28

Quality Assurance Report Precision

Reference Method: EPA 8270E
Batch ID: P408985

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301521	Heptachlor Epoxide	32.7	Spike	F	0 - 23
2301521	Hexachlorobenzene	48.7	Spike	F	0 - 22
2301521	Kepone	23.3	Spike	P	0 - 61
2301521	Lambda-Cyhalothrin	51.0	Spike	P	0 - 52
2301521	Methoprene	55.6	Spike	F	0 - 38
2301521	Methoxychlor	28.9	Spike	P	0 - 29
2301521	MGK-264	35.4	Spike	P	0 - 41
2301521	Mirex	45.7	Spike	F	0 - 39
2301521	Oxychlordane	49.3	Spike	F	0 - 33
2301521	Permethrin	39.4	Spike	F	0 - 39
2301521	Phenothrin	55.8	Spike	F	0 - 52
2301521	Piperonyl Butoxide	41.3	Spike	P	0 - 91
2301521	Prallethrin	48.2	Spike	F	0 - 34
2301521	Tau-Fluvalinate	50.7	Spike	F	0 - 41
2301521	Tetramethrin	50.7	Spike	F	0 - 43
2301521	Trans-Nonachlor	38.6	Spike	F	0 - 31
2301521	Triclosan Methyl	30.8	Spike	F	0 - 24
2301521	Trifluralin	35.4	Spike	F	0 - 23

Reference Method: EPA 8270E
Batch ID: P409108

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301326	Acetochlor	0.610	Spike	P	0 - 27
2301326	Alachlor	1.81	Spike	P	0 - 27
2301326	Ametryn	9.74	Spike	P	0 - 34
2301326	Atrazine	2.28	Spike	P	0 - 41
2301326	Atrazine Desethyl	7.37	Spike	P	0 - 38
2301326	Azinphos Methyl	8.81	Spike	P	0 - 62
2301326	Azoxystrobin	4.63	Spike	P	0 - 32
2301326	Bromacil	0.941	Spike	P	0 - 30
2301326	Butylate	18.8	Spike	P	0 - 59
2301326	Caffeine	9.40	Spike	P	0 - 60
2301326	Carbophenothion	5.19	Spike	P	0 - 25
2301326	Carfentrazone Ethyl	1.99	Spike	P	0 - 26
2301326	Chlorfenapyr	6.61	Spike	P	0 - 30
2301326	Chlorpyrifos Ethyl	3.36	Spike	P	0 - 27
2301326	Chlorpyrifos Methyl	2.68	Spike	P	0 - 26
2301326	Coumaphos	7.69	Spike	P	0 - 30
2301326	Cyanazine	6.37	Spike	P	0 - 58
2301326	Cyprodinil	4.93	Spike	P	0 - 30
2301326	Demeton	12.3	Spike	P	0 - 25
2301326	Diazinon	1.30	Spike	P	0 - 29
2301326	Difenoconazole	2.37	Spike	P	0 - 25
2301326	Dimethenamid	0.933	Spike	P	0 - 30
2301326	Dimethomorph	0.377	Spike	P	0 - 30
2301326	Disulfoton	10.2	Spike	P	0 - 33
2301326	Dithiopyr	4.75	Spike	P	0 - 22
2301326	EPTC	17.4	Spike	P	0 - 38
2301326	Ethion	8.77	Spike	P	0 - 79

Quality Assurance Report Precision

Reference Method: EPA 8270E
Batch ID: P409108

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301326	Ethoprop	3.96	Spike	P	0 - 23
2301326	Etridiazole	7.49	Spike	P	0 - 30
2301326	Fenamiphos	10.0	Spike	P	0 - 58
2301326	Fenbuconazole	1.92	Spike	P	0 - 37
2301326	Fipronil	0.0491	Spike	P	0 - 33
2301326	Fipronil Desulfinyl	5.02	Spike	P	0 - 26
2301326	Fipronil Sulfide	2.00	Spike	P	0 - 37
2301326	Fipronil Sulfone	0.798	Spike	P	0 - 50
2301326	Fluazifop-P-butyl	1.19	Spike	P	0 - 24
2301326	Fludioxonil	2.29	Spike	P	0 - 26
2301326	Flumioxazin	9.19	Spike	P	0 - 37
2301326	Flutolanil	1.33	Spike	P	0 - 26
2301326	Fluxapyroxad	22.8	Spike	P	0 - 34
2301326	Fonofos	9.70	Spike	P	0 - 27
2301326	Hexazinone	0.171	Spike	P	0 - 36
2301326	Imazalil	15.4	Spike	P	0 - 65
2301326	Iprodione	1.75	Spike	P	0 - 33
2301326	Loratadine	1.15	Spike	P	0 - 30
2301326	Malathion	0.886	Spike	P	0 - 44
2301326	Metalaxyl	7.66	Spike	P	0 - 38
2301326	Metolachlor	2.58	Spike	P	0 - 36
2301326	Metribuzin	5.90	Spike	P	0 - 63
2301326	Mevinphos	8.21	Spike	P	0 - 26
2301326	Molinate	13.1	Spike	P	0 - 40
2301326	Myclobutanil	1.01	Spike	P	0 - 21
2301326	Naled	4.89	Spike	P	0 - 26
2301326	Napropamide	1.98	Spike	P	0 - 30
2301326	Norflurazon	4.46	Spike	P	0 - 24
2301326	Oxadiazon	4.62	Spike	P	0 - 27
2301326	Oxyfluorfen	4.15	Spike	P	0 - 24
2301326	Parathion Ethyl	5.16	Spike	P	0 - 28
2301326	Parathion Methyl	7.69	Spike	P	0 - 27
2301326	Pendimethalin	1.21	Spike	P	0 - 31
2301326	Phenytol	5.16	Spike	P	0 - 30
2301326	Phorate	7.64	Spike	P	0 - 27
2301326	Prodiamine	2.78	Spike	P	0 - 52
2301326	Prometon	35.9	Spike	F	0 - 31
2301326	Prometryn	9.79	Spike	P	0 - 28
2301326	Pronamide	4.89	Spike	P	0 - 26
2301326	Propanil	2.58	Spike	P	0 - 30
2301326	Propargite	5.18	Spike	P	0 - 30
2301326	Propiconazole	2.40	Spike	P	0 - 31
2301326	Pyraflufen Ethyl	11.7	Spike	P	0 - 30
2301326	Pyridaben	11.3	Spike	P	0 - 30
2301326	Pyriproxyfen	14.5	Spike	P	0 - 50
2301326	Simazine	5.21	Spike	P	0 - 29
2301326	Stirophos	0.613	Spike	P	0 - 30
2301326	Sulfentrazone	14.5	Spike	P	0 - 33
2301326	Tebuconazole	33.5	Spike	P	0 - 44
2301326	Terbufos	2.59	Spike	P	0 - 29
2301326	Terbutylazine	5.05	Spike	P	0 - 27

Quality Assurance Report Precision

Reference Method: EPA 8270E
Batch ID: P409108

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301326	Thiobencarb	3.56	Spike	P	0 - 30
2301326	Triadimefon	3.26	Spike	P	0 - 30

Reference Method: EPA 8276
Batch ID: P408985

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2302073	Chlordane	12.3	Spike	P	0 - 25
2302073	Toxaphene	18.0	Spike	P	0 - 24

Reference Method: EPA 8321B
Batch ID: P409043

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301661	Acesulfame K	17.0	Spike	P	0 - 30
2301661	AMPA	6.47	Spike	P	0 - 30
2301661	Endothall	3.00	Spike	P	0 - 30
2301661	Glufosinate	17.9	Spike	P	0 - 30
2301661	Glyphosate	11.2	Spike	P	0 - 30
2301661	Sucralose	5.18	Spike	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P409047

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301321	3-Hydroxycarbofuran	20.0	Spike	P	0 - 30
2301321	Aldicarb	2.42	Spike	P	0 - 30
2301321	Aldicarb Sulfone	2.04	Spike	P	0 - 30
2301321	Aldicarb Sulfoxide	26.5	Spike	P	0 - 30
2301321	Carbaryl	4.15	Spike	P	0 - 30
2301321	Carbofuran	6.98	Spike	P	0 - 30
2301321	Methiocarb	5.84	Spike	P	0 - 30
2301321	Methomyl	4.63	Spike	P	0 - 30
2301321	Oxamyl	13.4	Spike	P	0 - 30
2301321	Propoxur	11.5	Spike	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P409048

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301732	2,4-D	0.393	Spike	P	0 - 30
2301732	2,4,5-T	11.7	Spike	P	0 - 30
2301732	Acetaminophen	7.07	Spike	P	0 - 30
2301732	Acetamiprid	15.9	Spike	P	0 - 30
2301732	Afidopyropen	14.0	Spike	P	0 - 30
2301732	Bentazon	1.54	Spike	P	0 - 30
2301732	Benzovindiflupyr	2.22	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P409048

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301732	Carbamazepine	1.98	Spike	P	0 - 30
2301732	Clothianidin	0.548	Spike	P	0 - 30
2301732	Dinotefuran	0.243	Spike	P	0 - 30
2301732	Diuron	5.16	Spike	P	0 - 30
2301732	Fenuron	2.99	Spike	P	0 - 30
2301732	Fluridone	2.46	Spike	P	0 - 30
2301732	Hydrocodone	0.721	Spike	P	0 - 30
2301732	Ibuprofen	7.74	Spike	P	0 - 30
2301732	Imazapyr	5.54	Spike	P	0 - 30
2301732	Imidacloprid	1.01	Spike	P	0 - 30
2301732	Linuron	2.84	Spike	P	0 - 30
2301732	Mandestrobin	1.10	Spike	P	0 - 30
2301732	MCP	2.78	Spike	P	0 - 30
2301732	Naproxen	16.5	Spike	P	0 - 30
2301732	Primidone	5.78	Spike	P	0 - 30
2301732	Pyraclostrobin	3.49	Spike	P	0 - 30
2301732	Silvex	8.67	Spike	P	0 - 30
2301732	Thiamethoxam	0.436	Spike	P	0 - 30
2301732	Tolfenpyrad	7.51	Spike	P	0 - 30
2301732	Triclopyr	8.70	Spike	P	0 - 30

Reference Method: SM 2320 B-2011
Batch ID: P409140

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301654	Total Alkalinity	0.494	Sample	P	0 - 5.2

Reference Method: SM 2540 C-2011
Batch ID: P409499

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301248	TDS	2.96	Sample	P	0 - 10

Reference Method: SM 2540 C-2011
Batch ID: P409502

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301760	TDS	3.30	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
Batch ID: P409145

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301654	Fluoride	0.960	Spike	P	0 - 2.1

Quality Assurance Report Precision

Reference Method: SM 5310 B-2011
Batch ID: P409765

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2301745	Organic Carbon	0.802	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Surrogates

Lab Sample ID: 2301770
Field Sample ID: G4NW0017

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Carbaryl-d7	87.8	P	30 - 160

Lab Sample ID: 2301771
Field Sample ID: G4NW0017

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	2,4-D-d3	83.7	P	30 - 160
EPA 8321B	Acetaminophen-d4	103	P	30 - 160
EPA 8321B	Carbamazepine-d10	107	P	30 - 160
EPA 8321B	Diuron-d6	91.9	P	30 - 160
EPA 8321B	Endothall-d6	126	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	97.2	P	30 - 160
EPA 8321B	Ibuprofen-d3	78.1	P	30 - 160
EPA 8321B	Primidone-d5	95.2	P	30 - 160
EPA 8321B	Sucralose-d6	109	P	30 - 160

Lab Sample ID: 2301774
Field Sample ID: G4NW0017

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270E	Decachlorobiphenyl	58.2	P	31 - 118
EPA 8270E	Tetrachloro-m-xylene	50.4	P	28 - 96.0
EPA 8276	Decachlorobiphenyl	59.5	P	30 - 97.0
EPA 8276	PCNB	69.5	P	45 - 119

Lab Sample ID: 2301775
Field Sample ID: G4NW0017

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270E	Caffeine-d9	79.1	P	20 - 200
EPA 8270E	Simazine-d10	134	P	71 - 140
EPA 8270E	Terbufos-d10	113	P	62 - 131

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A110055

Included Lab Sample IDs: 2301738, 2301741, 2301746, 2301747, 2301750, 2301753

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	100	P/P	90 - 110
O-Phosphate-P	99.9	100	P/P	90 - 110

Reference Method: DEP SOP: NU-094-1

Run ID: A110056

Included Lab Sample IDs: 2301736, 2301739, 2301742, 2301743, 2301748, 2301751

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	100	P/P	85 - 115

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A110058

Included Lab Sample IDs: 2301736, 2301739, 2301742, 2301743, 2301748, 2301751

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	101	P/P	95 - 105

Reference Method: SM 2320 B-2011

Run ID: A110099

Included Lab Sample IDs: 2301736, 2301739, 2301742, 2301743, 2301748, 2301751

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	96.9	99.8	P/P	90 - 110
Total Alkalinity	99.5	96.9	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A110099

Included Lab Sample IDs: 2301736, 2301739, 2301742, 2301743, 2301748, 2301751

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	101	P/P	90 - 110
Fluoride	101	101	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A110120

Included Lab Sample IDs: 2301778, 2301779, 2301780, 2301781

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aluminum	98.8	99.9	P/P	90 - 110
Aluminum	99.4	99.5	P/P	90 - 110
Aluminum	99.5	98.8	P/P	90 - 110
Antimony	101	101	P/P	90 - 110
Antimony	98.2	99.3	P/P	90 - 110
Antimony	99.3	101	P/P	90 - 110
Arsenic	100	100	P/P	90 - 110
Arsenic	101	100	P/P	90 - 110
Arsenic	103	101	P/P	90 - 110
Barium	98.4	98.9	P/P	90 - 110
Barium	98.9	99.5	P/P	90 - 110
Barium	99.4	98.4	P/P	90 - 110
Beryllium	97.6	98.0	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A110120

Included Lab Sample IDs: 2301778, 2301779, 2301780, 2301781

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Beryllium	98.0	98.3	P/P	90 - 110
Beryllium	98.7	97.6	P/P	90 - 110
Boron	103	104	P/P	90 - 110
Boron	104	104	P/P	90 - 110
Boron	97.7	103	P/P	90 - 110
Cadmium	101	101	P/P	90 - 110
Cadmium	99.3	99.9	P/P	90 - 110
Cadmium	99.9	101	P/P	90 - 110
Chromium	97.7	98.5	P/P	90 - 110
Chromium	98.5	99.1	P/P	90 - 110
Chromium	99.1	101	P/P	90 - 110
Cobalt	97.5	98.2	P/P	90 - 110
Cobalt	98.1	97.5	P/P	90 - 110
Cobalt	98.2	98.3	P/P	90 - 110
Copper	97.9	98.4	P/P	90 - 110
Copper	98.4	98.8	P/P	90 - 110
Copper	99.7	97.9	P/P	90 - 110
Lead	96.8	98.3	P/P	90 - 110
Lead	98.2	96.8	P/P	90 - 110
Lead	98.3	97.8	P/P	90 - 110
Manganese	97.1	98.5	P/P	90 - 110
Manganese	98.5	99.9	P/P	90 - 110
Manganese	99.9	101	P/P	90 - 110
Molybdenum	99.1	99.3	P/P	90 - 110
Molybdenum	99.3	99.3	P/P	90 - 110
Molybdenum	99.3	98.2	P/P	90 - 110
Nickel	98.8	99.0	P/P	90 - 110
Nickel	99.1	98.8	P/P	90 - 110
Nickel	99.9	99.1	P/P	90 - 110
Selenium	97.1	100	P/P	90 - 110
Selenium	98.4	97.8	P/P	90 - 110
Selenium	100	98.4	P/P	90 - 110
Silver	101	102	P/P	90 - 110
Silver	102	101	P/P	90 - 110
Silver	102	102	P/P	90 - 110
Thallium	104	104	P/P	90 - 110
Thallium	104	104	P/P	90 - 110
Thallium	104	104	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A110125

Included Lab Sample IDs: 2301737, 2301740, 2301744, 2301745, 2301749, 2301752

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110
NO2NO3-N	102	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270E
 Run ID: A110135
 Included Lab Sample IDs: 2301774

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aldrin	96.8		P	80 - 120
Alpha-BHC	103		P	80 - 120
Alpha-Chlordane	98.4		P	80 - 120
Beta-BHC	104		P	80 - 120
Beta-Cyfluthrin	96.4		P	80 - 120
Bifenthrin	98.0		P	80 - 120
Chlorothalonil	102		P	80 - 120
Cis-Nonachlor	98.8		P	80 - 120
Cypermethrin	94.5		P	80 - 120
Dacthal	98.6		P	80 - 120
DDD-p,p'	100		P	80 - 120
DDE-p,p'	99.6		P	80 - 120
DDT-p,p'	101		P	80 - 120
Delta-BHC	100		P	80 - 120
Deltamethrin	90.6		P	80 - 120
Dichlobenil	98.5		P	80 - 120
Dicofol	99.0		P	80 - 120
Dieldrin	101		P	80 - 120
Endosulfan I	97.2		P	80 - 120
Endosulfan II	97.2		P	80 - 120
Endosulfan Sulfate	90.3		P	80 - 120
Endrin	98.8		P	80 - 120
Endrin Aldehyde	95.9		P	80 - 120
Endrin Ketone	101		P	80 - 120
Esfenvalerate	94.1		P	80 - 120
Ethalfuralin	102		P	80 - 120
Fenpropathrin	96.8		P	80 - 120
Gamma-BHC	107		P	80 - 120
Gamma-Chlordane	98.4		P	80 - 120
Heptachlor	97.8		P	80 - 120
Heptachlor Epoxide	102		P	80 - 120
Hexachlorobenzene	101		P	80 - 120
Kepone	83.1		P	80 - 120
Lambda-Cyhalothrin	95.4		P	80 - 120
Methoprene	95.5		P	80 - 120
Methoxychlor	103		P	80 - 120
MGK-264	93.2		P	80 - 120
Mirex	95.4		P	80 - 120
Oxychlordane	96.2		P	80 - 120
Permethrin	98.4		P	80 - 120
Phenothrin	96.4		P	80 - 120
Piperonyl Butoxide	85.4		P	80 - 120
Prallethrin	94.0		P	80 - 120
Tau-Fluvalinate	95.9		P	80 - 120
Tetramethrin	86.9		P	80 - 120
Trans-Nonachlor	100		P	80 - 120
Triclosan Methyl	97.0		P	80 - 120
Trifluralin	98.3		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A110139

Included Lab Sample IDs: 2301779

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Manganese	98.9	99.3	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A110145

Included Lab Sample IDs: 2301778, 2301779, 2301780, 2301781

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Calcium	102	102	P/P	90 - 110
Calcium	99.4	100	P/P	95 - 105
Iron	102	103	P/P	95 - 105
Iron	104	105	P/P	90 - 110
Magnesium	101	102	P/P	95 - 105
Magnesium	104	104	P/P	90 - 110
Potassium	96.0	95.8	P/P	95 - 105
Potassium	97.2	96.8	P/P	90 - 110
Sodium	97.6	96.9	P/P	95 - 105
Sodium	97.9	97.5	P/P	90 - 110
Strontium	99.1	98.3	P/P	95 - 105
Strontium	100	100	P/P	90 - 110
Tin	100	101	P/P	95 - 105
Tin	102	102	P/P	90 - 110
Titanium	100	101	P/P	90 - 110
Titanium	98.5	98.5	P/P	95 - 105
Vanadium	101	101	P/P	90 - 110
Vanadium	98.7	98.7	P/P	95 - 105
Zinc	102	103	P/P	90 - 110
Zinc	98.1	99.8	P/P	95 - 105

Reference Method: EPA 8276

Run ID: A110147

Included Lab Sample IDs: 2301774

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chlordane	90.0		P	80 - 120
Toxaphene	101		P	80 - 120

Reference Method: EPA 8321B

Run ID: A110148

Included Lab Sample IDs: 2301771

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
2,4-D	85.7	83.4	P/P	50 - 150
2,4,5-T	71.3	83.0	P/P	50 - 150
3-Hydroxycarbofuran	107	109	P/P	60 - 150
Acetaminophen	130	122	P/P	60 - 160
Acetamiprid	108	110	P/P	50 - 150
Afidopyropen	103	96.6	P/P	50 - 150
Aldicarb	67.1	134	P/P	60 - 150
Aldicarb Sulfone	100	109	P/P	50 - 150
Aldicarb Sulfoxide	107	106	P/P	50 - 150
Bentazon	106	97.4	P/P	50 - 160

Quality Assurance Report Calibration Verification

Reference Method: EPA 8321B
Run ID: A110148
Included Lab Sample IDs: 2301771

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Benzovindiflupyr	110	108	P/P	50 - 160
Carbamazepine	110	104	P/P	60 - 150
Carbaryl	108	103	P/P	60 - 150
Carbofuran	104	105	P/P	50 - 150
Clothianidin	121	111	P/P	60 - 150
Dinotefuran	119	118	P/P	50 - 150
Diuron	115	121	P/P	60 - 160
Fenuron	118	115	P/P	60 - 150
Fluridone	116	116	P/P	60 - 160
Hydrocodone	104	112	P/P	60 - 150
Ibuprofen	80.0	69.4	P/P	50 - 160
Imazapyr	141	136	P/P	50 - 160
Imidacloprid	99.8	103	P/P	60 - 150
Linuron	109	114	P/P	60 - 150
Mandestrobin	116	116	P/P	50 - 150
MCPP	78.8	84.0	P/P	50 - 150
Methiocarb	107	113	P/P	50 - 150
Methomyl	111	110	P/P	50 - 150
Naproxen	73.5	65.4	P/P	50 - 160
Oxamyl	99.7	103	P/P	50 - 150
Primidone	107	104	P/P	60 - 150
Propoxur	100	103	P/P	50 - 150
Pyraclostrobin	113	112	P/P	60 - 150
Silvex	75.7	70.9	P/P	50 - 150
Thiamethoxam	101	112	P/P	50 - 150
Tolfenpyrad	101	98.4	P/P	60 - 150
Triclopyr	74.2	73.7	P/P	50 - 150

Reference Method: EPA 8321B
Run ID: A110153
Included Lab Sample IDs: 2301771

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
AMPA	97.0	92.5	P/P	60 - 160
Glufosinate	102	100	P/P	60 - 160
Glyphosate	93.6	89.9	P/P	60 - 160

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A110154
Included Lab Sample IDs: 2301736, 2301739, 2301742, 2301743, 2301748, 2301751

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	100	99.4	P/P	90 - 110
Chloride	101	100	P/P	90 - 110
Sulfate	100	99.8	P/P	90 - 110
Sulfate	99.8	99.1	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A110173

Included Lab Sample IDs: 2301737, 2301740, 2301744, 2301745, 2301749, 2301752

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	101	P/P	90 - 110
Total-P	101	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A110182

Included Lab Sample IDs: 2301737, 2301749

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.7	99.5	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A110183

Included Lab Sample IDs: 2301771

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acesulfame K	102	80.9	P/P	60 - 160
Endothall	139	125	P/P	60 - 160
Sucralose	97.7	104	P/P	60 - 160

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A110188

Included Lab Sample IDs: 2301744, 2301745, 2301752

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	94.0	97.0	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A110212

Included Lab Sample IDs: 2301740, 2301744, 2301745, 2301752

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	103	99.0	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A110218

Included Lab Sample IDs: 2301737, 2301749

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	94.5	93.7	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A110219

Included Lab Sample IDs: 2301778

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Lead	96.0	96.7	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270E
Run ID: A110253
Included Lab Sample IDs: 2301775

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acetochlor	101		P	80 - 120
Alachlor	101		P	80 - 120
Ametryn	97.9		P	80 - 120
Atrazine	98.6		P	80 - 120
Atrazine Desethyl	111		P	80 - 120
Azinphos Methyl	108		P	80 - 120
Azoxystrobin	104		P	80 - 120
Bromacil	99.7		P	80 - 120
Butylate	107		P	80 - 120
Caffeine	90.6		P	80 - 120
Carbophenothion	81.1		P	80 - 120
Carfentrazone Ethyl	93.5		P	80 - 120
Chlorfenapyr	106		P	80 - 120
Chlorpyrifos Ethyl	104		P	80 - 120
Chlorpyrifos Methyl	101		P	80 - 120
Coumaphos	94.9		P	80 - 120
Cyanazine	107		P	80 - 120
Cyprodinil	102		P	80 - 120
Demeton	103		P	80 - 120
Diazinon	100		P	80 - 120
Difenoconazole	104		P	80 - 120
Dimethenamid	101		P	80 - 120
Dimethomorph	105		P	80 - 120
Disulfoton	100		P	80 - 120
Dithiopyr	101		P	80 - 120
EPTC	102		P	80 - 120
Ethion	101		P	80 - 120
Ethoprop	100		P	80 - 120
Etridiazole	99.6		P	80 - 120
Fenamiphos	100		P	80 - 120
Fenbuconazole	94.1		P	80 - 120
Fipronil	101		P	80 - 120
Fipronil Desulfinyl	97.4		P	80 - 120
Fipronil Sulfide	105		P	80 - 120
Fipronil Sulfone	101		P	80 - 120
Fluazifop-P-butyl	102		P	80 - 120
Fludioxonil	90.7		P	80 - 120
Flumioxazin	98.7		P	80 - 120
Flutolanil	98.5		P	80 - 120
Fluxapyroxad	95.7		P	80 - 120
Fonofos	101		P	80 - 120
Hexazinone	102		P	80 - 120
Imazalil	95.3		P	80 - 120
Iprodione	101		P	80 - 120
Loratadine	113		P	80 - 120
Malathion	88.5		P	80 - 120
Metalaxyl	104		P	80 - 120
Metolachlor	103		P	80 - 120
Metribuzin	104		P	80 - 120
Mevinphos	105		P	80 - 120
Molinate	102		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270E
Run ID: A110253
Included Lab Sample IDs: 2301775

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Myclobutanil	102		P	80 - 120
Naled	106		P	80 - 120
Napropamide	103		P	80 - 120
Norflurazon	95.1		P	80 - 120
Oxadiazon	103		P	80 - 120
Oxyfluorfen	103		P	80 - 120
Parathion Ethyl	103		P	80 - 120
Parathion Methyl	100		P	80 - 120
Pendimethalin	105		P	80 - 120
Phenytol	96.6		P	80 - 120
Phorate	106		P	80 - 120
Prodiamine	106		P	80 - 120
Prometon	106		P	80 - 120
Prometryn	102		P	80 - 120
Pronamide	102		P	80 - 120
Propanil	98.6		P	80 - 120
Propargite	99.5		P	80 - 120
Propiconazole	97.4		P	80 - 120
Pyraflufen Ethyl	96.9		P	80 - 120
Pyridaben	102		P	80 - 120
Pyriproxyfen	114		P	80 - 120
Simazine	104		P	80 - 120
Stirophos	89.0		P	80 - 120
Sulfentrazone	96.1		P	80 - 120
Tebuconazole	99.0		P	80 - 120
Terbufos	97.4		P	80 - 120
Terbutylazine	103		P	80 - 120
Thiobencarb	104		P	80 - 120
Triadimefon	102		P	80 - 120

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A110275
Included Lab Sample IDs: 2301740

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	94.3	92.9	P/P	90 - 110

Reference Method: SM 5310 B-2011
Run ID: A110374
Included Lab Sample IDs: 2301737, 2301740, 2301744, 2301745, 2301749, 2301752

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	100	96.1	P/P	90 - 110
Organic Carbon	96.6	100	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			LCS	Precision SMP	MS
DEP SOP: NU-094-1	Color (true)	99.2					1.68	
EPA 180.1 Rev. 2.0	Turbidity						0.720	
EPA 200.7 Rev. 4.4	Calcium	102	102	106				3.21
	Iron	107	110	107	110			3.05
	Magnesium	102	104	110				3.71
	Potassium	97.5	96.9	101				3.24
	Sodium	97.7	97.4	103				2.03
	Strontium	98.7	99.8	97.9	101			1.86
	Tin	104	103	101	108			1.87
	Titanium	102	104	101	103			2.69
	Vanadium	102	105	102	104			3.44
	Zinc	101	104	101	105			3.04
EPA 200.8 Rev. 5.4	Aluminum	99.7	97.8	99.8	101			1.99
	Antimony	97.7	99.7	101	100			0.873
	Arsenic	101	102	101	99.6			0.130
	Barium	103	102	104	104			1.92
	Beryllium	100	101	100	101			0.962
	Boron	98.7	96.0	101	104			3.57
	Cadmium	98.8	98.5	100	101			1.57
	Chromium	99.0	100	102	101			1.72
	Cobalt	100	100	101	99.2			0.913
	Copper	101	101	102	102			0.838
	Lead	100	101	102	103			1.05
	Lead	101	101	100	99.6			0.322
	Manganese	97.6	98.6	99.5	100			0.796
	Molybdenum	100	102	103	100			0.319
	Nickel	100	100	101	101			0.523
	Selenium	99.8	98.7	99.3	96.3			0.656
	Silver	102	102	103	103			0.969
	Thallium	103	109	111	111			1.74
EPA 300.0 Rev. 2.1	Chloride	101	95.0	98.6			0.275	
	Sulfate	96.8	93.2	96.3			0.673	
EPA 350.1 Rev. 2.0	Ammonia-N	98.4	100	101				0.729
	Ammonia-N	97.4	95.0	96.2				1.10
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	99.5	102	111				5.08
	Kjeldahl Nitrogen	98.2						1.05
	Kjeldahl Nitrogen	99.7	101	96.2				3.59
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	100	102	101				0.451
EPA 365.1 Rev. 2.0	Total-P	105	102	104				1.93
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	100	100	102				1.87
EPA 8270E	Acetochlor	114	106	107				0.610
	Alachlor	103	96.4	98.2				1.81
	Aldrin	66.0	23.6	40.1				51.7
	Alpha-BHC	96.5	60.7	81.9				29.7
	Alpha-Chlordane	90.4	37.7	56.7				40.3
	Ametryn	112	152	138				9.74
	Atrazine	106	102	99.3				2.28
	Atrazine Desethyl	86.6	58.1	62.5				7.37
	Azinphos Methyl	72.2	72.1	78.8				8.81
	Azoxystrobin	109	98.6	105				4.63
	Beta-BHC	109	70.0	95.6				30.9
	Beta-Cyfluthrin	112	48.5	75.1				43.1
	Bifenthrin	92.6	36.7	62.5				51.9
	Bromacil	109	101	99.7				0.941

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision SMP	MS
			LCS			
EPA 8270E	Butylate	66.8	39.5	47.7		18.8
	Caffeine	85.8	115	126		9.40
	Carbophenothion	111	101	106		5.19
	Carfentrazone Ethyl	125	121	118		1.99
	Chlorfenapyr	103	85.7	91.5		6.61
	Chlorothalonil	113	91.9	169		59.2
	Chlorpyrifos Ethyl	113	105	109		3.36
	Chlorpyrifos Methyl	110	106	109		2.68
	Cis-Nonachlor	90.6	44.2	54.7		21.2
	Coumaphos	171	185	200		7.69
	Cyanazine	133	120	128		6.37
	Cypermethrin	86.7	47.1	70.5		39.7
	Cyprodinil	140	121	128		4.93
	Dacthal	95.8	53.6	73.7		31.5
	DDD-p,p'	90.3	36.3	54.1		39.3
	DDE-p,p'	89.0	32.7	48.2		38.3
	DDT-p,p'	88.4	40.2	60.6		40.4
	Delta-BHC	112	77.8	113		37.1
	Deltamethrin	132	60.7	87.2		35.9
	Demeton	96.8	86.5	97.8		12.3
	Diazinon	106	104	103		1.30
	Dichlobenil	85.9	34.8	52.9		41.1
	Dicofol	83.4	45.9	64.1		33.0
	Dieldrin	91.3	40.2	58.7		37.2
	Difenoconazole	108	84.6	86.6		2.37
	Dimethenamid	91.0	99.1	98.2		0.933
	Dimethomorph	111	102	102		0.377
	Disulfoton	89.0	93.8	104		10.2
	Dithiopyr	104	101	96.5		4.75
	Endosulfan I	90.6	43.1	62.4		36.4
	Endosulfan II	88.1	38.8	54.6		33.8
	Endosulfan Sulfate	91.6	52.6	69.3		27.4
	Endrin	89.2	38.0	56.4		39.0
	Endrin Aldehyde	91.3	32.2	50.5		44.4
	Endrin Ketone	101	52.9	75.7		35.4
	EPTC	61.1	31.3	37.2		17.4
	Esfenvalerate	109	53.3	77.5		36.9
	Ethalfuralin	87.6	38.3	57.3		39.8
	Ethion	103	93.8	85.9		8.77
	Ethoprop	85.4	100	104		3.96
	Etridiazole	69.1	52.0	56.1		7.49
	Fenamiphos	126	89.1	80.6		10.0
	Fenbuconazole	145	147	144		1.92
	Fenpropathrin	101	39.6	54.6		31.8
	Fipronil	110	109	109		0.0491
	Fipronil Desulfinyl	111	106	101		5.02
	Fipronil Sulfide	95.7	98.3	96.3		2.00
	Fipronil Sulfone	105	101	100		0.798
	Fluazifop-P-butyl	97.3	97.6	96.4		1.19
	Fludioxonil	111	93.0	95.1		2.29
	Flumioxazin	248	269	295		9.19
	Flutolanil	104				1.33
	Fluxapyroxad	127	99.3	70.5		22.8
	Fonofos	85.7	87.1	95.9		9.70

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 8270E	Gamma-BHC	87.0	72.7	96.3			27.9
	Gamma-Chlordane	85.9	36.8	55.4			40.3
	Heptachlor	80.1	31.3	50.7			47.3
	Heptachlor Epoxide	90.7	46.1	64.1			32.7
	Hexachlorobenzene	86.5	33.2	54.5			48.7
	Hexazinone	114	106	106			0.171
	Imazalil	88.8	104	89.3			15.4
	Iprodione	70.6	61.7	60.6			1.75
	Kepone	156	81.8	103			23.3
	Lambda-Cyhalothrin	111	53.5	90.1			51.0
	Loratadine	87.4	75.7	76.6			1.15
	Malathion	128	115	116			0.886
	Metalaxyl	99.7	97.8	90.6			7.66
	Methoprene	76.6	29.0	51.3			55.6
	Methoxychlor	103	58.6	78.4			28.9
	Metolachlor	111	102	98.6			2.58
	Metribuzin	111	113	107			5.90
	Mevinphos	89.0	99.9	108			8.21
	MGK-264	86.9	48.4	69.3			35.4
	Mirex	74.4	25.8	41.0			45.7
	Molinate	76.1	69.9	79.7			13.1
	Myclobutanil	109	103	102			1.01
	Naled	119	105	110			4.89
	Napropamide	103	95.0	93.1			1.98
	Norflurazon	119	122	116			4.46
	Oxadiazon	92.9	83.8	80.1			4.62
	Oxychlordane	87.6	52.4	86.7			49.3
	Oxyfluorfen	117	126	121			4.15
	Parathion Ethyl	106	92.2	97.1			5.16
	Parathion Methyl	120	109	118			7.69
	Pendimethalin	79.5	86.2	87.3			1.21
	Permethrin	112	53.0	79.0			39.4
	Phenothrin	87.5	38.7	68.7			55.8
	Phenytoin	113	114	108			5.16
	Phorate	98.5	116	107			7.64
	Piperonyl Butoxide	103	57.4	87.3			41.3
	Prallethrin	71.1	36.8	60.2			48.2
	Prodiamine	98.2	129	126			2.78
	Prometon	105	104	72.1			35.9
	Prometryn	94.7	93.8	85.0			9.79
	Pronamide	121	116	110			4.89
	Propanil	118	112	115			2.58
	Propargite	163	151	144			5.18
	Propiconazole	99.8	131	128			2.40
	Pyraflufen Ethyl	124	132	118			11.7
	Pyridaben	159	174	194			11.3
	Pyriproxyfen	61.7	55.1	47.2			14.5
	Simazine	123	105	111			5.21
	Stirophos	84.7	98.3	98.9			0.613
	Sulfentrazone	136	103	89.4			14.5
	Tau-Fluvalinate	101	41.1	69.0			50.7
	Tebuconazole	78.7	75.1	46.2			33.5
	Terbufos	102	118	121			2.59
	Terbuthylazine	101	99.0	94.1			5.05

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 8270E	Tetramethrin	80.3	42.1	70.6			50.7
	Thiobencarb	112	102	105			3.56
	Trans-Nonachlor	85.2	37.0	54.6			38.6
	Triadimefon	98.3	80.7	78.1			3.26
	Triclosan Methyl	92.0	37.8	51.5			30.8
EPA 8276	Trifluralin	85.4	39.8	56.9			35.4
	Chlordane	81.7	64.4	56.9			12.3
	Toxaphene	76.4	83.5	69.7			18.0
EPA 8321B	2,4-D	89.2	148	149			0.393
	2,4,5-T	75.5	102	114			11.7
	3-Hydroxycarbofuran	79.7	106	86.7			20.0
	Acesulfame K	99.1	100	84.0			17.0
	Acetaminophen	96.0	125	134			7.07
	Acetamiprid	98.2	106	90.2			15.9
	Afidopyropen	71.9	58.6	51.0			14.0
	Aldicarb	45.1	82.8	80.8			2.42
	Aldicarb Sulfone	93.7	116	119			2.04
	Aldicarb Sulfoxide	90.8	46.1	60.2			26.5
	AMPA	91.4	94.5	88.6			6.47
	Bentazon	105					1.54
	Benzovindiflupyr	93.1	92.0	94.0			2.22
	Carbamazepine	99.5					1.98
	Carbaryl	81.3	73.7	76.8			4.15
	Carbofuran	86.3	81.2	87.0			6.98
	Clothianidin	106	150	150			0.548
	Dinotefuran	105	146	146			0.243
	Diuron	92.4	88.9	84.5			5.16
	Endothall	136	159	164			3.00
	Fenuron	111	89.8	92.5			2.99
	Fluridone	97.5	106	102			2.46
	Glufosinate	97.9	99.0	82.7			17.9
	Glyphosate	91.2	90.4	80.7			11.2
	Hydrocodone	115	120	120			0.721
	Ibuprofen	79.1	71.8	77.6			7.74
	Imazapyr	149					5.54
	Imidacloprid	91.8					1.01
	Linuron	84.1	76.3	78.5			2.84
	Mandestrobin	99.1	97.2	98.3			1.10
	MCPP	108	138	142			2.78
	Methiocarb	70.4	68.6	72.8			5.84
	Methomyl	88.6	86.6	90.7			4.63
	Naproxen	98.5	89.4	106			16.5
	Oxamyl	95.6	94.1	108			13.4
	Primidone	99.4	133	121			5.78
	Propoxur	78.6	81.9	91.9			11.5
	Pyraclostrobin	92.4	94.0	97.3			3.49
	Silvex	69.8	102	111			8.67
	Sucralose	91.7	91.2	96.4			5.18
	Thiamethoxam	96.0	159	159			0.436
	Tolfenpyrad	52.0	54.7	58.9			7.51
	Triclopyr	75.5	110	120			8.70
SM 2320 B-2011	Total Alkalinity	99.0				0.494	
SM 2540 C-2011	TDS					2.96	
	TDS					3.30	

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
SM 4500 F-C-2011	Fluoride	97.5	99.1	97.8		0.960
SM 5310 B-2011	Organic Carbon	95.0	93.0	92.2		0.802

Reference Method Descriptions

Method	Description	Associated Samples
DEP SOP: NU-094-1	True Color measured at 450 nm (not valid for NPDES monitoring)	2301736, 2301739, 2301742, 2301743, 2301748, 2301751
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2301736, 2301739, 2301742, 2301743, 2301748, 2301751
EPA 200.7 Rev. 4.4	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2301778, 2301779, 2301780, 2301781
EPA 200.8 Rev. 5.4	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2301778, 2301779, 2301780, 2301781
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2301736, 2301739, 2301742, 2301743, 2301748, 2301751
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2301736, 2301739, 2301742, 2301743, 2301748, 2301751
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2301737, 2301740, 2301744, 2301745, 2301749, 2301752
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2301737, 2301740, 2301744, 2301745, 2301749, 2301752
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2301737, 2301740, 2301744, 2301745, 2301749, 2301752
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2301737, 2301740, 2301744, 2301745, 2301749, 2301752
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2301738, 2301741, 2301746, 2301747, 2301750, 2301753
EPA 8270E	Organochlorine pesticides, pyrethroids and tracers in water matrices by GC/MS/MS	2301774
EPA 8270E	Organonitrogen and phosphorus pesticides in water matrices by GC/MS/MS	2301775
EPA 8276	Organochlorine pesticides in water matrices by GC/MS with Negative Chemical Ionization	2301774
EPA 8321B	Carbamates in water matrices by HPLC/MS/MS	2301770
EPA 8321B	Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS	2301771
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2301736, 2301739, 2301742, 2301743, 2301748, 2301751
SM 2340B	Hardness by calculation, in aqueous matrices, reported as mg/L CaCO ₃	2301778, 2301779, 2301780, 2301781
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2301736, 2301739, 2301742, 2301743, 2301748, 2301751
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2301736, 2301739, 2301742, 2301743, 2301748, 2301751
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2301736, 2301739, 2301742, 2301743, 2301748, 2301751
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2301737, 2301740, 2301744, 2301745, 2301749, 2301752

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
DEP SOP: NU-094-1	01/26/2022			01/26/2022 15:36	Samantha L Bates	2301736, 2301739
	01/26/2022			01/26/2022 15:37	Samantha L Bates	2301742
	01/26/2022			01/26/2022 15:38	Samantha L Bates	2301743
	01/26/2022			01/26/2022 15:39	Samantha L Bates	2301748, 2301751
EPA 180.1 Rev. 2.0	01/26/2022			01/26/2022 14:59	Khloe J Berg	2301736, 2301739, 2301742, 2301743, 2301748, 2301751
EPA 200.7 Rev. 4.4	01/26/2022	01/28/2022 10:45	Vijayalakshmi Reddy	01/31/2022 12:56	Josef B Mick	2301781
	01/26/2022	01/28/2022 10:45	Vijayalakshmi Reddy	01/31/2022 16:20	Josef B Mick	2301778
	01/26/2022	01/28/2022 10:45	Vijayalakshmi Reddy	01/31/2022 16:26	Josef B Mick	2301779
	01/26/2022	01/28/2022 10:45	Vijayalakshmi Reddy	01/31/2022 16:33	Josef B Mick	2301780
EPA 200.8 Rev. 5.4	01/26/2022	01/28/2022 10:45	Vijayalakshmi Reddy	01/28/2022 17:02	Emily M Philpot	2301781
	01/26/2022	01/28/2022 10:45	Vijayalakshmi Reddy	01/28/2022 19:45	Emily M Philpot	2301780
	01/26/2022	01/28/2022 10:45	Vijayalakshmi Reddy	01/28/2022 20:57	Emily M Philpot	2301778
	01/26/2022	01/28/2022 10:45	Vijayalakshmi Reddy	01/28/2022 21:06	Emily M Philpot	2301779
EPA 300.0 Rev. 2.1	01/26/2022	01/28/2022 10:45	Vijayalakshmi Reddy	01/31/2022 14:21	Alexander Thompson	2301779
	01/26/2022	02/03/2022 11:00	Emily M Philpot	02/03/2022 19:13	McKinley C Carter	2301778
	01/26/2022			02/03/2022 14:43	James L Waggeberby	2301742
	01/26/2022			02/03/2022 16:31	James L Waggeberby	2301736
EPA 350.1 Rev. 2.0	01/26/2022			02/03/2022 17:08	James L Waggeberby	2301739
	01/26/2022			02/03/2022 17:20	James L Waggeberby	2301743
	01/26/2022			02/03/2022 17:32	James L Waggeberby	2301748
	01/26/2022			02/03/2022 17:44	James L Waggeberby	2301751
EPA 351.2 Rev. 2.0	01/26/2022			02/02/2022 13:42	Ping Hua	2301749
	01/26/2022			02/02/2022 13:56	Ping Hua	2301737
	01/26/2022			02/03/2022 14:45	Ping Hua	2301740
	01/26/2022			02/03/2022 14:56	Ping Hua	2301744
EPA 353.2 Rev. 2.0	01/26/2022			02/03/2022 14:57	Ping Hua	2301745
	01/26/2022			02/03/2022 15:00	Ping Hua	2301752
	01/26/2022	02/01/2022 09:14	Samantha L Bates	02/02/2022 14:44	Alexandra J Mattheus	2301744
	01/26/2022	02/01/2022 09:14	Samantha L Bates	02/02/2022 14:45	Alexandra J Mattheus	2301745
EPA 353.2 Rev. 2.0	01/26/2022	02/01/2022 09:14	Samantha L Bates	02/02/2022 14:47	Alexandra J Mattheus	2301752
	01/26/2022	02/02/2022 09:46	Samantha L Bates	02/03/2022 14:32	Alexandra J Mattheus	2301737
	01/26/2022	02/02/2022 09:46	Samantha L Bates	02/03/2022 14:34	Alexandra J Mattheus	2301749
	01/26/2022	02/04/2022 09:26	Samantha L Bates	02/07/2022 14:20	Alexandra J Mattheus	2301740
EPA 353.2 Rev. 2.0	01/26/2022			01/31/2022 11:14	Beverly Y. Sanders	2301744
	01/26/2022			01/31/2022 11:16	Beverly Y. Sanders	2301745
	01/26/2022			01/31/2022 11:17	Beverly Y. Sanders	2301749
	01/26/2022			01/31/2022 11:18	Beverly Y. Sanders	2301752
EPA 353.2 Rev. 2.0	01/26/2022			01/31/2022 11:27	Beverly Y. Sanders	2301740
	01/26/2022			01/31/2022 11:29	Beverly Y. Sanders	2301737

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 365.1 Rev. 2.0	01/26/2022	01/27/2022 17:15	Simonne.V. Calhoun	01/31/2022 16:21	Sarah A Nixon	2301737
	01/26/2022	01/27/2022 17:15	Simonne.V. Calhoun	01/31/2022 16:22	Sarah A Nixon	2301740
	01/26/2022	01/27/2022 17:15	Simonne.V. Calhoun	01/31/2022 16:23	Sarah A Nixon	2301744
	01/26/2022	01/27/2022 17:15	Simonne.V. Calhoun	01/31/2022 16:24	Sarah A Nixon	2301745
	01/26/2022	01/27/2022 17:15	Simonne.V. Calhoun	01/31/2022 16:25	Sarah A Nixon	2301749
	01/26/2022	01/27/2022 17:15	Simonne.V. Calhoun	01/31/2022 16:38	Sarah A Nixon	2301752
EPA 365.1 Rev. 2.0 dissolved	01/26/2022			01/26/2022 14:34	James L Waggeberby	2301741
	01/26/2022			01/26/2022 14:42	James L Waggeberby	2301738
	01/26/2022			01/26/2022 14:43	James L Waggeberby	2301746, 2301747
	01/26/2022			01/26/2022 14:44	James L Waggeberby	2301750
	01/26/2022			01/26/2022 14:45	James L Waggeberby	2301753
EPA 8270E	01/26/2022	01/28/2022 08:00	Fe-Val Siddell	01/31/2022 20:40	Vandana T. Nagar	2301774
	01/26/2022	01/31/2022 08:00	Rasheda Ghaffari	02/02/2022 20:35	Michael Poppell	2301775
EPA 8276	01/26/2022	01/28/2022 08:00	Fe-Val Siddell	02/01/2022 11:45	Arthur Maknenko	2301774
EPA 8321B	01/26/2022	01/28/2022 12:00	June Rhew	01/29/2022 09:42	Juyoung Kim	2301770
	01/26/2022	01/28/2022 14:00	Hoor Shaik	01/31/2022 16:04	Juyoung Kim	2301771
	01/26/2022	01/31/2022 09:00	Manjita Shrestha	01/31/2022 15:44	Manjita Shrestha	2301771
	01/26/2022	01/31/2022 09:00	Manjita Shrestha	01/31/2022 23:39	Manjita Shrestha	2301771
SM 2320 B-2011	01/26/2022			01/27/2022 20:48	Dale L. Simmons	2301736
	01/26/2022			01/27/2022 20:56	Dale L. Simmons	2301739
	01/26/2022			01/27/2022 21:05	Dale L. Simmons	2301742
	01/26/2022			01/27/2022 21:14	Dale L. Simmons	2301743
	01/26/2022			01/27/2022 21:57	Dale L. Simmons	2301748
	01/26/2022			01/27/2022 22:04	Dale L. Simmons	2301751
SM 2340B	01/26/2022			01/31/2022 12:56	Josef B Mick	2301781
	01/26/2022			01/31/2022 16:20	Josef B Mick	2301778
	01/26/2022			01/31/2022 16:26	Josef B Mick	2301779
	01/26/2022			01/31/2022 16:33	Josef B Mick	2301780
SM 2540 C-2011	01/26/2022			01/28/2022 14:02	Yijie Li	2301736, 2301739, 2301742, 2301743, 2301748, 2301751
SM 2540 D-2011	01/26/2022			01/28/2022 14:02	Yijie Li	2301736, 2301739, 2301742, 2301743, 2301748, 2301751
SM 4500 F-C-2011	01/26/2022			01/27/2022 20:48	Dale L. Simmons	2301736
	01/26/2022			01/27/2022 20:56	Dale L. Simmons	2301739
	01/26/2022			01/27/2022 21:05	Dale L. Simmons	2301742
	01/26/2022			01/27/2022 21:14	Dale L. Simmons	2301743
	01/26/2022			01/27/2022 21:57	Dale L. Simmons	2301748
	01/26/2022			01/27/2022 22:04	Dale L. Simmons	2301751
SM 5310 B-2011	01/26/2022			02/12/2022 21:13	Touraj Touran	2301745
	01/26/2022			02/12/2022 22:39	Touraj Touran	2301749

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
SM 5310 B-2011	01/26/2022			02/13/2022 01:10	Touraj Touran	2301737
	01/26/2022			02/13/2022 01:26	Touraj Touran	2301740
	01/26/2022			02/13/2022 01:49	Touraj Touran	2301744
	01/26/2022			02/13/2022 02:13	Touraj Touran	2301752